


BCA Assay w/ Cytation 5 (562 nm)

Source: doc_80a4341643e3_20260919-cryopreserved-ripa-fibroblasts-bone-marrowxs.xlsx | Date: 2026-09-16 19:25:03 | Reader: Cytation5 (S/N: 1710247)
Plate: 96 WELL PLATE | Temperature: 25.1 °C | Wavelength: 562 nm (Absorbance endpoint)

1. 4PL Calibration Curve & Parameters



4PL Model Equation:
$$y = \text{bottom} + \frac{\text{top} - \text{bottom}}{1 + (\text{EC}_{50}/x)^{\text{HillSlope}}}$$


Fitted Parameters:

- Bottom = 0.107951 • Top = 4.0000
- EC50 = 8.15889 µg/µL • Hill Slope = 1.04106

Goodness of Fit:


- R² = 0.999327 • RMSE = 0.006284 OD
- Standards: Wells C3:D10 duplicate series (2.0 to 0.0 µg/µL)
- Monotonic: Yes | Valid fit: Yes

2. Raw Absorbance Plate (OD 562 nm)



	1	2	3	4	5	6	7	8	9	10	11	12
A	1.66	1.782	1.556	1.786	1.353	2.045	1.232	1.027	0.958	1.168	0.871	0.05
B	1.408	1.289	0.823	0.958	0.646	0.956	0.711	0.528	0.524	0.636	0.497	0.049
C	1.038	1.099	0.82	0.641	0.498	0.391	0.311	0.212	0.16	0.102	0.051	0.047
D	0.871	0.918	0.859	0.723	0.502	0.397	0.321	0.217	0.162	0.103	0.048	0.048
E	0.656	0.697	0.34	0.185	0.146	0.338	0.181	0.202	0.048	0.047	0.048	0.048
F	0.443	0.463	0.052	0.049	0.049	0.048	0.047	0.049	0.048	0.05	0.049	0.049
G	0.313	0.329	0.088	0.05	0.048	0.048	0.052	0.048	0.049	0.05	0.048	0.047
H	0.161	0.162	0.049	0.051	0.048	0.048	0.052	0.049	0.05	0.049	0.048	0.048

3. Estimated Concentrations (µg/µL, Inverse 4PL)



	1	2	3	4	5	6	7	8	9	10	11	12
A												
B												
C			1.938	1.392	0.991	0.707	0.503	0.258	0.131	-0.016		
D			2.064	1.636	1.002	0.723	0.529	0.270	0.136	-0.013		
E			0.577	0.192	0.097	0.572	0.182	0.234				
F												
G			-0.051 ⚠									
H												

Notes & Quality Control:

- Standards: Wells C3:D10 duplicate series (2.000 to 0.000 µg/µL BSA, dotted black border).
- ⚠ Flagged Wells: Wells marked with ⚠ (red borders) exceed the 2.0 µg/µL calibration range or detector OVRFLW (>4.0 OD).
- Diagnostics: Negative values are signed fit diagnostics, not physical concentrations. Blank wells indicate baseline (<0.065 OD) or excluded rows/cols.
- Exclusions: Rows A & B and Columns 1 & 2 were excluded per protocol.