SCANSURE AI | CLINICAL ULTRASOUND ENHANCEMENT & ANOMALY REPORT

Offline AI-Assisted Review Packet for Remote Radiologist Evaluation

Study ID: demo_case03 | Generated: 2026-11-11 16:59:05 UTC | Workstation: jio-fiber (i5-13420H / RTX 3050 6GB)

CLINICAL SAFETY NOTICE (NON-DIAGNOSTIC AI ASSISTANCE):

ScanSure AI is an experimental AI-assisted tool. It does not provide a medical diagnosis. All findings require radiologist review.

1. ULTRASOUND ACQUISITION & IMAGE ENHANCEMENT SUMMARY

Source Video File: breast_ultrasound_normal_control03.mp4

Resolution & FPS: 320x240 @ 15.0 FPS (45 frames)

Mean PSNR / SSIM: 22.6 dB | SSIM 0.9244

Edge Preservation Index (EPI): 0.9438

Enhancement Pipeline: speckle_bilateral_clahe

Strength / CLAHE / Gamma: 75% | Clip 2.2 | Gamma 1.05

Speckle ENL Improvement: 47.94 -> 52.95 (+10.5%)

Video Frame Integrity: Exact Match Verified (45/45 frames)

2. AI-ASSISTED ANOMALY DETECTION & CONFIDENCE TRIAGE

Model Architecture: ScanSure-UNet-Lite-v1.2-BreastUS (1.2.0-busi-calibrated)

Supported Anatomical Scope: Breast Ultrasound (B-Mode Linear Array 5-12 MHz)

Frames With Findings: 0 / 45 **Peak Raw Score: 0.000 (Thr 0.55)** **Triage: No region detected above val**

3. ANNOTATED KEYFRAMES & REPRESENTATIVE REGION MEASUREMENTS

Frame #	Timestamp	Category / Classification	BBox [x,y,w,h]	Area (px)	Raw Score

No regions detected above the validated threshold across this ultrasound cine loop.

4. GENERATED ANNOTATED KEYFRAME FILES & RADIOLOGIST REVIEW SIGN-OFF

* frame_0000_annotated.png | frame_0000_comparison.png (Frame #0 (0.00s) - Baseline Enhanced)
* frame_0002_annotated.png | frame_0002_comparison.png (Frame #2 (0.13s) - Baseline Enhanced)
* frame_0004_annotated.png | frame_0004_comparison.png (Frame #4 (0.27s) - Baseline Enhanced)
* frame_0006_annotated.png | frame_0006_comparison.png (Frame #6 (0.40s) - Baseline Enhanced)

QUALIFIED RADIOLOGIST REVIEW & CLINICAL IMPRESSION:

[ ] Confirmed Simple Cyst [ ] Follow-Up Ultrasound Recommended [ ] Diagnostic Biopsy / Referral

Radiologist Name / Signature: _____ Date: _____