


# Spectral QC Summary Report

| Fluorophore     | Similarity | Status                 |
|-----------------|------------|------------------------|
| BV510           | 0.9976     | PASS                   |
| BV605           | 0.9968     | PASS                   |
| Spark Blue 550  | N/A        | No Reference Available |
| Alexa Fluor 700 | 0.9987     | PASS                   |
| 7-AAD           | N/A        | No Reference Available |


BV510  
Cosine Similarity: 0.9976

BV510 — Experiment    ■ Library Reference




BV605  
Cosine Similarity: 0.9968

BV605 — Experiment    ■ Library Reference




Spark Blue 550  
No Reference Available

Spark Blue 550 — Experiment    ■ Library Reference



Alexa Fluor 700  
Cosine Similarity: 0.9987

Alexa Fluor 700 — Experiment    ■ Library Reference



7-AAD  
No Reference Available

7-AAD — Experiment ■ Library Reference

