



Backscatter Calibration Certificate

RBRtridente bb.chl-a.fDOM, RA180, UW (6000dbar) s/n: 240124

Channel: 1

Calibrated Range: 0-0.05m-1sr-1

Initial accuracy (% of calibrated range): 5% of 0.05m-1sr-1 (0.0025m-1sr-1)

Reference Standards and Materials

Manufacturer	Model/Part No.	Description	Serial No.
Sequoia	LISST-Tau / SEQ-FP-LTAU-532NM	Transmissometer	1043
Sigma Aldrich	90517-10ML-F	Analytical polystyrene monodisperse standard, size: 0.1 um	CAS: 9003-53-6
RBR	RBRtridente / 12205	Three-channel fluorescence and backscatter sensor	233998

Reference	Reference Value (m-1sr-1)	Voltage Ratio, Vr	Measured (m-1sr-1)	Calibration Error (m-1sr-1)	Coefficients
					C0: -189.33875E-9
					C1: 7.3635845
Black cover	0.000	0.0000000	0.000	0.000	
Bead suspension	0.049	0.0066382	0.049	0.000	

Temperature: 21.8 °C

$$BB_{(m-1sr-1)} = C_0 + C_1 * V_r$$

Calibration Date: 2025-09-30

Issue Date: 2025-10-01

Operator:

dluong

Approver:

slangstaff