



fDOM Calibration Certificate

RBRtridente bb.chl-a.fDOM, RA180, UW (6000dbar) s/n: 240125

Channel: 3

Calibrated range: 0-500.0 ppb

Initial Accuracy: 5.0% FS (25.0 ppb)

Reference Standards and Materials

Manufacturer	Model/Part No.	Description	Serial No.
Sartorius	SECURA225D-1S	Analytical balance	0034903849
Troemner	7017-0T (Class U)	100g Test Weight	1000139797
Troemner	7016-0T (Class U)	200g Test Weight	1000139798
Thermo Fisher Scientific	AC163710100	Quinine sulfate dihydrate	CAS: 6119-70-6

Reference	Reference Value (ppb)	Voltage Ratio, Vr	Measured (ppb)	Calibration Error (ppb)	Coefficients
					C0: 7.12481E-3
Black cover	0.000	-0.00000001	0.000	0.000	C1: 88.94601E3
Quinine Solution	552.962	0.0062167	552.962	0.000	

Temperature: 21.2 °C

$$fDOM_{(ppb)} = C_0 + C_1 * V_r$$

Calibration Date: 2025-09-29

Issue Date: 2025-10-01

Operator:

dluong

Approver:

slangstaff