

Product: Oxygen Optode 4831UW DW
Serial No: 1300
Date: 06.10.2025

Certificate No: 2469392601300

This is to certify that this product has been pressure tested with the following instrument, and we confirm that no irregularities were found during the test:

Autoklav 800 bar – sn: 0210005

Pressure readings:

Pressure (Bar)	Pressure time (hour)
600	1

Date: 23 Oct 2025

Sign:

Laila A. Skålnes

Laila Skålnes, Production Engineer

Product: Oxygen Optode 4831UW DW
Serial No: 1301
Date: 06.10.2025

Certificate No: 2469402601301

This is to certify that this product has been pressure tested with the following instrument, and we confirm that no irregularities were found during the test:

Autoklav 800 bar – sn: 0210005

Pressure readings:

Pressure (Bar)	Pressure time (hour)
600	1

Date: 23 Oct 2025

Sign:

Laila A. Skålnes

Laila Skålnes, Production Engineer