Supplementary Materials of Distributionally Robust Image Classifiers for Stroke Diagnosis in Accelerated MRI

Boran Hao[1][0000−0001−7922−0513], Guoyao Shen[2][0000−0002−5479−8332],
Ruidi Chen[1][0000−0002−1508−1742], Chad W. Farris[3][0000−0002−1133−3834],
Stephan W. Anderson[3][0000−0002−5367−7459], Xin Zhang[2][0000−0002−4413−5084],
and Ioannis Ch. Paschalidis[1][0000−0002−3343−2913]

[1] Department of Electrical and Computer Engineering, Boston University, Boston
MA 02215, USA
yannisp@bu.edu
[2] Department of Mechanical Engineering and the Photonics Center, Boston
University, Boston MA 02215, USA
[3] Boston Medical Center and Boston University Chobanian & Avedisian School of
Medicine, Boston MA 02118, USA

1 MRI scanner parameters

Table 1: Example parameters for 1.5 T (from Subject 0001) and 3.0 T (from Subject 0002) Philips scanners. Parameters may change slightly in different scans.

	1.5 T Philips Scanners	3.0 T Philips Scanners
Image Collection System Name	Philips Achieva	Philips Ingenia
Magnetic Field Magnitude	1.5 Tesla	3.0 Tesla
Acquisition	2D Acquisition	2D Acquisition
Slice Thickness	5 mm	5 mm
Slice Gap	0.5 mm	0 mm
Repetition Time (TR)	4183 ms	4105 ms
Echo Time (TE)	68.8 ms	86.0 ms
Number of Averages	2	1
Number of Phase Encoding Steps	131	106
Echo Train Length	119	53
Flip Angle	90 degrees	90 degrees
FOV	240 × 240 mm	230 × 230 mm
Pixel size	0.94 × 0.94	1.2 × 1.2