

Supplementary Materials of Distributionally Robust Image Classifiers for Stroke Diagnosis in Accelerated MRI

Boran Hao¹[0000–0001–7922–0513], Guoyao Shen²[0000–0002–5479–8332],
Ruidi Chen³[0000–0002–1508–1742], Chad W. Farris⁴[0000–0002–1133–3834],
Stephan W. Anderson⁴[0000–0002–5367–7459], Xin Zhang²[0000–0002–4413–5084],
and Ioannis Ch. Paschalidis¹[0000–0002–3343–2913]

¹ Department of Electrical and Computer Engineering, Boston University, Boston
MA 02215, USA
yannisp@bu.edu

² Department of Mechanical Engineering and the Photonics Center, Boston
University, Boston MA 02215, USA

³ Amazon SCOT, Bellevue WA 98004, USA

⁴ Boston Medical Center and Boston University Chobanian & Avedisian School of
Medicine, Boston MA 02118, USA

1 MRI scanner parameters

Table 1: Example parameters for 1.5 T (from Subject 0001) and 3.0 T (from Subject 0002) Philips scanners. Parameters may change slightly in different scans.

	1.5 T Philips Scanners	3.0 T Philips Scanners
Image Collection System Name	Philips Achieva	Philips Ingenia
Magnetic Field Magnitude	1.5 Tesla	3.0 Tesla
Acquisition	2D Acquisition	2D Acquisition
Slice Thickness	5 mm	5 mm
Slice Gap	0.5 mm	0 mm
Repetition Time (TR)	4183 ms	4105 ms
Echo Time (TE)	68.8 ms	86.0 ms
Number of Averages	2	1
Number of Phase Encoding Steps	131	106
Echo Train Length	119	53
Flip Angle	90 degrees	90 degrees
FOV	240×240 mm	230×230 mm
Pixel size	0.94×0.94	1.2×1.2