

CALIBRATION CERTIFICATE

Department:CapacitorAmbient Temperature:Control No:EA-13-CP-106Relative Humidity:

Equipment Name: LC TESTER

Manufacturer: Pro-FACE DIGITAL
Model No.: PFXGP4301TADW
Serial No.: K00423LC042

Equipment Class: B

Location: BT-PRODUCTION (BFI2040)

Received Date: Calibration Date:

Calibration Due:

Procedure Used: CE102-QID-CL-048B

Reference Standard used during Calibration

| No | Standard Used | Control No. | Model / Serial | Calibration Due Date |
|----|---------------|-------------|----------------|----------------------|
| | | | | |
| | | | | |

The calibration was carried out with reference to the following calibration and measurement standards which are traceable to the NMC A*STAR (Singapore), NPL (UK), NIST (US) and/or other recognized national calibration laboratories.

Charging Voltage (CH1)

| Parameter | Nominal | Measured Value | Error | Min | Max | Expanded Uncertainty | Result |
|-----------|----------|-------------------|-------|----------|----------|-------------------------|--------|
| 1 V | 1.000 V | V | V | 0.970 V | 1.030 V | V | |
| 3 V | 3.000 V | V | V | 2.910 V | 3.090 V | V | |
| 5 V | 5.000 V | V | V | 4.850 V | 5.150 V | V | |
| 10 V | 10.000 V | V | V | 9.700 V | 10.300 V | V | |
| 20 V | 20.000 V | V | V | 19.400 V | 20.600 V | V | |
| 25 V | 25.000 V | V | V | 24.250 V | 25.750 V | V | |
| 30 V | 30.000 V | V | V | 29.100 V | 30.900 V | V | |
| 35 V | 35.000 V | V | V | 33.950 V | 36.050 V | V | |

Prepared By

Approved By

Document approvals are managed using e-Governance System, no signature required.

PT Panasonic Industrial Devices Batam

Puri Industrial Park 2000, Batam Centre Kelurahan III

Batam, Indonesia

Page 1 of 4 QID-CL-B21-007S



Charging Voltage (CH2)

| Parameter | Nominal | Measured Value | Error | Min | Max | Expanded Uncertainty | Result |
|-----------|----------|-------------------|-------|----------|----------|-------------------------|--------|
| 1 V | 1.000 V | V | V | 0.970 V | 1.030 V | V | |
| 3 V | 3.000 V | V | V | 2.910 V | 3.090 V | V | |
| 5 V | 5.000 V | V | V | 4.850 V | 5.150 V | V | |
| 10 V | 10.000 V | V | V | 9.700 V | 10.300 V | V | |
| 20 V | 20.000 V | V | V | 19.400 V | 20.600 V | V | |
| 25 V | 25.000 V | V | V | 24.250 V | 25.750 V | V | |
| 30 V | 30.000 V | V | V | 29.100 V | 30.900 V | V | |
| 35 V | 35.000 V | V | V | 33.950 V | 36.050 V | V | |

Charging Voltage Monitor

| Parameter | Nominal | Measured Value | Error | Min | Max | Expanded Uncertainty | Result |
|-----------|---------|-------------------|-------|---------|------------------------|----------------------|--------|
| 1 V | 1.00 V | V | V | 0.97 V | 1.03 V | V | |
| 3 V | 3.00 V | V | V | 2.95 V | 3.05 V | V | |
| 5 V | 5.00 V | V | V | 4.93 V | 5.07 V | V | |
| 10 V | 10.00 V | V | V | 9.88 V | 10.12 V | V | |
| 20 V | 20.00 V | V | V | 19.78 V | 20.22 V | V | |
| 25 V | 25.00 V | V | V | 24.73 V | 25.27 V | V | |
| 30 V | 30.00 V | V | V | 29.68 V | 3 <mark>0.3</mark> 2 V | V | |
| 35 V | 35.00 V | V | V | 34.63 V | 3 <mark>5.37</mark> V | V | |

Measurement Voltage (CH1)

| Parameter | Nominal | Measured Value | Error | Min | Max | Expanded Uncertainty | Result |
|-----------|----------|-------------------|-------|----------|----------|-------------------------|--------|
| 1 V | 1.000 V | V | V | 0.970 V | 1.030 V | V | |
| 3 V | 3.000 V | V | V | 2.910 V | 3.090 V | V | |
| 5 V | 5.000 V | V | V | 4.850 V | 5.150 V | V | |
| 10 V | 10.000 V | V | V | 9.700 V | 10.300 V | V | |
| 20 V | 20.000 V | V | V | 19.400 V | 20.600 V | V | |
| 25 V | 25.000 V | V | V | 24.250 V | 25.750 V | V | |
| 30 V | 30.000 V | V | V | 29.100 V | 30.900 V | V | |
| 35 V | 35.000 V | V | V | 33.950 V | 36.050 V | V | |

Measurement Voltage (CH2)

| Parameter | Nominal | Measured Value | Error | Min | Max | Expanded Uncertainty | Result |
|-----------|----------|-------------------|-------|----------|----------|-------------------------|--------|
| 1 V | 1.000 V | V | V | 0.970 V | 1.030 V | V | |
| 3 V | 3.000 V | V | V | 2.910 V | 3.090 V | V | |
| 5 V | 5.000 V | V | V | 4.850 V | 5.150 V | V | |
| 10 V | 10.000 V | V | V | 9.700 V | 10.300 V | V | |
| 20 V | 20.000 V | V | V | 19.400 V | 20.600 V | V | |
| 25 V | 25.000 V | V | V | 24.250 V | 25.750 V | V | |
| 30 V | 30.000 V | V | V | 29.100 V | 30.900 V | V | |
| 35 V | 35.000 V | V | V | 33.950 V | 36.050 V | V | |

Prepared By Approved By

Document approvals are managed using e-Governance System, no signature required.

PT Panasonic Industrial Devices Batam Puri Industrial Park 2000, Batam Centre Kelurahan III Batam, Indonesia

Page 2 of 4 QID-CL-B21-007S



Measurement Voltage Monitor

| Parameter | Nominal | Measured Value | Error | Min | Max | Expanded Uncertainty | Result |
|-----------|----------|-------------------|-------|----------|----------|-------------------------|--------|
| 1 V | 1.000 V | V | V | 0.970 V | 1.030 V | V | |
| 3 V | 3.000 V | V | V | 2.910 V | 3.090 V | V | |
| 5 V | 5.000 V | V | V | 4.930 V | 5.070 V | V | |
| 10 V | 10.000 V | V | V | 9.880 V | 10.120 V | V | |
| 20 V | 20.000 V | V | V | 19.780 V | 20.220 V | V | |
| 25 V | 25.000 V | V | V | 24.730 V | 25.270 V | V | |
| 30 V | 30.000 V | V | V | 29.680 V | 30.320 V | V | |
| 35 V | 35.000 V | V | V | 34.630 V | 35.370 V | V | |

LC Measurement (CH1) [Range: 20uA; Load Resistance: 1MΩ]

| Parameter | Nominal | Measured | Error | Min | Max | Expanded | Result |
|-----------|----------|----------|-------|------------------|----------|-------------|--------|
| | | Value | | | | Uncertainty | |
| 1 V | 1.00 μΑ | μΑ | μΑ | 0.94 μΑ | 1.06 μΑ | μΑ | |
| 3 V | 3.00 μΑ | μΑ | μΑ | 2.84 μΑ | 3.16 μΑ | μΑ | |
| 5 V | 5.00 μΑ | μΑ | μΑ | 4.74 μΑ | 5.26 μΑ | μΑ | |
| 10 V | 10.00 μΑ | μΑ | μΑ | 9.49 μΑ | 10.51 μΑ | μΑ | |
| 20 V | 20.00 μΑ | μΑ | μΑ | 18 .99 μΑ | 21.01 μΑ | μΑ | |

LC Measurement (CH2) [Range: 20uA; Load Resistance: 1MΩ]

| Parameter | Nominal | Measured Value | Error | Min | Max | Expanded Uncertainty | Result |
|-----------|----------|-------------------|-------|---------|----------|-------------------------|--------|
| 1 V | 1.00 μΑ | μΑ | μΑ | 0.94 μΑ | 1.06 μΑ | μΑ | |
| 3 V | 3.00 μΑ | μΑ | μΑ | 2.84 μΑ | 3.16 μΑ | μΑ | |
| 5 V | 5.00 μΑ | μΑ | μΑ | 4.74 μΑ | 5.26 μΑ | μΑ | |
| 10 V | 10.00 μΑ | μΑ | μΑ | 9.49 μΑ | 10.51 μΑ | μΑ | |
| 20 V | 20.00 A | A | А | 18.99 A | 21.01 A | A | |

LC Measurement (CH1) [Range: 200uA; Load Resistance: 100kΩ]

| Parameter | Nominal | Measured Value | Error | Min | Max | Expanded Uncertainty | Result |
|-----------|-----------|-------------------|-------|-----------|-----------|-------------------------|--------|
| 1 V | 10.00 μΑ | μΑ | μΑ | 9.49 μΑ | 10.51 μΑ | μΑ | |
| 3 V | 30.00 μΑ | μΑ | μA | 28.49 μΑ | 31.51 μΑ | μΑ | |
| 5 V | 50.00 μΑ | μΑ | μΑ | 47.49 μA | 52.51 μΑ | μΑ | |
| 10 V | 100.00 μΑ | μΑ | μΑ | 94.99 μΑ | 105.01 μΑ | μΑ | |
| 20 V | 200.00 μΑ | μΑ | μΑ | 189.99 μΑ | 210.01 μΑ | μΑ | |

LC Measurement (CH2) [Range: 200uA; Load Resistance: 100kΩ]

| Parameter | Nominal | Measured Value | Error | Min | Max | Expanded Uncertainty | Result |
|-----------|-----------|-------------------|-------|-----------|-----------|-------------------------|--------|
| 1 V | 10.00 μΑ | μΑ | μΑ | 9.49 μΑ | 10.51 μΑ | μΑ | |
| 3 V | 30.00 μΑ | μΑ | μΑ | 28.49 μΑ | 31.51 μΑ | μΑ | |
| 5 V | 50.00 μΑ | μΑ | μΑ | 47.49 μΑ | 52.51 μΑ | μΑ | |
| 10 V | 100.00 μΑ | μΑ | μΑ | 94.99 μΑ | 105.01 μΑ | μΑ | |
| 20 V | 200.00 μΑ | μΑ | μΑ | 189.99 μΑ | 210.01 μΑ | μΑ | |

Prepared By Approved By

Document approvals are managed using e-Governance System, no signature required.

PT Panasonic Industrial Devices Batam Puri Industrial Park 2000, Batam Centre Kelurahan III Batam, Indonesia

Page 3 of 4 QID-CL-B21-007S



LC Measurement (CH1) [Range: 1000uA; Load Resistance: 10kΩ]

| Parameter | Nominal | Measured Value | Error | Min | Max | Expanded Uncertainty | Result |
|-----------|------------|-------------------|-------|-----------|------------|-------------------------|--------|
| 1 V | 100.00 μΑ | μΑ | μΑ | 94.99 μΑ | 105.01 μΑ | μΑ | |
| 3 V | 300.00 μΑ | μΑ | μΑ | 284.99 μΑ | 315.01 μΑ | μΑ | |
| 5 V | 500.00 μΑ | μΑ | μΑ | 474.99 μΑ | 525.01 μΑ | μA | |
| 10 V | 1000.00 μΑ | μΑ | μΑ | 949.99 μΑ | 1050.01 μΑ | μΑ | |

LC Measurement (CH2) [Range: 1000uA; Load Resistance: $10k\Omega$]

| Parameter | Nominal | Measured Value | Error | Min | Max | Expanded Uncertainty | Result |
|-----------|------------|-------------------|-------|-----------|------------|-------------------------|--------|
| 1 V | 100.00 μΑ | μΑ | μΑ | 94.99 μΑ | 105.01 μΑ | μΑ | |
| 3 V | 300.00 μΑ | μΑ | μΑ | 284.99 μΑ | 315.01 μA | μΑ | |
| 5 V | 500.00 μΑ | μΑ | μΑ | 474.99 μA | 525.01 μA | μΑ | |
| 10 V | 1000.00 μΑ | μΑ | μΑ | 949.99 μΑ | 1050.01 μΑ | μΑ | |

Prepared By Approved By

Document approvals are managed using e-Governance System, no signature required.

PT Panasonic Industrial Devices Batam Puri Industrial Park 2000, Batam Centre Kelurahan III Batam, Indonesia

Page 4 of 4 QID-CL-B21-007S