

Asset_ID	Name	Serial_number	Model_number	Model_name	Manufacturer	Installation_date	Warranty_expires	Facility	Building	Floor	Department	Scrapping_date	Data_sheet
1	Syringe pump	2100587	SPLF2	MO-SPLF2	Holmark	2019-02-01	2024-02-01	Cairo hospital	1	1	ICU	0000-00-00	Syringe pump_1.pdf
2	Electric ICU Bed	3140687	AG-BY003C	AG-BY003C	AEgEA	2016-10-08	2021-10-08	Cairo hospital	1	1	ICU	0000-00-00	Electric ICU Bed_2.pdf
3	Defibrillator	5170687	M4735A	Heart start XL	Philips	2014-03-01	2022-03-01	Cairo hospital	1	1	ICU	0000-00-00	Defibrillator_3.pdf
4	CT scanner	6150487	CT6000ICT	CT6000ICT	Philips	2015-05-17	2026-05-17	Cairo hospital	2	1	Radiology	0000-00-00	CT scanner_4.pdf
5	MRI Machine	9145687	Siemens Aera 1.5T	Siemens Aera 1.5T	Siemens	2014-07-01	2027-07-01	Cairo hospital	2	1	Radiology	0000-00-00	MRI Machine_5.pdf
6	3D-4D Ultrasound	2175987	SAX8LV-EXP	Sonoae X8	Samsung	2018-04-07	2023-04-07	Cairo hospital	3	2	Obstetrics and Gynecology	0000-00-00	3D-4D Ultrasound_6.pdf
7	3D-4D Ultrasound	3140954	USS-SAR7C3U	Sonoace R7	Samsung	2019-08-03	2024-08-03	Cairo hospital	3	2	Obstetrics and Gynecology	0000-00-00	3D-4D Ultrasound_7.pdf
8	CARDIOHELP	5320187		CARDIOHELP System	MAQUET	2013-04-20	2018-04-10	Cairo hospital	2	1	Cardiology	2020-04-20	CARDIOHELP_8.pdf
9	Tourniquet	6740587	4000	A.T.S.4000 Tourniquet Systems	zimmer Biomet	2017-10-01	2024-10-01	Cairo hospital	3	1	Operating Theater	0000-00-00	Tourniquet_9.pdf
10	Heart-Lung machine	5940687	HL 20	HL 20 Vario Flex	MAQUET	2013-05-21	2020-05-21	Cairo hospital	3	1	Operating Theater	0000-00-00	Heart-Lung machine_10.pdf
11	Ventilator	8957157	SVM950	Servo-i Ventilator	MAQUET	2012-05-19	2016-05-19	Cairo hospital	1	1	ICU	0000-00-00	Ventilator_11.pdf
12	Vital signs monitor	2009807	VS30	EarlyVue	Philips	2019-09-08	2022-09-08	Cairo hospital	3	3	ICU	0000-00-00	Vital signs monitor_12.pdf
13	Electric steam sterilizer	2190787	SCF284/03	Philips Avent 3-in-1 electric steam sterilizer	Philips	2017-04-11	2020-04-11	Cairo hospital	1	2	Central sterile services department	0000-00-00	Electric steam sterilizer_13.pdf
14	Anesthesia Machine	9812345	GDOA847	GE Datex Ohmeda Aisys	Soma Tech Intl	2020-03-18	2022-03-18	Cairo hospital	3	1	Operating Theater	0000-00-00	Anesthesia Machine_14.pdf
15	Cardiograph	5463127	TC70	PageWriter	Philips	2017-05-06	2019-05-06	Cairo hospital	2	2	Cardiology	0000-00-00	Cardiograph_15.pdf
16	Electrosurgical unit	6100487	Shark II	Bovie Aaron Smoke Shark II	Soma Tech Intl	2014-07-13	2016-07-13	Cairo hospital	3	1	Operating Theater	0000-00-00	Electrosurgical unit_16.pdf
17	Vital signs monitor	6172487	863283	VS4	Philips	2018-03-02	2021-03-02	Cairo hospital	3	3	ICU	0000-00-00	Vital signs monitor_17.pdf
18	Vital signs monitor	6936487	863283	VS4	Philips	2020-04-02	2023-04-02	Cairo hospital	3	3	ICU	0000-00-00	Vital signs monitor_17.pdf
19	Vital signs monitor	9385487	67NCTX-B	Connex 6700	Welch Allyn	2017-10-02	2020-10-02	Cairo hospital	3	3	ICU	0000-00-00	Vital signs monitor_19.pdf
20	Electric ICU Bed	3040687	ES-96HD	ES-96HD	JosonCare	2018-10-20	2023-10-20	Cairo hospital	1	1	ICU	0000-00-00	Electric ICU Bed_20.pdf
21	Electric ICU Bed	3040649	ES-96HD	ES-96HD	JosonCare	2018-11-15	2023-11-15	Cairo hospital	1	1	ICU	0000-00-00	Electric ICU Bed_20.pdf
22	Electric ICU Bed	8360649	ES-12DW	ES-12DW	JosonCare	2017-02-15	2022-02-15	Cairo hospital	1	1	ICU	0000-00-00	Electric ICU Bed_22.pdf
23	Ventilator	9134729	980	Puritan Bennett 980	Puritan Bennett	2017-08-19	2021-08-19	Cairo hospital	1	1	ICU	0000-00-00	Ventilator_23.pdf
24	Ventilator	9705255	Newport_HT70_Plus	HT70	Covidien	2018-01-01	2022-01-01	Cairo hospital	1	1	ICU	0000-00-00	Ventilator_24.pdf
25	3D C-Arm	2800649	ARCADIS Orbic 3D	3000T	Siemens	2017-02-15	2022-02-15	Cairo hospital	2	1	Radiology	0000-00-00	3D C-Arm_25.pdf
26	Portable X-Ray	1362149	Optima	XR200MAX	Philips	2016-03-01	2020-03-01	Cairo hospital	2	1	Radiology	0000-00-00	Portable X-Ray_26.pdf

Part_Number	Name	Asset_ID	Vendor	Cost_USD	Quantity
10128486	QUENCH ELBOW ASSY-FLANGED	5	siemens	435	4
10128499	Aux. Vent blank(BD END)	5	siemens	640	8
10128723	Aux. Vent Blank(elbow end)	5	siemens	270	15
10128733	De-Mountable pin fitting kit	5	siemens	890	1
10128734	QV elbow fitting kit	5	siemens	2700	1
10128742	air and sea Freight assy	5	siemens	750	5
10128924	Li battery	5	siemens	250	10
10128974	de-coupler fitting kit	5	siemens	375	1
10129376	4k filtered PRV adapter	5	siemens	360	6
10129669	OR98 pipe set kit	5	siemens	500	3
10129671	FCL pin kit	5	siemens	2000	2
10129672	quench valve	5	siemens	1500	1
10144747	Anti-Siphon Plate Adjustment Knob	1	Holmark	50	6
10152525	Servo Duo Guard filter	11	MAQUET	100	8
31014789	Aerogen Solo Nebulizer	11	MAQUET	150	2
41014789	HLS Module Advanced	8	MAQUET	500	2
51014732	Guide Rod Collar Clamp	1	Holmark	15	7
58998123	Li battary	4	PHILIPS	140	9
61014859	Bubble sensor	8	MAQUET	200	3
110040031	Couch Down position switch	4	PHILIPS	550	3
110065761	ROTOR BELT	4	PHILIPS	100	6
110190131	patient table limit swich cab	4	PHILIPS	900	2
566016252	external 4x CD ROM	4	PHILIPS	765	15
566017651	BRAKE SOLENOID	4	PHILIPS	520	1
566019761	fuse 150A/700V	4	PHILIPS	30	11
566308451	CLAMP ASSY	4	PHILIPS	700	3
567474001	portuguese	4	PHILIPS	390	2

SSN	Name	Phone_number	Company
534890245	Mahmoud	318459362	Soma Tech Intl
546896213	Ashraf	1111647852	MAQUET
632598751	Hany	100025487	Siemens
1023589637	Samy	1125896472	Philips
1100235612	Yasser	211063542	MAQUET
1230470140	Ayman	147258369	Samsung
1283421801	Khaled	132990153	Samsung
1365987450	Khaled	1289751264	zimmer Biomet
1429932802	Ehab	321045128	MAQUET
1550221561	Haitham	182246042	Zimmer Biomet
1628129635	Kareem	291583547	Soma Tech Intl
1723091235	Basem	172532981	AEgEA
1723642532	Mahmoud	1215301432	Soma Tech Intl
1882044351	Eslam	182251618	Siemens
1902213572	Eyad	201872417	AEgEA
2019335814	Yassin	207135512	Zimmer Biomet
2047080647	Mahmoud	101234798	Holmark
2147483647	Ahmed	101234798	Philips

SSN	Order_number	Number_of_hours
2147483647	1	7
1230470140	2	10
2147483647	3	10
1365987450	4	8
546896213	5	8
632598751	6	8
2047080647	7	5
546896213	8	4
632598751	9	9
1365987450	10	3
1230470140	11	3
546896213	12	4
2147483647	13	5
1365987450	14	1
2047080647	15	3
2147483647	16	3
546896213	17	5
1230470140	18	7
2147483647	19	8
1365987450	20	4
546896213	21	4
632598751	22	6
546896213	23	3
2147483647	24	4
1230470140	25	4
546896213	26	2
632598751	27	4
2147483647	28	4
1365987450	29	9
546896213	30	4
1230470140	31	4
2147483647	32	6
1365987450	33	2
632598751	34	4
546896213	35	4
1365987450	36	4
2047080647	37	3
2147483647	38	5
1365987450	39	4

Order_number	Asset_ID	Name	Status	Creation_date	Repair/PM	Due_date	PM_date	PM_frequency	Priority	Description	Repair_cost
1	1	Inspection, PM	Completed	2020-04-15	PM	2020-05-01	2020-05-01	Monthly	Low	Checked power, battery, alarms, tubes, cleaning and calibration	
2	5	Inspection, PM	Completed	2019-06-05	PM	2019-07-02	2019-07-01	Monthly	High	Inspecting and recording the MRI's cryogen levels.	
3	4	fix error	complete	2020-03-31	Repair	2020-04-06	2020-05-17	Yearly	High	fatal error	700
4	4	fix rotor error	In progress	2020-05-04	Repair	2020-05-30	2020-05-17	Yearly	Low	rotor error	300
5	1	fix	Not started	2020-05-02	Repair	2020-05-14	2020-06-01	Monthly	High	Guide Rod Collar Clamp of the syringe pump is broken, please fix.	
6	3	Inspection, PM	Not Started	2020-05-08	PM	2020-05-20	2020-05-20	Yearly	High	Checking electrodes, battery, synchronizer, connectors	
7	6	Inspection, PM	Not Started	2020-04-03	PM	2020-04-10	2020-04-07	Yearly	Low	Checking battery, probes, lamps and cleaning filters, transducers, keyboard	
8	14	Inspection, PM	Completed	0000-00-00	PM	2019-03-18	2019-03-18	Yearly	High	Checked battery, indicators, alarms, fuse, tubes, fluid levels	
9	7	Inspection, PM	Not Started	2020-05-08	PM	2020-08-03	2020-08-03	Yearly	Low	Checking battery, probes, lamps and cleaning filters, transducers, keyboard	
10	6	fix image artifacts	Completed	2020-03-20	Repair	2019-04-10	2021-05-07	Yearly	Low	Artifacts in the images, fixed by replacing transducer	
11	3	fix battery failure	completed	2020-04-03	Repair	2020-04-15	2020-05-20	Yearly	Low	Battery failure, fixed by replacing battery	15000
12	5	fix helium low level	completed	2020-02-05	Repair	2020-02-22	2020-07-01	Yearly	Low	Low helium level, fixed by helium filling	50
13	11	Inspection, PM	Completed	2019-01-20	PM	2019-02-20	2019-02-19	Monthly	High	Checked air filters, modes of ventilation, alarms, cleaning filters	
14	11	Inspection, PM	Completed	2019-12-20	PM	2019-01-05	2019-02-19	Monthly	High	Checked air filters, modes of ventilation, alarms, cleaning filters	
15	1	Inspection, PM	Completed	2020-02-22	PM	2020-03-01	2020-03-01	Monthly	Low	Check power, battery, alarms, tubes, cleaning and calibration	
16	1	Inspection, PM	Completed	2020-03-20	PM	2020-04-01	2020-04-01	Monthly	Low	Check power, battery, alarms, tubes, cleaning and calibration	
17	15	Inspection, PM	Completed	2019-12-06	PM	2020-01-06	2020-01-06	Monthly	High	Check connectors, electrodes, fuse, lubricate paper drive, replace battery	
18	13	Inspection, PM	Completed	2020-03-20	PM	2020-04-11	2020-04-11	Weekly	High	Checked and cleaned sensors, steam lines, replaced filter, verified temperature	
19	12	fix alarm failure	Not Started	2020-05-20	Repair	2020-06-15	2019-09-08	Yearly	Low	Alarm failure	
20	10	fix alarm	Completed	2020-01-02	Repair	2020-01-07	2020-01-21	Monthly	High	Temperature sensor not working, fixed by replacing sensor	100
21	10	Inspection, PM	Completed	2020-03-01	PM	2020-03-22	2020-03-21	Monthly	High	Checked battery, mixer, alarms, connectors, calibration	
22	10	Inspection, PM	Completed	2020-05-01	PM	2020-05-22	2020-05-21	Monthly	High	Checked battery, mixer, alarms, connectors, lubricated motors	
23	9	Inspection, PM	Completed	2018-09-01	PM	2018-10-02	2018-10-01	Yearly	High	Checked motors, tubes, switches, alarms, pressure, replaced tube	
24	9	Inspection, PM	Completed	2019-09-01	PM	2019-10-02	2019-10-01	Yearly	High	Checked motors, tubes, switches, alarms, pressure, replaced battery	
25	2	Inspection, PM	Completed	2019-09-30	PM	2019-10-08	2019-10-08	Yearly	Low	Checked mattress, actuators, motors	
26	2	Inspection, PM	Completed	2018-09-28	PM	2018-10-08	2018-10-08	Yearly	Low	Checked mattress, actuators, motors	
27	19	Inspection, PM	Completed	2018-09-20	PM	2018-10-02	2018-10-02	Yearly	Low	Checked wires, battery, SpO2, alarms, cleaned connectors	
28	19	Inspection, PM	Completed	2019-09-20	PM	2019-10-02	2019-10-02	Yearly	Low	Checked wires, battery, SpO2, alarms, changed battery	
29	19	fix monitor problem	Not Started	2020-05-05	Repair	2020-05-30	2020-10-02	Yearly	Low	Monitor problem	
30	17	Inspection, PM	Completed	2019-02-22	PM	2019-03-02	2019-03-02	Yearly	Low	Checked wires, battery, SpO2, alarms, cleaned connectors, changed wire connector	
31	17	Inspection, PM	Completed	2020-02-22	PM	2020-03-02	2020-03-02	Yearly	Low	Checked wires, battery, SpO2, alarms	
32	17	Inspection, PM	Completed	2021-02-22	PM	2021-03-02	2021-03-02	Yearly	Low	Checked wires, battery, SpO2, alarms	
33	23	Inspection, PM	Completed	2018-07-20	PM	2018-08-19	2018-08-19	Monthly	High	Checked air filters, modes of ventilation, alarms, cleaned filters, tubes	
34	23	Inspection, PM	Completed	2019-07-23	PM	2019-08-19	2019-08-19	Monthly	High	Checked air filters, modes of ventilation, alarms, cleaned filters, tubes	
35	20	Inspection, PM	Completed	2019-09-30	PM	2019-10-20	2019-10-20	Yearly	Low	Checked mattress, actuators, motors, lubricated motors	
36	21	Inspection, PM	Completed	2019-10-30	PM	2019-11-15	2019-11-15	Yearly	Low	Checked mattress, actuators, motors, lubricated motors	
37	22	Inspection, PM	Completed	2018-02-01	PM	2018-02-15	2018-02-15	Yearly	Low	Checked mattress, actuators, motors	
38	22	Inspection, PM	Completed	2019-02-01	PM	2019-02-15	2019-02-15	Yearly	Low	Checked mattress, actuators, motors, lubricated motors	
39	22	Inspection, PM	Completed	2020-02-04	PM	2020-02-15	2020-02-15	Yearly	Low	Checked mattress, actuators, motors, lubricated motors	

Order_num ber	Part_used
4	Li battary
3	ROTOR BELT