

# **VR-406: Haptic Synchronization for Robotic Teleoperation**



CDE4301 Interim Report

AY 2025/26

**Submitted by**

Matthew Kurniawan, A0264690X

Lu BingYuan, A0259353X

## **Supervisor**

Dr. Cai Shaoyu

## **Acknowledgements**

We would like to express our sincere gratitude to Dr. Liu Zhenbang, Consultant at the Department of Urology at Tan Tock Seng Hospital (TTSH), for his invaluable insights, clinical expertise, and willingness to engage with us during the site visit and interviews. His perspectives on the challenges and opportunities in robotic surgery training have been instrumental in shaping this research direction.

We also acknowledge Dr Cai ShaoYu, our supervisor, for his invaluable input and guidance in this project. Dr Cai has gone the extra mile of using personal resources and connections to allow us to get extra help in our project. We also thank prof Khoo Eng Tat for his valuable input and expertise in this topic allowing us to scope our project further as we move forward to prototyping.

# Table of Contents

Acknowledgements	2
1. Introduction	5
1.1 The Rise of Robotic Surgery	5
1.2 Clinical Benefits of Minimally Invasive Surgery	5
1.3 The Emerging Training Crisis	6
2. Problem Clarification	6
2.1 Tan Tock Seng Hospital Context	6
2.2 Observed Complexities in Robotic Surgery	7
2.3 Current Training Methods and Their Limitations	9
2.4 The Pre-Training Gap at TTSH	9
2.5 The Motor-Cognitive Skills Gap	10
2.6 Problem Statement	12
3. Solution Exploration	14
3.1 Specific Skills Assessment	14
3.1.1 Motor Skills Assessment Framework	14
3.1.2 Cognitive Skills Assessment Framework	15
3.2 Literature Review	18
3.2.1 Haptic Feedback	18
3.2.2 Collaborative Training	22
3.3.3 Differentiating Features	25
3.3 Value Proposition	25
4. Methods	27
4.1 System Design	27
4.1.1 Scope Limitations	27
4.1.2 Design Requirements	27
4.1.2 System Specifications	27
4.1.3 System and Software Architecture	28
4.2 Tasks	28
4.3 Metrics	29
4.4 Study and Data collection	29
4.5 Result Analysis	30
4.6 Validation Plan	30
On-Premise Validation:	31
Future Enhancements (Off-Premise):	31
5. Prototyping	31
5.1 Prototyping phases	31
5.2 Virtual Task Implementation	31
5.3 Current Prototypes	32

6. Limitations and Future Work	33
6.1 Current Limitations	33
6.2 Future Enhancements	33
References	34
Appendices	34

## **1. Introduction**

### **1.1 The Rise of Robotic Surgery**

Robotic-assisted surgery has become the standard of care for many procedures, fundamentally transforming surgical practice over the past two decades. The global robotic surgery market

demonstrates explosive growth: from USD 11.83 billion in 2024, the market is projected to reach USD 54.66 billion by 2034, representing a compound annual growth rate (CAGR) of 16.54% (Towards Healthcare, 2024). This extraordinary market expansion reflects both the clinical validation of minimally invasive approaches and the increasing adoption of robotic platforms across surgical specialties.

In urological surgery specifically, robotic approaches have achieved dominance: 62% of prostatectomies worldwide are now performed robotically, with laparoscopic procedures accounting for 37% and open surgery representing only 1% of cases (Williams et al., 2025).

## 1.2 Clinical Benefits of Minimally Invasive Surgery

Minimally invasive surgery offers substantial and well-documented clinical advantages over traditional open surgery across multiple important patient outcomes:

- **Surgical Site Infections:** 0.8–1.5% (MIS) vs. 3.4–9.33% (Open) — representing approximately 70% reduction (Nahas et al., 2024; Shi, 2023; Aeschbacher et al., 2021)
- **Blood Loss:** 120–250 mL (MIS) vs. 400–719 mL (Open) — approximately 60% reduction (Nahas et al., 2024)
- **Hospital Stay:** 2.1–6.7 days (MIS) vs. 4.4–13.2 days (Open) — approximately halved length of hospitalization (Nahas et al., 2024)

These substantial risk reductions drive increasing demand for minimally invasive procedures. Beyond reducing morbidity, minimally invasive approaches also offer significant cosmetic and functional benefits such as reduction of postoperative pain, faster recovery, and improved patient satisfaction (Gurung et al., 2022; Delongchamps et al., 2013).

## 1.3 The Emerging Training Crisis

With market growth at 16.54% CAGR and robotic approaches now accounting for 62% of urological procedures globally, there is an urgent need for a proportional increase in trained

robotic surgeons (Williams et al., 2025). However, surgical training infrastructure has not kept pace with this demand. Current training methods are inadequate: they are either expensive and inaccessible, lacking realism, and require dedicated training facilities (Roswell Park, 2025). This creates a critical pre-training gap between fellowship completion and independent practice.



*Figure 1: daVinci current training methods*

## 2. Problem Clarification

### 2.1 Tan Tock Seng Hospital Context

To understand the real-world challenges in robotic surgical training, we conducted a site visit and interviews with Dr. Liu Zhenbang, Consultant in the Department of Urology at Tan Tock Seng Hospital (TTSH) in Singapore. TTSH operates robotic-assisted surgical programs,



**Dr. Liu Zhenbang**

**Designation**

Consultant

**Credentials**

MBBS, MRCS (Edin), FAMS (Urology)

**Languages Spoken**

English, Mandarin, Cantonese

specifically daVinci Surgical Systems, for urological procedures including partial nephrectomy, radical nephrectomy, and radical prostatectomy.



*Figure 2: Dr Liu Zhenbang's profile and our visit group photo*

## 2.2 Observed Complexities in Robotic Surgery

During observation of a robotic-assisted partial nephrectomy procedure, we witnessed the remarkable complexity of da Vinci manipulation. This procedure requires:

1. Extreme time constraints:

Complete nephron-sparing removal (including clamping of the renal artery, precise tumor excision, and renorrhaphy) must be accomplished in approximately 20 minutes before warm ischemia compromises kidney function

2. Complex bimanual coordination:

Simultaneous control of multiple instruments of 3 robotic arms and 1 camera arm

3. Precision and delicacy:

Unnecessary force application risks organ damage, bleeding, or suture breakage; excessive hesitation prolongs ischemia time

Observation also revealed the complex surgical team dynamics: beyond the primary surgeon, the operating theater includes a surgical first assistant, surgical technologist, circulating nurse,

anesthesiologist, and often a resident learning the system. The resident in particular faces the challenge of transitioning from observer to active participant.



Figure 3: Operation room observation




Figure XX: TTSH doctor operating daVinci simulator

## 2.3 Current Training Methods and Their Limitations

Generally, three formal training modalities exist (Diaz et al., 2020) (Sridhar et al., 2017):

Training Type	Definition	Advantages	Limitations
Dry Lab Training	Practice on inanimate models using instruments and console	Low cost; Widely accessible; Safe environment	Totally unrealistic compared to living tissue  No cognitive training; Limited value after basics mastered  Cannot replicate bleeding, tissue response, or anatomical variations
Wet Lab Training	Practice using biological tissue (human cadavers or animal models)	Most realistic tissue properties; Enables bleeding scenarios; Improves early learning curve	High cost per case; Limited accessibility; Ethical concerns; Requires specialized facilities; No active expert guidance
VR Simulation (da Vinci Skills Simulator)	Computer-based virtual reality training	Autonomous use; automated objective feedback; standardized curriculum; scalable; no consumables	Very high initial cost (hundreds of thousands of dollars); Lower fidelity than real surgery; Limited cognitive training

## 2.4 The Pre-Training Gap at TTS defense

However, TTS defense lacks all three formal training methods due to space and budget limitations! The dedicated da Vinci robots cannot be used for training purposes as they are reserved for clinical surgery. Therefore, the current reality for novice surgeons at TTS defense is:

<b>Training Phase</b>	<b>Description</b>
<b>Observation phase</b>	Junior consultants observe surgeries and assist as first surgical assistant
<b>Fellowship accessible) (if</b>	Surgeons obtain external fellowship training, including dry lab and da Vinci simulation (typically 1 year)
<b>Proctorship phase</b>	Surgeons return to TTSH and perform surgeries under expert supervision; minimum <b>30 proctored cases</b> required
<b>Independent practice</b>	Surgeon practices independently after successful completion of the proctorship

However, even for fortunate surgeons like Dr. Liu who completed external fellowship training, he reports that the training was "unrealistic and insufficient." Upon returning to TTSH, the opportunity constraints are severe: proctorship sessions occur only approximately once weekly or less frequently, depending on patient case availability and surgeon scheduling.

## 2.5 The Motor-Cognitive Skills Gap

The transition from observation to proctorship represents an enormous responsibility jump. Dr. Liu describes the gap clearly:

<b>Level</b>	<b>Cognitive Characteristics</b>	<b>Motor Characteristics</b>

<b>Novice doctors</b>	Delayed in locating surgical landmarks (e.g., identifying the renal artery)	Slow, small, jittery movements with frequent hesitation and correction cycles
<b>Expert surgeons</b>	Immediate identification of anatomical landmarks	Swift, precise movements executed confidently and efficiently

Research confirms the critical importance of both motor and cognitive skills:

1. Motor skills: Surgeon technical skill explains 25.8–27.5% of variation in patient complication rates (Stulberg et al., 2020)

Figure. Association Between Surgeon Technical Skill Score and Risk-Adjusted Postoperative Colectomy Outcomes




Figure 4: Inverse relationship between surgeon technical skills with patient outcome

2. Cognitive skills: Higher nontechnical skills scores (NOTSS) independently predict decreased complications by 5.1% and decreased mortality by 1.1% (Abahuje et al.,

**TABLE 5.** Robust Regression Nonadjusted and Adjusted

	Unadjusted		Adjusted	
	*Coef (95%CI)	p-value	Coef (95%CI)	p-value
Any complication	-1.7 (-5.2; 1.7)	0.29	-5.1 (-8.1; -2.0)	0.003
Mortality	-0.8 (-1.6; -0.01)	0.04	-1.1 (-1.8; -0.2)	0.01
Readmission	0.4 (-0.4; 1.3)	0.35	0.4 (-0.5; 1.4)	0.33
Death or serious morbidity	-1.2 (-3.0; 0.5)	0.14	-1.3 (-2.8; 0.1)	0.06
Return to the operating room	-1.0 (-1.8; -0.2)	0.01	-1.1 (-1.9; -0.4)	0.005
Surgical site infection	-0.2 (-0.9; 0.4)	0.45	0.1 (-0.5; 0.7)	0.73
Skills-related composite	-4.6 (-7.4; -1.8)	0.002	-5.3 (-8.3; -2.4)	0.001
Skills-unrelated composite	-5.3 (-7.6; -3.1)	< 0.001	-5.6 (-7.9; -3.3)	< 0.001

2024)

*Figure 5: Correlation coefficient of non-technical skills with patient outcome*

## 2.6 Problem Statement

Despite the exponential growth in robotic surgery demand (16.54% CAGR) and the critical clinical benefits of minimally invasive approaches, a substantial gap exists in surgical training infrastructure. Some of the gaps are

1. Absent motor skill training: TTS defense does not have any training module that allows for motor skills training.
2. Inadequate cognitive skill training: Current training methods in TTS defense allow surgeon-in-training to participate and observe the surgery, but without proper practice and assessment, it is likely to be inadequate.
3. Existing opportunity barriers: The only practice TTS defense doctors receive are from fellowship programs, which depend on the Ministry of Health (MOH)'s sponsorship. As such, not every doctor would get to train on robotic surgical systems skills.
4. Existing pre-training gap: Between fellowship/simulator training and proctorship, novice surgeons lack the integrated motor-cognitive expertise needed for safe, efficient practice.

5. Prolonged learning curves: Trainees require 30+ proctored patient cases before achieving independent practice, extending training timelines and limiting patient access to care

### 3. Solution Exploration

#### 3.1 Specific Skills Assessment

##### 3.1.1 Motor Skills Assessment Framework

Global operative performance questions
Respect for tissue
Time and motion
Instrument handling
Flow of operation
Exposure
Tissue planes
Completeness of dissection
Overall technical skill

Figure XX: Global Operative Performance Metric (Stulberg et al., 2020)

This framework, employed by Stulberg et al. (2020), and the JIGSAWS dataset (2014) study linking technical skills to patient outcomes, focuses on observable motor behaviors including:

For our scope, we specifically focus on:

1. Motion:
  - a. Path efficiency
  - b. Movement smoothness
  - c. Economy of movement
2. Time:
  - a. Task completion duration
  - b. Decision latency
  - c. Idle periods

These metrics are directly measurable through haptic interface sensors and provide objective, quantitative assessment of motor proficiency development.

### 3.1.2 Cognitive Skills Assessment Framework

Hospital .....		Trainer name .....	Date .....	
		Trainee name .....	Operation .....	
Category	Category rating*	Element	Element rating*	Feedback on performance and debriefing notes
Situation Awareness		Gathering information		
		Understanding information		
		Projecting and anticipating future state		
Decision Making		Considering options		
		Selecting and communicating option		
		Implementing and reviewing decisions		
Communication and Teamwork		Exchanging information		
		Establishing a shared understanding		
		Co-ordinating team activities		
Leadership		Setting and maintaining standards		
		Supporting others		
		Coping with pressure		

\* 1 Poor; 2 Marginal; 3 Acceptable; 4 Good; N/A Not Applicable

- |              |  |
|--------------|--|
| 1 Poor       | Performance endangered or potentially endangered patient safety, serious remediation is required                             |
| 2 Marginal   | Performance indicated cause for concern, considerable improvement is needed  |
| 3 Acceptable | Performance was of a satisfactory standard but could be improved   |
| 4 Good       | Performance was of a consistently high standard, enhancing patient safety; it could be used as a positive example for others |
| N/A          | Not Applicable   |

Figure XX: Nontechnical Skills for Surgeons (NOTSS) (Abahuje et al., 2024)

The NOTSS framework, validated in the prospective study demonstrating direct links between cognitive skills and patient mortality, encompasses multiple cognitive domains. For our training system, we focus on three critical components:

1. Gathering Information
  - a. Visual scanning patterns to identify anatomical landmarks
  - b. Recognition of tissue characteristics and pathology
  - c. Assessment of instrument positioning

2. Understanding Information
  - a. Interpretation of visual cues
  - b. Situational awareness
  - c. Recognition of normal vs. abnormal states
3. Projecting/Anticipating Future States
  - a. Predicting tissue behavior under manipulation
  - b. Planning next movements based on current surgical field state
  - c. Forward-thinking to maintain procedural flow and efficiency

### 3.1.3 Skills Validation

Dr. Liu's clinical example illustrates the need for these skills: when the renal artery is located posterior to the kidney:

Level	Cognitive Characteristics	Motor Characteristics
<b>Expert surgeon</b>	<b>Immediately</b> identifies the artery location based on anatomical knowledge and visual cues	Executes <b>one swift, precise</b> movement to grasp the artery without injury
<b>Novice surgeon</b>	<b>Delayed</b> in locating the artery; uncertain about depth and position	Performs <b>many slow, small, jittery</b> movements with multiple correction attempts

This real-world scenario demonstrates that cognitive knowledge (knowing where the artery is) must be coupled with motor skill (executing the precise grasp) to achieve expert performance.



Training one without the other leaves a critical competency gap.

Figure XX: Dr Liu validating the need for motor and cognitive skills in renal artery extraction

## 3.2 Literature Review

### 3.2.1 Haptic Feedback

#### 3.2.1.1 Evidence for Haptic Feedback Effectiveness

Having identified motor skills as a critical training target, particularly proprioception movement, we systematically reviewed evidence on haptic feedback effectiveness in surgical training.

##### 1. Meta-Analysis Evidence (Bergholz et al., 2023)

A comprehensive meta-analysis examining 51 studies found that 67% of studies (34/51) demonstrated statistically significant performance improvements with haptic feedback across three key metrics: task completion time, accuracy of movement, applied force control

##### 2. Force Reduction Benefits (Azher et al., 2024)

A systematic review specifically examining force-related outcomes revealed dramatic benefits:

- 83% reduction in average applied forces with haptic feedback
- 69% reduction in peak forces with haptic feedback

This force reduction is critically important in robotic surgery where excessive force risks suture thread breakage, may cause organ damage or bleeding, and visual force estimation alone is insufficient for novices

##### 3. Dr. Liu's Clinical Validation

When asked specifically about haptic feedback for training, Dr. Liu affirmed: "Especially motor skills can be taught more effectively through using haptic feedback. Although there are variations of manipulation among surgeons, it would be beneficial for the novice surgeons to learn from the experts instead of figuring the movement on their own."

This clinical endorsement from an experienced robotic urologist validates the relevance of haptic training for real-world surgical practice.

### 3.2.1.2 Dual-User Haptic Feedback Study (Zhang et al., 2023)

#### Study Design and Methodology:

Zhang et al. (2023) developed a novel dual-user haptic training system specifically for laparoscopic surgical training. The study investigated whether haptic feedback would have differential impacts based on task characteristics—specifically hypothesizing that “haptic feedback would facilitate orientation tasks more than cutting tasks”.


“On-Demand” haptic feedback mode, in which experts intervened when trainees made errors or struggled, providing corrective forces without verbal communication.

#### Results:


Metric / Task	Finding
<b>Overall time</b>	No statistically significant difference
<b>Grasping time (orientation task)</b>	Haptic group showed significantly reduced grasping time compared to control group
<b>Cutting time (executive task)</b>	No statistically significant difference

#### Key Finding:


“Haptic feedback has a more substantial impact on orientation tasks than on cutting tasks in laparoscopic surgery training.”



*Figure XX: Laparoscopy haptic experiment setup*



*Figure XX: Laparoscopy haptic experiment task*



*Figure XX: Laparoscopy haptic experiment result*

### 3.2.1.3 Interpretation and Limitations

Haptic feedback benefits orientation tasks more because:

- Fast, feedforward tasks (cutting) rely on pre-planned motor commands, while
- Slow, feedback-dependent tasks (grasping, positioning) continuously incorporate sensory information to adjust trajectory in real-time

Therefore, haptic guidance can directly influence feedback-controlled movements (orientation, positioning, grasping), effectively "teaching" optimal motor patterns through physical guidance.

### Study Limitations:

1. Simple Task: The circle-cutting task was simple and does not represent actual surgical procedures. Similar learning rates between groups may be due to task simplicity; more complex tasks might show greater haptic benefits
2. Lack of Cognitive Training: Focus solely on motor skill transfer. Does not address cognitive decision-making component identified as critical


### 3.2.2 Collaborative Training

#### 3.2.2.1 Experimental motivation and setup

Ganesh et al. (2014): "Two is Better Than One" investigated the fundamental motor responses and adaptations that govern physical interactions between humans. They developed a novel interactive learning paradigm that differs critically from previous joint action studies. Pairs of individuals were physically connected during a motor task without conscious knowledge of the connection. This eliminated cognitive complications from conscious coordination, which enable investigation of pure reactive motor adaptations driven solely by haptic signals

#### Critical Design Feature:

The compliant connection allowed independent movement while transmitting partner forces. Subjects had to actively track the target, they could not simply relax and follow their partner. Subjects were not informed about the nature of forces and were not consciously aware of the connection with their partner (verified by post-experiment questionnaire).



*Figure XX: Collaborative haptic experiment setup and task*


### 3.2.2.2 Key Findings: Mutual Benefits Regardless of Relative Performance

#### Finding 1: Connection with a Partner Improves Performance

The central discovery: "Connection with a partner improved the task performance in an individual irrespective of whether the partner's performance was better or worse than the individual's own performance." (Ganesh et al., 2014)


- When connected to a better-performing partner (positive x-axis): Individual performance improved
- When connected to a worse-performing partner (negative x-axis): Individual performance still improved

This counterintuitive finding demonstrates that physical interaction is fundamentally beneficial,



not merely a transfer from expert to novice.

*Figure XX: Improvement of performance through collaborative training*



*Figure XX: Improvement based on modes of interaction*

### 3.2.2.3 Critical Insights

- Physical connection is inherently beneficial regardless of skill difference. Reactive motors adapt driven by haptic feedback.

- Mutual adaptation outperforms passive or pre-recorded guidance. Non-reactive systems cannot reproduce these learning effects, even when the expert partner makes small errors
- The physical connection enables novices to feel expert adjustments and experts to feel novice deviations, creating a feedback loop that produces lasting motor learning, not temporary assistance.

### 3.3.3 Differentiating Features

Feature	Zhang et al. (2023)	Ganesh et al. (2014)	Our System
Physical Connection	Yes	Yes	Yes
Surgical Context	Yes	No	<b>Yes (robotic surgery)</b>
Haptic Feedback Type	On-demand correction	Continuous elastic link	Variable-stiffness collaborative
Cognitive Training	None	None	<b>Integrated (NOTSS-based)</b>
Task Complexity	Simple cutting (circle)	Simple (tracking)	<b>Meaningful surgical subtasks</b>
Two-Way Interaction	Limited	Full mutual adaptation	Full mutual adaptation

## 3.3 Value Proposition

FOR: Robotic surgery novice doctors

WHO: Require pre-training before proctorship

OUR: Collaborative Haptic Teaching System

IS: A guided practice platform

THAT: Enables expert surgeons to actively guide novice hand movements to transfer cognitive and motor skills, bridging observation with real surgery

## 4. Methods

### 4.1 System Design

#### 4.1.1 Scope Limitations

For this project, we do not seek to replace the existing simulation training systems (Bernie and Sanchez, 2016), but aim to validate that haptic synchronisation is an effective medium of skill transfer in similar adjacent training systems. With relative budget and time constraints, a scoped down setup that can still validate the above hypothesis is needed.

This involves reducing the degrees of freedom in motion and the number of controllers operated by a single user.

#### 4.1.2 Design Requirements

To enable two-user synchronised action, a minimum of two handheld controllers is required. Each controller must be able to perform the 3 following functions.

- a) Operate a virtual arm. This is the tool that the user pilots remotely, similar to how a simulator performs.
- b) Record movements precisely. The controller must communicate the movements of the user in real time to the other synchronised controller.
- c) Provide haptic feedback. The controller must exert different forces on the user's hand based on the movement data it is receiving from the other paired controller.

Aside from the controllers, the two users are collaborating on a single surgical task, hence they must share the same visual response on a single screen.

#### 4.1.2 System Specifications

To fulfill the earlier design specifications, our haptic controller of choice is the 3D Systems Touch which allows users to feel and manipulate virtual objects. For our experimental setup, we will use two devices, one for the expert and one for the novice, connected to a central computer running the simulation environment. The workspace of the Phantom Omni (3D Systems Inc.) is approximately 160W x 120H x 70D mm, and the device can exert a maximum force of 3.3 N.

The Phantom Omni provides 6 degrees of freedom (DOF) for position and orientation tracking, simplified to a 4+1 DOF model to reduce complexity. This includes 3 translational and 1 rotational DOF for movement, plus a gripper (end effector) control. While this is a reduction from the 7 DOF available in the da Vinci surgical system, it is sufficient to operate the core tasks of our experiment. To provide a more realistic feel of a surgical instrument, a custom gripper will be designed and 3D printed to be attached to the stylus of the Phantom Omni.

#### 4.1.3 System and Software Architecture

The system architecture is designed to support real-time collaborative haptic interaction between the expert and the novice. The software will be developed using the Unity engine for its strong support for 3D graphics, physics simulation, and compatibility with the OpenHaptics SDK. The architecture is modular and consists of several key components:

##### **Simulation Environment:**

Developed in Unity, the environment will contain the virtual surgical task, including a deformable tissue model created using a mesh-based approach, a needle model, and a thread model. Collision detection will be implemented to handle the interactions between the virtual instrument, the needle, the thread, and the tissue.

##### **Haptic Rendering and Collaborative Control:**

This module is responsible for calculating and rendering the forces that the users feel. It takes the position and orientation data from the haptic devices and computes the appropriate force feedback. The force feedback is proportional to the difference in movement between the expert and the novice ( $\text{feedback} \propto \text{expert} - \text{novice movement}$ ) to create a closed loop. This will be implemented using a proportional-derivative (PD) controller to generate the guidance force, creating a virtual spring-damper system that gently guides the novice's hand towards the expert's path.

### 4.2 Tasks

The surgical tasks for the experiment are designed to be representative of basic surgical skills and to have varying levels of difficulty. This allows us to evaluate the effectiveness of our

collaborative haptic training system across a range of complexities. The tasks are broken down into four main actions:

1. Loading needle: pick up the needle with the instrument.
2. Puncturing tissue: puncture the virtual tissue with the needle at a specific point.
3. Driving needle through tissue: guide the needle through the tissue along a curved path.
4. Knot tying: tie a surgical knot.

These tasks were jointly chosen with Dr Liu because they are fundamental to many urology surgical procedures and they allow for objective measurement of performance. To establish a benchmark for performance, we will record the movements of expert surgeons performing these tasks as a standard against which the performance of the novices can be compared.

### 4.3 Metrics

To evaluate the effectiveness of our collaborative haptic training system, we will use a combination of general and specific performance metrics. These metrics will be used to compare the performance of the study group (collaborative training) with the control group (guided simulation).

General Metrics:

1. Time to completion: The total time taken to complete each task.
2. Wasted movements: The total path length of the instrument, with a focus on movements that are not productive towards completing the task.
3. Shaking / Trembling: The amount of high-frequency movement of the instrument, which is indicative of a lack of control.

Specific Metrics:

1. Suture interval: The distance between sutures in the suturing task.
2. Puncture accuracy: The accuracy of the needle puncture in the puncturing task.

### 4.4 Study and Data collection

The experiment will be designed as a between-subjects study, with participants randomly assigned to either the control group or the study group. The participants will be novice users

with no prior experience in robotic surgery. The training protocol will consist of pre-test, iterative training and post-test. In the pre-test, participants will perform the tasks independently once to establish a baseline. In the repeated training session, the control group will receive guided simulations, while the study group will engage in collaborative training with an expert. They will finally be evaluated on their individual performance in the post-test.

## 4.5 Result Analysis

The collected data will be analyzed to compare the performance of the two groups. We will use statistical tests, such as t-tests to determine if there are any significant differences between the groups. We will also analyze the learning curves of the participants to see how their performance improves over time.

Our primary hypothesis is that synchronized collaborative training will lead to a steeper learning curve, meaning that the study group will show a faster rate of improvement in these metrics. This is illustrated in the speculated graph of time to completion versus repetitions from our presentation, where the "Synchronised" group's performance improves more rapidly than the "Baseline" group.




Figure XX: Hypothesised difference in training performance over iterations

## 4.6 Validation Plan

The validation of our system will be conducted in two phases: an on-premise experiment and a plan for future off-premise enhancements.

## On-Premise Validation:

Within our experimental setup, we will conduct a study with two groups: a control group and a study group. The control group will receive training through guided simulations, where they will follow a pre-recorded path of an expert. The study group will engage in live collaborative training with an expert. Both groups will then perform the tasks independently, and their performance will be measured using the metrics defined in the previous section. This will be repeated over a number of trials to measure the learning rate.

## Future Enhancements (Off-Premise):

Beyond our immediate experimental setup, possible future enhancements could provide more extensive validation of our system. This would involve a qualitative assessment of how easily novice surgeons who have used our system pick up Da Vinci surgical skills during their fellowship. A quantitative assessment could involve tracking the number of surgeries they need to perform to become qualified.

## 5. Prototyping

### 5.1 Prototyping phases

The development of the prototype will be divided into two main phases. The first phase will focus on creating a baseline system for independent practice while the second phase will focus on implementing the collaborative features, including the networking module and the two-way force feedback.

### 5.2 Virtual Task Implementation

The virtual tasks will be implemented as a series of challenges with increasing difficulty. We will start with the simplest task, loading the needle, and progressively add the more complex tasks. The appropriate metrics for each task will be determined and prepared for collection to evaluate user performance.

### 5.3 Current Prototypes

A python script was created to read out controller movement from existing robotic surgery dataset JIGSAW to appreciate the motions in a virtual space.




Figure XX: Visualisation of controller movement from JIGSAW

Preliminary object building was also prototyped on the OpenHaptics software, to create a simple membrane with variable tension. With implementation of custom physics, the property of such a membrane can be altered to simulate a human tissue.




Figure XX: Simple membrane built on OpenHaptics SDK

## **6. Limitations and Future Work**

### **6.1 Current Limitations**

Our current prototype has several limitations that should be addressed in future work. The reduced degrees of freedom of the Phantom Omni compared to the da Vinci system is a significant limitation. The single-handed operation is also a simplification of the real surgical environment, where surgeons typically use both hands. The tasks are also simplified and do not fully capture the complexity of real surgical procedures. The virtual environment, while providing a good approximation, cannot fully replicate the properties of real tissue. Finally, the small sample size of our planned experiment will limit the generalizability of our findings.

### **6.2 Future Enhancements**

There are several avenues for future enhancement of our system. The most obvious next step is to integrate our system with a more realistic surgical simulator, such as the da Vinci Simulator. This would allow us to validate our findings in a more realistic environment. We also plan to extend our system to support dual-handed collaborative operation. This would require the use of four haptic devices, two for the expert and two for the novice. We also plan to develop more complex and realistic surgical tasks. In the long term, we envision our system being used for longitudinal studies to track the skill development of surgeons throughout their fellowship training.

## References

- 3D Systems, Inc. (n.d.). *Touch haptic devices*. 3D Systems.  
<https://www.3dsystems.com/haptics-devices/touch>
- Abahuje, E., Cong, L., Iroz, C. B., Barsuk, J. H., Stey, A., Likosky, D. S., Lagu, T., Johnson, J. K., & Halverson, A. (2024). A Prospective Study to Assess the Relationship Between Nontechnical Skills for Surgeons (NOTSS) and Patient Outcomes. *Journal of Surgical Education*, 81(11), 1568–1576. <https://doi.org/10.1016/j.jsurg.2024.07.022>
- Aeschbacher, P., Nguyen, T.-L., Dorn, P., Kocher, G. J., & Lutz, J. A. (2021). Surgical Site Infections Are Associated With Higher Blood Loss and Open Access in General Thoracic Practice. *Frontiers in Surgery*, 8. <https://doi.org/10.3389/fsurg.2021.656249>
- Aggravi, M., Estima, D. A. L., Krupa, A., Misra, S., & Pacchierotti, C. (2021). Haptic Teleoperation of Flexible Needles Combining 3D Ultrasound Guidance and Needle Tip Force Feedback. *IEEE Robotics and Automation Letters*, 6(3), 4859–4866.  
<https://doi.org/10.1109/Lra.2021.3068635>
- Bergholz, M., Ferle, M., & Weber, B. M. (2023). The benefits of haptic feedback in robot assisted surgery and their moderators: a meta-analysis. *Scientific Reports*, 13(1), 19215.  
<https://doi.org/10.1038/s41598-023-46641-8>
- Bernier, G. V., & Sanchez, J. E. (2016, May 10). *Surgical Simulation: The Value of Individualization - A SAGES Technology and Value Assessment*. Society of American Gastrointestinal and Endoscopic Surgeons.
- <https://www.sages.org/publications/tavac/surgical-simulation-value-individualization/>
- Cancer (IARC), T. I. A. for R. on. (n.d.). *Global Cancer Observatory*. Gco.iarc.fr.  
<https://gco.iarc.who.int/media/globocan/factsheets/cancers/27-prostate-fact-sheet>
- Diaz, A., Gidi, G., Faes-Petersen, R., Giordano, F., & Villegas Tovar, E. (2020). The road to becoming a certified robotic surgeon. *World Journal of Advanced Research and*

*Reviews*, 2020(01), 2581–9615.

<https://wjarr.com/sites/default/files/WJARR-2020-0253.pdf>

Ganesh, G., Takagi, A., Osu, R., Yoshioka, T., Kawato, M., & Burdet, E. (2014). Two is better than one: Physical interactions improve motor performance in humans. *Scientific Reports*, 4. <https://doi.org/10.1038/srep03824>

Nahas, W. C., Rodrigues, G. J., Rodrigues Gonçalves, F. A., Sawczyn, G. V., Barros, G. G., Cardili, L., Guglielmetti, G. B., Fazoli, A. J. D. C., Cordeiro, M. D., Cassão, V. D. A., Chade, D. C., Neves De Oliveira, L. C., Murta, C. B., Pontes Júnior, J., Trindade, E. M., Bastos, D. A., Sarkis, A. S., Mitre, A. I., Trinh, Q.-D., & Coelho, R. F. (2024). Perioperative, Oncological, and Functional Outcomes Between Robot-Assisted Laparoscopic Prostatectomy and Open Radical Retropubic Prostatectomy: A Randomized Clinical Trial. *Journal of Urology*, 212(1), 32–40.

<https://doi.org/10.1097/ju.0000000000003967>

Roswell Park. (2025). *Fundamental Skills of Robot-Assisted Surgery (FSRS)*. Roswell Park Comprehensive Cancer Center.

<https://www.roswellpark.org/education/professional-training/atlas-program/testing-training/fundamental-skills-robot-assisted>

Sayed Azher, Mills, A., He, J., Taliah Hyjazie, Tokuno, J., Quaiattini, A., & Harley, J. M. (2024). Findings Favor Haptics Feedback in Virtual Simulation Surgical Education: An Updated Systematic and Scoping Review. *Surgical Innovation*, 31(3), 331–341.

<https://doi.org/10.1177/15533506241238263>

Shi, Z. (2023). Laparoscopic vs. open surgery: A comparative analysis of wound infection rates and recovery outcomes. *International Wound Journal*, 21(3).

<https://doi.org/10.1111/iwj.14474>

- Song, Y., Guo, S., Zhang, L., & Yu, M. (2017). Haptic feedback in robot-assisted endovascular catheterization. *IEEE International Conference on Mechatronics and Automation (ICMA)*.  
<https://doi.org/10.1109/icma.2017.8015851>
- Sridhar, A. N., Briggs, T. P., Kelly, J. D., & Nathan, S. (2017). Training in Robotic Surgery—an Overview. *Current Urology Reports*, 18(8). <https://doi.org/10.1007/s11934-017-0710-y>
- Stulberg, J. J., Huang, R., Kreutzer, L., Ban, K., Champagne, B. J., Steele, S. R., Johnson, J. K., Holl, J. L., Greenberg, C. C., & Bilimoria, K. Y. (2020). Association Between Surgeon Technical Skills and Patient Outcomes. *JAMA Surgery*, 155(10), 960.  
<https://doi.org/10.1001/jamasurg.2020.3007>
- Towards Healthcare. (2024). *Robotic Surgery Market Size Booms by 16.54% CAGR till 2034*.  
Towardshealthcare.com; Towards Healthcare.  
<https://www.towardshealthcare.com/insights/robotic-surgery-market-sizing>
- Williams, Z., Saad, J., Ayeni, F. E., Wang, H., Zhong, W., Gendy, R., Arianayagam, M., Canagasingham, B., Goolam, A., Jeffery, N., Kam, J., Khadra, M., Ko, R., Mehan, N., Varol, C., & Thangasamy, I. (2025). Comparing outcomes of robotic-assisted radical prostatectomy by specialists and trainees using a modular training approach. *Journal of Robotic Surgery*, 19(1). <https://doi.org/10.1007/s11701-025-02277-6>
- Zhang, Y., Wang, O., Wang, Y., Tavakoli, M., & Zheng, B. (2023). Enhancing skill learning with dual-user haptic feedback: insights from a task-specific approach. *Frontiers in Robotics and AI*, 10. <https://doi.org/10.3389/frobt.2023.1286282>