


SMART INDIA HACKATHON 2026



TITLE PAGE

- **Problem Statement ID** - 26181
- **Problem Statement Title** - AI-Powered Personal Health Companion & Edge Disaster Monitor
- **Theme** - MedTech
- **PS Category** - Hardware
- **Team ID** - 01
- **Team Name (Registered on portal)** - CHIPSTERS



⚡ 20 ns FPGA Timing | 🧠 91% INT8 TinyML | 📄 92% MIMIC-III Validated

❖ Proposed Solution (Describe your Idea/Solution/Prototype)

✗ How Current Technology Fails During Disasters

- **Wearables Are Environmentally Blind:** Commercial smartwatches track fitness steps and resting heart rate, but have zero awareness if ambient air is toxic (AQI 500+) or if 48°C heat is causing heat exhaustion.
- **Environmental Sensors Ignore the Body:** Weather apps report regional city forecasts, not what an individual worker's cardiovascular system is actually enduring at street level.
- **The Cloud Dependency Trap:** Cloud AI health platforms require mobile networks. During floods, cyclones, or power grid failures, cellular towers collapse and cloud health apps die completely.
- **Software Jitter Destroys HRV:** Operating system scheduling delays on standard CPUs jitter by 5-20 ms, severely corrupting heart rate variability (HRV) metrics.

✓ EdgeGuard Proposed Solution & Innovation

- **Detailed Solution Explanation:** Simultaneously acquires dual-wavelength optical PPG pulses along with hyper-local ambient temperature, relative humidity, and laser PM2.5 particulate matter on the wrist.
- **How It Addresses the Problem:** Fuses bodily vitals with environmental stress in real time to calculate cardiac drift and respiratory strain, surfacing cardiac strain and respiratory stress as it develops.
- **Hardware Innovation (FPGA Silicon):** Custom Verilog 8-tap filter runs on 50 MHz hardware with 20 ns cycle resolution, completely eliminating operating system jitter without DSP multipliers or BRAM.
- **Uniqueness & Zero Cloud Dependency:** Quantized INT8 neural network fits into only 619 bytes of INT8 weights, providing 100% offline protection during infrastructure blackouts.

💡 **Layman Core:** An offline medical guardian that connects what's happening outside (disaster) with what's happening inside (your heart) — flagging risk before a person would notice symptoms.

❖ Technologies Used & Implementation Methodology (4-Stage Autonomous Dataflow)

Stage 01: Sensing
Physical Acquisition

- **MAX30102 PPG**
Dual Red & IR optical pulse oximeter for arterial blood flow & SpO₂.
- **BME280 Climate**
Ambient temperature & relative humidity for Steadman Heat Index.
- **PMSA003 PM2.5**
Laser scattering sensor measuring hazardous fine particulate smog.

Stage 02: FPGA RTL
50 MHz Custom Silicon


- **Dual 8-Tap Filter**
O(1) running-sum noise filter with bit-shift (>> 3). 0 DSP, 0 BRAM.
- **Systolic Peak FSM**
4-state finite state machine with 250 ms refractory lock-out window.
- **20 ns Cycle Timer**
Continuous 50 MHz cycle counter latching IBI with zero software jitter.

Stage 03: Clinical AI
On-Device TinyML

- **mNEWS2 Engine**
Royal College of Physicians clinical early warning triage protocol.
- **Karlen/Elgendi SQI**
Signal Quality Index rejecting motion noise and false beats.
- **6→24→16→3 Model**
619-parameter INT8 micro-network (88.47% synthetic validation accuracy).

Stage 04: Edge Action
Alerts & Telemetry

- **Local OLED Display**
128x64 display of live HR, SpO₂, RR, and multi-hazard risk tiers.
- **Haptic Alarm**
Vibration motor & emergency buzzer for immediate worker evacuation.
- **WiFi / MQTT Mesh**
Optional emergency beacon broadcasting vitals to disaster rescue teams.



- **Verified Cycle-Accurate Timing:** GTKWave trace proves O(1) noise filter convergence and systolic peak latching with 20 ns resolution.
- **Zero OS Scheduling Jitter:** Unlike software CPUs with 5-20 ms latency, the 50 MHz FPGA timer gives 100% medically reliable Heart Rate Variability (HRV).
- **Hardware Technologies:** Synthesizable Verilog-2001 (Renesas ForgeFPGA + Xilinx Zynq-7000), C99 bare-metal / FreeRTOS, INT8 TinyML, Win32 GDI GUI.


Feasibility Analysis: Hardware, Silicon & Economics

- **Vendor-Agnostic Verilog RTL:** Identical synthesizable core Verilog runs on both AMD Xilinx (6-input LUT) and Renesas ForgeFPGA (5-input LUT) architectures without redesign.
 - **Ultra-Low Silicon Footprint:** Synthesized on Renesas ForgeFPGA requiring 443 / 1120 LUT5s (39.6% logic utilization) with 0 DSP multipliers and 0 Block RAM.
 - **Static Timing Closure (+5.603 ns WNS):** Baseline Zynq-7000 build, tag v1.0-zynq-SIH: +5.603 ns setup slack, 69.45 MHz Fmax against a 50 MHz target clock.
 - **Economic Viability & Bulk BOM:** • 12 line items: ForgeFPGA, ESP32-S3, 3 sensors, OLED
 - Projected under \$15 at 100k-unit volume (distributor estimates)
 - No cloud, no subscription, no data plan
- => Parts list: hardware/shrikefi/pcb/ BOM

Potential Challenges, Risks & Mitigation Strategies

- **Challenge 1: Motion & Optical Noise:** Mitigation -> Dual 8-tap running-sum digital filter on FPGA plus Karlen/Elgendi Signal Quality Index (SQI) algorithm automatically discards corrupted beats.
- **Challenge 2: Micro-FPGA Resource Limits:** Mitigation -> Pure logic O(1) bit-shift division ($\gg 3$) requires 0 DSP48 multipliers and 0 Block RAM, running on the lowest-cost micro-FPGAs.
- **Challenge 3: Grid & Cellular Tower Outages:** Mitigation -> Fully on-device offline TinyML; 619 bytes of INT8 weight storage. Operates autonomously during catastrophic floods and storms without internet.
- **Challenge 4: Clinical Real-World Accuracy:** Mitigation -> Benchmarked against 16,387 gold-standard patient records from the MIT MIMIC-III ICU database (94.11% triage accuracy on real HR/SpO2 inputs).
- **Commercial Roadmap Target:** Migration path identified to Qualcomm Snapdragon Wear W5+ Gen 1 Low-Power Island (target platform; power not yet measured).

IMPACT AND BENEFITS



20 ns

beat-to-beat timing
resolution (hardware)

94.11%

triage accuracy vs
16,387 ICU records

619 B

offline INT8 network
(6 -> 24 -> 16 -> 3)

12

components in the
entire wearable BOM

SOCIAL

Offline triage during network and grid blackouts.
Health data never leaves the device - no cloud, no phone.


ECONOMIC

12-component commodity BOM; no subscription.
No server, no data plan, no recurring cost.

ENVIRONMENTAL

Zero-cloud: no data-centre or network load.
0 DSP and 0 BRAM in the FPGA fabric - pure logic.

Implements the NHS early-warning score used in ICUs - running on a wrist, with no network.



RESEARCH AND REFERENCES



REFERENCE (paper, dataset or standard)	IMPLEMENTED IN
CLINICAL & PHYSIOLOGICAL FOUNDATIONS (peer-reviewed)	
RCP (2017) NEWS2: National Early Warning Score 2 - NHS acute-illness severity	clinical_vitals_engine.c
Moran DS et al. (1998) A physiological strain index to evaluate heat stress - Am J Physiol	disaster_risk_engine.c
Brook RD et al. / AHA (2010) Particulate matter air pollution and cardiovascular disease - Circulation	disaster_risk_engine.c
Steadman RG (1979) The assessment of sultriness, Parts I & II - J Appl Meteorol	disaster_risk_engine.c
Karlen W et al. (2012) PPG signal quality estimation using repeated Gaussian filters - Physiol Meas	ppg_sqi.c
Elgendi M (2016) Optimal signal quality index for photoplethysmogram signals - Bioengineering	ppg_sqi.c
Charlton PH et al. (2018) Breathing rate estimation from the ECG and PPG - IEEE Rev Biomed Eng	ppg_respiratory_rate.c
Si M et al. (2019) Low-cost particle sensor calibration using machine learning - Atmos Meas Tech	pm25_calibration_int8.c
DATASETS & STANDARDS	
Johnson AEW et al. (2016) MIMIC-III, a freely accessible critical care database - Sci Data	data/mimic/
WHO (2021) Global air quality guidelines: PM2.5 hazard thresholds	disaster_risk_engine.c
ARM AMBA AXI and ACE Protocol Specification (AXI4-Lite)	axi_ppg_accelerator.v

Every reference above maps to a file in this repository - these were implemented, not just cited.

Toolchain (how it was built, not a reference): Vivado ML 2026.1 | Renesas ForgeFPGA Workshop v6.55 | Icarus Verilog | GTKWave | GCC | ESP-IDF v5.5.5