Using deep learning to diagnose knee injuries on magnetic resonance images: current potential and limitations

Nicolai Sandau, MD
Stig Brorson, MD PhD DMSc

Centre for Evidence-Based Orthopedics,
Dept. of Orthopedic Surgery,
Zealand University Hospital

UNIVERSITY OF COPENHAGEN

REGION ZEALAND
ZEALAND UNIVERSITY HOSPITAL

Background

RESEARCH ARTICLE

Deep-learning-assisted diagnosis for knee magnetic resonance imaging: Development and retrospective validation of MRNet

Nicholas Bien[1☯*], Pranav Rajpurkar[1☯], Robyn L. Ball[2], Jeremy Irvin[1], Allison Park[1], Erik Jones[1], Michael Bereket[1], Bhavik N. Patel[3], Kristen W. Yeom[3], Katie Shpanskaya[3], Safwan Halabi[3], Evan Zucker[3], Gary Fanton[4], Derek F. Amanatullah[4], Christopher F. Beaulieu[3], Geoffrey M. Riley[3], Russell J. Stewart[3], Francis G. Blankenberg[3], David B. Larson[3], Ricky H. Jones[3], Curtis P. Langlotz[3], Andrew Y. Ng[1‡], Matthew P. Lungren[3‡]

1 Department of Computer Science, Stanford University, Stanford, California, United States of America, **2** Quantitative Sciences Unit, Department of Medicine, Stanford University, Stanford, California, United States of America, **3** Department of Radiology, Stanford University, Stanford, California, United States of America, **4** Department of Orthopedic Surgery, Stanford University, Stanford, California, United States of America

Methods: Data

- 1250 cases
  - Training: 1130
  - Validation: 120

$$\begin{pmatrix} 1 & 203 & 130 & 26 & 62 & 43 \\ 206 & 51 & 120 & 40 & 18 & 5 \\ 0 & 53 & 231 & 102 & 158 & 191 \\ 31 & 41 & 99 & 156 & 224 & 139 \\ 209 & 79 & 153 & 156 & 168 & 89 \\ 72 & 132 & 83 & 65 & 134 & 230 \end{pmatrix}$$

Methods: Deep learning model



Results: ACL


AUC: 0.97
True Positive Rate (Sensitivity)
False Positive Rate (1- specificity)
0
0.1
0.2
0.3
0.4
0.5
0.6
0.7
0.8
0.9
1
Bien et al.
Model
Radiologists

Results: Meniscus



Results: GradCAM



Limitations

- Lack of generalizability
  - Patient population
  - Different scanners
  - Other pathologies
- Potential solution: More high quality data

Conclusions

- Deep learning methods has the potential to aid radiologists and orthopedic surgeons in the diagnosis of meniscus and ACL injuries on MRI.
- More high quality data is needed to improve generalizability.

Thank you!

Nicolai Sandau, MD
Centre for Evidence-Based Orthopedics,
Dept. of Orthopedic Surgery,
Zealand University Hospital,
Denmark
nicsa@regionsjaelland.dk