

PROPRIETARY NOTICE
THIS DRAWING CONTAINS CONFIDENTIAL INFORMATION PROPRIETARY TO VARIAN. IT MUST NOT BE REPRODUCED OR DISCLOSED TO OTHERS OR USED IN ANY OTHER WAY, IN WHOLE OR PART EXCEPT AS AUTHORIZED IN WRITING BY VARIAN.

REVISIONS				
ZONE	REV	DESCRIPTION		APPROVAL/DATE
	A	RELEASE PER ECO #22443	GJK 8/24/98	JC 8/26/98

D

C

B

A

D

C

B

A

POWER SUPPLY CONTROLLER
ASSY E15004060
SCHEM E13004060

ADDRESS = 16

POWER SUPPLY
ASSY E21000001
(XENTEK XT18-15)

PS1

INPUT

ØA	BLK
ØB	RED
ØC	ORN
RTN	GRN

E16000141

HALL PROBE
AMPLIFIER

HALL PROBE
READBACK



3A95

AC POWER
P1

HALL PROBE



FOR REF ONLY

TB2




J2
ANALOG
I/O

J1
POWER



208
VOLTS

S1 SETTING ON
E15004060



ADDRESS = 16

						DESCRIPTION VENDOR & PART NUMBER		PART NUMBER		DWG SIZE		FIND NO.		REV CHG	
PN SUFFIX & QUANTITY				UI											
PARTS LIST															
WORKMANSHIP PER VARIAN WORKMANSHIP STANDARDS				ALL DIM ARE IN INCHES UNSPECIFIED TOL ARE: .XXX ± .XX ± FRAC ± ANG ± DO NOT SCALE THIS DRAWING RADIUS OR CHAMFER ALL SHARP EDGES .015 MAX ALL MACH. SURFACES 63 RMS CONCENTRICITY .005 TIR PERPENDICULARITY WITHIN TOLERANCE ZONE		APPROVALS		DATE		varian ION IMPLANT SYSTEMS, GLOUCESTER, MA SCHEMATIC DIAGRAM ANALYZER MAGNET P.S. WITH GAUSS FEEDBACK					
						DRAWN		8/24/98							
		CHECKED		8/24/98											
		ELEC ENGR		8/24/98											
		MECH ENGR													
		PROJ ENGR				SIZE D E10000142 REV A									
		MFG ENGR													
		QUAL ENGR													
NEXT ASSY				E220		USED ON				SCALE		NONE		SHEET 1 OF 1	
APPLICATION				FINISH		NONE									