

CXR Batch QA Report — MTV-INT-RAD-003


Batch Statistics

Metric	Value
Total studies	300
Acceptable	181
Borderline	118
Repeat	1
Mean composite score	0.64
Std composite score	0.07

Inter-Rater Agreement (κ)

Axis	κ	Agreement %	Interpretation
sharpness	0.2605	79.3%	fair
exposure	0.5841	75.7%	moderate
rotation	0.3019	84.0%	fair
coverage	0.328	80.0%	fair
inspiration	-0.0494	64.7%	poor
artifact	0.453	74.0%	moderate
metadata	-0.0071	98.3%	poor
global_rating	0.5248	72.0%	moderate
CEILING (global)	0.5248		

Calibration Plot



Failure Mode Summary

See reports/failure_catalogue.md for full analysis.