

Histograms should not be used to segment functional lung MRI

Nicholas J. Tustison¹, Talissa A. Altes², Kun Qing³, James C. Gee⁴, G. Wilson Miller¹, John P. Mugler III¹, Jaime F. Mata¹

¹Department of Radiology and Medical Imaging, University of Virginia, Charlottesville, VA

²Department of Radiology, University of Missouri, Columbia, MO

³Department of Radiation Oncology, City of Hope, Duarte, CA

⁴Department of Radiology, University of Pennsylvania, Philadelphia, PA

Corresponding author:
Nicholas J. Tustison, DSc
Department of Radiology and Medical Imaging
University of Virginia
ntustison@virginia.edu

Abstract

Magnetic resonance imaging using hyperpolarized gases, most notably He-3 and Xe-129, has made possible the novel visualization of airspaces, such as the human lung. The advent and refinement of these imaging capabilities has furthered the development of various avenues of research into the growth, development, and disease associated with the pulmonary system. In conjunction with the improvements associated with image acquisition, multiple image analysis approaches have been proposed and developed for quantifying such images with much research effort devoted to semantic segmentation, or voxelwise classification, into clinically-oriented categories based on ventilation levels. Given the functional nature of these images and the consequent complexity of the segmentation task, many of these algorithmic approaches reduce the complex spatial image intensity information to intensity-only considerations, specifically those associated with the intensity histogram. This results in the loss of important spatial cues for identifying unique imaging features, specifically ventilation defects (as spatial objects) which have been identified as correlating with lung pathophysiology. In this work, we demonstrate the interrelatedness of the most common approaches for ventilation-based segmentation of hyperpolarized gas lung imaging which rely on the intensity histogram for driving voxelwise classification. We also illustrate the underlying assumptions associated with each approach and how these assumptions lead to suboptimal performance. We then illustrate how a deep learning-based solution is constructed in a multi-scale, hierarchically feature-based (i.e., spatial) manner which circumvents the problematic issues associated with existing histogram-based approaches. Importantly, we provide this newly reported deep learning functionality and evaluation framework as open-source through the well-known Advanced Normalization Tools (ANTs) library.

Introduction

Early hyperpolarized gas pulmonary imaging research reported findings in qualitative terms.

Descriptions:

- “ ^3He MRI depicts anatomical structures reliably” (1)
- “hypointense areas” (2)
- “signal intensity inhomogeneities” (2)
- “wedge-shaped areas with less signal intensity” (2)
- “patchy or wedge-shaped defects” (3)
- “ventilation defects” (4)
- “defects were pleural-based, frequently wedge-shaped, and varied in size from tiny to segmental” (4)

Subsequently, initial attempts at quantification were limited to enumerating the number of “ventilation defects” or estimating ventilation defect percentage (as a percentage of total lung volume).

Additional sophistication:

- linear binning
- (semi-automated) k-means
- some percentage of the global mean intensity
- Gaussian mixture modeling with Markov Random Field prior-based smoothing

Results

Discussion

Methods

References

1. Bachert P, Schad LR, Bock M, et al.: Nuclear magnetic resonance imaging of airways in humans with use of hyperpolarized ^3He . *Magn Reson Med* 1996; 36:192–6.
2. Kauczor HU, Hofmann D, Kreitner KF, et al.: Normal and abnormal pulmonary ventilation: Visualization at hyperpolarized he-3 mr imaging. *Radiology* 1996; 201:564–8.
3. Kauczor HU, Ebert M, Kreitner KF, et al.: Imaging of the lungs using ^3He mri: Preliminary clinical experience in 18 patients with and without lung disease. *J Magn Reson Imaging*; 7:538–43.
4. Altes TA, Powers PL, Knight-Scott J, et al.: Hyperpolarized ^3He MR lung ventilation imaging in asthmatics: Preliminary findings. *J Magn Reson Imaging* 2001; 13:378–84.