

## Oscilloscope Scanning of Visible Spectra

W. E. Deal, William Bradshaw, and F. A. Matsen

Citation: *The Journal of Chemical Physics* **16**, 638 (1948); doi: 10.1063/1.1746966

View online: <http://dx.doi.org/10.1063/1.1746966>

View Table of Contents: <http://scitation.aip.org/content/aip/journal/jcp/16/6?ver=pdfcov>

Published by the AIP Publishing

---

### Articles you may be interested in

[Wiener noise power spectra of radiological television systems using a digital oscilloscope](#)

Med. Phys. **17**, 58 (1990); 10.1118/1.596528

[Vibrating string resonance spectra on the oscilloscope](#)

Am. J. Phys. **50**, 570 (1982); 10.1119/1.12809

[Spectra of Polycrystalline Phthalocyanines in the Visible Region](#)

J. Chem. Phys. **48**, 2674 (1968); 10.1063/1.1669501

[Application of Storage Oscilloscope to Scanning Electron Probe Microanalyzer](#)

Rev. Sci. Instrum. **35**, 1724 (1964); 10.1063/1.1719295

[Polarized Visible Spectra of Crystalline Trisoxalatometallates](#)

J. Chem. Phys. **35**, 1809 (1961); 10.1063/1.1732149

---



### Erratum: Ultraviolet Absorption Spectrum of Hydrogen Peroxide

[J. Chem. Phys. 16, 225 (1948)]

R. B. HOLT, C. K. McLANE, AND O. OLDENBERG  
*Lyman Physical Laboratory, Harvard University,  
 Cambridge, Massachusetts*

**T**HROUGH an error a wrong cut was used for Fig. 1 in this article. The following figure should have appeared in place of the one used.





FIG. 1. Absorption coefficient for hydrogen peroxide vapor (average values).

### Oscilloscope Scanning of Visible Spectra


W. E. DEAL, WILLIAM BRADSHAW, AND F. A. MATSEN  
*The University of Texas, Austin, Texas*  
 April 23, 1948

**A**N infra-red spectrograph with cathode-ray presentation has been developed by Daly and Sutherland and King, Temple, and Thompson.<sup>1</sup> On the vertical deflection system of a long-persistence tube was impressed the amplified output from a bolometer past which an infra-red spectrum was swept by means of a fourteen-second period rotation of the prism. A linear horizontal trace was employed.

In this laboratory the same principle has been applied in the visible region to obtain "instantaneous" spectra.<sup>2</sup> A



1a



1b

FIG. 1. Oscilloscope Trace: a—Emission spectrum of tungsten bulb; b—Dilute potassium permanganate in path. Gain for b twice that for a.

glass prism was rotated at speeds from 1800 to 3600 r.p.m., the spectrum being detected by a photo-multiplier tube. A shutter attached to the prism table interrupts a beam of light producing a trigger pulse for a driven sweep circuit.

In Fig. 1a is reproduced the emission spectrum of a tungsten bulb; in Fig. 1b a cell containing dilute potassium permanganate solution has been placed in the optical path. These spectra were photographed from a P-11 phosphor tube with  $\frac{1}{10}$ -second exposure time on Eastman Super XX film. A movie camera will be used to study spectral sequences in rapid chemical reactions. A complete report will appear elsewhere.

Work is in progress extending the technique to the near infra-red region using photo-conductive detectors and to the ultraviolet using quartz optics.

<sup>1</sup> E. B. Baker and C. D. Robb, *Rev. Sci. Inst.*, **14**, 362 (1943); E. F. Daly and G. B. M. Sutherland, *Proc. Phys. Soc. (London)* **59**, 77 (1947). See also J. King, R. B. Temple, and H. W. Thompson, *Nature* **158**, 196 (1946).

<sup>2</sup> See R. C. Herman and S. Silverman, *J. Opt. Soc. Am.* **38**, 209 (1948) for a photographic method involving translation of the spectrographic plate.