

# Integrated Capillary Electrophoresis Microsystem for Multiplex Analysis of Human Respiratory Viruses

Numrin Thaitrong,<sup>†</sup> Peng Liu,<sup>‡</sup> Thomas Briese,<sup>§</sup> W. Ian Lipkin,<sup>§</sup> Thomas N. Chiesl,<sup>†</sup> Yukiko Higa,<sup>†</sup> and Richard A. Mathies<sup>\*†,‡</sup>

Department of Chemistry and UCSF/UC Berkeley Joint Graduate Group in Bioengineering, University of California, Berkeley, California 94720, United States, and Center for Infection and Immunity, Mailman School of Public Health, Columbia University, New York, New York 10032, United States

We developed a two-layer, four-channel polymerase chain reaction (PCR)-capillary electrophoresis microdevice that integrates nucleic acid amplification, sample cleanup and concentration, capillary electrophoretic separation, and detection for multiplex analysis of four human respiratory viral pathogens, influenza A, influenza B, coronavirus OC43, and human metapneumovirus. Biotinylated and fluorescently labeled double-stranded (ds) deoxyribonucleic acid (DNA) amplification products are generated in a 100 nL PCR reactor incorporating an integrated heater and a temperature sensor. After amplification, the products are captured and concentrated in a cross-linked acrylamide gel capture matrix copolymerized with acrydite-functionalized streptavidin-capture agents. Thermal dehybridization releases the fluorescently labeled DNA strand for capillary electrophoresis injection, separation, and detection. Using plasmid standards containing the viral genes of interest, each target can be detected starting from as few as 10 copies/reactor. When a two-step reverse transcription PCR amplification is employed, the device can detect ribonucleic acid (RNA) analogues of all four viral targets with detection limits in the range of 25–100 copies/reactor. The utility of the microdevice for analyzing samples from nasopharyngeal swabs is demonstrated. When size-based separation is combined with four-color detection, this platform provides excellent product discrimination, making it readily extendable to higher-order multiplex assays. This portable microsystem is also suitable for performing automated assays in point-of-care diagnostic applications.

Accurate and rapid detection of the pathogens responsible for infectious human respiratory disease is crucial for outbreak control, disease surveillance, and vaccine development, as well as for appropriate and efficient use of antibiotics and antiviral

therapies.<sup>1–3</sup> Considering the wide range of existing and emerging microbial pathogens causing respiratory illnesses, many of which begin with similar influenza-like symptoms,<sup>4</sup> multiplex diagnostic assays, i.e., assays that simultaneously screen for many pathogens, are imperative for clinical medicine and public health.<sup>3,5–7</sup> Multiplex polymerase chain reaction (PCR), in which distinct primer pairs are used to amplify more than one target sequence in a single reaction, is one of the most appealing methods in molecular diagnostics because of its high sensitivity, rapid turnaround time, and high throughput.<sup>6–10</sup> Through careful primer design and the use of hot start polymerase enzymes, high specificity and sensitivity can be achieved.<sup>6,11</sup> A high degree of multiplexing, ranging from 2- to 22-plex for detection of respiratory pathogens including bacteria and viruses, has been reported.<sup>1,5,7,12–16</sup>

Advancements in miniaturization and microfluidic integration technology provide a pathway for the widespread practice of multiplex PCR diagnostics in clinical settings. When multiple processes are integrated onto a single microfabricated device,

- (1) Briese, T.; Palacios, G.; Kokoris, M.; Jabado, O.; Liu, Z.; Renwick, N.; Kapoor, V.; Casas, I.; Pozo, F.; Limberger, R.; Perezbrena, P.; Ju, J.; Lipkin, W. I. *Emerging Infect. Dis.* **2005**, *11*, 310–313.
- (2) Lin, B.; Blaney, K. M.; Malanoski, A. P.; Ligler, A. G.; Schnur, J. M.; Metzgar, D.; Russell, K. L.; Stenger, D. A. *J. Clin. Microbiol.* **2007**, *45*, 443–452.
- (3) Woo, P. C.; Chiu, S. S.; Seto, W. H.; Peiris, M. *J. Clin. Microbiol.* **1997**, *35*, 1579–1581.
- (4) Kelly, H.; Birch, C. *Aust. Fam. Physician* **2004**, *33*, 305–309.
- (5) Gröndahl, B.; Puppe, W.; Hoppe, A.; Kühne, I.; Weigl, J. A.; Schmitt, H. J. *J. Clin. Microbiol.* **1999**, *37*, 1–7.
- (6) Elnifro, E. M.; Ashshi, A. M.; Cooper, R. J.; Klapper, P. E. *Clin. Microbiol. Rev.* **2000**, *13*, 559–570.
- (7) Marshall, D. J.; Reisdorf, E.; Harms, G.; Beaty, E.; Mose, M. J.; Lee, W.; Gern, J. E.; Nolte, F. S.; Shult, P.; Prudent, J. R. *J. Clin. Microbiol.* **2007**, *45*, 3875–3882.
- (8) Henrickson, K. J. *Pediatr. Infect. Dis. J.* **2004**, *23*, S6–10.
- (9) Whelen, A. C. *Annu. Rev. Microbiol.* **1996**, *50*, 349–373.
- (10) Kehl, S. C.; Kumar, S. *Clin. Lab. Med.* **2009**, *29*, 661–671.
- (11) Chou, Q.; Russel, M.; Birch, D. E.; Raymond, J.; Bloch, W. *Nucleic Acids Res.* **1992**, *11*, 1717–1723.
- (12) Boivin, G.; Côté, S.; Déry, P.; Serres, G. D.; Bergeron, M. G. *J. Clin. Microbiol.* **2004**, *42*, 45–51.
- (13) Leung, T. F.; To, M. Y.; Yeung, A. C. M.; Wong, Y. S.; Wong, G. W. K.; Chan, P. K. S. *Chest* **2010**, *137*, 348–354.
- (14) Miyashita, N.; Saito, A.; Kohno, S.; Yamaguchi, K.; Watanabe, A.; Oda, H.; Kazuyamag, Y.; Matsushima, T., and the CAP study group *Respir. Med.* **2004**, *6*, 542–550.
- (15) Stockton, J.; Ellis, J. S.; Saville, M.; Clewley, J. P.; Zambon, M. C. *J. Clin. Microbiol.* **1998**, *36*, 2990–2995.
- (16) van Elden, L. J. R.; Nijhuis, M.; Schipper, P.; Schuurman, R.; van Look, A. M. *J. Clin. Microbiol.* **2001**, *39*, 196–200.

\* Corresponding author. Address: Department of Chemistry, MS 1460, University of California, Berkeley, CA, 94720. Phone: (510) 642-4192. Fax: (510) 642-3599. E-mail: ramathies@berkeley.edu.

<sup>†</sup> Department of Chemistry, University of California.

<sup>‡</sup> UCSF/UC Berkeley Joint Graduate Group in Bioengineering, University of California.

<sup>§</sup> Columbia University.

rapid, low-volume, automated, and sensitive assays can be performed.<sup>17–22</sup> Because of the superior performance offered by integrated microdevices, a variety of pathogens including the respiratory pathogen *B. pertussis*,<sup>23</sup> SARS-coronavirus,<sup>24</sup> dengue-2 virus,<sup>25</sup> influenza A virus,<sup>26</sup> and BK virus<sup>27</sup> have already been successfully detected with on-chip approaches. Nevertheless, most of these assays have been limited to single- or duplex analysis<sup>23–27</sup> performed on preliminary PCR-capillary electrophoresis (CE) microdevices, which do not include sample treatment and processing.<sup>20,23–27</sup> Recently, our group developed an integrated PCR-CE device with a preamplification immunomagnetic bead-based cell capture for highly specific and sensitive detection of *E. coli*.<sup>28</sup> In addition, we have reported a novel integrated PCR-CE device with gel-phase oligonucleotide affinity capture for postamplification cleanup.<sup>29</sup> When oligonucleotide capture probes are copolymerized in a photodefinable cross-linked polyacrylamide matrix directly between the PCR reactor and the CE separation channel, PCR products complementary to those probes are specifically hybridized and preconcentrated. This integrated capture/injection technique exhibits a 3- to 5-fold increase in signal intensity and a 6-fold increase in resolution compared to that using externally polymerized, noncross-linked capture matrixes.<sup>29</sup> The technique, however, is limited in its ability to scale up for multiplex analysis because of the complexity in designing unique and specific capture probes with similar thermodynamic properties for multiple targets.

To expand the multiplexing capability of our devices to the level needed for practical pathogen detection, we developed a universal gel capture purification method that relies on the biotin–streptavidin recognition.<sup>30</sup> PCR products, one strand biotinylated and the other fluorescently labeled, are captured in their native double-stranded (ds) state by a photopolymerized streptavidin capture gel, followed by electrokinetic washing and thermal melting of the bound deoxyribonucleic acid (DNA). The fluorescently labeled strands are released and injected for electrophoresis, leaving the biotinylated strands bound to the capture gel. This technique has previously been used to demonstrate on-chip

sample cleanup for forensic short tandem repeat analysis<sup>30,31</sup> but is applied here for the first time to multiplex pathogen detection in an integrated format.

We present a new two-layer integrated PCR-CE microdevice with streptavidin/biotin-mediated postamplification sample processing technology for rapid and parallel detection of respiratory viruses responsible for acute respiratory diseases (ARD). The pathogens include influenza A virus (FLUAV), human metapneumovirus (hMPV), human coronavirus OC43 (hCoV-OC43), and influenza B virus (FLUBV). Coupled with the multichannel capillary array electrophoresis portable scanner (McCAEPS) instrument previously developed in our group,<sup>32</sup> the entire process of nucleic acid amplification, capture, and CE analysis is streamlined and automated, thus minimizing contamination and sample loss between steps.


## EXPERIMENTAL SECTION

**Microdevice Design and Fabrication.** The integrated microdevice shown in Figure 1A contains four identical units forming two symmetrical doublet structures similar to the PCR-CE device previously developed in our group.<sup>33,34</sup> In addition to PCR amplification and CE separation, the new microdevice is designed to accomplish postamplification cleanup/concentration, as well as inline injection of amplified products. Each genetic analysis system consists of a 100 nL reaction chamber, an integrated heater and resistive temperature detector (RTD), a capture and injection region containing a 500  $\mu\text{m}$  long double T-junction, and a 10 cm long CE separation channel. The expanded view of the analyzing units is shown in Figure 1B. To facilitate automated operation using the portable microcapillary array electrophoresis ( $\mu\text{CAE}$ ) instrument (Figure 1C), the two doublets are designed to share common waste and cathode ports, while all four lanes share a common anode.

Microchannel features are patterned photolithographically on the surface of a four-inch glass wafer and sealed by thermally bonding a second wafer to form enclosed structures. The top wafer contains Ti/Pt heaters fabricated on its top-side and the etched fluidic structure containing reactors, capture regions, and separation channels on the bottom, while the bottom wafer contains microfabricated Ti/Pt RTDs on its top side. The microfabrication procedure of the microdevice follows the protocols developed previously.<sup>33–35</sup> Briefly, the fluidic layer is created from a 500  $\mu\text{m}$  thick Borofloat glass wafer coated with a 2000 Å layer of amorphous silicon on one side and 200 Å Ti/2000 Å Pt on the other side. Fluidic structures are etched on the amorphous silicon side to a depth of 40  $\mu\text{m}$  using 49% hydrofluoric (HF) acid. Heaters are fabricated on the Ti/Pt side of the same wafer by electroplating gold leads followed by ion beam etching of the Ti/Pt to form the serpentine heater elements. All access holes are then drilled using a CNC mill. The RTD layer is formed using a 762  $\mu\text{m}$  Borofloat glass wafer coated with 200 Å Ti and 2000 Å Pt. Features are

- (17) Manz, A.; Graber, N.; Widmer, H. M. *Sens. Actuators, B* **1990**, *1*, 244–248.
- (18) Dittich, P. S.; Tachikawa, K.; Manz, K. *Anal. Chem.* **2006**, *78*, 3887–3907.
- (19) Liu, P.; Mathies, A. *Trends Biotechnol.* **2009**, *27*, 572–581.
- (20) Chen, L.; Manz, A.; Day, P. J. *R. Lab Chip* **2007**, *7*, 1413–1423.
- (21) Bienvenue, J. M.; Legendre, L. A.; Ferrance, J. P.; Landers, J. P. *Forensic Sci. Int.: Genet.* **2010**, *4*, 178–186.
- (22) Ligler, F. S. *Anal. Chem.* **2009**, *81*, 519–526.
- (23) Prakash, A. R.; De la Rosa, C.; Fox, J. D.; Kaler, K. V. I. S. *Microfluid. Nanofluid.* **2007**, *4*, 451–456.
- (24) Zhou, Z. M.; Liu, D. Y.; Zhong, R. T.; Dai, Z. P.; Wu, D. P.; Wang, H.; Du, Y. G.; Xia, Z. N.; Zhang, L. P.; Mei, X. D.; Lin, B. C. *Electrophoresis* **2004**, *25*, 3032–3039.
- (25) Huang, F. C.; Liao, C. S.; Lee, G. B. *Electroporesis* **2006**, *27*, 3297–3305.
- (26) Pal, R.; Yang, M.; Lin, R.; Johnson, B. N.; Srivastava, N.; Razzacki, S. Z.; Chomistek, K. J.; Heldsinger, D. C.; Haque, R. M.; Ugaz, V. M.; Thwar, P. K.; Chen, Z.; Alfano, K.; Yim, M. B.; Krishnan, M.; Fuller, A. O.; Larson, R. G.; Burke, D. T.; Burns, M. A. *Lab Chip* **2005**, *5*, 1024–1032.
- (27) Kaigala, G. V.; Huskins, R. J.; Preiksaitis, J.; Pang, X. L.; Pilarski, L. M.; Backhouse, C. J. *Electrophoresis* **2006**, *27*, 3753–3763.
- (28) Beyor, N.; Yi, L.; Seo, T. S.; Mathies, R. A. *Anal. Chem.* **2009**, *81*, 3523–3528.
- (29) Thaitrong, N.; Toriello, N. M.; Del Bueno, N.; Mathies, R. A. *Anal. Chem.* **2009**, *8*, 1371–1377.
- (30) Yeung, S. H. I.; Liu, P.; Del Bueno, N.; Greenspoon, S. A.; Mathies, R. A. *Anal. Chem.* **2009**, *81*, 1371–1377.

- (31) Liu, P.; Scherer, J. R.; Greenspoon, S. A.; Chiesl, T. N.; Mathies, R. A. *Forensic Sci. Int.: Genet.* **2010**, doi: 10.1016/j.fsigen.2010.10.009.
- (32) Scherer, J. R.; Liu, P.; Mathies, R. A. *Rev. Sci. Instrum.* **2010**, *81*, doi: 10.1063/1.3502457.
- (33) Liu, C. N.; Toriello, N.; Mathies, R. A. *Anal. Chem.* **2006**, *78*, 5474–5479.
- (34) Toriello, N. M.; Liu, C. N.; Mathies, R. A. *Anal. Chem.* **2006**, *78*, 7997–8003.
- (35) Liu, P.; Seo, T. S.; Beyor, N.; Shin, K. J.; Scherer, J. R.; Mathies, R. A. *Anal. Chem.* **2007**, *79*, 1881–1889.



**Figure 1.** (A) Layout of the integrated PCR-CE microdevice. Each of the four analyzers is composed of an RTD temperature sensor (green) and integrated heater (red and orange) for thermal cycling, a 100 nL PCR reaction chamber, etched structures for analyte capture and purification, and 10 cm long CE separation channels with a common cathode and waste structures. (B) Magnified view of one reactor pair with temperature control elements and capture/purification structures. (C) The portable 12 × 12 × 10 in. multichannel capillary array electrophoresis portable scanner (McCAEPS) instrument contains all necessary elements for microchip control and operation as well as fluorescence excitation and detection.

formed by etching the coated glass wafer in hot aqua regia (3:1 HCl/HNO<sub>3</sub>) for 5 min. The two wafers are then permanently bonded together in a vacuum furnace at 650 °C for 6 h.

**Sample Preparation.** The multiplex PCR panel was adapted from the 22-plex MassTag PCR assay.<sup>1</sup> Gene targets, primer sequences, and expected product fragment sizes are listed in Table 1. Primer sequences were designed on the basis of conserved genomic regions so that each pathogen target could be robustly detected. To ensure stringent multiplex assays, the primers chosen had been verified to have similar melting temperatures with minimal cross-hybridization interactions to avoid primer dimers. To enable streptavidin/biotin-mediated capture and cleanup, the forward primers were fluorescently labeled at the 5' end with

FAM, TAMRA, or JOE, while the 5'-end of the reverse primers were functionalized with biotin (Integrated DNA Technologies (IDT), Coralville, IA). Each PCR mix was prepared to a volume of 25  $\mu$ L, containing up to 8  $\mu$ L of up to four different types of plasmid standards, 0.2 U of HotStarTaq DNA polymerase, 0.5 mM MgCl<sub>2</sub>, 0.5 mM dNTPs, 2.5  $\mu$ L of multiplex PCR buffer containing 6 mM MgCl<sub>2</sub>, pH 8.7 (Qiagen, Valencia, CA), 2.5  $\mu$ L of Q-solution (Qiagen), and an additional 1.25  $\mu$ L of multiplex master mix containing HotStar Taq DNA polymerase, multiplex PCR buffer, and dNTP mix (Qiagen) and 0.2, 0.4, 0.8, and 0.8  $\mu$ M (each) concentrations of the forward and reverse primers for FLUBV, hMPV, hCoV-OC43, and FLUAV, respectively. These primer ratios and concentrations were optimized to give similar product yield for all targets at the same starting concentrations. The same assay conditions were applied for standards and swab samples, allowing successful detection of the respective targets. Standard plasmids containing the gene target sequence were prepared from virus or assembled from synthetic polynucleotides. The plasmid standards were diluted in 2.5 ng/ $\mu$ L human placenta DNA to mimic random nucleic acid background present in clinical samples and stored at -20 °C. cDNA standards generated from nasopharyngeal swabs (NPS) that contained the respective agents were used to assess performance of the final assay. Briefly, total nucleic acid from 250  $\mu$ L of NPS suspended in 1 mL of standard viral transport media had been extracted using a NucliSENS easyMAG system (Biomérieux, Inc.). Reverse transcription was performed with 10  $\mu$ L of total nucleic acid using Superscript II with hexamer priming according to the manufacturer's protocol (Invitrogen, Carlsbad, CA). Three microliters of cDNA sample were added to the PCR cocktail to a final volume of 20  $\mu$ L. Off-chip reactions were performed in parallel to the on-chip experiments using an Eppendorf Mastercycler thermocycler following the same thermocycling protocol. The amplified product was analyzed by conventional capillary electrophoresis using an Applied Biosystems 96 capillary 3730 xl DNA Analyzer. The final sample mixture was prepared using a 9:1:1 ratio of PCR Hi-Di Formamide (Applied Biosystems)/unpurified PCR sample/500 GeneScan ROX size standard (Applied Biosystems).

For each one-step RT-PCR reaction, a 20  $\mu$ L RT-PCR master mix is composed of 0.5  $\mu$ L of OneStep RT-PCR enzyme mix containing Ominiscript Reverse Transcriptase, Sensiscript Reverse Transcriptase, and HotStarTaq DNA Polymerase (Qiagen), 5  $\mu$ L of OneStep RT-PCR buffer containing 12.5 mM MgCl<sub>2</sub> (Qiagen), 0.2 U of HotStarTaq DNA polymerase, 0.5 mM MgCl<sub>2</sub>, 0.5 mM dNTPs, 2.5  $\mu$ L of multiplex PCR buffer containing 6 mM MgCl<sub>2</sub>, pH 8.7 (Qiagen), 5  $\mu$ L of Q-solution (Qiagen), and 10 U of RNase inhibitor (Promega, Madison, WI). To prevent the formation of secondary structure during the reaction, the ribonucleic acid (RNA) template and primer mixture were prepared separately: the 5  $\mu$ L primer-template mixture contains 2.5  $\mu$ L of the RNA standards and 2.5  $\mu$ L of 0.5, 2.0, 1.5, and 2.0  $\mu$ M (each) forward and reverse primers for FLUBV, hMPV, hCoV-OC43, and FLUAV, respectively. The primer-template mixture was heat-denatured at 75 °C for 5 min and

(36) Schweiger, B.; Zadow, I.; Heckler, R.; Timm, H.; Pauli, G. J. *Clin. Microbiol.* 2000, 38, 1552-1558.



**Table 1. Primer Sequences for Multiplex Respiratory Virus Detection**


pathogen/gene target	primer	$T_m$ (°C)	sequence	size (bp)
FLUBV/hemagglutinin protein (HA gene)	BHA-U188-1	52	5'-FAM-AGA CCA GAG GGA AAC TAT GCC-3' <sup>36</sup>	159
	BHA-L347-1	54	5'-biotin-CTG TCG TGC ATT ATA GGA AAG CA-3' <sup>36</sup>	
hMPV/nucleocapsid protein (N gene)	MPV01.2	52	5'-TAMRA-AAC CGT GTA CTA AGT GAT GCA CTC-3' <sup>37</sup>	212
	MPV02.2	60	5'-biotin-CAT TGT TTG ACC GGC CCC ATA A-3' <sup>37</sup>	
hCoV-OC 43/nucleocapsid protein (N gene)	Co43-270F	55	5'-JOE-TGT GCC TAT TGC ACC AGG AGT-3' <sup>1</sup>	238
	Co43-508R	54	5'-biotin-CCC GAT CGA CAA TGT CAG C-3' <sup>1</sup>	
FLUAV/matrix protein (M gene)	AM-U151	56	5'-FAM-CAT GGA ATG GCT AAA GAC AAG ACC-3' <sup>36</sup>	246
	AM-L397	55	5'-biotin-AAG TGC ACC AGC AGA ATA ACT GAG-3' <sup>36</sup>	

then immediately chilled on ice to allow the primers to bind to the RNA templates. Finally, the RT-PCR master mix was added to the primer-annealed templates. In vitro transcribed RNA standards were diluted in 2.5 ng/ $\mu$ L yeast t-RNA (Sigma) to prevent RNA degradation due to frequent freeze–thaw cycles and stored aliquoted at  $-80^\circ\text{C}$ . The materials and methods employed for customized DNA ladder synthesis can be found in the Supporting Information (Section I).

**Matrix Synthesis.** A 5% linear polyacrylamide (LPA) gel with 6.15 M urea in 1 $\times$  TTE (50 mM Tris, 50 mM TAPS acid, and 2 mM EDTA) was used as a sieving matrix for all separations. The synthesis protocol of the gel was reported previously.<sup>38</sup> The cross-linked polyacrylamide capture gel was prepared in situ by cophotopolymerizing a 19:1 acrylamide/bis monomer mix with a 5' acrydite-modified streptavidin capture probe. Following the previously developed protocol,<sup>30</sup> a 500  $\mu$ L solution was prepared to 5% (v/v) bis-acrylamide, 6.15 M urea, 1 $\times$  TTE, 2  $\mu$ g/ $\mu$ L streptavidin–acrylamide (Invitrogen), 0.000 6% riboflavin (w/v) and 0.125% TEMED (v/v) in an opaque 2 mL scintillation vial with Teflon closure (National Scientific, Rockwood, TN). This solution was used to form a 500  $\mu$ m capture gel plug within the double-T junction in the dynamically coated microchannel as illustrated in Figure 2. To prevent the monomer solution from contaminating the reactor with PCR and RT-PCR inhibitors such as EDTA and acrylamide monomers, 8% LPA was first loaded from the coinjector into the channel separating the reactor from the capture and separation region (Figure 2, step 1). The monomer solution was then introduced into the microchannels. Using an inverted microscope equipped with a mercury arc lamp (Nikon) centered on 365 nm,<sup>29</sup> the double-T junction was irradiated (10 mW/cm<sup>2</sup>) for 5 min, polymerizing the solution in this region. After photopolymerization, unreacted monomer solution was removed and the microchannels were rinsed with deionized water (Figure 2, step 2).

**Microdevice Preparation.** The Ti/Pt RTDs were calibrated following methods previously described.<sup>33</sup> To passivate the glass surface, all microchannels were first treated with 1 M HCl for 15 min. The channels were then rinsed with deionized water prior to treatment with 0.25% w/v polyDuramide dynamic coating solution.<sup>39</sup> After an hour, the dynamic coating solution was evacuated and the channels were rinsed with water and dried by vacuum.

To prepare the microchip for a PCR-based assay, photopolymerized capture plugs were created following the protocol described above. A 0.5  $\times$  2 cm piece of PDMS (0.25 in. thick, Bisco Silicones, Carol Stream, IL) with a 3 mm diameter opening was first placed over each access hole of the microdevice to serve as a longer reservoir preventing solution evaporation during the run. The separation matrix was then loaded from the anode to waste and from the cathode to the coinjection reservoir, sandwiching the capture plug previously synthesized (Figure 2, step 3). The reactor wall was passivated by treatment with 1% w/v heat-denatured bovine serum albumin (BSA) at room temperature for 10 min to prevent enzyme adhesion during the reaction. The



**Figure 2.** Schematic of microdevice preparation and operation. (1) To avoid contamination of the reactor, a 8% linear polyacrylamide (LPA) gel (gray) is loaded into the coinjector prior to introducing the bis/acrylamide monomer solution (brown). (2) A streptavidin-copolymerized 5% bis/acrylamide gel plug 40  $\mu$ m deep  $\times$  120  $\mu$ m wide  $\times$  500  $\mu$ m long is formed within the microchannel through in situ photopolymerization (gold). (3) The plug is sandwiched by a separation matrix manually loaded from both anode and cathode. (4) The assay is started by filling the reactor with PCR (or RT-PCR) cocktail (blue), immediately followed by thermocycling. (5) After the amplification is completed, the sample is electrophoresed (red arrow) through the capture plug, where biotinylated PCR products are captured. (6) The captured dsDNA is dehybridized at 67  $^\circ\text{C}$ , releasing the fluorescently labeled ss-DNA for size-based separation at 300 V/cm.

- (37) Mackay, I. M.; Jacob, K. C.; Woolhouse, D.; Waller, K.; Syrmis, M. W.; Whitley, D. M.; Siebert, D. J.; Nissen, M.; Sloots, T. P. J. *Clin. Microbiol.* **2003**, *41*, 100–105.
- (38) Ruiz-Martinez, M. C.; Berka, J.; Belenkii, A.; Foret, F.; Miller, A. W.; Karger, B. L. *Anal. Chem.* **1993**, *65*, 2851–2858.
- (39) Albarghouthi, M. N.; Buchholz, B. A.; Huiberts, P. J.; Stein, T. M.; Barron, A. E. *Electrophoresis* **2002**, *23*, 1429–1440.

reactor was thoroughly rinsed with nuclease-free water before the introduction of PCR solution.

**Microdevice Operation.** Each reactor was filled by placing PCR (or RT-PCR) cocktail at the reactor inlet and applying a vacuum briefly at the reactor outlet (Figure 2, step 4). Immediately after loading, a drop of 5% LPA was added at both ports to completely seal the reactor. Then, the microchip was transferred to a portable detection instrument containing all electrical and detection interfaces for subsequent analysis steps. Amplification was initiated with a 10 min polymerase activation at 95 °C, followed by 15 step-down cycles consisting of a 10 s denaturation step at 95 °C, a 20 s annealing step starting at 65 °C with a reduction of 1 °C per cycle, and a 30 s extension at 72 °C. Following the step-down process, the final cycle (annealing at 50 °C) was repeated for 35 cycles followed by a final extension at 72 °C for 5 min. The thermocycling protocol for the one-step RT-PCR is similar, only with an additional step of cDNA synthesis at 42 °C for 20 min before the polymerase activation.

To perform postamplification sample cleanup and concentration, the sample was electrophoresed from the reactor to the waste at room temperature under an electric field of 25 V/cm for 10 min (Figure 2, step 5). While biotinylated dsDNA products were retained by the streptavidin-copolymerized capture matrix, excess salts, dNTPs, and unreacted forward primers were washed toward the waste. Following thermal melting of retained dsDNA products at 67 °C, the fluorescently labeled strands were directly injected into the separation channel at a separation field of 300 V/cm between the cathode and the anode (Figure 2, step 6). Sample detection was performed using laser-induced fluorescence 1 cm from the anode. After each run, the gels were pushed out with water. The microchannels and reactors were cleaned with piranha solution (3:1 H<sub>2</sub>SO<sub>4</sub>/H<sub>2</sub>O<sub>2</sub>) at 65 °C for 10 min, followed by a 5 min water rinse. *Caution: Piranha solution is highly hazardous and corrosive. The solution should be handled with extreme caution and proper personal protective equipment (a full face shield and heavy rubber gloves) should be worn while handling the chemical.*

**Instrumentation and Data Processing.** The 12 × 12 × 10 in. McCAEPS instrument (Figure 1B) contains an optical system including a 488 nm diode laser (75 mW, Sapphire 488-75, Coherent, Santa Clara, CA), an objective, and a four-color PMT (Hamamatsu H9797, Bridgewater, NJ) for fluorescence detection.<sup>35</sup> In addition, the device features pneumatic controls, electronics for temperature control, and four high voltage power supplies for CE. Thermal cycling is controlled through a LabVIEW program (National Instruments, Austin, TX) with a proportional/integral/differential (PID) module. More detail on the instrument can be found at <http://www.cchem.berkeley.edu/ramgrp/scanner>. Fluorescence crosstalk was corrected using BaseFinder v.6.1 (Giddings Lab, University of North Carolina), and processed traces were analyzed using PeakFit v4. (Systat Software, Inc., San Jose, CA). Fragment analysis data obtained from the conventional Applied Biosystems instrument was processed by GelQuest (SequentiX, Germany).

## RESULTS AND DISCUSSION

**Microdevice and Assay Design.** The two-layer glass microdevice is designed for simple operation and automation. The doublet structure with common cathode, waste, and anode

structures is employed to reduce the number of access holes, minimizing the number of electrical connections during electrophoresis and facilitating easier and faster gel loading. Unlike four-layer integrated PCR-based microdevices previously developed, this microdevice contains no removable PDMS-glass microvalves or pumps. The reactors are sealed by adding ~500 µL of a viscous 5% LPA gel over the reactor inlets and outlets. To prevent the reaction cocktail from leaking out of the reactor during thermocycling, excess (~500 µL) 5% linear polyacrylamide (LPA) gel is added to the coinjectors. The absence of PDMS microvalves vastly simplifies chip preparation and operation. By removing the need to apply an active pressure to seal the microvalves, we eliminate the potential for displacement of the PCR solution out of the reactor during thermocycling due to air leakage through the porous PDMS membrane.

Furthermore, the sequence-independent streptavidin/biotin-mediated capture process employed here provides a universal target capture technique applicable to a variety of multiplex assays. In contrast to the oligonucleotide-affinity capture format employed in previous work,<sup>29,40</sup> streptavidin/biotin-based capture eliminates complications in designing sequence-specific capture probes with similar thermodynamic properties, as well as the tedious process of optimizing capture conditions. Because of the denaturing condition introduced by 6.15 M urea in both capture and separation matrixes, full-strand melting and release can be performed at a moderate temperature of 67 °C. As a result, the potential for failed electrophoresis due to acrylamide gel breakdown is avoided.<sup>41</sup>


**Multiplex Amplification.** Multiplex amplification of gene targets from FLUAV, FLUBV hMPV, and hCoV-OC43 was carried out on the integrated microdevice using plasmid standards, each containing the cloned respective viral target sequence. In the multiplex PCR experiments, each target was amplified from 100 copies of plasmid standard. The higher annealing temperature during the initial step-down process promotes assay specificity by ensuring that the primers only hybridize to their complementary target in these critical early cycles.<sup>42</sup> A high level of PCR discrimination is crucial in pathogen detection assays because the pathogen target templates are commonly present in much lower abundance than the ubiquitous host genomic and mRNA backgrounds. For this reason, a step-down PCR scheme was employed. Because a total of 50 cycles were performed, additional amounts of *Taq* DNA polymerase, enzyme-stabilizing MgCl<sub>2</sub>, and dNTPs were added to ensure that the reaction is not reagent limited.

The effects of BSA passivation on PCR efficiency and robustness were examined by comparing amplifications in nonpassivated and BSA-passivated microreactors. As indicated by the S/N ratios in Figure 3A, varying PCR yields with declining yields as a function of amplicon size were obtained when reactions were performed in a bare, untreated reactor (*n* = 3) using 100 copies of each target. In contrast, all four products were generated in comparable or greater amounts (*n* = 3) when reactions were performed in BSA-coated reactors. These results confirm that BSA passivation helps increase reaction yield by likely occupying sites on the glass

(40) Toriello, N. M.; Liu, C. N.; Blazej, R.; Thaitrong, N.; Mathies, R. A. *Anal. Chem.* **2007**, *79*, 8549–8556.

(41) Muller, G. *Polym. Bull.* **1981**, *5*, 31–37.


(42) Don, R. H.; Cox, P. T.; Wainwright, B. J.; Baker, K.; Mattick, J. S. *Nucleic Acids Res.* **1991**, *19*, 4008.



**Figure 3.** On-chip multiplex PCR analysis of FLUBV, hMPV, hCoV-OC43, and FLUAV. (A) Bar graph shows the effects of BSA passivation on reaction yield obtained from multiplex on-chip PCR amplification ( $n = 3$ ). (B) Representative electropherograms of multiplex PCR products obtained from an uncoated reactor (a) and BSA-coated reactor (b). (c) Negative control containing no plasmid template. (d) Custom-made ROX-labeled sizing reference.

reactor wall and preventing adherence of the polymerase.<sup>43</sup> This effect is also seen in the individual electropherograms of representative 4-plex PCR amplification in passivated and nonpassivated reactors shown in Figure 3B for the 159-bp FAM-labeled FLUBV, 212-bp TAMRA-labeled hMPV, 238-bp JOE-labeled hCoV-OC43, and 246-bp FAM labeled FLUAV dsDNA amplification products. In a nonpassivated reactor, the products are biased toward smaller fragments whereas all products were generated in approximately equal yield when the reactor was coated with BSA. A negative control verifies the absence of carryover from run to run.

A ROX-labeled DNA ladder serving as a sizing reference was captured and injected in a similar fashion. The average resolution/base for the ladder between 95 and 300 nucleotides (nt) is  $0.26 \pm 0.08$ . The separation performance of our system characterized by




**Figure 4.** Assay specificity and sensitivity determination. (A) Representative electropherograms show that each plasmid standard (100 copies each) exclusively generates the correct product in the presence of all four primer pairs. (B) Sensitivity determination: standard plasmids for FLUBV, hMPV, hCoV-OC43, and FLUAV are amplified at 100, 50, 10, and 0 copies/reactor ( $n = 3$  for each data point). The S/N ratios for each of the four products at the three starting template concentrations are shown.

peak capacity,<sup>44</sup> defined as the maximum number of analytes detected by the system, can be calculated on the basis of the separation of the DNA ladder. With the average number of theoretical plates of  $2 \times 10^4$  plates/cm, the peak capacity in the 90–300 s time interval was determined to be 126. However, since the peak width of the PCR products is significantly larger than that of the reference ladder peaks, the actual peak capacity when analyzing real samples will be lower. Nevertheless, this theoretical calculation suggests that our microsystem is highly suitable for multiplex analysis assays. Successful amplification and detection of the four viral targets using both size-based separation and four-color spectral discrimination establish the feasibility of the device and system for multiplexed viral detection.

**Assay Sensitivity.** To determine the sensitivity of the microdevice for each virus in the multiplex panel, each individual plasmid was amplified separately in the presence of all primer pairs included in the panel. Representative electropherograms in Figure 4A show successful on-chip amplification from 100 copies/reactor. The detection limit of the system was determined in a similar fashion by amplifying each of the four templates following serial dilution to 100, 50, and 10 copies. The average S/N ratios

(43) Zhang, C.; Xing, D. *Nucleic Acids Res.* **2007**, *34*, 4223–4237.


(44) Grushka, E. *Anal. Chem.* **1970**, *42*, 1142–1147.



**Figure 5.** Analysis of in vitro transcribed RNA standards processed on-chip using one-step RT-PCR. (A) Representative electropherograms show that each RNA standard (100 copies each) can be accurately identified in the presence of all four primer pairs using the microdevice. A negative control containing no template verifies the absence of carryover between analyses. (B) Sensitivity analysis, as a function of initial RNA template concentrations, shows that the microdevice can detect hCoV-OC43, FLUBV, FLUAV, and hMPV RNA at 25, 50, 100, and 100 copies, respectively, with  $S/N > 5$  ( $n = 3$ ).

are plotted Figure 4B. A linear increase of  $S/N$  for all targets indicates that the amplifications are performed in the exponential phase.<sup>45,46</sup> The system demonstrates high sensitivity, with  $S/N > 10$  for 10 copies ( $n = 3$ ) of all four standard plasmids. The calculated detection limits of FLUBV, hMPV, hCoV-OC43, and FLUAV obtained by extrapolation of the linear regression for  $S/N = 3$  are 0.65, 0.63, 0.60, and 0.66, respectively, indicating the feasibility of single-copy viral detection. In addition, the four templates exhibit comparable  $S/N$  for each starting concentration with a standard deviation of 26% for 100 copies, 13% for 50 copies, and 6% for 10 copies. This observation confirms that similar amplification efficiency was achieved for all targets at the optimized primer concentration ratios.

**Analysis of RNA Samples.** Because many viruses have RNA genomes, the capability of the microdevice to analyze RNA samples is crucial. Therefore, RT-PCR amplification using in vitro transcribed RNA standards was investigated, and the sensitivity



**Figure 6.** PCR-based analysis NPS derived cDNA with the PCR-CE microdevice. Total nucleic acids derived from nasopharyngeal swabs (NPS) and reverse transcribed into cDNA were amplified and analyzed on-chip. Electropherograms reveal that sample #1 is positive for FLUBV, #2 is positive for hCoV-OC43, and #3 is positive for FLUAV, while sample #4 is negative for all the four viruses tested. Shaded bands show the expected locations of FLUBV (159 nt), hMPV (212 nt), hCoV-OC43 (238 nt), and FLUAV (246 nt). The accuracy of the results was verified by running parallel PCR reactions off-chip.

of the assay in detecting each RNA target was determined. In the presence of all four primer pairs, each RNA standard was reverse-transcribed on-chip at 42 °C for 20 min and amplified in the same reactor using the previously described step-down PCR conditions. Representative electropherograms in Figure 5A show detection and identification of all transcripts at 100 copies/reactor. A negative RT-PCR control containing no templates verifies the absence of nonspecific priming. The sensitivity of the device in detecting each viral RNA transcript was determined by performing RT-PCR on serially diluted RNA standards in the range of

(45) Sachse, K. *Methods in Molecular Biology: PCR Detection of Microbial Pathogens*; Sachse, K., Frey, J., eds.; Humana Press: Totowa, New Jersey, 2003; pp 10–12.

(46) Lantz, O.; Bonney, E.; Griscelli, F.; Taoufik, Y. *Methods in Molecular Medicine: Quantitative PCR Protocol*; Kochanowski, B., Reischl, U., eds.; Humana Press: Totowa, New Jersey, 1999; pp 91–94.



25–10 000 copies/reactor. Each standard was reverse-transcribed and amplified separately in the presence of all four primer pairs as described previously. Figure 5B shows a comparison of S/N for the amplified products over the range of starting template concentrations ( $n = 3$ ). While FLUBV and hCoV-OC43 can be detected down to 50 and 25 copies, respectively, FLUAV and hMPV can be detected at 100 copies (with  $S/N > 5$ ). As determined by extrapolation to  $S/N = 3$ , the calculated limits of detection for FLUBV, hCoV-OC43, FLUAV, and hMPV are 4.8, 6.3, 10, and 167 copies, respectively. A lower sensitivity obtained from the RNA samples compared to their DNA counterparts resulting from the known limited efficiency of reverse transcription has been previously observed.<sup>1</sup> In particular, for a multiplex one-step RT-PCR where more than one virus-specific RT primer is present, the optimal choice of RT conditions can become more difficult than in hexamer primed reactions.<sup>6</sup> Nevertheless, the results obtained here demonstrate the ability of the microdevice to process RNA samples, which opens the door to other RNA-based applications such as detection and genotyping of other viruses, HIV viral load testing, and gene expression studies.

**Demonstration of Microdevice Detection with Nasopharyngeal Swabs (NPS).** To evaluate the capability of the microsystem to analyze real diagnostic samples, four cDNA samples derived from NPS were PCR amplified and analyzed on-chip. Total nucleic acid extracts from NPS reverse-transcribed using random hexamers were used for multiplex amplification without further purification or concentration. Samples were aliquoted into two groups: the first group was processed on-chip in the same manner as the standard plasmids, while the second group was amplified and analyzed off-chip as a control.

The electropherograms in Figure 6 indicate successful detection of FLUBV, hCoV-OC43, and FLUAV samples, while one sample was negative for all four viruses tested. The identity of the pathogens was determined by matching the migration time of the product peaks with those obtained from the standards of known identity. Whereas the performance for FLUAV, FLUBV, and hCoV-OC43 was excellent and the negative control yielded no amplification signal, detection of hMPV was not achieved with NPS samples (data not shown). The hMPV primer pair, as already indicated by results with in vitro transcribed RNA, appears to

perform less efficiently than the other three primer pairs in this platform and will need to be specifically redesigned for the microsystem application. Although nonspecific PCR noise in amplification #2 was minimal, the more pronounced background seen in amplification #3 may lead to false-positive or false-negative calls, especially when more primer pairs are introduced in the assay. For an expanded panel, the cutoff levels and the detection limit will need to be empirically redetermined and optimized.

## CONCLUSIONS

We have developed a new user-friendly two-layer integrated PCR-CE microdevice for rapid multiplex detection of FLUBV, hMPV, hCoV-OC43, and FLUAV. The postamplification purification process employs streptavidin/biotin-mediated recognition for simultaneous capture of multiple targets. The PCR amplification, sequence-independent product capture and concentration, and CE analysis on the automated portable genetic analysis instrument were streamlined to within 2 h. The microdevice is capable of processing both DNA and RNA samples with high specificity and sensitivity in clinical samples. This novel integrated microdevice has potential as a platform for robust, automated, high-throughput, and sensitive point-of-care diagnostics.

## ACKNOWLEDGMENT

We thank Vishal Kapoor and Aaloki Shah for excellent technical assistance, as well as helpful suggestions and valuable discussions. We also acknowledge Dr. James Scherer for the design and fabrication of the portable genetic analyzer. Microfabrication was performed at the University of California Berkeley Microfabrication Laboratory. This work was supported by Samsung Electronics Co. Ltd., Award 20082064, National Institutes of Health (NIH) Grants #CA77664 and #AI57158 (Lipkin, Northeast Biodefense Center).

## SUPPORTING INFORMATION AVAILABLE

Additional information as noted in text. This material is available free of charge via the Internet at <http://pubs.acs.org>.

Received for review August 5, 2010. Accepted November 1, 2010.

AC1020744