OIML Member State
P. R. China

OIML Certificate No.
R76/2006-A-CN2-26.02

OIML CERTIFICATE

ISSUED UNDER SCHEME A

OIML Issuing Authority	Name: National Institute of Metrology, China Address: No.18, Bei San Huan Dong Lu, Chaoyang Dist., Beijing, P.R.China Person responsible: Mr. Li Wentao
Applicant	Name: Jiangxi Zonjli High-Tech Co., Ltd. Address: No.579 HuoJu Road, Nanchang Hi-Tech Industrial Development Zone, Jiangxi P.R.China
Manufacturer	Name: Jiangxi Zonjli High-Tech Co., Ltd. Address: No.579 HuoJu Road, Nanchang Hi-Tech Industrial Development Zone, Jiangxi P.R.China
Identification of the certified type	Weighing Indicator DZ2008-D1 series (Further characteristics on page 2)

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R76 – Edition 2006

For accuracy class: (III)

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

OIML Issuing Authority: **National Institute of Metrology, China**

Date: 3 September 2026

OIML Issuing Authority China 中国 OIML 证书发证机构

No.18, Bei San Huan Dong Lu,
Chaoyang Dist, Beijing,
100029, P.R.China
+86 10 64525646
oimlia@nim.ac.cn
www.nim.ac.cn

Li Wentao
Director
National Institute of Metrology, China

OIML Member State
P. R. China

OIML Certificate No.
R76/2006-A-CN2-26.02

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation report:

No. TER-R76/2006-CN2-26.02 dated 31 August 2026 that includes 26 pages

and the associated OIML test reports:

No. LSmm2026-02141.1 dated 19 August 2026 that includes 19 pages
No. LSmm2026-02141.2 dated 19 August 2026 that includes 17 pages

OIML Certificate History

Revision No.	Date	Description of the modification
0	3 September 2026	Initial issue

Characteristics of the Instrument

Type	DZ2008-D1	
	DZ2008-D1-F	DZ2008-D1-M
Accuracy class	Ⓘ	
Fraction of mpe, p_i	0.0	
Excitation voltage	DC15V，≤400mA	
Maximum number of verification scale intervals, n	≤ 6000	
Temperature range	-10℃~+40℃	
Power supply	AC:100 V~240 V,50Hz DC or rechargeable battery:12V	
Software version	V20XXX note:V20 refers to the legally relevant part of the software and will not be changed, while XXX refers to the non-legally relevant part the software and ranges from 000 to 999.	