OIML Member State Germany	**OIML Certificate No.** **R76/2006-A-DE1-26.03**

OIML CERTIFICATE ISSUED UNDER SCHEME A

OIML Issuing Authority

Name:	Physikalisch-Technische Bundesanstalt, Conformity Assessment Body
Address:	Bundesallee 100, 38116 Braunschweig, GERMANY
Person responsible:	Dr.-Ing. Prof. h. c. Frank Härtig

Applicant

Name:	SBOT Technologies LLC
Address:	50 West 23rd Street 9th floor New York, 10010 United States of America

Manufacturer

Name:	SBOT Technologies LLC
Address:	50 West 23rd Street 9th floor New York, 10010 United States of America

Identification of the certified type *(the detailed characteristics will be defined in the additional pages)*

Non-automatic weighing instrument for direct sales to the public
Type: Caper Cart 3, Caper Cart 3 Lite

Designation of the module *(if applicable)*

Not applicable

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 Edition (year): 2006

For accuracy class (if applicable): III

OIML Certificate No.
R76/2006-A-DE1-26.03

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation report:

No. KBS-2-4.3-TER-06.02.03/0397#0001-0008 dated 02.04.2026 that includes 11 pages

The technical documentation relating to the identified type is contained in documentation file:

No. ZDS-R76/2006-A-DE1-26.03 dated 02.04.2026 that includes 2 pages

OIML Certificate History

Revision No.	Date	Description of the modification
R76/2006-A-DE1-26.03	02.04.2026	Initial issue

Digital signature

The Issuing Authority

PTB
Physikalisch-Technische Bundesanstalt
Nationales Metrologieinstitut

digital gesiegelt
Physikalisch-Technische Bundesanstalt
02.04.2026
13:18:16 +02

Dr.-Ing. Oliver Mack

Member of Conformity Assessment Body

Date: 02.04.2026

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

Variant		**Caper Cart 3** **Caper Cart 3 Lite**
Accuracy class		(III)
Minimum load *Min*	g	100
Maximum capacity *Max*	kg	40
Verification scale interval *e*	g	5
Number of verification scale intervals *n*		≤ 8.000
Tare range (subtractive)	• Max	≤ 100%
Preset tare range	• Max	≤ 50%
Power supply voltage		24 V DC