	DANAK PROD Reg.no. 7026	FORCE Certification

OIML Member State Denmark	**OIML Certificate No.** **R76/2006-A-DK2-24.10 Revision 2**

OIML CERTIFICATE ISSUED UNDER SCHEME A

OIML Issuing Authority

Name:	**FORCE Certification A/S**
Address:	Park Allé 345, 2605 Brøndby, Denmark
Person responsible:	Per Rafn Crety

Applicant

Name:	**Stryker Medical**
Address:	3800 E Centre St. Portage, Michigan 49002 USA

Manufacturer **Stryker Medical**

Identification of the certified type *(the detailed characteristics will be defined in the additional pages)*

ProCuity MedSurg Bed

Designation of the module *(if applicable)*

Non-automatic weighing instrument

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1, Edition (year): 2006

For accuracy class (if applicable**): IIII**

FORCE Certification A/S · Park Alle 345 2605 Brøndby Tel+45 43 25 01 77 info@forcecertification.com www.forcecertification.com
Task no.: 126-28015.20 and ID no.: FC-OIML-17982-2 forcecertification.com/en/weighing

OIML Certificate No.
R76/2006-A-DK2-24.10 Revision 2

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated OIML reports:

Type examination report: No. 120-24467.10, dated 23 September 2020, that includes 65 pages

Type evaluation report: No. 124-29915.90.10, dated 05 August 2024, that includes 18 pages

Type examination report: No. 126-28015.10, dated 30 September 2026, that includes 17 pages

Type evaluation report: No. 126-28015.20, dated 30 September 2026, that includes 25 pages

The technical documentation relating to the identified type is contained in documentation file:

120-24467, 124-29915, 126-28015

OIML Certificate History

Revision No.	Date	Description of the modification
Initial version	22 August 2024	-
Revision 1	02 September 2024	Correction of typo error
Revision 2	07 October 2026	Voltage span updated for 100-240 VAC, 50/60 Hz, information on load cells used in the approved system added

Identification, signature and stamp
The OIML Issuing Authority
FORCE Certification A/S

Date: 07 October 2026

FORCE Certification
Digitally signed by Rasmus Møller Jensen
Date: 2026.10.07 11:31:33 +02'00'

Rasmus Møller Jensen
Certification Manager

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

Descriptive annex

Characteristics

Type:	ProCuity MedSurg Bed
Accuracy Class:	IIII
Weighing ranges:	Single-interval
Maximum capacity (Max):	250 kg
Minimum capacity (Min):	20 kg
Verification scale interval (e=):	2 kg
Maximum numbers of verification scale intervals	n=125
Maximum tare effect:	- 60 kg
Operating temperature range:	5°C to 35°C
Power supply:	100 – 240 VAC, 50/60 Hz

Software

The legal software for the scale has its own version and checksum.
These can be found in the setting menu.

Allowed version are

v01.08.000.01 with checksum 0xE86A93DA
v01.09.000.01 with checksum 0x77DAA10D
v01.10.000.01 with checksum 0x22920465

Devices

- Semi-automatic subtractive tare device
- Exchange equipment function*)
- Weight logging device
- Operator information messages
- Gravity compensation device (sealed)
- Stable equilibrium and Net indicators.

*) Lock the present net weight and take any change in the gross weight as a change in tare weight.

Interfaces

The scale has no interfaces.

Load cells

4 pcs. K-TSTB/300LBSTR1-1 manufactured by HBM, Stryker P/N 300900370057