OIML Member State Denmark		**OIML Certificate No.** **R76/2006-A-DK2-25.04**

OIML CERTIFICATE ISSUED UNDER SCHEME A

OIML Issuing Authority

Name:	**FORCE Certification A/S**
Address:	Venlighedsvej 4, 2970 Hørsholm, Denmark
Person responsible:	Per Rafn Crety

Applicant

Name:	**Dini Argeo S.r.l.**
Address:	Via della Fisica 20, 41042 Spezzano di Fiorano Modena Italy

Manufacturer:

Dini Argeo S.r.l.

Identification of the certified type:
(the detailed characteristics will be defined in the additional pages)

3590E / 3590X / CPWE / DFW / DGT / KS

Designation of the module: *(if applicable)*

Non-automatic weighing indicators / weight transmitter

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1, Edition (year): 2006

For accuracy class (if applicable**):**

III or IIII

FORCE Certification A/S · Park Alle 345 2605 Brøndby Tel+45 43 25 01 77 info@forcecertification.com www.forcecertification.com
Task no.: 124-30692.90.10 and ID no.: FC-OIML-22907-1 forcecertification.com/en/weighing

OIML Certificate No.
R76/2006-A-DK2-25.04

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated OIML reports:
Type examination report: No. SN 1193, dated 19 December 2011, that includes 21 pages
Type examination report: No. SN 1228, 20 September 2012, that includes 11 pages
Type examination report: No. SN 1229, 20 September 2012, that includes 12 pages
Type examination report: No. SN 1230, 20 September 2012, that includes 12 pages
Type examination report: No. SN 1231, 20 September 2012, that includes 11 pages
Type examination report: No. SN 1232, 20 September 2012, that includes 24 pages
Type examination report: No. SN 1282, 12 May 2014, that includes 21 pages
Type examination report: No. TR 767, 25 October 2018, that includes 28 pages
Type examination report: No. 125-32351.10, dated 07 September 2026, that includes 24 pages

Type evaluation report: No. PO2483, dated 05 February 2019, that includes 17 pages.
Type evaluation report: No. 124-30692.90.10, dated 02 October 2026, that includes 68 pages

The technical documentation relating to the identified type is contained in documentation file:

124-30692, 117-22690, 119-22344, 123-29241, 124-30850, 125-32351

OIML Certificate History

Revision No.	Date	Description of the modification
Initial Revision	5 October2026	Certificate issued.

Identification, signature and stamp
The OIML Issuing Authority
FORCE Certification A/S

Date: 5 October2026

FORCE Certification
Digitally signed by Rasmus Møller Jensen
Date: 2026.10.05 16:17:15 +02'00'

Rasmus Møller Jensen
Certification Manager

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

Descriptive annex

Characteristics

- Accuracy class: III and IIII
- Single interval or multi-interval (up to 3 intervals), multi-range (up to 3 ranges)
- Maximum number of verification scale intervals: 10,000
- Maximum capacity: $n_i \times e_i$
- Scale interval: ≤ 10000 (III), ≤ 1000 (IIII)
- Minimum capacity: 20*e
- Maximum tare effect: -Max
- Circuit for remote sense: present on the model with 6-wire connection
- Temperature range.: -10 °C to +40 °C
- Power supply: 110-240 VAC, 50/60 Hz (3590E, 3590X, DGT, DFW), or 110-240 VAC, 50/60 Hz using external adapter (3590E, 3590X, DGT, DFW), or 6-12 VDC (3590E) / 8-36 VDC(CPWE) / 12-24 VDC (DGT) / 12 VDC (DFW). Internal rechargeable battery (optional).

Software

The software is held on the Flash Memory and cannot be modified by the user. The calibration and legally relevant parameters are protected via physical or software means.

A jumper located on the main board prevents all access to the legally relevant parameters.

Alternatively, software sealing may be used to protect the calibration and legally relevant parameters. Two non-editable counters, designated CAL and CONFIG, are incremented each time the calibration and legally relevant parameters respectively are modified, with access to these parameters being password-protected.

The software identification is fully described in the user manual, and can be displayed at power up or via the software menu.

The legally relevant software is identified by two parts: prefix/version.

The prefix shows the instrument model and shall be:

3590E: 01 or 10

3590X: 10

CPWE: 01

DFW: 02 or 20

DGT: 09

The version shows the legally relevant software version shall be 01 for all series.

The prefix/version may be followed by a suffix indicating the software program version and other options installed which may be freely modified and are regarded as the non-legally relevant software.

Since the code may be longer than the digits available on the display, it is shown in two parts.

Devices

The family of indicators may have the following devices:

- Semi-automatic zero setting (≤ 4% Max)
- Zero tracking (≤ 4% Max)
- Semi-automatic subtractive tare weighing
- Pre-set tare
- Recall of Gross indication when tare is active
- Determination of stability of equilibrium
- Indication of stability of equilibrium
- Multi-range and multi-interval function
- Checking of display
- Printing
- Alibi storage device
- Gravity compensation
- Event counters
- Manual checkweighing
- Real time clock
- Counting
- Command via external device (PC)
- Accumulation
- Battery level indicator
- Remote control
- LCD or LED
- Peak Hold (non-legal function)
- Gross, Net, Tare, Preset tare, Print, Zero, Motion, Accumulation, Over/Under weight and Network indicators
- Range in use indicators (multi-range variant)
- Connection to up to 4 load receptors, with load receptor number indicator

OIML Certification System

Interfaces

The indicator may have the following load cell interfaces.

- 4 or 6 wire analogue load cell interface (1-4 interfaces)
- Serial interface for connection to digital load cell(s)
- Serial interface for connection to weight transmitter

The instrument may have the following protected interfaces:

- RS-232
- RS-485
- Ethernet
- Profibus
- Profinet
- DeviceNet
- CANopen
- Ethernet IP
- Ethercat
- USB
- WiFi
- Bluetooth
- RF (radio frequency)
- DC voltage input
- Control inputs/outputs
- Opto-isolated inputs
- Photomosfet outputs
- SENSOR (Digital in)
- Analogue output and input
- Analogue output and input

OIML Certification System

The instrument may use any analogue load cell(s) that may be used for instruments under this certificate of type examination provided the following conditions are met:

1) There is a respective Part / Evaluation / Test Certificate (EN 45501) or an OIML Certificate of Conformity (R60:2000 or R60:2017)

2) The certificate contains the load cell types and the necessary load cell data required for the manufacturer's declaration of compatibility of modules (OIML R76:2006 Annex F), and any particular installation requirements). A load cell marked NH is allowed only if humidity testing to EN 45501 has been conducted on this load cell.

3) The compatibility of load cells and indicator is established by the manufacturer by means of the compatibility of modules form, contained in the above OIML R76:2006 document, or the like, at the time of EC verification or declaration of EC conformity of type.

4) The load transmission must conform to one of the examples shown in the WELMEC 2.4 Guide for load cells.

The instrument may use one of the following digital load cells:

Manufacturer	Load cell type	Cert. No.
Flintec	RC3D digital load cell *)	TC6586
HBM	C16i *)	D09-00.46
Scaime	CB50X-DL *)	TC7078
Dini Argeo	RCD *)	TC7547
Dini Argeo	RCPTD **)	TC8786
Dini Argeo	RSBT-D	0200-WL-16434
Utilcell / Dini Argeo	740D / RCUD **)	E-04.02.C06
Utilcell	740D / 740CPD	TC13091
Sensocar S.A.	SP-D **)	E-02.02.C02 5th addition
Sensocar S.A.	CS-D **)	E-16.02.C01
Keli Sensing Technology	ZSF-D / ZSW-D *)	TC8339
Keli Sensing Technology	QS-D	TC12299
Rice Lake Weighing Systems	RL5416D / RL5416DC **)	TC11637
Rice Lake Weighing Systems	RL5426D / RL5426DC **)	TC11699
Rice Lake Weighing Systems	RL5418D	TC13289

*) up to 32 digital load cells may be connected to the indicator
**) up to 64 digital load cells may be connected to the indicator.

When digital load cells are used the number of verification scale intervals is limited to the number specified in the applicable Test or Parts Certificate.

The instrument may be connected to any peripheral device that has been issued with a Parts, Test or Evaluation Certificate issued by a Notified Body responsible for type examination under Directive 2014/31/EU in any Member State and bears the CE marking of conformity to the relevant directives.

A peripheral device without a Parts (Test) certificate may be connected under the following conditions:

- it bears the CE marking for conformity to the EMC Directive;
- it is not capable of transmitting any data or instruction into the weighing instrument, other than to release a printout, checking for correct data transmission or validation;
- it prints weighing results and other data as received from the weighing instrument without any modification or further processing; and
- it complies with the applicable requirements of EN:45501, i.e. 4.2, 4.4, 4.6 and 4.7.

A printing device may print additional information such as date or number to identify the printed weighing result(s) or sets of weighing results.

The following set of peripheral devices “printers” can be used in conjunction with any of the approved indicators covered by this certificate,

- DP24, DP24S1N, DP190.
- LP542, LP542S, LP542Plus, LP842Plus, LP942Plus, LP1042Plus (also called LPxx).
- TDP248M, TDP643Plus, TDP245, TTP244MEPlus, TT246MPlus (also called SMTxx).
- TM295, LX300, LX350.
- TG2460H.
- CUSTOM PLUS or TPR.
- DATAMAX: DMXM4206 e EX2.
- PR80 (alternative designation PRINT!).

The indicator may be connected to a tilt switch or a tilt sensor.