OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-26.37 revision 0
Project number 4127757

Issuing authority	NMi Certin B.V. Person responsible: M.Ph.D. Schmidt
Applicant and Manufacturer	Shinko Denshi Co., Ltd. 1-52-1 Itabashi, Itabashi-ku Tokyo 173-0004 Japan
Identification of the certified type	A **Non-automatic weighing instrument** Type : ALG-RCE series or ALG-CE series
Characteristics	See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class (I) or (II)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
29 September 2026

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
the Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.

OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-26.37 revision 0
Project number 4127757

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. 26-01-A_R76-06 dated 9 April 2026 that includes 56 pages;
- No. 26-01-B_R76-06 dated 9 April 2026 that includes 16 pages;
- No. 26-01-C_R76-06 dated 9 April 2026 that includes 16 pages;
- No. 26-01-D_R76-06 dated 9 April 2026 that includes 11 pages;
- No. 26-01-E_R76-06 dated 9 April 2026 that includes 12 pages;
- No. NMi-4127757-01 dated 29 September 2026 that includes 11 pages;
- No. NMi-4127757-02 dated 29 September 2026 that includes 21 pages;
- No. NMi-4127757-03 dated 29 September 2026 that includes 18 pages;
- No. NMi-4127757-04 dated 29 September 2026 that includes 16 pages.

Characteristics of the non-automatic weighing instrument:
The construction of the non-automatic weighing instrument is recorded in the Documentation folder no. T13297-1.

Type	All types
Weighing range	Single interval
Tare	T ≤ -100%
Power supply voltage	100 – 240 V AC 50/60 Hz (by AC/DC Power Supply), Or 2x 1,5 V AA battery
Software identification Checksum	A28B

Accuracy class	(I)	(II)
Maximum capacity	1200 g ≤ Max ≤ 10000 g 6000 ct ≤ Max ≤ 50000 ct	220 g ≤ Max ≤ 15000 g 1100 ct ≤ Max ≤ 31000 ct
Verification scale interval	e ≥ 0,01 g	e ≥ 0,01 g
Actual scale interval	e = 10 d	
Maximum number of scale intervals	n ≤ 120000 divisions	n ≤ 62000 divisions
Temperature range	+10 °C / +30 °C	+5 °C / +35 °C

OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-26.37 revision 0
Project number 4127757

Type	ALG8201RCE / ALG8201CE	ALG15001RCE / ALG15001CE
Accuracy class	(II)	
Maximum capacity	Max = 8200 g	Max = 15000 g
Verification scale interval	e = 1 g	
Actual scale interval	e = 10 d	
Maximum number of scale intervals	n = 8200 divisions	n = 15000 divisions
Temperature range	+5 °C / +35 °C	

The software identification is displayed at start-up.

Revision History

Revision	Date	Change(s)
0	2026-09-29	Initial issue.