OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-26.41 revision 0
Project number 4058795

Issuing authority	NMi Certin B.V. Person responsible: M.Ph.D. Schmidt
Applicant	Mettler-Toledo GmbH Im Langacher 44 CH-8606 Greifensee Switzerland
Manufacturer	Mettler-Toledo (Changzhou) Measurement Technology Ltd. No.111, West TaiHu Road Changzhou, Jiangsu, 213125 China
Identification of the certified type	An **Indicator** Type : ACT305, ACT300
Characteristics	See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class (III) or (IIII)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
16 September 2026

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.

OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-26.41 revision 0
Project number 4058795

The conformity was established by the results of tests and examinations provided in the associated reports:
- No. NMi-4058795-01 dated 16 September 2026 that includes 24 pages;
- No. NMi-4058795-02 dated 16 September 2026 that includes 23 pages;
- No. NMi-4058795-03 dated 16 September 2026 that includes 14 pages;
- No. NMi-4058795-04 dated 16 September 2026 that includes 15 pages;
- No. NMi-4058795-05 dated 16 September 2026 that includes 39 pages.

Characteristics of the indicator:
The construction of the indicator is recorded in the Documentation folder no. TC13213-1.

Configuration	ACT305 Panel / DIN	ACT300 Panel only
Accuracy class	(III) or (IIII)	
Weighing range	Single interval	
Maximum number of scale intervals (one weighing range)	n ≤ 10000	n ≤ 6000
Load cell excitation voltage	5 V DC	
Minimum signal input voltage	U_{min} = 0 mV	
Minimum input voltage per verification scale interval	0,3 µV	0,5 µV
Minimum load cell resistance	43 Ω	58 Ω
Maximum load cell resistance	1245 Ω	
Fraction of the maximum permissible error	0,5	
Load cell interface	6-wire with sense technology, may be configured as 4-wire	
Maximum value of the cable length per cross wire section between the indicator and the junction box or load cells	1606,9 m/mm^2 In case sense technology is not used the load cells are connected directly without junction box or extension cable	
Temperature range	-10 °C / +40 °C	
Power supply voltage	10 – 30 V DC (not suitable for a road vehicle power supply)	
Software identification	Version number: 1.xx.xx.xxxx (x is a number between 0 and 9 representing the non-legally relevant part of the software)	

Software:
- The identification number will be displayed after pressing the key sequence:
  - DIN version: Settings (arrow key) →  → Info Recall;
  - Panel version: Settings (arrow key) → Info → Info Recall.

- The indicator has embedded software.

Revision History

Revision	Date	Change(s)
0	2026-09-16	Initial issue.