Member State
Switzerland

OIML Certificate No.
R076/2006-A-CH1-18.02 Rev. 3

OIML-CS CERTIFICATE ISSUED UNDER SCHEME A

Issuing authority

Name	**Federal Institute of Metrology METAS** Conformity Evaluation Body METAS-Cert
Address	Lindenweg 50, 3003 Bern-Wabern, Switzerland
Person responsible	Gulian Couvreur, Head of METAS-Cert

Applicant

Name	Mettler-Toledo GmbH
Address	Im Langacher 44, 8606 Greifensee, Schweiz

Manufacturer The manufacturer of the certified pattern is the Applicant

Identification of the certified type

XPR / XSR Analytical Balances

Type **XPR / XSR**

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

METAS digitally signed
15.09.2026

OIML R 76-1, edition 2006

for accuracy class (I)

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

Metrological characteristics

Accuracy class	(I)
Max	3.2 g to 320 g
d	≥ 1 µg
e	1 mg
n	≤ 320 000
Tare-balancing range	Until 100 % of Max
Preset tare value	Until 100 % of Max
Temperature range	10 °C / 30 °C

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation reports:

No.	***Date***	***Including pages***
CH-MIER-26012-00	2026-09-03	13

The technical documentation relating to the identified type is contained in documentation file:

Name	***Date***	***Including pages***
CH-W1-18026-04_LERD	2026-09-03	5

OIML Certificate History:

Revision No.	**Date**	**Description of the modification**
00	2018-08-28	First issue
01	2020-10-26	Editorial changes
02	2024-05-21	Change of design, load receptor and housing
03	2026-09-03	Software Validation (Welmec 7.2) Software update

The OIML Issuing Authority CH1

3003 Bern-Wabern, 2026-09-03

Approved by Gulian Couvreur, Head of sector
METAS-Cert

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full