# OIML Certificate

**OIML Member State**
The Netherlands

Number R76/2006-A-NL1-26.40 revision 0
Project number 4058795


| | |
|---|---|
| Issuing authority | NMi Certin B.V.<br>Person responsible: M.Ph.D. Schmidt |
| Applicant | Mettler-Toledo GmbH<br>Im Langacher 44<br>CH-8606 Greifensee<br>Switzerland |
| Manufacturer | Mettler-Toledo (Changzhou) Measurement Technology Ltd.<br>No.111, West TaiHu Road<br>Changzhou, Jiangsu, 213125<br>China |
| Identification of the certified type | An **Analog data processing device**<br>Type : ACT300 |
| Characteristics | See next page |

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76-1**:2006 for accuracy class (III) or (IIII)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
16 September 2026

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.

# OIML Certificate

**OIML Member State**
The Netherlands

Number R76/2006-A-NL1-26.40 revision 0
Project number 4058795


The conformity was established by the results of tests and examinations provided in the associated reports:
- No. NMi-4058795-01 dated 16 September 2026 that includes 24 pages;
- No. NMi-4058795-03 dated 16 September 2026 that includes 14 pages;
- No. NMi-4058795-05 dated 16 September 2026 that includes 39 pages.

**Characteristics of the ADPD:**
The construction of the ADPD is recorded in the Documentation folder no. TC13212-1.

| | |
|---|---|
| Configuration | ACT300 |
| Accuracy class | (III) or (IIII) |
| Weighing range | Single interval |
| Maximum number of scale intervals (one weighing range) | $n \leq 6000$ |
| Load cell excitation voltage | 5 V DC |
| Minimum signal input voltage | $U_{min}$ = 0 mV |
| Minimum input voltage per verification scale interval | 0,5 µV |
| Minimum load cell resistance | 58 Ω |
| Maximum load cell resistance | 1245 Ω |
| Fraction of the maximum permissible error | 0,5 |
| Load cell interface | 6-wire with sense technology, may be configured as 4-wire |
| Maximum value of the cable length per cross wire section between the analog data processing device and the junction box or load cells | 1606,9 m/mm$^2$<br>In case sense technology is not used the load cells are connected directly without junction box or extension cable |
| Temperature range | -10 °C / +40 °C |
| Power supply voltage | 10 – 30 V DC<br>(not suitable for a road vehicle power supply) |
| Software identification | Version number: 1.xx.xx.xxxx<br>(x is a number between 0 and 9 representing the non-legally relevant part of the software) |

Software:
- The identification number can be displayed on the device that displays the primary indications;
- The analog data processing device has embedded software.

# OIML Certificate

**OIML Member State**
The Netherlands

Number R76/2006-A-NL1-26.40 revision 0
Project number 4058795


**Revision History**

| Revision | Date | Change(s) |
|---|---|---|
| 0 | 2026-09-16 | Initial issue. |