apoplex medical technologies

Requirementproposal

Document: G1.01.Requirementproposal
Release Date: 19.05.2022
Owner: Head of R&D
Template Version: 2

1. Responsible

Lukas Oremek is the contact person for this requirement. The contact person accompanies this requirement over the entire life cycle.

2. Identifier

ARC-Req1_1 : Receive ECG files

3. Description

3.1. What should be improved/solved with the requirement?

- Customer of apoplex record ecg data
- To analyse the ecg data, it must be transferred to apoplex

3.2. Acceptance critera for a solution

- apoplex received ecg data from customer
- The received data must be uniquely identifiable
- It must be ensured that the transmitted ecg data can be assigned to a customer and a patient without any patient information
- It must be ensured that the data received is supported and complete

4. Exemplary use case

4.1. Use Case ARC-Req1_1-Uc1 (First Use Case)

- Personas
  - Employee in a private medical practice
  - Supported ECG recorder
- Preconditions
  - ECG data was recorded on the recorder
- Procedure
  - Employee transfer the ecg data
- Expected result
  - received ecg data are uniquely identified and can be used for further analysis

1. Approval

Approver	Date	Signature
PO	25.05.2022	
Requirement Responsible	25.05.2022	
R&D	25.05.2022	