

Performance test of linear regressors using inertial information combined with sEMG to minimize the limb position effect in proportional and simultaneous control of lower arm prosthetics.



Irene Uriarte, Martin Garenfeld, Oliver Damsgaard, Simon Bruun. Aalborg University, School of Medicine and Health

AIM

 f_c =10Hz.

the variance applied to

The aim for this project are expressed in the following two hypotheses:

- Simultaneous and proportional control of two DOF's of the wrist in different limb positions, can be achieve trough the use of linear regression as control system.
- Combining ssurface EMG and IMU's can minimize the limb position effect when using regression as control system.

were recorded with Myo armband, positioned on the right fore-

The sEMG was recorded by eight channels in a frequency range

0-200Hz. IMU data was recorded using the buildt in accelerom-

eter in the Myo armband. sEMG data were filtered through a

second-order Butterworth high-pass filter, with cutoff frequency

Features are extracted using a sliding-window of 40 samples with

an overlapping of the 50%. Two time domain features are ex-

tracted; Mean absolute value (MAV) and logarithmic variance.

MAV represent the amplitud of the signal. It is defined as the

 $MAV = \frac{1}{N} \sum_{i=1}^{N} |x_i|$

where N is the length of the signal, and x_i is the signal of i sam-

ples. The logarithmic variance is a nonlinear transformation of

 $\log(\sigma^2) = \log(\frac{\sum\limits_{i=1}^{N}(x_i - \mu)^2}{N})$

where N is the length of the signal, x_i is the i^th sample of the

signal and μ is the mean. PCA is applied to qualitatively deter-

mine the separability of the feature data. Data is evaluated for

differncies in feature data clusters and significant outliers. If the

feature clusters are distinguishable from each other and have no

significant outliers, the data is of high quality and will be used

to train the regressors. Only the first three principal components

identified through PCA will be used to train the regressors. The

 $Y = \alpha + \beta X + \epsilon$

regressors are implemented through simple linear regression:

average of the absolute values of the sEMG signal:

MATERIALS & METHODS

arm of the subjects while standing.

INTRODUCTION

The development of EMG controlled prosthetics have advanced greatly in recent years. More complex prosthetics are demanded and more advanced control mothods has been developed. Most control methods so far has utilized pattern recognition which only enables control of one degree of freedom at a time. Most studies have conducted tests in only one limb position, not considering the limb position effect on EMG signals. [1] This study aims to overcome the limb position effect by combining EMG with inertial information in the training sessions of the regressor to obtain simultaneous and proportional control of EMG prosthesis.

RESULTS 2

Surface EMG data was collected from four able-bodied subjects. Subjects were instructed to performed four different hand gestures. This study only focus on two DOF, which are, flexion and lobortis nisl accumsan. extension, radial and ulnar deviation of the wrist. sEMG signals

Treatments	Response 1	Response
Treatment 1	0.0003262	0.562
Treatment 2	0.0015681	0.910
Treatment 3	0.0009271	0.296

Table 1: Table caption

gravida tempor. Mauris feugiat elit et augue placerat ultrices. Morbi accumsan enim nec tortor consectetur non commodo.

Table 2: Table caption

Donec faucibus purus at tortor egestas eu fermentum dolor facilisis. Maecenas tempor dui eu neque fringilla rutrum. Mauris

Treatments	Response 1	Response
Treatment 1	0.0003262	0.562
Treatment 2	0.0015681	0.910
Treatment 3	0.0009271	0.296

Nulla ut porttitor enim. Suspendisse venenatis dui eget eros

Treatments	Response 1	Response 2
Treatment 1	0.0003262	0.562
Treatment 2	0.0015681	0.910
Treatment 3	0.0009271	0.296

RESULTS 1

Placeholder

Image

Sed fringilla tempus hendrerit. Vestibulum ante ipsum primis in faucibus orci luctus et ultrices posuere cubilia Curae; Etiam ut elit sit amet metus lobortis consequat sit amet in libero. Lorem ipsum dolor sit amet, consectetur adipiscing elit. Phasellus vel sem magna. Nunc at convallis urna. isus ante. Pellentesque condimentum dui. Etiam sagittis purus non tellus tempor volutpat. Donec et dui non massa tristique adipiscing. Quisque vestibulum eros

Figure 1: Figure caption


Aliquam auctor, metus id ultrices porta, risus enim cursus sapien, quis iaculis sapien tortor sed odio. Mauris ante orci, euismod vitae tincidunt eu, porta ut neque. Aenean sapien est, viverra vel lacinia nec, venenatis eu nulla. Maecenas ut nunc nibh, et tempus libero. Aenean vitae risus ante. Pellentesque condimentum dui. Etiam sagittis purus non tellus tempor volutpat. Donec et dui non massa tristique adipiscing.

Placeholder

Image

Figure 2: Figure caption

CONCLUSION



- Pellentesque eget orci eros. Fusce ultricies, tellus et pellentesque fringilla, ante massa luctus libero, quis tristique purus urna nec nibh. Phasellus fermentum rutrum elementum. Nam quis justo lectus.
- Vestibulum sem ante, hendrerit a gravida ac, blandit quis magna.
- Donec sem metus, facilisis at condimentum eget, vehicula ut massa. Morbi consequat, diam sed convallis tincidunt, arcu
- Nunc at convallis urna. isus ante. Pellentesque condimentum dui. Etiam sagittis purus non tellus tempor volutpat. Donec et dui non massa tristique adipiscing.

where Y is the dependent variable or response, X is the independent variable or the predictor, β is the regression coefficient or the slope, and α is the Y intercept (predicted value of Y at X=0), ϵ is the error. The regressor accuracy of control is tested qualitatively through superimposition of the output of the regressors build for each feature onto the actual data for the intensities of the movements. The regressor accuracy is quantitatively tested through a target reaching task measuring time to reach 16 targets. The scores are compared through stastistical t-test, comparing scores between limb positions and between only using sEMG data and including IMU data.

All data processing is performed in Matlab (2017).

REFERENCES

- Anders Fougner, Erik Scheme, Adrian D.C. Chan, Kevin Englehart, and Øyvind Stavdahl. Resolving the limb position effect in myoelectric pattern recognition. IEEE Transactions on Neural Systems and Rehabilitation Engineering, 19, 2011.
- [2] J. M. Smith and A. B. Jones. *Book Title*. Publisher, 7th edition, 2012.
- A. B. Jones and J. M. Smith. Article Title. Journal title, 13(52):123–456, March 2013.

CONTACT INFORMATION

Web www.smh.aau.dk Email 17gr7404@hst.aau.dk