

UncertRadio: Software for determining characteristic limits in accordance to DIN EN ISO 11929 for radioactivity measurements

Günter Kanisch¹, Florian Ober^{2¶}, and Marc-Oliver Aust³

¹ Hamburg, Germany, formerly Thünen-Institute ² Max Rubner-Institute, Kiel, Germany ³

Thünen-Institute, Bremerhaven, Germany ¶ Corresponding author

DOI: [10.xxxxxx/draft](https://doi.org/10.xxxxxx/draft)

Software

- [Review](#) ↗
- [Repository](#) ↗
- [Archive](#) ↗

Editor: [Kelly Rowland](#) ↗ 

Reviewers:

- [@villaa](#)
- [@alberto743](#)
- [@fanurs](#)

Submitted: 27 August 2024

Published: unpublished

License

Authors of papers retain copyright and release the work under a Creative Commons Attribution 4.0 International License ([CC BY 4.0](#)).

Summary

In Germany, radioactive substances in the environment are monitored in accordance with the Treaty establishing the European Atomic Energy Community (EURATOM) of 1957 and the German Radiation Protection Act ([Bundesrepublik Deutschland, 2017](#)). The Evaluation of radiometric measurements requires the estimation of associated uncertainties as defined in the ISO GUM ([Joint Committee for Guides in Metrology, 2008b](#)). In addition, German law requires that the characteristic limits (decision threshold (DT) and the detection limit (DL)) are determined on the basis of this uncertainty in accordance with ISO 11929-1:2019 ([2019a](#)) to ISO 11929-4:2019 ([2022](#)).

Statement of need

To the best of the authors' knowledge, UncertRadio is the only publicly available software to determine the characteristic limits in a user-centralized way. UncertRadio can be used for a variety of applications from alpha, beta and gamma radiation measurements including dosimetry for up to three radionuclides simultaneously. Therefore, it is especially suited for modern liquid scintillation measurement procedures of e. g. strontium isotopes. The user only needs to define the evaluation model by providing a set of equations in text-form to calculate the output quantity value. The required partial derivatives are calculated internally.

There are two main analytical approaches used within the software:

- Procedures without linear unfolding: The basic evaluation model is linear in the net count rate. Thus, the output value can be calculated directly ([Kanisch, 2016a](#)).
- Procedures utilizing linear unfolding methods: The model additionally includes linear least squares procedures for fitting e. g. time-dependent decay or build-up curves ([Kanisch, 2016b](#)).

Additionally, the model can be evaluated with a Monte Carlo simulation following ISO 11929-2 ([2019b](#)) (see [Figure 1](#)). This represents the method of propagating whole distributions, which has advantages if the distributions of the input variable distributions significantly deviate from the normal distribution; see ISO GUM Supplements 1 ([2008a](#)) and 2 ([2011](#)).




Figure 1: Example of results obtained with the Monte Carlo simulation (upper row). The results (output quantity, decision threshold and detection limit) calculated in accordance with ISO 11929-2 (International Organization for Standardization, 2019b) using a weighted linear least-squares (WLS) approach are also included.

Scientific references

UncertRadio has been used in several scientific publications. It is referenced by the standard ISO 11929-1:2019 (2019a) to ISO 11929-4:2022 (2022) and actively used by the German authorities for monitoring environmental radioactivity and external radiation.

In Kanisch (2016a), an overview considering evaluations without using linear unfolding has been presented. Two significant linear relationships in the model equations for the net count rate (common in evaluation models) were identified providing a generalized approach for the determination of the characteristic limits. Kanisch (2016b) extends the evaluation models to include linear unfolding methods utilizing a weighted linear least-squares (WLS) approach for the first stage of the model. This step is solved using matrix-algebra which also takes parameters with uncertainties in the design matrix into account.

Applications, examples and quality control

UncertRadio includes a set of approximately 70 example projects, which are structured text files available in both English and German language. They are mostly based on real-world applications but also cover the examples in ISO 11929-4 (2022). These examples illustrate the structure of the set of equations for various measurement models. A short overview of all

example projects is given in [section 2.5](#) of the UncertRadio documentation. These examples contributed to the validation of UncertRadio. To verify if UncertRadio is working correctly, all examples can be run automatically by selecting “Options/QC batch test” in the main menu. Since Version 2.6 this can also be done in the terminal by running:

```
./UncertRadio run_tests
```

Availability and documentation

UncertRadio is available free for download as compiled windows binaries since 2014.

Recently, it was decided to make the source code available as open source software under the GNU General Public License 3. UncertRadio is written in modern Fortran utilizing many Fortran 2003 and 2008 features, e. g. the C-interoperability.

The graphical user interface (GUI) is build with GTK 3 in combination with gtk-fortran ([Magnin et al., 2019](#)), which provides the required Fortran bindings. PLplot is implemented for the graphical presentations ([Alan W. Irwin et al., 2019](#)).

Many of the utilized numerical procedures are derived from the work of Miller ([2004](#)) , Burkardt ([2024](#)) and Brandt ([1999](#)). A fortran function parser ([Schmehl, 2008](#)) is included for interpreting user-defined equations.

The UncertRadio source code is available on [GitHub](#). Detailed building instructions are provided within the README file. UncertRadio works both on Linux and Windows and comes with language packages for English, French and German. Nevertheless, only Windows binaries are currently provided for download. They are available on GitHub and on the homepage of the [Thünen-Institute](#).

Until Version 2.5.3 UncertRadio contained an extensive compiled HTML help (chm) file for the description of the program features in German and English language. However, since Version 2.6 these files have been replaced by a modern python sphinx ([Sphinx Documentation, 2025](#)) based documentation, but is lacking some german translations. Thus, the old chm files are still available in the repository or upon request. The individual help topics are available from within UncertRadio using various help buttons. In addition, the complete documentation is available on [GitHub Pages](#). The current development goals and open issues can be found in the [README](#) file.

Code problems can be reported in the [issues](#) tab on GitHub. The authors are happy to help within their capabilities. Feedback and contributions via [pull](#) request are greatly appreciated.

Conflict of interest

The authors declare no financial conflicts of interest.

Acknowledgements

The authors thank all those who have contributed to improving UncertRadio by reporting problems and suggesting new features. We also thank D. Schrammel (KIT, Karlsruhe, Germany) for many comparison test calculations as well as Prof. R. Michel (formerly IRS, Hannover, Germany), for testing and applying UncertRadio during the preparation of ISO 11929-4 ([2022](#)).

Furthermore, the authors especially thank Marco Lechner and Josef Peter, who have contributed with their knowledge and constructive suggestions to make UncertRadio open source.

References

- 89
- 90 Alan W. Irwin, Andrew Ross, Geoffrey Furnish, Hazen Babcock, António Tomé, Arjen Markus,
91 Andrew Roach, Hezekiah M. Carty, Doug Hunt, James Dishaw, Jerry Bauck, Maurice
92 LeBrun, Phil Rosenberg, & Werner Smekal. (2019). *PLplot*. <http://plplot.org/index.php>
- 93 Brandt, S. (1999). *Datenanalyse: Mit Statistischen Methoden Und Computerprogrammen*.
94 Spektrum Akademischer Verlag. ISBN: 978-3-8274-0158-8
- 95 Bundesrepublik Deutschland. (2017). *Act on protection against the harmful effects of*
96 *ionising radiation (StrlSchG) (German Radiation Protection Act)* [Federal Act]. <https://www.gesetze-im-internet.de/strlschg/>
- 97
- 98 Burkardt, J. (2024). *John Burkardt's Home Page*. <https://people.sc.fsu.edu/~jburkardt/>
- 99 International Organization for Standardization. (2019a). *Determination of the characteristic*
100 *limits (decision threshold, detection limit and limits of the coverage interval) for mea-*
101 *surements of ionizing radiation — Fundamentals and application; Part 1: Elementary*
102 *applications* (ISO 11929-1:2019(E)). <https://www.iso.org/standard/69579.html>
- 103 International Organization for Standardization. (2019b). *Determination of the characteristic*
104 *limits (decision threshold, detection limit and limits of the coverage interval) for mea-*
105 *surements of ionizing radiation — Fundamentals and application; Part 2: Advanced applications*
106 *(ISO 11929-1:2019(E))*. <https://www.iso.org/standard/69580.html>
- 107 International Organization for Standardization. (2022). *Determination of the characteristic*
108 *limits (decision threshold, detection limit and limits of the coverage interval) for mea-*
109 *surements of ionizing radiation — Fundamentals and application; Part 4: Guidelines to*
110 *applications* (ISO 11929-4:2022(E)). <https://www.iso.org/standard/84497.html>
- 111 Joint Committee for Guides in Metrology. (2008a). *Evaluation of measurement data -*
112 *Supplement 1 to the GUM: JCGM 101:2008*. International Bureau of Weights and
113 Measures. <https://www.bipm.org/en/publications/guides/gum.html>
- 114 Joint Committee for Guides in Metrology. (2008b). *Guide to the expression of uncertainty in*
115 *measurement: JCGM 100:2008*. International Bureau of Weights and Measures. <https://www.bipm.org/en/publications/guides/gum.html>
- 116
- 117 Joint Committee for Guides in Metrology. (2011). *Evaluation of measurement data - Sup-*
118 *plement 2 to the GUM: JCGM 102:2011*. International Bureau of Weights and Measures.
119 <https://www.bipm.org/en/publications/guides/gum.html>
- 120 Kanisch, G. (2016a). Generalized evaluation of environmental radioactivity measurements with
121 UncertRadio Part I: Methods without linear unfolding. *Applied Radiation and Isotopes*,
122 110, 28–41. <https://doi.org/10.1016/j.apradiso.2015.12.003>
- 123 Kanisch, G. (2016b). Generalized evaluation of environmental radioactivity measurements with
124 UncertRadio Part II: Methods with linear unfolding. *Applied Radiation and Isotopes*, 110,
125 74–86. <https://doi.org/10.1016/j.apradiso.2015.12.046>
- 126 Magnin, V., Tappin, J., Hunger, J., & Lisle, J. D. (2019). Gtk-fortran: A GTK+ binding to
127 build Graphical User Interfaces in Fortran. *Journal of Open Source Software*, 4(34), 1109.
128 <https://doi.org/10.21105/joss.01109>
- 129 Miller, A. (2004). *Alan Miller's Fortran Software*. <https://jblevins.org/mirror/amiller/>
- 130 Schmehl, R. (2008). *Fortran 95 function parser*. <https://fparser.sourceforge.net/>
- 131 *Sphinx documentation*. (2025). <https://www.sphinx-doc.org/en/master/index.html>