

lys_instr: A Python Package for Automating Scientific Measurements

Ziqian Wang^{1,2}, Hidenori Tsuji², Toshiya Shiratori³, and Asuka Nakamura^{2,3}

¹ Institutes of Innovation for Future Society, Nagoya University, Japan  ² RIKEN Center for Emergent Matter Science, Japan  ³ Department of Applied Physics, The University of Tokyo, Japan   Corresponding author

DOI: [10.21105/joss.10906](https://doi.org/10.21105/joss.10906)

Software

- [Review](#) 
- [Repository](#) 
- [Archive](#) 

Editor: [Evan Spotte-Smith](#)  

Reviewers:

- [@Anatoly1010](#)
- [@peterhys](#)

Submitted: 19 November 2025

Published: 21 September 2026

License

Authors of papers retain copyright and release the work under a Creative Commons Attribution 4.0 International License ([CC BY 4.0](#)).

Summary

Modern experiments increasingly demand automation frameworks that coordinate diverse scientific instruments while remaining flexible and customizable. Developing custom measurement systems, however, often requires explicit management of low-level device communication and concurrency, resulting in substantial development overhead. We present `lys_instr`, a Python package that addresses these challenges through an object-oriented, three-layer architecture for instrument control, workflow coordination, and GUI construction. It enables researchers to construct responsive, asynchronous measurement systems by reusing instrument abstractions, workflow logic, and GUI components across hardware configurations. Integrated with the `lys` platform ([Nakamura, 2023](#)), `lys_instr` provides a common environment for experiment control, data acquisition, and multidimensional visualization.

Statement of need

Modern scientific research often relies on systematic measurements to elucidate physical phenomena. Such experiments may span wide parameter spaces, require long acquisition times, or coordinate multiple instruments, making efficient automation increasingly valuable. Measurement workflows may also incorporate informatics-driven, condition-based optimization, further motivating reusable and interoperable programmatic interfaces.

However, building such a measurement system remains time-consuming for researchers. At the low level, instrument-control implementations must accommodate diverse communication protocols (e.g., TCP/IP, VISA, and serial), often tying higher-level code to device-specific details and limiting device interchangeability and reuse across systems. At the high level, coordinating workflows that combine conditional logic, iterative processes, and advanced algorithms across multiple libraries frequently leads to redundant implementations, reducing development efficiency. For example, capturing temperature-dependent images with a camera and acquiring temperature-dependent spectra share the same general workflow: iterative temperature adjustment followed by data acquisition. Nevertheless, this logic is often implemented separately for different experiments. Moreover, for long or complex measurements, graphical user interfaces (GUIs) provide real-time visualization, device-state monitoring, and support for manual intervention. Implementing responsive GUIs for these low- and high-level functions, however, requires familiarity with GUI frameworks and event-handling mechanisms. These requirements impose substantial development overhead and highlight the need for a control framework that balances architectural flexibility with reduced implementation complexity.

State of the field

Scientific instrument-control software includes platforms with different priorities. LabVIEW (National Instruments Corporation, 2024) and MATLAB's Instrument Control Toolbox (The MathWorks, Inc., 2024) integrate instrument communication, workflows, and GUI development. QCoDeS (Nielsen et al., 2025) and PyMeasure (PyMeasure Developers, 2024) provide instrument abstractions and experimental procedures. Atomize (Melnikov et al., 2025) and fsc2 (Törting, 2019) support script-based experiments through reusable device operations. Bluesky (Allan et al., 2019) separates hardware abstraction from reusable orchestration plans and optional GUI clients, while PyMoDAQ (Weber, 2021) combines actuator and detector plugins with dashboards and scan extensions. Qudi (Binder et al., 2017) explicitly separates hardware, experiment logic, and GUI layers. Taken together, these systems show that hardware, workflow, and GUI abstractions are well established; the relevant differences lie in how they are integrated and which experimental workflows they target.

Among these systems, Qudi provides the closest point of comparison for `lys_instr`. Its mature modular architecture separates hardware, experiment logic, and GUI modules, while its logic layer supports application-specific data processing and analysis. A central design goal of `lys_instr`, however, is to support experiments that generate increasingly complex multidimensional datasets, for which immediate inspection, analysis, and visualization during acquisition are both challenging and valuable. Rather than reimplementing general multidimensional visualization and analysis capabilities within Qudi, `lys_instr` is designed to integrate directly with `lys`, a general-purpose platform that already provides them. This integration allows acquired data to be inspected, analyzed, and visualized in the same environment used for instrument control. In addition, `lys_instr` provides reusable controller and detector abstractions and GUI-configurable nested scan workflows, allowing measurement components to be reused across hardware configurations.

Software design

The three layers and their relationships are summarized in Figure 1. Each layer applies established object-oriented design patterns (Gamma et al., 1994), including *Template Method*, *Bridge*, *Composite*, *Observer*, and *Mediator*, to define how components are extended, composed, and connected.




Figure 1: Schematic of the code architecture of `lys_instr`.

1. Base Layer: Instrument Abstraction

This layer defines standardized abstract interfaces for two principal roles in a control-and-detect workflow: *controllers*, which adjust experimental parameters such as external fields,

temperature, or physical position, and *detectors*, which acquire data such as images or spectra. Following the *Template Method* pattern, the base interfaces define the device lifecycle, state monitoring, synchronization, and asynchronous execution, while subclasses implement the primitive operations required for device-specific communication. Higher layers interact with different instruments through the same methods, independently of their concrete implementations. Because background monitoring and acquisition are managed by the interfaces, other devices and GUIs can remain responsive, and device-specific subclasses can be integrated directly into higher-level workflows without reimplementing concurrency logic.

2. Top Layer: Workflow Coordination

This layer constructs hardware-independent workflows from the Base Layer interfaces. In a *scan*, controllers vary experimental parameters while detectors acquire data at each step. Following the *Bridge* pattern, scan execution depends on controller and detector abstractions rather than concrete hardware implementations. The *Composite* pattern allows individual scan processes to be nested into multi-parameter workflows while retaining a common execution model. Prebuilt GUI components invoke the same abstract methods for direct device control and receive state and data updates through event-driven signals, following the *Observer* pattern and avoiding direct dependencies on concrete devices. Consequently, workflow and GUI logic can be reused across different hardware configurations without rebuilding standard control and coordination components.

3. Connection Layer: Control-System Assembly

This layer assembles components from the Base and Top Layers into an interactive control system, including device interfaces, workflows, data storage, visualization, and GUI components. Following the *Mediator* pattern, it centrally manages connections within and across layers, linking GUI components to their corresponding interfaces and coordinating data flow without direct dependencies among individual components. Concrete hardware details therefore remain confined to the Base Layer, allowing system configurations and GUI layouts to be adapted without application-level implementation of inter-device communication or thread coordination. Prebuilt templates provide starting configurations that can be adapted to individual experimental setups.

Together, these layers allow new instruments to be integrated through device-specific subclasses while reusing asynchronous execution, workflow coordination, and GUI behavior. Workflows with different coordination models can be supported through additional components or customized interfaces.

Example GUI configuration

`lys_instr` allows users to assemble GUIs suited to individual experimental setups; [Figure 2](#) shows one possible configuration. Here, the `lys_instr` window is embedded in the `lys` platform, with Sector A for data storage, Sector B for detector control, and Sector C for controller operation and scan configuration. Multidimensional, nested scan sequences can be defined through the Scan tab, while `lys` tools in the outer window tabs allow on-the-fly customization of acquired-data visualization.




Figure 2: Example GUI configuration in `lys_instr`. The main window, embedded in the `lys` window, contains three sectors: storage (A), detector control (B), and controller operation and scan configuration (C). The Scan tab in (C) enables configuration of multidimensional, nested experimental workflows.

Research impact statement

`lys_instr` has been used in experiments reported in several peer-reviewed publications. It has been used to automate ultrafast electron diffraction (UED) and ultrafast transmission electron microscopy (UTEM) systems by coordinating experimental parameters and data acquisition in pump-probe experiments using ultrafast laser excitation and pulsed electron beams (Koga et al., 2024; Nakamura et al., 2020, 2021, 2022, 2023; Shimojima et al., 2021, 2023a, 2023b). It has also been used to control electromagnetic lenses and electron deflectors as part of complex microscopy workflows involving electron-beam precession (Hayashi et al., 2026; Shiratori et al., 2024).

The same interface-based workflows have been used to control transmission electron microscopes from multiple manufacturers at RIKEN Center for Emergent Matter Science and Nagoya University, demonstrating portability across hardware configurations. Integration with sister packages in the `lys` family, including `lys_em` (lys_em Developers, 2026) and `lys_fem` (lys_fem Developers, 2026), allows `lys_instr` components to participate in multi-instrument workflows while preserving modularity and extensibility.

AI usage disclosure

Generative AI tools provided debugging suggestions during the final stages of software development. The authors implemented and reviewed all changes. Core non-GUI behavior was tested with unit tests, while GUI and hardware workflows were validated with real instruments.

Acknowledgements

We acknowledge valuable comments from Takahiro Shimojima and Kyoko Ishizaka. This work was partially supported by Grant-in-Aid for Scientific Research (KAKENHI) Grants No. 21K13889, No. 26K17092, and No. 25K00057, and JST PRESTO Grant No. JPMJPR24JA.

References

- Allan, D., Caswell, T., Campbell, S., & Rakitin, M. (2019). Bluesky's Ahead: A multi-facility collaboration for an a la carte software project for data acquisition and management. *Synchrotron Radiation News*, 32(3), 19–22. <https://doi.org/10.1080/08940886.2019.1608121>
- Binder, J. M., Stark, A., Tomek, N., & others. (2017). Qudi: A modular python suite for experiment control and data processing. *SoftwareX*, 6, 85–90. <https://doi.org/10.1016/j.softx.2017.02.001>
- Gamma, E., Helm, R., Johnson, R., & Vlissides, J. (1994). *Design patterns: Elements of reusable object-oriented software*. Addison-Wesley. ISBN: 978-0201633610
- Hayashi, S., Han, D., Tsuji, H., Ishizaka, K., & Nakamura, A. (2026). Development of precession lorentz transmission electron microscopy. *Ultramicroscopy*, 280, 114276. <https://doi.org/10.1016/j.ultramic.2025.114276>
- Koga, J., Chiashi, Y., Nakamura, A., Akiba, T., Takahashi, H., Shimojima, T., Ishiwata, S., & Ishizaka, K. (2024). Unusual photoinduced crystal structure dynamics in TaTe₂ with double zigzag chain superstructure. *Applied Physics Express*, 17(4), 042007. <https://doi.org/10.35848/1882-0786/ad3b61>
- lys_em Developers. (2026). *Lys_em: Python package for electron microscopy control and data analysis*. https://github.com/a-tock/lys_em
- lys_fem Developers. (2026). *Lys_fem: Python package for finite-element modeling and experiment integration*. https://github.com/lys-devel/lys_fem
- Melnikov, A., Vedkal, A., Ishchenko, A., & Veber, S. (2025). Atomize: A modular software for control and automation of scientific and industrial instruments. *Journal of Open Research Software*, 13, 26. <https://doi.org/10.5334/jors.594>
- Nakamura, A. (2023). lys: Interactive Multi-Dimensional Data Analysis and Visualization Platform. *Journal of Open Source Software*, 8(92), 5869. <https://doi.org/10.21105/joss.05869>
- Nakamura, A., Shimojima, T., Chiashi, Y., Kamitani, M., Sakai, H., Ishiwata, S., Li, H., & Ishizaka, K. (2020). Nanoscale Imaging of Unusual Photoacoustic Waves in Thin Flake VTe₂. *Nano Letters*, 20(7), 4932–4938. <https://doi.org/10.1021/acs.nanolett.0c01006>
- Nakamura, A., Shimojima, T., & Ishizaka, K. (2021). Finite-element Simulation of Photoinduced Strain Dynamics in Silicon Thin Plates. *Structural Dynamics*, 8(2), 024103. <https://doi.org/10.1063/4.0000059>
- Nakamura, A., Shimojima, T., & Ishizaka, K. (2022). Visualizing Optically-Induced Strains by Five-Dimensional Ultrafast Electron Microscopy. *Faraday Discussions*, 237, 27–39. <https://doi.org/10.1039/D2FD00062H>
- Nakamura, A., Shimojima, T., & Ishizaka, K. (2023). Characterizing an Optically Induced Sub-micrometer Gigahertz Acoustic Wave in a Silicon Thin Plate. *Nano Letters*, 23(7), 2490–2495. <https://doi.org/10.1021/acs.nanolett.2c03938>
- National Instruments Corporation. (2024). *LabVIEW*. National Instruments Corporation. <https://www.ni.com/en-us/shop/labview.html>
- Nielsen, J. H., Astafev, M., Nielsen, W. H. P., & others. (2025). *QCoDeS: Modular data acquisition framework* (Version v0.54.2). <https://doi.org/10.5281/zenodo.17459861>
- PyMeasure Developers. (2024). *PyMeasure: Scientific measurement library for instruments, experiments, and live-plotting*. (Version 0.14.0). Zenodo. <https://doi.org/10.5281/zenodo.11241567>

- Shimojima, T., Nakamura, A., & Ishizaka, K. (2023a). Development of Five-Dimensional Scanning Transmission Electron Microscopy. *Review of Scientific Instruments*, 94(2), 023705. <https://doi.org/10.1063/5.0106517>
- Shimojima, T., Nakamura, A., & Ishizaka, K. (2023b). Development and Applications of Ultrafast Transmission Electron Microscopy. *Microscopy*, 72(4), 287–298. <https://doi.org/10.1093/jmicro/dfad021>
- Shimojima, T., Nakamura, A., Yu, X., Karube, K., Taguchi, Y., Tokura, Y., & Ishizaka, K. (2021). Nano-to-Micro Spatiotemporal Imaging of Magnetic Skyrmion's Life Cycle. *Science Advances*, 7(25), eabg1322. <https://doi.org/10.1126/sciadv.abg1322>
- Shiratori, T., Koga, J., Shimojima, T., Ishizaka, K., & Nakamura, A. (2024). Development of ultrafast four-dimensional precession electron diffraction. *Ultramicroscopy*, 267, 114064. <https://doi.org/10.1016/j.ultramic.2024.114064>
- The MathWorks, Inc. (2024). *MATLAB instrument control toolbox*. The MathWorks, Inc. <https://www.mathworks.com/products/instrument.html>
- Törring, J. T. (2019). *fsc2: A program for controlling spectrometers*. <https://users.physik.fu-berlin.de/~jtt/fsc2.phtml>
- Weber, S. J. (2021). PyMoDAQ: An open-source python-based software for modular data acquisition. *Review of Scientific Instruments*, 92(4), 045104. <https://doi.org/10.1063/5.0032116>