

Cross Modality Image Translation In Medical Imaging Using Generative Frameworks

Giulia Romoli^{a,*}, Alessia Capoccia^b, Filippo Ruffini^{a,b}, Francesco Di Feola^a, Luca Boldrini^{c,d}, Arturo Chiti^{e,f}, Renato Cuocolo^g, Tugba Akinci D'Antonoli^{h,i}, Fatemeh Darvizeh^j, Marcello Di Pumpo^{k,l}, Bradley J. Erickson^m, Liu Fangⁿ, Deborah Fazzini^j, Paola Feraco^o, Fabrizia Gelardi^e, Francesco Gossetti^p, Ana Isabel Hernáiz Ferrer^j, Michail E. Klontzas^{q,r}, Seyedmehdi Payabvash^s, Katrine Riklund^t, Sara N. Strandberg^t, Valerio Guarrasi^b and Paolo Soda^{a,b,*}

^aDepartment of Diagnostics and Intervention, Radiation Physics, Biomedical Engineering, Umeå University, Universitetstorget, 4, Umeå, 90187, Sweden

^bUnit of Artificial Intelligence and Computer Systems, Department of Engineering, Università Campus Bio-Medico di Roma, Via Alvaro del Portillo, 21, Roma, 00128, Italy

^cFondazione Policlinico Universitario "A. Gemelli" IRCCS, Largo Agostino Gemelli, 8, Roma, 00168, Italy

^dUniversità Cattolica del Sacro Cuore, Largo Francesco Vito, 1, Roma, 00168, Italy

^eIRCCS Ospedale San Raffaele, via Olgettina, 60, Milano, 20132, Italy

^fVita-Salute San Raffaele University, Via Olgettina, 58, Milano, 20132, Italy

^gDepartment of Medicine, Surgery and Dentistry, University of Salerno, Via Salvador Allende, 43, Baronissi, 84081, Italy

^hDivision of Diagnostic and Interventional Neuroradiology, Department of Radiology, University Hospital Basel, Spitalstrasse, 21, Basel, 4031, Switzerland

ⁱDepartment of Pediatric Radiology, University Children's Hospital Basel, Basel, Switzerland

^jCentro Diagnostico Italiano (CDI), Via Simone Saint Bon, 20, Milano, 20147, Italy

^kDepartment of Life Science and Public Health, Università Cattolica del Sacro Cuore, Largo Francesco Vito, 1, Roma, 00168, Italy

^lItalian Society for Artificial Intelligence in Medicine (SIAM), Roma, Italy

^mMayo Clinic, 200 First Street SW, Rochester, 55905, Minnesota, USA

ⁿAthinoula A. Martinos Center for Biomedical Imaging, 149 13th Street, Charlestown, 02129, Massachusetts, USA

^oCentro interdipartimentale di scienze mediche (CISMed), Università degli studi di Trento, Via S. Maddalena, 1, Trento, 38123, Italy

^pSapienza Università di Roma, Piazzale Aldo Moro, 5, Roma, 00185, Italy

^qArtificial Intelligence and Translational Imaging (ATI) Lab, Department of Radiology, School of Medicine, University of Crete, 71003, Greece

^rDivision of Radiology, Department of Clinical Science, Intervention and Technology (CLINTEC), Karolinska Institute, Huddinge, Sweden

^sColumbia University Medical Center, 530 West 166th St, New York, 10032, New York, USA

^tDepartment of Diagnostics and intervention, Diagnostic radiology, Umeå University, Universitetstorget, 4, Umeå, 90187, Sweden

ARTICLE INFO

Keywords:

Oncological imaging
Image-to-image translation
Virtual scanning
Generative models
Benchmarking
Visual Turing test

ABSTRACT

Medical image-to-image (I2I) translation enables virtual scanning, i.e. the synthesis of a target imaging modality from a source one without additional acquisitions. Despite growing interest, most proposed methods operate on 2D slices, are evaluated on isolated tasks with different experimental set-ups and lack clinical validation. The primary contribution of this work is a reproducible, standardized comparative evaluation of 3D I2I translation methods in oncological imaging, designed to standardize preprocessing, splitting, inference, and multi-level evaluation across heterogeneous clinical tasks. Within this framework, we compare seven generative models, three Generative Adversarial Networks (GANs: Pix2Pix, CycleGAN, SRGAN) and four latent generative models (Latent Diffusion Model, Latent Diffusion Model+ControlNet, Brownian Bridge, Flow Matching), across eleven datasets spanning three anatomical regions (head/neck, lung, pelvis) and four translation directions (cone-beam CT to CT, MRI to CT, CT to PET, MRI T2-weighted to T2-FLAIR), for a total of 77 experiments under uniform training, inference, and evaluation conditions. The results show that GANs outperform latent generative models across all tasks, with SRGAN achieving statistically significant superiority. Our lesion-level analysis reveals that all models struggle with small lesions and that, in CT to PET synthesis, models reproduce lesion shape more reliably than absolute uptake-related intensity. We also performed a Visual Turing test administered to 17 physicians, including 15 radiologists, which shows near-chance classification accuracy (56.7%), confirming that synthetic volumes are largely indistinguishable from real acquisitions, while exposing a dissociation between quantitative metrics and clinical preference.

1. Introduction

Cancer is one of the most significant global health challenges, with imaging playing a key role in screening, diagnosis and treatment. Despite progress, access to advanced imaging scanners, including Magnetic Resonance Imaging (MRI), Positron Emission Tomography (PET), and Computed Tomography (CT), is still limited [1], which contributes to substantial disparities in cancer care. While high-income countries often face the overuse of low-value imaging, resulting in increased financial costs and pressure on radiology capacity, many low- and middle-income countries still lack adequate access to scanners and trained personnel [2]. This dual imbalance results either in unnecessary examinations and longer waiting lists, or in delayed diagnosis and treatment, ultimately leading to advanced disease at diagnosis, as well as avoidable morbidity, mortality and public health inequities. Beyond accessibility, imaging also entails a procedural burden for patients, including exposure to ionizing radiation and the administration of contrast agents or radio-tracers, with cumulative risks over repeated examinations.

Against this background, recent advances in generative Artificial Intelligence (AI) have facilitated progress in Image-to-Image (I2I) translation, a paradigm in which a model learns to transform an image from a source domain into a corresponding image in a target domain. In medical imaging, I2I translation provides the methodological foundation for *virtual scanning*, namely the synthesis of a target imaging modality from an already available source modality without requiring an additional physical acquisition. I2I translation across MRI, CT, and PET is especially compelling in oncology because these modalities provide complementary anatomical, structural, and functional information for cancer diagnosis, treatment planning, and response assessment [3, 4].

Although I2I approaches are advancing rapidly, they still face three major limitations that hinder their clinical applicability. The first concerns data dimensionality: most existing I2I methods are inherently two-dimensional, operating on individual slices rather than full image volumes [5]. While computationally efficient, two-dimensional processing neglects inter-slice context, with discontinuities that compromise volumetric anatomical coherence, essential for clinically reliable synthesis. Three-dimensional models are better suited to preserving spatial consistency and capturing long-range anatomical dependencies. However, their adoption is limited by the high computational and memory costs of processing high-resolution medical volumes.

The second limitation lies in the scarcity of studies addressing inter-modality translation. Unlike intra-modality synthesis, where the source and target images are acquired using similar principles, inter-modality tasks require models to overcome significant differences in imaging physics, contrast mechanisms and semantic content [3]. This makes cross-modality synthesis more challenging, but also more clinically valuable, particularly in oncology, where complementary imaging information is often crucial for patient management.

The third limitation relates to evaluation practices [6]: current methods are usually evaluated using a small number of datasets that are often selected in an ad hoc manner, limiting clarity on how well the methods generalize across patient populations, anatomical regions and translation tasks. Even when public datasets are used, experimental protocols, including pre-processing steps and train-test splits, are frequently under-specified, thereby limiting reproducibility and preventing rigorous cross-study comparisons. Moreover, most studies evaluate individual tasks in isolation, with little effort toward unified benchmarking under uniform experimental conditions. These shortcomings are particularly significant in oncology, where virtual scanning must achieve more than just generating visually plausible images. It must also preserve clinically meaningful information, such as lesion conspicuity, tumor boundary fidelity and anatomical consistency, in regions that directly impact diagnosis, treatment planning and monitoring. Addressing these limitations is therefore essential to move virtual scanning beyond proof-of-concept demonstrations and towards the development of robust, clinically useful tools for augmented cancer care, that align with the broader vision of *augmented intelligence*, in which AI is designed to support, rather than replace, clinical expertise [7].

To address the challenges mentioned so far, this work proposes a unified methodology for 3D I2I translation that isolates model performance from pipeline variability. Indeed, by standardizing pre-processing, volumetric inference, and evaluation across diverse oncological datasets, it ensures that results reflect architectural merit rather than experimental heterogeneity. Beyond standard quantitative metrics, the benchmark incorporates a dedicated lesion-level

*Corresponding author

 giulia.romoli@umu.se (G. Romoli); paolo.soda@umu.se (P. Soda)

ORCID(s): 0000-0003-1624-1709 (G. Romoli); 0000-0003-2621-072X (P. Soda)

analysis and a reader study with 17 clinical experts, enabling a systematic and reproducible evaluation of state-of-the-art generative models. Specifically, we present here the following contributions:

- C1) We design and release an open-source, modular benchmarking framework for 3D medical I2I translation. The framework provides a unified pipeline, from data ingestion and pre-processing through model training, sliding-window inference, and multi-level evaluation, that enables reproducible comparison of diverse generative models on volumetric medical data. Its plug-in architecture allows new models to be integrated, while the shared pre-processing, patch extraction, and stitching infrastructure remains fixed, isolating the effect of the generative component from confounding pipeline differences.
- C2) Using this framework, we conduct a comparative analysis of seven state-of-the-art generative models for medical 3D I2I translation across eleven datasets, focusing on three key anatomical regions: head/neck, lung, and pelvis; and four translation directions (cone-beam CT to CT, MRI to CT, CT to PET, MRI T2-weighted to T2-FLAIR), for a total of 77 experiments under identical training, inference, and evaluation conditions.
- C3) We evaluate results using metrics that capture not only image quality but also structural fidelity and potential clinical relevance; in particular, we perform a targeted evaluation of lesion regions to assess how faithfully pathological features are preserved in the generated images, stratified by lesion size and translation task.
- C4) We develop and release a web-based Visual Turing test platform for interactive volumetric evaluation, and administer a reader study with 17 physicians, including 15 radiologists, to assess whether the generated volumes are perceptually indistinguishable from real medical scans and to quantify the dissociation between standard image quality metrics and clinical preference.

2. Related works

In the last decade, I2I translation in medical imaging has been driven by Generative Adversarial Networks (GANs, [8]), which represent the most widely adopted framework due to their ability to produce high-fidelity, structurally consistent outputs. Over the past few years, diffusion-based models have emerged as a promising alternative, offering superior sample diversity and training stability, despite requiring higher computational costs [9]. A third approach, based on Variational Autoencoders (VAEs), has also been investigated. However, VAEs are generally regarded as more suitable for reconstruction and representation learning, rather than I2I translation, due to their propensity to generate blurry outputs, limiting practical applicability [10]. Accordingly, the remainder of this section focuses on GANs and diffusion-based methods.

Generative Adversarial Networks for Medical Image Translation. Our analysis of the literature shows that 2D slice-wise approaches dominate GAN-based medical I2I translation, where Pix2Pix [11] and CycleGAN [12] are the two prevailing paradigms, addressing paired and unpaired settings, respectively. Numerous studies apply them to cross-modality synthesis such as MRI-to-CT [13, 14], CT-to-MRI [15], and multi-contrast MRI generation [16, 17], processing each axial cross-section independently. Despite their effectiveness, 2D approaches ignore inter-slice spatial dependencies and produce visible discontinuities along coronal and sagittal planes. Since the majority of imaging modalities produced in clinical settings are inherently 3D, volumetric processing is better suited to preserve anatomical coherence. Some works have therefore extended GAN architectures to operate directly on 3D volumes. For instance, EA-GAN [18] incorporates 3D convolutions with edge-aware learning to better preserve fine structural boundaries, while 3D-MedTranCSGAN [19] employs a cascade of 3D-U-Blocks, enabling iterative refinement through successive encoding-decoding pairs over the full volume. These approaches process entire volumes in a single forward pass, thereby capturing global context. However, extending GAN architectures to three dimensions introduces substantial computational costs due to the cubic growth in memory and operations, as well as challenges related to information compression, as encoders must distill volumetric spatial structure into compact latent representations without losing clinically relevant detail. A solution is dividing the volume into patches that are processed independently and reassembled at inference time. Guarrasi et al. [20], for example, train 3D variants of Pix2Pix and CycleGAN on volumetric patches, employing sliding-window inference with overlap-averaging.

Beyond memory considerations, most existing methods operate at a single spatial resolution and thus fail to explicitly leverage the hierarchical, multi-scale structure of medical volumetric data. For this reason, Ha et al. [21] propose SRGAN, a multi-resolution 3D GAN framework in which both generator and discriminator operate across multiple spatial scales. The generator adopts a Dense-Attention UNet architecture, combining residual dense blocks

for hierarchical feature extraction with attention mechanisms to fuse information across resolutions. The discriminator, also UNet-based, performs voxel-wise classification at each scale. Experiments across diverse modalities (MRI-to-CT, intra-modal MRI and CT) and body regions (brain and pelvis) confirm the effectiveness of this framework, which outperforms EA-GAN in translation quality.

From denoising diffusion to latent image translation. While being an effective paradigm for I2I translation in medical imaging, GAN-based pipelines still face well-documented limitations, like training instability, mode collapse, and sensitivity to architectural and hyperparameter choices that can degrade output diversity and reliability [8]. These shortcomings have motivated the exploration of alternative generative frameworks, among which denoising diffusion probabilistic models have recently emerged as a promising direction. Denoising Diffusion Probabilistic Models (DDPMs, [22]) ground their generative mechanism in the progressive diffusion from noise to target data distribution, where a neural network learns to reverse a Markovian noising chain step-by-step [23, 24]. In its unconditional form, this framework generates samples from pure noise without external guidance. For I2I translation, however, the noise-estimation network must incorporate information from the source modality, which conditions the denoising trajectory. Therefore, the model still maps noise toward the target distribution, but the source image steers this mapping to produce a synthetic output modality that is also anatomically coherent with the source.

The DDPM denoising formulation introduces a practical bottleneck: multi-step sampling demands far more computation than single-pass GAN inference, and this cost multiplies when moving from 2D slices to full 3D volumes. Whether this overhead translates into better synthesis quality remains task-dependent. Moschetto et al. [25] investigated this question in the first unified benchmark comparing GANs, conditional DDPMs, and flow matching models for 2D T1w-to-T2w brain MRI translation across three public datasets. Their results show that Pix2Pix outperforms all diffusion and flow variants in both structural fidelity and inference speed, a gap the authors attribute to the limited size of the 2D datasets used [26, 27]. In volumetric settings, where inter-slice consistency and long-range spatial context become critical, diffusion models narrow the gap with GANs and, in several tasks, surpass them. Pan et al. [28] proposed MC-IDDPM, the first conditional improved DDPM for 3D MRI-to-CT synthesis. On private brain and prostate datasets, MC-IDDPM outperforms GAN-based and standard DDPM baselines across all image-quality metrics, though inference remains orders of magnitude slower than single-pass alternatives.

The cubic memory cost associated with voxel-space diffusion has driven two principal alternatives: 2D slice-based architectures augmented with inter-slice consistency mechanisms [29, 30]; and 3D Latent Diffusion Models (LDMs), which mitigate memory demands by operating in a compressed latent space instead of the original voxel domain. A notable example of the latter is ALDM [31], a 3D LDM which performs denoising in a VQGAN-compressed latent space [32], enabling full-volume MRI translation without patch cropping. On BraTS2021, ALDM surpasses all 2D and 3D GAN baselines and produces artifact-free sagittal views that slice-wise methods fail to preserve. Sargood et al. [33] extended latent diffusion to cross-modality tasks with CoCoLIT, a ControlNet-conditioned [34] framework for 3D MRI-to-amyloid-PET synthesis. On access-restricted brain datasets, CoCoLIT outperforms all baselines in both image fidelity and amyloid-positivity classification, demonstrating that latent diffusion can capture clinically meaningful pathological signals in the synthesized modality.

Clinical and oncological relevance. These works trace a clear progression from 2D slice-wise approaches to native 3D solutions, which eliminate inter-slice artifacts but at a higher computational cost. Although a 3D generative model can reproduce the statistical distribution of a target modality with high fidelity, the synthesized image must preserve the underlying anatomical and functional information, a requirement that becomes especially critical in inter-modality I2I translation. Pan et al. [28] validated MRI-to-CT synthesis using MC-IDDPM for brain and prostate tumors, showing that dose distributions computed on generated CT scans match those from planning CTs within clinically acceptable margins. Graf et al. [35] demonstrated that 3D Pix2Pix and Denoising Diffusion Implicit Models (DDIM) for MRI-to-CT translation of the spine produce segmentation masks accurate enough to support biomechanical modeling, a prerequisite for surgical planning and vertebral fracture assessment in metastatic disease. The translation problem changes substantially when the target modality is functional rather than anatomical, like in MRI-to-PET and CT-to-PET synthesis. Rajagopal et al. [36] showed that synthetic FDG-PET generated from whole-body MRI can replace real PET sinograms in SUV quantification experiments with synthetic lesion insertion, supporting the development and qualification of PET/MRI reconstruction algorithms without requiring new patient acquisitions. Sargood et al. [33] demonstrated that latent diffusion can synthesize amyloid-PET scans from structural MRI with sufficient accuracy to classify amyloid positivity in Alzheimer's disease screening, a task with direct implications for early oncological




Figure 1: I2I translation tasks. Overview of paired I2I translation tasks selected for this study, grouped by anatomical region (lung, A; head/neck, B; and pelvis, C). Triangle vertices represent the three imaging modalities (CT, MRI, and PET). Inter-modality translations are represented by arrows between vertices, while intra-modality ones are indicated by self-loops. Arrow colors are assigned based on clinical relevance and public dataset availability: green arrows denote tasks included in this benchmark, with the corresponding dataset name reported alongside; yellow arrows indicate clinically relevant directions for which no publicly accessible paired dataset was identified; red arrows denote synthesis directions considered clinically not relevant in the corresponding anatomical context.

differential diagnosis of brain lesions. Guarrasi et al. [20] addressed CT-to-PET translation by segmenting whole-body CT volumes into anatomical regions and training region-specific Pix2Pix, an approach that acknowledges the heterogeneity of tracer bio-distribution across body regions.

Open Challenges in Medical I2I Translation. Despite these advances, there are still four issues with medical imaging translation. First, not all I2I tasks carry equal oncological relevance. MRI-to-CT synthesis for radiotherapy planning and MRI/CT-to-PET translation for staging and response monitoring represent the most clinically motivated directions, while intra-modal MRI and CT synthesis serve complementary but narrower roles [37]. Second, a pronounced anatomical bias persists. Brain dominates the landscape, followed by spine and lung, while whole-body approaches either decompose the anatomy into separate regions or reveal that a single model fails to capture full-body heterogeneity [20]. Third, dataset availability further constrains progress, with reproducibility limited by the predominance of private institutional data. Only a few studies rely on multiple public benchmarks. Initiatives such as the SynthRAD challenge [38] represent a step toward standardized evaluation, but broader multi-site, multi-task datasets with clinical endpoint validation remain a critical gap for translating these methods into oncological practice. Finally, generalization remains limited. Only a few works attempt multiple translation tasks within a single model, and none has demonstrated stable accuracy across modality pairs with different physical characteristics. The methodological novelty of this work lies in addressing this comparability gap by controlling the main experimental degrees of freedom across tasks, providing a structure that can be reused, extended, and stress-tested as new datasets and models become available.

3. Materials and Methods

This work establishes a standardized benchmark for 3D I2I translation in oncology, which considers three imaging modalities (CT, MRI, and PET) and three anatomical regions (lung, head/neck, and pelvis). We selected these anatomical regions for their high incidence and mortality rates [39, 40], and because multi-modal, 3D imaging protocols are routinely employed in their clinical workup. Lung cancer alone is responsible for nearly one in five cancer deaths in the United States. Brain cancer ranks among the most lethal malignancies relative to their incidence [41]. Pelvic tumors, including prostate, colon-rectal, and cervical, represent some of the most frequently diagnosed malignancies. Across the selected anatomical regions, clinical protocols routinely combine CT, MRI, and PET to obtain complementary anatomical, structural, and metabolic information, producing co-registered, multi-modal volumes that facilitate the construction of paired training datasets for 3D image synthesis. Tumors primarily diagnosed through 2D imaging, such as breast cancer (mammography) and skin cancer (dermoscopy), fall outside the scope of this benchmark despite their high incidence, although emerging 3D modalities such as breast MRI represent a natural future extension. For more details about the selected imaging modalities, and the epidemiology and clinical management of these cancers, the interested readers can refer to Supplementary Materials, Sections A and B.

We graphically represent the imaging modalities and the I2I translation tasks for each body region in Figure 1, where triangle's vertices correspond to the three modalities (CT, MRI, and PET), while edges represent translations

Table 1

Overview of the datasets used in this study. Each row specifies the I2I translation task, representative downstream applications, the source dataset, the anatomical region, and the total number of patients available to train the I2I model, with the test set size in parentheses. Details on the data splitting strategy are provided in Supplementary Materials, Section C.

Task (Source-to-Target)	Applications	Dataset	District	Patients (Test)
MRI T1w-to-CT	MRI-only radiotherapy planning: eliminate CT acquisition for dose calculation in brain, head/neck, and pelvic cancers	Synthrad23 [42]	Pelvis	180 (45)
			Head/Neck	180 (45)
		Synthrad25 [38]	Lung	182 (46)
			Head/Neck	156 (39)
CBCT-to-CT	Online adaptive radiotherapy: enable dose recalculation from daily CBCT without additional full-dose CT	Synthrad23 [42]	Pelvis	180 (45)
			Head/Neck	180 (45)
		Synthrad25 [38]	Lung	258 (65)
			Head/Neck	260 (65)
MRI T2w-to-T2f	Multi-parametric MRI protocol completion: recover missing FLAIR sequences due to acquisition failures or time constraints in brain tumor imaging	BraTS23 [43]	Brain	1251 (313)
CT-to-PET	Reduce radiotracer exposure in lung cancer follow-up	autoPET [44]	Lung	150 (38)
		EnhancePET [45]		583 (146)

between the connected pair, with arrow direction indicating the source-to-target mapping. Not all directions are equally relevant: we classify a direction in each edge as either a benchmark task with available public paired data (green arrows), a clinically relevant direction currently lacking public paired data (yellow arrows), or a clinically not relevant direction (red arrows). Public paired datasets were specifically searched with the condition that they contain oncological cases. More specifically, here we consider an I2I task as clinically justified when its acquisition entails additional cost, or radiation exposure or logistical burden for the patient. Conversely, we mark a direction as not clinically meaningful when the target modality is the principal imaging modality for the specific anatomical region (like MRI for pelvis and head/neck), or when the source lacks the structural information needed to reliably infer the target (e.g., PET-to-CT or PET-to-MRI). In this benchmark, we restrict our evaluation to the core tasks (green arrows), as the availability of public paired data ensures a reproducible comparison across methods. We identified 5 publicly accessible datasets providing oncological cases with paired multi-modal volumes (Table 1): SynthRAD2023 [42], and its extension SynthRAD2025 [38]; BraTS2023 (BraTS23, [43]); autoPET [44]; and *ENHANCE.PET 1.6k* (EnhancePET, [45]). These collections are widely adopted in the medical image synthesis literature and, together, span two inter-modality (MRI-to-CT, CT-to-PET) and two intra-modality (Cone Beam CT, CBCT-to-CT, T2w-to-T2f) translation directions. Further information on such datasets are in Supplementary Materials, Section C. Although CT-to-PET translation is equally well-motivated in the pelvis and in the head/neck, no publicly available paired datasets were identified for these regions. Intra-modal MRI translation in the pelvis, although clinically motivated for multi-parametric prostate and rectal cancer protocols ([3]), lacks public paired data and is therefore not included. In the thorax, MRI plays a secondary role, being primarily reserved for characterizing mediastinal invasion, detecting brain metastases, and evaluating adrenal lesions during staging workups, limiting the clinical demand for CT-to-MRI translation. PET synthesis from MRI remains an emerging research direction for which no publicly accessible paired dataset is currently available.

For each dataset, we reserved 25% of subjects as a fixed test set (Table 1). From the remaining training partition, 5 subjects were held out as a shared validation subset and used across all corresponding task configurations.

3.1. Models selection

We consider seven generative models spanning two architectural families: GANs, including Pix2Pix [11], CycleGAN [12], and SRGAN [21]; and latent generative models, including LDM-Palette (LDM, [46]), LDM-Palette augmented with ControlNet [34] (LDM+ControlNet), Brownian Bridge diffusion model (BBridge, [47]) and flow matching (FlowM, [48]); all receive as input volumetric patches of size $96 \times 96 \times 96$ voxels. The seven models were




Figure 2: The proposed benchmark experiments. Each of the 11 dataset configurations (left) is evaluated against all 7 generative models (centre) using 2 evaluation metrics (right), yielding 77 experimental combinations in total.

selected as widely adopted for I2I translation in the medical imaging literature. Pix2Pix and CycleGAN are GAN baselines in the vast majority of medical I2I studies [3], while SRGAN is representative of perceptual-loss-based GANs. Among latent generative models, LDM-based approaches (with and without ControlNet conditioning) represent the current standard for diffusion-based medical image synthesis; the Brownian Bridge diffusion model has gained increasing interest in medical MRI translation tasks [49]; and flow matching has recently emerged as a competitive alternative in medical imaging benchmarks [50]. By combining the 11 dataset configurations with the 7 generative models, we obtain a total of 77 experiments, as illustrated in Figure 2.

A central challenge in volumetric medical image synthesis is that full-resolution 3D volumes cannot be processed in a single forward pass due to GPU memory constraints. While patch-based processing alleviates memory constraints, it introduces stitching artifacts, visible discontinuities at patch boundaries caused by the absence of surrounding context during individual patch prediction [51]. To address this issue, all models use a sliding-window during testing and inference, which overcomes the stitching problem through two complementary mechanisms: overlapping patches and Gaussian blending. We applied the same pre-processing pipeline to all datasets, extracting patches of size $96 \times 96 \times 96$ voxels per volume, for both source and target modalities (see Supplementary Materials, Section D for further details). Patches are extracted with an overlap fraction of 0.625, so that each voxel in the volume is covered by multiple overlapping patches. The contributions of these patches are then combined using a Gaussian blending with a sigma scale of 0.125, which assigns higher weights to predictions at the center of each patch and progressively lower weights towards the edges, producing smooth and spatially consistent predictions across the full volume without hard boundaries between patches. The Gaussian blending with a sigma scale of 0.125 follows the default configuration of MONAI’s sliding-window inference [91], which applies Gaussian importance weighting to suppress boundary artifacts as originally proposed by Isensee et al. [51] in nnU-Net.

The three GAN models used in this benchmark (Pix2Pix, CycleGAN, and SRGAN), follow the standard adversarial training paradigm in which a generator and a discriminator network are optimized simultaneously. The four latent generative models of this benchmark include diffusion and flow matching. They operate in the compressed latent space computed by a shared variational autoencoder (VAE), pre-trained independently for each dataset and kept frozen during the training of the downstream model. The VAE encodes $96 \times 96 \times 96$ voxel patches into $24 \times 24 \times 24$ latent volumes,

achieving a $\times 4$ spatial compression factor. As diffusion models, we implement LDM-Palette both with and without ControlNet conditioning. The model learns to iteratively denoise a target latent space conditioned on the source, using a DDPM noise schedule during training and DDIM sampling at inference. The BBridge diffusion model replaces the standard Gaussian noise prior with a Brownian Bridge process that interpolates directly between source and target latent spaces, eliminating the need to start from pure noise. Finally, flow matching takes a different approach, learning a deterministic velocity field that transports the source latent towards the target latent along straight-line trajectories via numerical ODE integration, without any stochastic component.

We trained all the models for 1000 epochs on a single NVIDIA A40 GPU (48 GB VRAM), using a batch size of three randomly sampled patches of $96 \times 96 \times 96$ voxels per iteration (as in [21]). Furthermore, to ensure reproducibility and fair comparison, we did not perform hyperparameter tuning for any model, as hyperparameter optimization is highly dataset-specific and is considered out of the scope of this work. This choice is also aligned with the *No Free Lunch* theorems for optimization, which state that no single algorithm or parameter configuration can be optimal across all possible problem domains [94]. In line with this theoretical result, it is also worth noting that empirical evidence has further shown that hyperparameter tuning does not necessarily yield substantial improvements over default configurations [95]. Full architectural details and training hyperparameters for each model are provided in Supplementary Materials, Section E.

3.2. Comparative methodology

Beyond the individual models described above, a central contribution of this work is the benchmarking framework itself, which we release as an open-source repository to serve as a reusable and extensible research instrument for the medical image synthesis community. We describe here the design principles that distinguish the framework from a collection of independently trained models. Specifically, this benchmark extends the line of investigation of [25] along four dimensions: (i) dimensionality (3D volumetric vs. 2D slice-wise processing); (ii) scope (4 translation directions across 3 body regions vs. a single intra-MRI task); (iii) scale (77 experiments across 11 dataset configurations vs. 3 datasets); and (iv) evaluation depth (lesion-level analysis and a Visual Turing test vs. standard image quality metrics only).

Unified pre-processing pipeline. All datasets pass through a single pre-processing pipeline (Supplementary Material, Section D, Figure 10) comprising body masking, voxel resampling, intensity clipping, normalization, spatial padding, foreground mask computation, mask intersection, and patch extraction. Each transformation records its parameters, enabling exact inversion at inference time so that generated volumes are mapped back to the original image space without residual interpolation artifacts. This invertibility is critical for clinical evaluation, where generated images must be interpretable in the native coordinate system of the source acquisition.

Shared patch-based 3D processing. All seven models operate on identical $96 \times 96 \times 96$ voxel patches extracted with the same sampling strategy. This eliminates a common confound in cross-architecture comparisons, where differences in input resolution, patch size, or sampling strategy can bias results independently of the generative model itself.

Sliding-window inference with Gaussian blending. At inference time, all models share the same stitching procedure: patches are extracted with an overlap fraction of 0.625 and reassembled using Gaussian blending (sigma scale of 0.125). This ensures that boundary artifacts are handled identically across architectures, preventing one model from appearing superior simply because of a more favorable stitching strategy.

Shared VAE backbone for controlled latent-model comparisons. The four latent generative models (LDM, LDM+ControlNet, BBridge, FlowM) share a single pre-trained VAE encoder-decoder, frozen during all downstream training. This design isolates the effect of the generative component (diffusion, flow matching, Brownian bridge) from the effect of the autoencoder, enabling fair attribution of performance differences to the generative mechanism rather than to differences in latent-space quality. It also exposes the VAE reference reconstruction level as an explicit, measurable upper bound on latent-model fidelity (Section 4.1).

Plug-in model interface and end-to-end reproducibility. The framework defines a standardized model interface, where any new generative architecture that conforms to it can be benchmarked against all existing baselines on all datasets without modifying the data pipeline, training loop, inference stitcher, or evaluation code. The complete pipeline is encoded in configuration files specifying data paths, model hyperparameters, and training schedules (Supplementary Material, Figure 11). Reproducing any of the 77 experiments requires a single command-line invocation with the corresponding configuration file.

3.3. Evaluation metrics

We quantitatively evaluate model performance using Peak Signal-to-Noise Ratio (PSNR) and Structural Similarity Index Measure (SSIM) [52]. For the autoPET and BraTS23 datasets, for which tumor masks are available, the same metrics are additionally computed within lesion regions and stratified according to lesion size. It is worth noting that such metrics do not fully capture perceptual image quality [53, 54]. For this reason, we also developed a Visual Turing test, which has emerged in the literature as an experimental tool to assess the quality perceived by medical experts [55]. While used in some works in 2D I2I medical synthesis [56, 57, 58, 59], to the best of our knowledge none has investigated the perceived quality of 3D synthetic medical images in intra- and inter-modality image translation. To this aim, we administered this test through a custom web-based platform with interactive volumetric visualization. The evaluation covered all synthesis tasks across the four best-performing generative models selected on the basis of quantitative metrics, two for each architectural family.

The platform, which we release as a standalone open-source tool alongside the benchmarking code, is designed to be reusable for future reader studies in medical image synthesis. It supports arbitrary NIfTI volumes, configurable question types, per-volume windowing controls, linked multi-planar navigation, and anonymous response collection. Studies are configured through a JSON specification file that defines the question sequence, volume assignments, and sanity-check placement, without requiring any modification to the platform code. To our knowledge, this is the first open-source platform specifically designed for interactive volumetric Visual Turing tests in medical imaging, addressing a gap in evaluation infrastructure that has limited the adoption of reader studies in the 3D synthesis literature.

In the Visual Turing test, we presented images and answer options in randomized order, for a total of 50 questions divided into three parts, each covering all translation tasks and anatomical regions to ensure balanced representation, as detailed in Table 2. **Part 1** presented 15 single images (7 real and 8 synthetic) and asked participants to classify them as real or AI-generated, as a binary test following the paradigm used in prior synthetic medical imaging studies [56, 57]. **Part 2** presented 18 side-by-side image pairs, and participants selected the one they preferred for diagnostic purposes, or indicated that no meaningful difference could be identified. We selected this pairwise comparison design in place of a Likert-scale rating because it yields substantially higher inter- and intra-observer agreement in subjective image quality assessment [54, 58, 59]. Both images in each pair were synthetic and always drawn from the same patient but generated by different models, so as to eliminate inter-patient variability from the comparison. Three pairs of identical images were embedded as sanity-checks to monitor response consistency. **Part 3** presented 17 image triplets, and participants assigned a rank from 1 (most realistic) to 3 (least realistic) to each one, providing an ordinal quality signal across models under matched conditions. In each triplet, images were either real or generated by different models. As for Part 2, all images within a triplet belonged to the same patient to eliminate inter-patient variability. Two sanity-check triplets were included in this part, each containing one real image and two outputs from the same model, to identify response bias in case of > 1 rank separation between two identical outputs. For each question, our tool displays each volume using the axial, sagittal, and coronal views, alongside a 3D surface reconstruction rendered using NiiVue v0.50.0 [92]. Adjustable brightness, contrast, and slice controls were provided for each volume (see Supplementary Material, Section F).

A total of 17 clinicians participated in the study via a personalised link, with free navigation across all questions: 15 radiologists and 2 physicians with a different clinical specialisation, from 12 different institutions and 5 different countries. In terms of experience, 12 participants reported more than 10 years of clinical practice, 3 between 5 and 10 years, and 2 less than 5 years. Regarding primary imaging modality experience, 8 reported working with CT and MRI, 4 with all three modalities (CT, MRI, and PET), 2 with MRI only, 1 with CT only, 1 with PET only, and 1 with CT and PET. All responses were anonymous and no personal data was collected.

4. Results and Discussion

In this section we analyze the results along three complementary levels, i.e., quantitative performance, lesion-level fidelity, and perceptual realism. Indeed, the first examines in section 4.1 how the selected models shape translation performance across modalities and anatomical regions. The second level of analysis shifts the evaluation from whole-volume metrics to fidelity in pathological regions (section 4.2), where synthesis errors should carry direct consequences on the clinical evaluation. The third direction of our investigation focuses on perceptual credibility under expert review by comparing quantitative rankings with the human perception collected via the Visual Turing test administered to the physicians (section 4.3).

Table 2

Distribution of the 50 questions across the three parts of the visual Turing test, organized by translation task and anatomical region. Part 1 (real vs. synthetic identification, 15 questions), Part 2 (pairwise preference, 15+3 sanity-check questions), and Part 3 (quality rating, 15+2 sanity-check questions). Each part covers all four translation categories and three anatomical regions to ensure balanced task representation across evaluation modes. The pelvis region has fewer questions across all parts, as fewer translation tasks in the benchmark map to this anatomical region. Similarly, CBCT-to-CT is assigned fewer questions than the other translations, reflecting the narrower domain gap of this intra-modality task.

Translation	Part 1				Part 2				Part 3			
	Head/Neck	Pelvis	Lung	Total	Head/Neck	Pelvis	Lung	Total	Head/Neck	Pelvis	Lung	Total
T1w-to-CT	1	2	1	4	1	2	1	4	1	2	1	4
CBCT-to-CT	1	1	1	3	1	1	1	3	1	1	1	3
T2w-to-T2f	4	—	—	4	4	—	—	4	4	—	—	4
CT-to-PET	—	—	4	4	—	—	4	4	—	—	4	4
Sanity-checks	—	—	—	—	1	1	1	3	1	—	1	2
Total	6	3	6	15	7	4	7	18	7	3	7	17

Table 3

PSNR (dB) and SSIM results across datasets, tasks, and models (mean $\pm$ std). Bold values indicate the best result per row; underlined values indicate the second best. The rightmost column reports the VAE reconstruction level, i.e. the fidelity of the autoencoder alone before any generative modeling, which represents a reference for all latent-based architectures.

	Dataset	Task	District	SRGAN	Pix2Pix	CycleGAN	BBridge	LDM	LDM+CN	FlowM	VAE
PSNR	Synthrad23	T1w-to-CT	Pelvis	29.15 $\pm$ 1.64	27.48 $\pm$ 1.95	27.04 $\pm$ 1.77	26.08 $\pm$ 1.44	23.88 $\pm$ 2.11	24.25 $\pm$ 2.08	25.94 $\pm$ 1.37	31.07 $\pm$ 1.57
		CBCT-to-CT		29.83 $\pm$ 2.29	28.40 $\pm$ 1.60	28.61 $\pm$ 2.72	24.84 $\pm$ 1.16	23.76 $\pm$ 1.30	23.73 $\pm$ 1.47	24.78 $\pm$ 1.21	24.22 $\pm$ 3.58
		T1w-to-CT	Head/Neck	27.48 $\pm$ 1.47	23.03 $\pm$ 0.95	24.77 $\pm$ 1.54	24.99 $\pm$ 0.85	24.96 $\pm$ 0.80	25.24 $\pm$ 0.88	24.82 $\pm$ 0.83	29.33 $\pm$ 0.88
		CBCT-to-CT		28.88 $\pm$ 2.05	26.98 $\pm$ 1.84	<u>28.21 $\pm$ 2.60</u>	25.43 $\pm$ 1.65	23.98 $\pm$ 2.20	24.25 $\pm$ 2.30	25.68 $\pm$ 1.84	26.66 $\pm$ 1.35
	BraTS23	T2w-to-T2f	Brain	25.28 $\pm$ 2.90	23.89 $\pm$ 2.72	23.52 $\pm$ 3.02	23.00 $\pm$ 2.67	20.84 $\pm$ 1.97	20.83 $\pm$ 1.95	23.55 $\pm$ 2.56	28.97 $\pm$ 1.55
	autoPET	CT-to-PET	Lung	34.21 $\pm$ 3.19	31.68 $\pm$ 2.68	<u>33.41 $\pm$ 2.84</u>	27.78 $\pm$ 1.16	28.33 $\pm$ 1.26	28.50 $\pm$ 1.50	27.89 $\pm$ 1.11	35.10 $\pm$ 1.78
	Synthrad25	T1w-to-CT	Lung	25.74 $\pm$ 1.51	23.17 $\pm$ 1.84	<u>24.39 $\pm$ 1.67</u>	22.68 $\pm$ 1.46	20.88 $\pm$ 1.64	20.87 $\pm$ 1.76	21.83 $\pm$ 1.82	26.84 $\pm$ 3.01
		CBCT-to-CT		28.07 $\pm$ 1.98	26.54 $\pm$ 1.76	<u>27.49 $\pm$ 1.95</u>	24.96 $\pm$ 1.97	24.97 $\pm$ 2.01	25.07 $\pm$ 2.03	24.90 $\pm$ 2.12	26.35 $\pm$ 2.56
		T1w-to-CT	Head/Neck	26.32 $\pm$ 3.01	22.77 $\pm$ 1.89	<u>24.45 $\pm$ 2.39</u>	21.78 $\pm$ 1.47	22.13 $\pm$ 1.50	22.18 $\pm$ 1.53	21.42 $\pm$ 1.45	22.66 $\pm$ 1.77
		CBCT-to-CT		29.50 $\pm$ 1.70	27.10 $\pm$ 1.53	<u>28.65 $\pm$ 1.66</u>	26.45 $\pm$ 1.53	26.54 $\pm$ 1.77	26.66 $\pm$ 1.77	26.55 $\pm$ 1.92	29.61 $\pm$ 2.19
SSIM	EnhancePET	CT-to-PET	Lung	42.77 $\pm$ 7.53	41.16 $\pm$ 6.84	<u>41.21 $\pm$ 7.41</u>	33.71 $\pm$ 3.09	33.79 $\pm$ 3.25	35.77 $\pm$ 2.63	35.63 $\pm$ 2.38	39.46 $\pm$ 4.78
	Synthrad23	T1w-to-CT	Pelvis	0.88 $\pm$ 0.02	0.80 $\pm$ 0.04	<u>0.80 $\pm$ 0.03</u>	0.75 $\pm$ 0.05	0.69 $\pm$ 0.06	0.70 $\pm$ 0.06	0.73 $\pm$ 0.05	0.87 $\pm$ 0.05
		CBCT-to-CT		0.88 $\pm$ 0.03	0.74 $\pm$ 0.04	<u>0.85 $\pm$ 0.05</u>	0.76 $\pm$ 0.05	0.74 $\pm$ 0.07	0.74 $\pm$ 0.07	0.76 $\pm$ 0.05	0.75 $\pm$ 0.13
		T1w-to-CT	Head/Neck	0.85 $\pm$ 0.04	0.66 $\pm$ 0.03	<u>0.77 $\pm$ 0.05</u>	0.77 $\pm$ 0.04	0.76 $\pm$ 0.04	0.77 $\pm$ 0.04	0.77 $\pm$ 0.04	0.91 $\pm$ 0.04
		CBCT-to-CT		0.89 $\pm$ 0.04	0.81 $\pm$ 0.06	<u>0.88 $\pm$ 0.06</u>	0.77 $\pm$ 0.09	0.71 $\pm$ 0.11	0.74 $\pm$ 0.12	0.78 $\pm$ 0.09	0.81 $\pm$ 0.07
	BraTS23	T2w-to-T2f	Brain	0.83 $\pm$ 0.05	<u>0.76 $\pm$ 0.04</u>	0.75 $\pm$ 0.04	0.55 $\pm$ 0.07	0.52 $\pm$ 0.05	0.52 $\pm$ 0.06	0.59 $\pm$ 0.06	0.72 $\pm$ 0.06
	autoPET	CT-to-PET	Lung	0.91 $\pm$ 0.02	0.84 $\pm$ 0.03	<u>0.89 $\pm$ 0.02</u>	0.72 $\pm$ 0.06	0.75 $\pm$ 0.05	0.76 $\pm$ 0.05	0.74 $\pm$ 0.05	0.94 $\pm$ 0.03
	Synthrad25	T1w-to-CT	Lung	0.73 $\pm$ 0.05	0.57 $\pm$ 0.09	<u>0.66 $\pm$ 0.07</u>	0.59 $\pm$ 0.05	0.49 $\pm$ 0.08	0.49 $\pm$ 0.09	0.56 $\pm$ 0.06	0.81 $\pm$ 0.06
		CBCT-to-CT		0.80 $\pm$ 0.05	0.69 $\pm$ 0.08	<u>0.78 $\pm$ 0.06</u>	0.65 $\pm$ 0.10	0.65 $\pm$ 0.11	0.65 $\pm$ 0.11	0.65 $\pm$ 0.11	0.74 $\pm$ 0.08
		T1w-to-CT	Head/Neck	0.76 $\pm$ 0.08	0.58 $\pm$ 0.09	<u>0.68 $\pm$ 0.06</u>	0.55 $\pm$ 0.05	0.54 $\pm$ 0.05	0.55 $\pm$ 0.05	0.53 $\pm$ 0.05	0.69 $\pm$ 0.11
		CBCT-to-CT		0.87 $\pm$ 0.04	0.75 $\pm$ 0.07	<u>0.84 $\pm$ 0.04</u>	0.71 $\pm$ 0.08	0.70 $\pm$ 0.09	0.71 $\pm$ 0.09	0.71 $\pm$ 0.09	0.80 $\pm$ 0.09
SSIM	EnhancePET	CT-to-PET	Lung	0.95 $\pm$ 0.03	0.93 $\pm$ 0.04	<u>0.94 $\pm$ 0.04</u>	0.75 $\pm$ 0.08	0.75 $\pm$ 0.08	0.83 $\pm$ 0.10	0.83 $\pm$ 0.09	0.90 $\pm$ 0.06

4.1. Quantitative Performance Across Modalities and Regions

Here we analyse whether a single generative framework can reliably serve multiple imaging workflows, or whether task-specific model selection remains necessary. To this aim, Table 3 compares GAN-based and latent generative architectures using PSNR and SSIM, supported by pairwise statistical testing, across all 11 translation tasks for the seven architectures under comparison. The radar charts in Figure 3 offer a complementary visual reading of the same results across all task-anatomy configurations simultaneously.

It is worth noting that SRGAN achieves the highest scores on every task for both metrics (it traces the outermost polygon across all tasks in both radar plots). Among the other GAN-based methods, CycleGAN ranks second or third, performing particularly well on intra-modality CBCT-to-CT tasks where the domain gap is narrower (from Table 3, 28.61 vs. 29.83 dB on Synthrad23, pelvis; 28.65 vs. 29.50 dB on Synthrad25, head/neck). Pix2Pix shows more variable performance across regions, with its accuracy being highly task-dependent. On pelvis tasks, it remains competitive, reaching 28.40 dB on Synthrad23 CBCT-to-CT and 0.80 SSIM on MRI-to-CT, but drops sharply on head/neck tasks, falling to 23.03 dB PSNR and 0.66 SSIM on Synthrad23 MRI-to-CT and to 22.77 dB PSNR and 0.58 SSIM on Synthrad25 MRI-to-CT. This inconsistency is especially evident in SSIM, where Pix2Pix ranges from 0.93




Figure 3: Quantitative performance. Radar charts (PSNR on the right and SSIM on the left) comparing seven I2I synthesis models across eleven task-anatomy configurations.

on EnhancePET down to 0.57 on Synthrad25 MRI-to-CT (lung), whereas CycleGAN exhibits a narrower range across the same tasks (0.94 to 0.66).

Latent generative models generally fall below their GAN counterparts, with the gap being most pronounced on structurally complex tasks such as BraTS23 T2w-to-T2f (e.g., SSIM of 0.52–0.59 versus 0.83 for SRGAN). PSNR gains rarely exceed 0.5 dB (e.g., 24.25 vs. 23.98 dB on Synthrad23 CBCT-to-CT, head/neck; 22.18 vs. 22.13 dB on Synthrad25 MRI-to-CT, head/neck) and SSIM differences are of 0.01 or less across the majority of tasks (e.g., 0.77 vs. 0.76 on Synthrad23 MRI-to-CT, head/neck; 0.71 vs. 0.70 on Synthrad25 CBCT-to-CT, head/neck). The only exception is the EnhancePET CT-to-PET task, where LDM+ControlNet reaches 35.77 dB and 0.83 SSIM compared to 33.79 dB and 0.75 for the plain LDM.

A key factor underlying the systematic underperformance of latent models is the reconstruction level imposed by the VAE backbone. As reported in Table 3, even before any generative modeling takes place, the VAE reconstruction alone introduces measurable fidelity loss across all datasets. For instance, on Synthrad23 MRI-to-CT (head/neck), the VAE achieves a PSNR of 29.33 dB and an SSIM of 0.91, values that represent a reference no latent generative model can on average exceed. The effect is particularly severe on datasets with high anatomical variability: on Synthrad25 MRI-to-CT (head/neck), the VAE reconstruction reaches only 22.66 dB PSNR and 0.69 SSIM, leaving little room for the downstream generative model to operate without further degrading quality. This confirms that the autoencoder backbone acts as an inherent bottleneck for tasks requiring fine-grained accuracy in pixel space [63]. We notice that, for some datasets, the VAE reconstruction is lower than the translation achieved by certain latent models (e.g., Synthrad23 CBCT-to-CT, pelvis: VAE has a PSNR of 24.22 dB vs. BBridge, 24.84 dB). This occurs because the VAE reconstruction level measures the autoencoder’s ability to reconstruct a single target volume, whereas some translation tasks map from different source modalities. When source and target are structurally similar (as in CBCT-to-CT), the latent model can learn a near-identity mapping that partially compensates for VAE reconstruction artifacts. The VAE reference reconstruction level therefore represents an upper bound on average but may be exceeded for individual tasks where the source provides strong structural guidance.

We observe that the addition of ControlNet to the LDM baseline yields only marginal improvements in most translations, suggesting that conditioning alone does not compensate for the information loss inherent in latent-space compression. Therefore, while latent models require larger architectures to handle compressed 3D representations, this added complexity does not translate into better performance, indicating that the limiting factor lies not in the generative component but in the fidelity of the encoding–decoding stage.

The pairwise Wilcoxon signed-rank analysis reported in Table 4 provides statistical grounding for these observations. Across both PSNR and SSIM, GAN-based approaches dominate the ranking, with SRGAN achieving the best task-level ranking across all 11 configurations. The result indicates that the hierarchical, multi-scale structure of medical

Table 4

Pairwise Wilcoxon signed-rank test on model rankings (PSNR and SSIM). Each cell reports the number of tasks N (out of 11) in which the model on the row achieved a higher average rank than the model on the column. Asterisks (*) denote statistically significant differences ($p < 0.05$), as determined by the one-tailed Wilcoxon signed-rank test.

(a) PSNR							
vs.	Pix2Pix	CycleGAN	SRGAN	BBridge	LDM	LDM+CN	FlowM
Pix2Pix	—	2	0	10*	10*	10*	10*
CycleGAN	9	—	0	10*	10*	10*	9*
SRGAN	11*	11*	—	11*	11*	11*	11*
BBridge	1	1	0	—	6	5	6
LDM	1	1	0	5	—	3	4
LDM+CN	1	1	0	6	8*	—	6
FlowM	1	2	0	5	7	5	—

(b) SSIM							
vs.	Pix2Pix	CycleGAN	SRGAN	BBridge	LDM	LDM+CN	FlowM
Pix2Pix	—	1	0	8	10*	10*	9
CycleGAN	10*	—	0	10*	11*	10*	10*
SRGAN	11*	11*	—	11*	11*	11*	11*
BBridge	3	1	0	—	9*	6	4
LDM	1	0	0	2	—	2	2
LDM+CN	1	1	0	5	9*	—	2
FlowM	2	1	0	7	9*	9	—

volumetric data is a critical inductive bias that single-scale architectures fail to capture. This finding generalizes the observations of Ha et al. [21], who demonstrated the advantage of multi-resolution processing on a smaller set of tasks, and extends them to inter-modality and functional imaging. Within the latent family, pairwise differences are rarely statistically significant, further confirming the absence of a clear winner among these approaches.

The dominance of GANs is modulated by task difficulty. CBCT-to-CT synthesis, as an intra-modality task, outperforms its MRI-to-CT counterpart within matched anatomical regions and datasets, reflecting the structural proximity between source and target domains (e.g., SRGAN reaches 29.83 vs. 29.15 dB on Synthrad23, pelvis; 29.50 vs. 26.32 dB on Synthrad25, head/neck). MRI-to-CT translation proves more challenging, with the performance gap between GAN-based and latent architectures widening as anatomical complexity increases. The sharpest divergence between the two families emerges on the T2w-to-T2f task: while GAN-based models maintain high structural similarity, all latent models drop substantially, a result that aligns with the VAE reconstruction level discussed above, as the subtle contrast differences between T2-w and T2-f sequences are particularly vulnerable to information loss in the encoding–decoding stage.

CT-to-PET synthesis requires a separate consideration. Although these tasks yield the highest absolute PSNR and SSIM values across all experiments (see EnhancePET results in radar plots, Figure 3), these scores are misleading: the inherently sparse signal of PET images, with large near-zero background regions, artificially inflates both pixel-wise and structural fidelity metrics independently of the actual synthesis quality. This is a well-known limitation of standard image quality metrics [53, 54], particularly pronounced for functional modalities such as PET, where clinically meaningful information is concentrated in a small fraction of voxels.

Figure 4 compares the images generated from the two best-performing GAN-based (SRGAN and CycleGAN) and latent generative models (BBridge and FlowM), across all tasks and datasets. For each task, the input and target images are displayed alongside the model predictions and the corresponding error maps, computed as the pixel-wise difference between the synthesized output and the ground truth. A task-specific baseline is also included, representing a prediction with no learned structure: a uniform water image (0 HU) for CT tasks and a mean intensity image for the other modalities. Across all tasks, GAN-based methods exhibit lower and more spatially confined errors compared to latent models. Models' performance is influenced by the heterogeneous tissue composition of the anatomical regions: in anatomically complex regions, such as the nasal cavities, sinuses, and dental areas in head/neck tasks, or the bone–soft




Figure 4: Error maps. Visual comparison across I2I translation tasks, for the two best-performing GAN-based (SRGAN and CycleGAN) and latent generative models (BBridge and FlowM). For each task, we display the target and input images (first column, first and second row respectively); the corresponding model predictions (first row); and the associated error maps with respect to the reference target (second row), computed as the pixel-wise difference between the synthesized output and the ground truth. The last column includes a task-specific baseline, representing a prediction with no learned structure. The baseline image is a uniform water image (0 HU) for CT tasks and a mean intensity image for the other modalities.

tissue interfaces in the pelvis, errors are larger than in more homogeneous regions like the brain parenchyma, where residuals remain low and uniform. The largest errors occur in correspondence with lesion regions, whose irregular morphology and appearance make them particularly difficult to reproduce, and at image boundaries, when transitioning to the background. For instance, in the T2w-to-T2f brain task (BraTS2023), all models exhibit elevated residuals at the tumor site compared to the surrounding healthy tissue. This difficulty is further amplified in inter-modality tasks such as CT-to-PET, where the model must translate structural information into functional metabolic signal, representing a full domain shift.

4.2. Lesion-Level Fidelity Evaluation

Over the entire image volume discussed in the previous section, it is important to analyse to what extent the generative models work in pathological regions, because generation artifacts in tumor lesions could compromise the diagnostic reliability of the synthesized images. To this aim, we carry out a lesion-level evaluation, which requires the availability of ground-truth lesion segmentation masks to isolate the region of interest and compute metrics exclusively within it, which are included in the BraTS23 dataset (T2w-to-T2f task) and the autoPET dataset (CT-to-PET task). It is worth noting that such two datasets differ in their lesion characteristics: BraTS23 contains relatively large brain tumors (median diameter 51.2 mm, interquartile range, IQR, 37.9–62.3 mm), while autoPET is dominated by smaller pulmonary lesions (median diameter 19.3 mm, IQR 15.2–30.3 mm), making the latter a considerably more demanding benchmark for lesion-level synthesis.




Figure 5: Lesion analysis from BraTS23. PSNR and SSIM vs lesion size group for the MRI T2w-to-T2f task (BraTS dataset, median lesion diameter: 51.2 mm, IQR: 37.9–62.3 mm).




Figure 6: Lesion analysis from autoPET. PSNR and SSIM vs lesion size group for the CT-to-PET task (autoPET dataset, median lesion diameter: 19.3 mm, IQR: 15.2–30.3 mm).

Figures 5 and 6 report PSNR and SSIM as a function of lesion size for the two datasets, respectively. In the BraTS23 dataset (Figure 5), both metrics improve with increasing lesion size across all models, confirming that larger lesions are easier to synthesize due to their greater spatial extent and structural visibility and this benefit is uniform across both metrics. In the autoPET dataset (Figure 6), a different pattern emerges: indeed, while SSIM improves with lesion size, suggesting that models successfully capture the morphological structure of larger lesions, PSNR decreases. This dissociation between structural and pixel-level metrics can be explained by the nature of the task itself: the model captures the lesion's shape but is less reliable in predicting absolute intensity values, indicating that structural fidelity does not necessarily translate into faithful uptake quantification in CT-to-PET translation. The challenge is compounded by the smaller absolute size of these lesions (with a median diameter roughly three times smaller than in BraTS23), which leaves the generative models with considerably less spatial context to infer the metabolic signal.

We deem that such results highlight two complementary aspects of lesion-level synthesis. First, smaller lesions are more difficult to generate across all models and metrics, as their limited spatial extent reduces the amount of learnable information available to the network. Second, in the inter-modality domain shift (CT-to-PET translation), when the model converts structural information into functional metabolic signal, the prediction of intensity values remains unreliable, even when the lesion morphology is correctly reproduced (as SSIM scores suggest), highlighting a fundamental gap between structural and functional image synthesis.

4.3. Perceptual-level Evaluation

We now turn to the analysis of perceptual realism, which compares quantitative rankings with human perception collected via the Visual Turing test administered to 17 physicians. This investigation is needed because to probe whether the models that score highest on PSNR and SSIM also produce the volumes that clinicians judge most realistic, and




Figure 7: Summary of results from Visual Turing test, Part 1. Summary of classification performance in Visual Turing test, Part 1. Each column reports the rate of correctly (blue) and incorrectly (red) classified images, separately for real and AI-generated cases (top and bottom row, respectively). **Best**: physician with the highest balanced accuracy (R3). **Worst (Real)** and **Worst (AI-gen)**: physicians with the lowest accuracy on real and AI-generated images, respectively (R12 and R1), illustrating opposite classification biases. **Average**: mean performance across all physicians.

whether expert readers can distinguish synthetic acquisitions from real ones. To this aim, in the following we used synthetic volumes generated by the two best-performing GAN- (SRGAN and CycleGAN) and latent-based generative models (BBridge and FlowM).

4.3.1. Visual Turing test, Part 1

In the first part of the test we asked participants to classify volumes as real or AI-generated. The classification accuracy is 57%, only marginally above chance level (50%), indicating that the synthesized images are largely indistinguishable from real acquisitions. Misclassification errors were balanced in both directions: 43% of real images were incorrectly labeled as AI-generated, while 44% of synthetic images were mistaken for real ones. We also note that accuracy varies across modalities: indeed, MRI images are the most challenging to classify correctly (45%), followed by PET (50%) and CT (67%).

The performance across physicians is summarized in Figure 7. One of them, denoted as R3, clearly stood out from the rest, achieving high accuracy on both classes (R3, 71% on real and 88% on AI-generated images) and emerging as the only reader above chance on both tasks. By contrast, no single physician performed uniformly poorly across both classes; rather, the two worst cases each reflected opposite classification biases. Physician R12 correctly identified only 14% of real images, systematically labeling them as AI-generated, whereas R1 correctly identified only 12% of AI-generated images, systematically labeling them as real. Reporting both cases illustrates how low accuracy on one class can coexist with above-average performance on the other, revealing two distinct failure modes within the cohort. We also observe that neither the years of experience nor the physician's subspecialty had a measurable impact on classification accuracy, with comparable performance observed across all experience levels and specialty groups.

4.3.2. Visual Turing test, Part 2

The second part of the Visual Turing test asked participants to look at synthetic volumes pairs, and select the one they preferred, or indicate that no difference could be identified. Figure 8 shows results of direct comparisons between the two models of the GAN family (SRGAN and CycleGAN) and the two models of the latent generative family (BBridge and FlowM), respectively.

Within the GAN family, CycleGAN was preferred over SRGAN in direct comparisons (46.9% vs. 26.9%, with 26.2% reporting no difference), despite SRGAN achieving higher PSNR and SSIM scores in the quantitative evaluation. This suggests a dissociation between quantitative metrics and perceived clinical quality. CycleGAN may produce outputs that, while numerically less accurate, appear more natural to the clinical eye. This preference is not uniform across tasks. CycleGAN is most strongly favored in CT-to-PET synthesis (78.1% vs. 12.5%) and in T2w-to-T2f (62.5% vs. 25.0%), suggesting that its outputs are perceived as more visually consistent in these settings. In CBCT-to-CT

(pelvis), CycleGAN maintains an advantage over SRGAN (43.8% vs. 12.5%), though with an equally high proportion of “no difference” responses (43.8%), indicating that the two models produce outputs of comparable perceptual quality in that setting. The only notable exception is MRI-to-CT (pelvis), where SRGAN was preferred by 59.4% of readers against 25.0% for CycleGAN, suggesting that in tasks requiring accurate bone and soft-tissue contrast reproduction, its higher structural fidelity translates into a perceptually meaningful advantage. Finally, CBCT-to-CT (brain) constitutes a borderline case, with 50.0% of readers reporting no perceptual difference and the remaining preferences split evenly between the two models (25.0% each).

Among latent models, FlowM was preferred over BBridge (60.6% vs. 22.5%, with 16.9% reporting no difference), in agreement with its quantitative advantage. This preference is not uniform across tasks. FlowM is most strongly favored in MRI-to-CT (lung) synthesis, where it was selected by 93.8% of readers with no preference for BBridge and only 6.2% reporting no difference. It maintains a clear advantage in both T2w-to-T2f (brain) (75.0% vs. 18.8%, 6.2% no difference) and CBCT-to-CT (lung) (56.2% vs. 31.2%, 12.5% no difference). In MRI-to-CT (brain), FlowM retains a nominal lead (37.5% vs. 18.8%), but the largest share of responses falls in the “no difference” category (43.8%), suggesting comparable perceptual quality in that setting. The only borderline case is CT-to-PET synthesis, where BBridge is marginally preferred (43.8% vs. 40.6%, 15.6% no difference), making the two models nearly indistinguishable to expert readers. Furthermore, we monitored the response consistency in this part through three sanity-check pairs of identical images. The large majority of readers correctly identified no perceptual difference (75%), while 25% indicated a preference despite the two images being identical, suggesting a moderate degree of response bias in a minority of participants.

It is worth noting that such pairwise comparisons reveal on the one side, that the model that achieves the highest quantitative scores is not necessarily the one preferred by clinicians. This finding reinforces the need for perceptual evaluation as a complement to standard metrics when assessing generative models. On the other side, the degree of agreement between metrics and clinical preference appears to depend on the nature of the translation task. Indeed, in CBCT-to-CT synthesis, an intra-modality task with narrow domain gap between source and target, the Visual Turing test reveals that clinicians frequently perceive no meaningful difference between competing models, e.g. in the GAN comparison, 50.0% of “no difference” in the brain and 43.8% in the pelvis. This suggests that outputs from different models have similar acceptable quality for humans and model selection is not critical. This perception agrees also with the quantitative evaluation (Table 3), with CBCT-to-CT translation having the smallest inter-architecture differences.

In the MRI-to-CT translation, the inter-modality gap amplifies differences between architectures. In MRI-to-CT (pelvis), SRGAN is preferred by 59.4% of readers (Figure 8), this advantage being confirmed by the PSNR and SSIM scores, suggesting that accurate bone and soft-tissue contrast reproduction translates into a tangible clinical benefit. Among latent models, FlowM dominates in MRI-to-CT (lung) with 93.8% preference, while in MRI-to-CT (brain) the distinction becomes less clear, with 43.8% of readers reporting no difference, indicating that the perceptual gap narrows in regions with less anatomical complexity.

Finally, the CT-to-PET synthesis is the most revealing case. In the Visual Turing test, CycleGAN is strongly preferred over SRGAN (78.1% vs. 12.5%), while FlowM and BBridge have similar perceptual scores (40.6% vs. 43.8%), despite measurable differences in their quantitative scores. This suggests that clinicians evaluate PET image quality primarily on the basis of structural plausibility rather than pixel-wise intensity accuracy, effectively judging whether lesion morphology and anatomical context appear realistic while being less sensitive to deviations in absolute SUV values.

4.3.3. Visual Turing test, Part 3

The last part of the Visual Turing test asked participants to rank triplets of volumes from most to least realistic (rank 1 to 3, respectively), and the results are shown in Figure 9 over all tasks and physicians. Real volumes are ranked as most realistic in 58% of cases, i.e., that in over 42% of the comparisons a synthetic image was judged more realistic than the corresponding real acquisition. Among synthetic models, CycleGAN gets the highest frequency of most realistic assignments (rank 1, 26.0%), marginally surpassing SRGAN (22.4%), further confirming the dissociation between quantitative performance and clinical perception. SRGAN and FlowM emerged as the most reliable second-ranked models, each assigned to rank 2 in approximately half of the comparisons. BBridge is ranked last, receiving a rank 3 assignment in over 72% of cases, confirming its weakest perceptual quality among the evaluated architectures.

The two sanity-check triplets included in this section contained one real image and two outputs from the same model. In the CT-to-PET triplet, 70.6% of readers ranked an SRGAN-generated image as the most realistic, failing to identify the real acquisition, in agreement with the difficulty of discriminating synthetic PET images observed




Figure 8: Results from Visual Turing test, Part 2. Pairwise preference results for GAN models (left) and latent generative models (right), for each task and as an overall aggregate ("Average").




Figure 9: Results from Visual Turing test, Part 3. Three-way ranking results (Visual Turing test, Part 3). Each panel reports the percentage of rank 1 (most realistic), rank 2, and rank 3 (least realistic) assignments across all triplets, aggregated over all tasks and physicians.

throughout the test. In contrast, in the T2w-to-T2f triplet, only 17.6% of readers were fooled, with the large majority correctly assigning the highest realism rank to the real image, suggesting that MRI synthesis artifacts remain more perceptible to expert readers.

These Visual Turing test results confirm again that the models with the highest scores on standard metrics do not necessarily produce the images that physicians judge as most realistic, as we noticed for CycleGAN and SRGAN in the pairwise comparison. Furthermore, the near-chance accuracy in the binary classification task, combined with the frequent ranking of synthetic images above real ones, confirms that current generative models, particularly GAN-based architectures, have reached a level of visual fidelity that challenges expert discrimination.

5. Limitations and future directions

While this work provides a comprehensive comparison of generative architectures for volumetric medical image translation, it is important to acknowledge the limitations in terms of data coverage, architectural scope and evaluation methodology, that point to directions for future research.

First, the scarcity of public cancer datasets with paired multi-modal volumes [60] limits the translation directions that could be evaluated (Figure 1). Annotated datasets are even scarcer: of the eleven configurations included, only two have lesion segmentation masks, restricting the lesion-level analysis to a subset of tasks and preventing a systematic assessment of pathological fidelity across all translation directions. Expanding the availability of publicly shared, multi-modal datasets with voxel-level pathological annotations remains a prerequisite for moving the field beyond isolated proof-of-concept evaluations.

Second, on the architectural side, the quantitative results indicate that a multi-resolution framework is better suited to the biological structure of the input data for medical image translation. Indeed, SRGAN [21] achieves the best performance thanks to its multi-resolution framework, which processes data at multiple spatial scales and captures anatomical information that single-scale architectures miss. Extending this principle to latent generative models, for instance through multi-scale VAE encoders capable of preserving features across compression levels, or through hierarchical latent spaces that retain both global anatomy and local detail, represents a promising direction for future research.

Third, inter-modality tasks such as CT-to-PET, where source and target encode fundamentally different physical quantities, expose the limits of current approaches. Models can reproduce lesion morphology but are less reliable in predicting metabolic intensity, as SUV values depend on physiological information that CT attenuation maps do not carry, and addressing tasks of this complexity will likely require domain-specific innovations such as physics-informed losses or hybrid architectures that incorporate auxiliary physiological priors.

Fourth, the evaluation framework itself presents open challenges. As demonstrated by the Visual Turing test, standard full-reference metrics such as PSNR and SSIM do not reliably reflect human perception [3, 60], a finding that is supported by recent literature, highlighting that these metrics were designed for natural images and exhibit well-documented shortcomings when applied to medical data, including insensitivity to blurring, dependence on normalization methods, and poor correlation with human judgment [61, 62]. Developing new evaluation metrics specifically designed for 3D medical volumes, capable of capturing local anatomical accuracy, functional signal fidelity, and sensitivity to pathological features, remains a critical open challenge in the field [60].

6. Conclusion

Virtual scanning offers a promising approach to many of the challenges facing oncology imaging today. By enriching the diagnostic information available from routine examinations while avoiding additional scans, it has the potential to reduce patient burden, lower examination costs, ease pressure on imaging services, and help bridge inequities in access to advanced imaging, particularly in resource-constrained settings. Within this context, this work presents a systematic benchmark of seven generative architectures for 3D medical I2I translation, evaluated across 77 experiments spanning five datasets, four modalities, and three anatomical regions (head/neck, lung, pelvis). Its main value lies in enabling controlled and reproducible comparison across heterogeneous modality pairs, anatomical regions, and evaluation perspectives. By standardizing the core experimental approach and integrating whole-volume, lesion-level, and physician-centered assessment, the framework is transferable across studies and informative for clinically motivated model selection.

We compare GAN and latent generative architectures using PSNR and SSIM, supported by pairwise statistical testing, to identify which design principles, such as multi-scale processing, latent-space compression, or direct image-domain generation, drive quantitative accuracy in volumetric synthesis. We find that the SRGAN model achieves the highest PSNR and SSIM across all tasks, with its multi-scale, multi-resolution framework proving to be a decisive advantage for volumetric synthesis, while CycleGAN is the second-best GAN. Latent generative models underperform their GAN counterparts, with the VAE reconstruction being the main bottleneck that limits fidelity. These results suggest that multi-scale architectures are better suited for volumetric medical image synthesis, where anatomically relevant features span a wide range of spatial scales, whereas latent-space compression does not preserve this hierarchy of detail.

The lesion-level analysis revealed that all models struggle with smaller lesions, and exposed a critical dissociation in CT-to-PET synthesis. While structural similarity improves with lesion size, pixel-level accuracy simultaneously degrades, indicating that current architectures can reproduce lesion morphology but fail to predict metabolic intensity.

The Visual Turing test administered to 17 physicians demonstrated that quantitative superiority does not guarantee human expert preference. CycleGAN was favored over SRGAN in the majority of tasks despite its lower scores, and the degree of agreement between metrics and perception proved strongly task-dependent.

The joint analysis of such findings with the related limitations suggests that future efforts in medical image synthesis should pursue three parallel directions: i) expanding the availability of paired multi-modal datasets with pathological annotations; ii) developing domain-adapted architectures that incorporate multi-scale processing and task-specific priors; iii) designing evaluation frameworks that integrate quantitative metrics, lesion-level fidelity, and perceptual assessment to bridge the persistent gap between pixel-level accuracy and diagnostic utility.

We therefore view this study not only as a comparative analysis, but also as a reusable evaluation protocol that can support future algorithmic development, dataset expansion, and clinically grounded validation in medical image synthesis. The full codebase, including model implementations, training pipelines, evaluation scripts, and the Turing test platform, will be published upon acceptance.

7. Acknowledgments

This work was supported by Strategic University Projects “IDEA: Kempe Foundation under grant no. JCSMK24-0094; and “AI-powered Digital Twin for next-generation lung cancer care” financed by Università Campus Bio-Medico di Roma 2023 (GEN0469). Resources are provided by the National Academic Infrastructure for Supercomputing in Sweden (NAISS) and the Swedish National Infrastructure for Computing (SNIC) at Alvis @ C3SE, partially funded by the Swedish Research Council through grant agreements no. 2022-06725 and no. 2018-05973. The funders had no role in study design, data collection and analysis, decision to publish, or preparation of the manuscript.

CRedit authorship contribution statement

Giulia Romoli: Conceptualization, Investigation, Methodology, Software, Writing - First draft, Writing – review and editing. **Alessia Capoccia:** Data curation, Investigation, Methodology, Software, Writing - First draft, Writing – review and editing. **Filippo Ruffini:** Conceptualization, Investigation, Methodology, Software, Writing – review and editing. **Francesco Di Feola:** Conceptualization, Investigation, Methodology, Software. **Luca Boldrini:** Validation. **Arturo Chiti:** Validation. **Renato Cuocolo:** Validation. **Tugba Akinci D’Antonoli:** Validation. **Fatemeh Darvizeh:** Validation. **Marcello Di Pumpo:** Validation. **Bradley J. Erickson:** Validation. **Liu Fang:** Validation. **Deborah Fazzini:** Validation. **Paola Feraco:** Validation. **Fabrizia Gelardi:** Validation. **Francesco Gossetti:** Validation. **Ana Isabel Hernáiz Ferrer:** Validation. **Michail E. Klontzas:** Validation. **Seyedmehdi Payabvash:** Validation. **Katrine Riklund:** Validation. **Sara N. Strandberg:** Validation. **Valerio Guarrasi:** Conceptualization, Methodology, Supervision, Writing – review and editing. **Paolo Soda:** Conceptualization, Funding acquisition, Project administration, Methodology, Supervision, Writing – review and editing.

References

- [1] WHO, WHO compendium of innovative health technologies for low-resource settings 2024, World Health Organization WHO, 2024.
- [2] E. Kjelle, et al., Cost of low-value imaging worldwide: a systematic review, *Applied health economics and health policy* 22 (2024) 485.
- [3] S. Dayarathna, et al., Deep learning-based synthesis of MRI, CT and PET: Review and analysis, *Computer Methods and Programs in Biomedicine* 257 (2024) 108173.
- [4] L. Doan, et al., Bridging modalities with ai: a review of ai advances in multimodal biomedical imaging, *Communications Engineering* 5 (2026) 30.
- [5] M. Sherwani, S. Gopalakrishnan, A systematic literature review: deep learning techniques for synthetic medical image generation and their applications in radiotherapy, *Frontiers in Radiology* 4 (2024) 1385742.
- [6] X. Fu, et al., A systematic review of generative artificial intelligence techniques for synthetic medical image datasets: Quality, models, public availability and applications, *Computer Methods and Programs in Biomedicine* (2026) 109331.
- [7] A. Rofena, et al., Augmented intelligence for multimodal virtual biopsy in breast cancer using generative artificial intelligence, *Journal of Biomedical Informatics* (2025) 104971.
- [8] S. Kazemina, et al., GANs for medical image analysis, *Artificial Intelligence in Medicine* 109 (2020) 101938.
- [9] A. Kazerouni, et al., Diffusion models in medical imaging: A comprehensive survey, *Medical Image Analysis* 88 (2023) 102846.
- [10] G. Bredell, et al., Explicitly minimizing the blur error of variational autoencoders, in: *The Eleventh International Conference on Learning Representations*, 2023, pp. 1–16.
- [11] P. Isola, et al., Image-to-image translation with conditional adversarial networks, in: *Proceedings of the IEEE Conference on Computer Vision and Pattern Recognition*, 2017, pp. 1125–1134.
- [12] J. Zhu, et al., Unpaired image-to-image translation using cycle-consistent adversarial networks, in: *Proceedings of the IEEE International Conference on Computer Vision*, 2017, pp. 2223–2232.
- [13] D. Nie, et al., Medical image synthesis with deep convolutional adversarial networks, *IEEE Transactions on Biomedical Engineering* 65 (2018) 2720–2730.
- [14] J. Wolterink, et al., Deep MR to CT synthesis using unpaired data, in: *International Workshop on Simulation and Synthesis in Medical Imaging*, Springer, 2017, pp. 14–23.
- [15] Y. Liu, et al., Magnetic resonance image synthesis from brain computed tomography images based on deep learning methods for magnetic resonance-guided radiotherapy, *Quantitative Imaging in Medicine and Surgery* 10 (2020) 1358.
- [16] S. Dar, et al., Image synthesis in multi-contrast MRI with conditional generative adversarial networks, *IEEE Transactions on Medical Imaging* 38 (2019) 2375–2388.

- [17] A. Chatsias, et al., Multimodal MR synthesis via modality-invariant latent representation, *IEEE Transactions on Medical Imaging* 37 (2018) 803–814.
- [18] B. Yu, et al., Ea-GANs: Edge-aware generative adversarial networks for cross-modality MR image synthesis, *IEEE Transactions on Medical Imaging* 38 (2019) 1750–1762.
- [19] S. Poonkodi, M. Kanchana, 3D-MedTranCSGAN: 3D medical image transformation using CSGAN, *Computers in Biology and Medicine* 153 (2023) 106541.
- [20] V. Guarrasi, et al., Whole-body image-to-image translation for a virtual scanner in a healthcare digital twin, in: *Proceedings of the IEEE 38th International Symposium on Computer-Based Medical Systems (CBMS)*, 2025, pp. 528–534.
- [21] J. Ha, et al., Multi-resolution guided 3D GANs for medical image translation, in: *IEEE/CVF Winter Conference on Applications of Computer Vision (WACV)*, 2025, pp. 4342–4351.
- [22] J. Ho, A. Jain, P. Abbeel, Denoising diffusion probabilistic models, *Advances in neural information processing systems* 33 (2020) 6840–6851.
- [23] P. Dhariwal, A. Nichol, Diffusion models beat GANs on image synthesis, in: *Advances in Neural Information Processing Systems*, volume 34, 2021, pp. 8780–8794.
- [24] A. Kazerouni, et al., Diffusion models in medical imaging: A comprehensive survey, *Medical Image Analysis* 88 (2023) 102846.
- [25] A. Moschetto, et al., Benchmarking gans, diffusion models, and flow matching for t1w-to-t2w mri translation, in: *International Conference on Image Analysis and Processing*, Springer, 2025, pp. 429–440.
- [26] Q. Bertrand, A. Gagneux, M. Massias, R. Emonet, On the closed-form of flow matching: Generalization does not arise from target stochasticity, *arXiv preprint arXiv:2506.03719* (2025).
- [27] M. Akbar, W. Wang, A. Eklund, Beware of diffusion models for synthesizing medical images – a comparison with GANs in terms of memorizing brain MRI and chest x-ray images, *Machine Learning: Science and Technology* 6 (2025) 015022.
- [28] S. Pan, et al., Synthetic CT generation from MRI using 3D transformer-based denoising diffusion model, *Medical Physics* 51 (2024) 2538–2548.
- [29] K. Choo, Y. Jun, M. Yun, S. Hwang, Slice-consistent 3D volumetric brain CT-to-MRI translation with 2D Brownian bridge diffusion model, in: *Medical Image Computing and Computer Assisted Intervention – MICCAI 2024*, Springer, 2024, pp. 657–667.
- [30] X. Zhu, et al., Introducing 3D representation for dense volume-to-volume translation via score fusion, in: *International Conference on Machine Learning*, 2025, pp. 1–22.
- [31] J. Kim, H. Park, Adaptive latent diffusion model for 3D medical image to image translation: Multi-modal magnetic resonance imaging study, in: *Proceedings of the IEEE/CVF Winter Conference on Applications of Computer Vision*, 2024, pp. 7604–7613.
- [32] P. Esser, R. Rombach, B. Ommer, Taming transformers for high-resolution image synthesis, in: *Proceedings of the IEEE/CVF Conference on Computer Vision and Pattern Recognition*, 2021, pp. 12873–12883.
- [33] A. Sargood, et al., CoCoLIT: ControlNet-conditioned latent image translation for MRI to amyloid PET synthesis, in: *Proceedings of the AAAI Conference on Artificial Intelligence*, volume 40, 2026, pp. 8778–8786.
- [34] L. Zhang, A. Rao, M. Agrawala, Adding conditional control to text-to-image diffusion models, in: *Proceedings of the IEEE International Conference on Computer Vision (ICCV)*, 2023, pp. 3836–3847.
- [35] R. Graf, et al., Denoising diffusion-based MRI to CT image translation enables automated spinal segmentation, *European Radiology Experimental* 7 (2023) 70.
- [36] A. Rajagopal, et al., Synthetic PET via domain translation of 3-D MRI, *IEEE Transactions on Radiation and Plasma Medical Sciences* 7 (2023) 333–343.
- [37] M. Bahloul, et al., Advancements in synthetic CT generation from MRI: A review of techniques and trends in radiation therapy planning, *Journal of Applied Clinical Medical Physics* (2024).
- [38] A. Thummerer, et al., Synthrad2025 grand challenge dataset: Generating synthetic cts for radiotherapy from head to abdomen, *Medical Physics* 52 (2025) e17981.
- [39] F. Bray, et al., Global cancer statistics 2022: GLOBOCAN estimates of incidence and mortality worldwide for 36 cancers in 185 countries, *CA: A Cancer Journal for Clinicians* 74 (2024) 229–263.
- [40] R. Siegel, et al., Cancer statistics, 2026, *CA: A Cancer Journal for Clinicians* 76 (2026).
- [41] S. Karimi, et al., Glioblastoma: Clinical presentation, multidisciplinary management, and long-term outcomes, *Cancers* 17 (2025).
- [42] A. Thummerer, et al., SynthRAD2023 grand challenge dataset: Generating synthetic CT for radiotherapy, in: *Proceedings of the Medical Image Computing and Computer Assisted Intervention (MICCAI) Challenges*, 2023, pp. 4664–4674.
- [43] A. Kazerooni, et al., The ASNR-MICCAI brain tumor segmentation (BraTS) challenge 2023: Intracranial meningioma, in: *Proceedings of MICCAI*, 2023, pp. 1–11.
- [44] S. Gatidis, et al., A whole-body FDG-PET/CT dataset with manually annotated tumor lesions, *Scientific Data* 9 (2022) 601.
- [45] D. Ferrara, et al., Sharing a whole-/total-body [18f] fdg-pet/ct dataset with ct-derived segmentations: an enhance. pet initiative, *Scientific Data* (2026).
- [46] C. Saharia, et al., Palette: Image-to-image diffusion models, in: *ACM SIGGRAPH 2022 Conference Proceedings*, 2022, pp. 1–10.
- [47] B. Li, K. Xue, B. Liu, Y. Lai, BBDM: Image-to-image translation with Brownian bridge diffusion models, in: *Proceedings of the IEEE Conference on Computer Vision and Pattern Recognition (CVPR)*, 2023, pp. 1952–1961.
- [48] Y. Lipman, R. Chen, H. Ben-Hamu, M. Nickel, M. Le, Flow matching for generative modeling, in: *Proceedings of the International Conference on Learning Representations (ICLR)*, 2023, pp. 1–28.
- [49] M. Valls, P. Bourdon, C. Fernandez, G. Herpe, D. Helbert, Prob-bbmd: A probabilistic brownian bridge diffusion model for mri sequence image-to-image translation, *Computerized Medical Imaging and Graphics* (2026) 102745.
- [50] M. Yazdani, Y. Medghalchi, P. Ashrafi, I. Hachililoglu, D. Shahriari, Flow matching for medical image synthesis: Bridging the gap between speed and quality, in: *Medical Image Computing and Computer Assisted Intervention – MICCAI*, 2025, pp. 216–226.

- [51] F. Isensee, P. Jaeger, S. Kohl, J. Petersen, K. Maier-Hein, nnU-Net: a self-configuring method for deep learning-based biomedical image segmentation, *Nature Methods* 18 (2021) 203–211.
- [52] F. Di Feola, L. Tronchin, P. Soda, A comparative study between paired and unpaired image quality assessment in low-dose ct denoising, in: 2023 IEEE 36th International Symposium on Computer-Based Medical Systems (CBMS), IEEE, 2023, pp. 471–476.
- [53] M. Roberts, et al., Imaging evaluation of a proposed 3d generative model for mri to ct translation in the lumbar spine, *The Spine Journal* (2023).
- [54] C. Tang, et al., Incorporating radiologist knowledge into mri quality metrics for machine learning using rank-based ratings, *Journal of Magnetic Resonance Imaging* (2024).
- [55] V. Guarasi, et al., Multimodal explainability via latent shift applied to covid-19 stratification, *Pattern Recognition* 156 (2024) 110825.
- [56] Y. Myong, et al., Evaluating diagnostic content of AI-generated chest radiography: A multi-center visual turing test, *PLoS ONE* (2023).
- [57] M. Jang, et al., Image turing test and its applications on synthetic chest radiographs by using the progressive growing generative adversarial network, *Scientific Reports* (2023).
- [58] A. Phelps, et al., Pairwise comparison versus likert scale for biomedical image assessment, *American Journal of Roentgenology* (2015).
- [59] E. Hoeijmakers, et al., How subjective CT image quality assessment becomes surprisingly reliable: pairwise comparisons instead of likert scale, *European Radiology* (2024).
- [60] L. Friedrich, et al., Deep learning for medical image-to-image translation: Methods, datasets, and evaluation, *npj Digital Medicine* 7 (2024) 114.
- [61] A. Breger, et al., A study of why we need to reassess full reference image quality assessment with medical images, *Journal of Imaging Informatics in Medicine* 38 (2025) 3444–3469.
- [62] M. Dohmen, M. Klemens, I. Baltruschat, T. Truong, M. Lenga, Similarity and quality metrics for MR image-to-image translation, *Scientific Reports* 15 (2025) 3853.
- [63] R. Rombach, A. Blattmann, D. Lorenz, P. Esser, B. Ommer, High-resolution image synthesis with latent diffusion models, in: *Proceedings of the IEEE Conference on Computer Vision and Pattern Recognition (CVPR)*, 2022, pp. 10684–10695.
- [64] E. Haacke, R. Brown, M. Thompson, R. Venkatesan, *Magnetic Resonance Imaging: Physical Principles and Sequence Design*, Wiley-Liss, 1999.
- [65] J. Bushberg, J. Seibert, E. Leidholdt, J. Boone, *The Essential Physics of Medical Imaging*, 3 ed., Lippincott Williams & Wilkins, 2011.
- [66] J. Barentsz, et al., ESUR prostate MRI guidelines, *European Radiology* 22 (2012) 746–757.
- [67] R. Beets-Tan, et al., Magnetic resonance imaging for clinical management of rectal cancer, *European Radiology* 28 (2018) 1465–1475.
- [68] P. Wen, et al., Updated response assessment criteria for high-grade gliomas, *Journal of Clinical Oncology* 28 (2010) 1963–1972.
- [69] D. Louis, et al., The 2021 WHO classification of tumors of the central nervous system, *Neuro-Oncology* 23 (2021) 1231–1251.
- [70] F. Fazekas, et al., MR signal abnormalities at 1.5 T in Alzheimer’s dementia and normal aging, *American Journal of Neuroradiology* 14 (1993) 1237–1242.
- [71] American Cancer Society, *Cancer facts & figures 2026*, 2026.
- [72] R. Stupp, et al., Radiotherapy plus concomitant and adjuvant temozolomide for glioblastoma, *New England Journal of Medicine* 352 (2005) 987–996.
- [73] A. A. Aizer, et al., Brain metastases: A society for neuro-oncology (SNO) consensus review on current management and future directions, *Neuro-Oncology* 24 (2022) 1613–1646.
- [74] M. Singh, et al., Epidemiology of brain metastases, *Neurosurgery Clinics of North America* 31 (2020) 481–495.
- [75] Z. S. Mayo, et al., Radiation necrosis or tumor progression? A review of the radiographic modalities used in the diagnosis of cerebral radiation necrosis, *Journal of Neuro-Oncology* 161 (2023).
- [76] M. Spadea, M. Maspero, P. Zaffino, J. Seco, Deep learning based synthetic-CT generation in radiotherapy and PET: A review, *Medical Physics* 48 (2021) 6537–6566.
- [77] S. De Pietro, et al., The role of MRI in radiotherapy planning: a narrative review “from head to toe”, *Insights into Imaging* 15 (2024) 255.
- [78] M. Maspero, et al., Deep learning for CT synthesis in radiotherapy, *Bioengineering* 12 (2025) 1297.
- [79] G. Cordier, et al., Generative adversarial networks to synthesize missing T1 and FLAIR MRI sequences for use in a multisequence brain tumor segmentation model, *Radiology* 299 (2021) E209–E219.
- [80] National Lung Screening Trial Research Team, Reduced lung-cancer mortality with low-dose computed tomographic screening, *New England Journal of Medicine* 365 (2011) 395–409.
- [81] H. de Koning, et al., Reduced lung-cancer mortality with volume CT screening in a randomized trial, *New England Journal of Medicine* 382 (2020) 503–513.
- [82] A. Wolf, et al., Screening for lung cancer: 2023 guideline update from the American Cancer Society, CA: *A Cancer Journal for Clinicians* 74 (2024) 50–81.
- [83] L. Umutlu, et al., ESR essentials: staging and restaging with FDG-PET/CT in oncology—practice recommendations by the European Society for Hybrid, Molecular and Translational Imaging, *European Radiology* 35 (2025) 1894–1902.
- [84] B. Turkbey, et al., Prostate imaging reporting and data system version 2.1: 2019 update of prostate imaging reporting and data system version 2, *European Urology* 76 (2019) 329–340.
- [85] S. Klein, et al., Elastix: a toolbox for intensity-based medical image registration, *IEEE Transactions on Medical Imaging* 29 (2010) 196–205.
- [86] U. Baid, et al., The RSNA-ASNR-MICCAI BraTS 2021 benchmark on brain tumor segmentation and radiogenomic classification, *arXiv preprint arXiv:2107.02314* (2021).
- [87] O. Morin, et al., Exploring adult glioma through MRI: A review of publicly available datasets to guide efficient image analysis, *Neuro-Oncology Advances* 7 (2025) vdae197.
- [88] H. Li, et al., The brain tumor segmentation (BraTS) challenge 2023: Brain MR image synthesis for tumor segmentation (BraSyn), *arXiv preprint arXiv:2305.09011* (2023).

- [89] T. Rohlfing, N. Zahr, E. Sullivan, A. Pfefferbaum, The SRI24 multichannel atlas of normal adult human brain structure, *Human Brain Mapping* 31 (2010) 798–819.
- [90] P. Kinahan, J. Fletcher, Positron emission tomography-computed tomography standardized uptake values in clinical practice and assessing response to therapy, *Seminars in Ultrasound, CT and MRI* 31 (2010) 496–505.
- [91] MONAI Consortium, MONAI: Medical open network for AI, arXiv preprint arXiv:2211.02701 (2022).
- [92] NiiVue, Niivue, 2026. URL: <https://github.com/niivue/niivue>, accessed: March 2026.
- [93] D. Kingma, M. Welling, Auto-encoding variational Bayes, in: *Proceedings of the International Conference on Learning Representations (ICLR)*, 2014, pp. 1–14.
- [94] D. Wolpert, W. Macready, No free lunch theorems for optimization, *IEEE transactions on evolutionary computation* 1 (2002) 67–82.
- [95] A. Arcuri, G. Fraser, Parameter tuning or default values? an empirical investigation in search-based software engineering, *Empirical Software Engineering* 18 (2013) 594–623.

A. Medical Imaging Modalities

Magnetic Resonance Imaging (MRI), Computed Tomography (CT), and Positron Emission Tomography (PET) are among the most widely used medical imaging modalities, providing complementary anatomical, functional, and metabolic information about tissues and organs throughout the body. In oncology, these modalities enable clinicians to localize tumors, characterize their extent and biological activity, and monitor disease progression or regression in response to therapy. The tasks proposed in this benchmark span these three modalities, selected for their clinical relevance, the complementary nature of the information they provide, and the availability of large, paired public datasets, thereby offering a diverse and clinically grounded testing bench for cross-modality synthesis research. This section provides an overview of each modality, with emphasis on the clinical role, the limitations, and the complementary strengths that motivate cross-modality synthesis tasks.

A.1. Magnetic Resonance Imaging

Magnetic Resonance Imaging (MRI) is an imaging technique that exploits the magnetic properties of hydrogen nuclei in biological tissues to generate high-contrast images of soft tissues, making it particularly valuable for brain, pelvic, and head/neck oncology [64, 65]. A strong external magnetic field aligns hydrogen nuclear spins, which are then perturbed by radio-frequency pulses; the subsequent relaxation signal is spatially encoded to reconstruct volumetric images. MRI high contrast arises from differences in hydrogen nuclei concentration across tissues, which directly affect signal intensity, and the different relaxation processes through which excited nuclei return to equilibrium. After excitation, hydrogen nuclei return to their equilibrium state through two independent relaxation processes: longitudinal relaxation (T_1), describing how quickly magnetization recovers along the direction of the external field; and transverse relaxation (T_2), describing magnetization decay in the perpendicular plane. Since relaxation times vary across tissues, adjusting acquisition parameters allows to emphasize either T_1 or T_2 differences, producing T1-weighted (T1w) or T2-weighted (T2w) images with enhanced tissue contrast. In T1w images, fat appears bright, fluid appears dark, and grey/white matter boundaries are clearly delineated, making this sequence well suited for anatomical detail. T1w imagery, optionally with gadolinium-based contrast enhancement, is the standard for delineating tumor margins in head/neck and pelvic cancer. In T2w images, fluid-filled structures such as edema and cerebrospinal fluid appear bright. Since tumors are typically fluid-rich, this produces high contrast between tumor and healthy tissue, which is particularly valuable for prostate cancer staging [66], rectal cancer delineation [67], and assessment of peritumoral edema in neuro-oncology [68]. Fluid-Attenuated Inversion Recovery (FLAIR, T2f hereafter) is a T2w variant in which a preparatory inversion pulse suppresses the signal from cerebrospinal fluid, making peri-ventricular and cortical lesions more clearly visible. T2f is the reference modality for detecting gliomas [69, 68] and white matter lesions [70].

In this benchmark, T2w is used as the source for FLAIR synthesis (intra-MRI, **T2w-to-T2f** task), evaluated on the BraTS23 dataset (Section C.3). Synthesizing T2f from T2w images is clinically relevant, as it can provide complementary information without requiring additional acquisition. Indeed, the bright cerebrospinal fluid signal is suppressed in FLAIR images, thereby improving the contrast of lesions that would otherwise be masked, which is of particular importance in glioma characterization.

A.2. Computed Tomography

Computed Tomography (CT) is a X-ray-based imaging modality that reconstructs three-dimensional volumetric images by measuring the attenuation of a rotating X-ray beam through the body. Attenuation values are expressed in Hounsfield Units (HU), calibrated so that air corresponds to -1024 HU and water to 0 HU. CT images provide high

spatial resolution (0.5 – 1 mm), excellent bone-tissue contrast, and rapid acquisition times (seconds to tens of seconds for whole-body coverage), establishing CT as the preferred modality for cancer screening and radiotherapy planning. Despite these advantages, CT presents two significant limitations that motivate the use of complementary modalities: limited soft-tissue contrast, and patients' exposure to ionizing radiation. CT image contrast arises from differences in X-ray attenuation, which depends on electron density. Since electron density varies minimally among soft tissues, soft-tissue discrimination in CT remains limited. MRI overcomes this limitation by exploiting differences in T_1 , T_2 , and proton density, enabling the differentiation of structures that would appear nearly identical on CT. Furthermore, CT exposes patients to ionizing radiation, which is a significant limitation when conducting repeated examinations during cancer follow-up. Cone-Beam CT (CBCT), an acquisition variant that uses a cone-shaped X-ray beam and a flat-panel detector installed on a linear accelerator, partially addresses this issue by delivering a substantially lower radiation dose per acquisition compared to conventional CT. This makes CBCT suitable for repeated imaging protocols. However, this reduced dose, combined with the cone-beam geometry, results in higher image noise, increased scatter artifacts, and reduced contrast.

The limitations of CT acquisitions justify the attempt to synthesize CT images from other modalities. Two tasks are included in this benchmark, evaluated using the SynthRAD2023 (Section C.1) and SynthRAD2025 (Section C.2) datasets: **MRI-to-CT** synthesis, where a CT scan is generated from a T1w MRI scan; and **CBCT-to-CT** synthesis, where a low-dose CBCT is translated into a CT, reducing the need for additional full-dose CT acquisitions.

A.3. Positron Emission Tomography

Positron Emission Tomography (PET) is an imaging modality that maps the spatial distribution of a positron-emitting radio-tracer administered to the patient. In oncology, the most widely used tracer is ^{18}F -fluorodeoxyglucose (FDG), a glucose analogue that accumulates preferentially in metabolically active tissues. Because malignant tumors typically exhibit elevated glucose uptake, FDG-PET provides functional information complementary to the anatomical detail of CT or MRI. PET image intensity is quantified as the Standardised Uptake Value (SUV), defined as $\text{SUV} = \frac{Cw}{D \cdot 10^6}$, where C is the tissue radioactivity concentration (in kBq/mL), D is the injected dose (in MBq) and w is the patient body weight (in kg). SUV accounts for inter-patient differences in injected dose and body mass, enabling comparison across acquisitions.

In clinical practice, PET is generally acquired together with CT in hybrid PET/CT scanners, providing co-registered functional-anatomical volumes. The main limitations of PET include the radiation dose delivered by the radio-tracer (typically 5-8 mSv for a standard FDG-PET/CT), the need for a cyclotron facility for tracer production, and the relatively high cost of the examination. Synthesizing PET from CT could reduce the need for radio-tracer administration and extends access to functional oncological assessment in resource-limited settings. However, generating accurate SUV quantification from purely structural information remains an open research challenge. In this benchmark, the **CT-to-PET** translation task has been evaluated on the autoPET (Section C.4) and EnhancePET (Section C.5) datasets.

B. Anatomical Region Selection and Clinical Rationale

The global impact of cancer underscores the need for advanced imaging solutions across multiple clinical contexts. This section provides a detailed epidemiological and clinical overview of the three anatomical regions selected for this benchmark: head/neck, lung, and pelvis. These regions were chosen for their high incidence and mortality rates, accounting for a substantial portion of cancer-related deaths. Moreover, the clinical protocols associated with these regions generally exploit 3D imaging modalities (see Section A) to get co-registered, multi-modal volumes, facilitating the construction of paired training datasets for 3D medical image synthesis framed as a I2I translation.

Although breast cancer is among the most prevalent malignancies worldwide, with approximately 2.3 million new cases in 2022 [39], its primary screening modality is X-ray mammography, which is inherently a 2D planar acquisition. Since this benchmark is specifically designed to evaluate 3D volumetric image synthesis solutions, breast cancer was excluded from the current framework. Similarly, skin cancer, despite its high incidence, is primarily diagnosed through dermoscopy, a 2D surface imaging technique. Breast MRI and breast CT are emerging 3D modalities and their inclusion represents a natural future extension of this work.

The following subsections describe the epidemiology and the principal imaging modalities of each anatomical region, together with the clinical rationale for selecting specific synthesis tasks. Rather than exhaustively covering all possible modality combinations, this benchmark focuses on synthesis directions with direct clinical applicability.

A synthesis task is considered meaningful when the target modality either cannot be acquired alongside the source modality, or when its acquisition entails additional cost or increased radiation exposure for the patient.

B.1. Head/neck and brain

Epidemiology: Brain cancer, although relatively rare, is among the most lethal in proportion to its incidence. It is the primary cause of cancer death in children, adolescents, and young adults (aged 20-39 years), while ranking fourth among women in the same age group [40]. In the United States, approximately 24,740 malignant brain and spinal cord tumors are expected to be diagnosed in 2026, with an estimated 18,350 deaths [71]. Glioma is the most common primary malignant brain tumor, with glioblastoma multiforme (GBM), the most aggressive subtype, having a median survival of approximately 14.6 months [72, 41]. Besides primary tumors, brain metastases represent an even more frequent occurrence [73]. Lung cancer is the most common source, accounting for approximately 80% of brain metastases identified at the time of primary cancer diagnosis [74].

Imaging Modalities: MRI is the standard for brain tumor detection, staging, and treatment planning, due to its superior soft-tissue contrast. In clinical practice, multiple MRI sequences are routinely acquired for each patient [43]. T1w images, with and without gadolinium-based contrast enhancement, are used for delineating tumor boundaries and detecting blood-brain barrier breakdown, while T2w images provide high contrast between tumor and peritumoral edema. FLAIR suppresses cerebrospinal fluid signal, making it the reference sequence for detecting infiltrative tumor margins and periventricular lesions. When radiotherapy is recommended, CT is additionally acquired to derive electron density maps for dose calculation, as treatment planning systems require HU values not available from MRI. During radiotherapy, CBCT is acquired before each treatment fraction to verify patient positioning and monitor anatomical changes for adaptive planning. As a complementary imaging modality, FDG-PET is used selectively in neuro-oncology, primarily for differentiating tumour recurrence from treatment-related changes such as radiation necrosis [75].

Clinical Rationale for Image Synthesis Tasks: MRI is the primary diagnostic and treatment planning modality for brain cancer, while CT is required exclusively for radiotherapy dose calculation. **MRI-to-CT** synthesis is therefore clinically well-motivated: replacing the CT acquisition with a synthetic counterpart derived from MRI would eliminate the need for an additional CT session, thereby sparing patients from radiation exposure and removing the systematic co-registration errors that arise when aligning independently acquired volumes [37, 76, 77]. Conversely, **CT-to-MRI** synthesis is of limited clinical value. Since MRI is the reference modality, it is typically acquired first for diagnostic purposes, and CT is only added subsequently for radiotherapy dose planning. There is therefore no realistic clinical scenario where a patient has a brain CT but lacks an MRI, making the reverse synthesis direction unnecessary. **CBCT-to-CT** synthesis is motivated by the well-known limitations of CBCT images, limited by scatter-induced artifacts, truncated projections, and inaccurate HU values, that prevent its direct use for dose computation and online plan adaptation [76, 78]. Translating CBCT into synthetic CT would enable accurate dose calculation directly from the images acquired at each treatment fraction, supporting adaptive radiotherapy without requiring additional full-dose CT acquisitions. Intra-MRI translation between T2-weighted and FLAIR contrasts (**T2w-to-T2f**) is clinically relevant because multi-parametric MRI protocols are routinely acquired for brain tumor characterization, yet in clinical practice one or more sequences are frequently missing due to acquisition failures, patient motion, or time constraints [79]. Synthesizing FLAIR from the T2w volume preserves the diagnostic completeness of the protocol and enables the use of automated segmentation models trained on complete multi-parametric inputs. FDG-PET is used selectively in neuro-oncology and its synthesis from CT or MRI, while an emerging research direction, is not yet part of routine clinical workflows. To the best of our knowledge, no public paired dataset are available for brain CT-to-PET or MRI-to-PET translations.

B.2. Lung

Epidemiology: Lung cancer is the most lethal and most frequently diagnosed cancer globally, with Europe and North America among the regions with the highest incidence and mortality rates [39]. In the United States, an estimated 229,410 new cases and 124,990 deaths are expected in 2026, making lung cancer responsible for nearly one in five cancer deaths [40], more than colorectal and pancreatic cancer combined [40]. The five-year survival rate remains below 20% in most countries, largely because the majority of cases are diagnosed at an advanced stage [40]. Non-small-cell lung cancer (NSCLC) accounts for approximately 85% of all lung cancer cases, with adenocarcinoma and squamous cell carcinoma as the predominant histological subtypes [39].

Imaging Modalities: CT is the primary modality for lung cancer screening, diagnosis, and radiotherapy planning. Low-dose CT (LDCT) screening reduces lung cancer mortality by approximately 20% compared to chest radiography [80, 81], and is recommended for high-risk individuals by oncology guidelines [82]. CT provides high spatial resolution and contrast between pulmonary nodules and surrounding lung parenchyma, making it essential for lesion detection and characterisation. During radiotherapy, CBCT is used for image guidance and adaptive planning, given the respiratory motion that can interfere with thoracic treatments. For staging and treatment response assessment, FDG-PET/CT is the standard for NSCLC. The functional information provided by FDG-PET enables assessment of tumor metabolic activity, lymph node involvement, and distant metastases, completing the anatomical details of CT [83]. MRI plays a secondary role, primarily used for characterizing mediastinal invasion, detecting brain metastases (as discussed in Section B.1), and evaluating adrenal lesions in staging workups.

Clinical Rationale for Image Synthesis Tasks: CT is the primary modality for lung cancer staging and radiotherapy planning, while FDG-PET provides complementary metabolic information for staging and treatment response assessment. **CT-to-PET** synthesis is clinically motivated as it could reduce or avoid the need for radiotracer administration, thereby limiting the associated radiation dose and operational complexity, particularly in follow-up or treatment-monitoring settings where repeated acquisitions are required [3, 76]. PET-to-CT synthesis is not clinically meaningful, as CT provides anatomical information that cannot be reliably inferred from metabolic PET data alone. **MRI-to-CT** synthesis in the thorax follows the same rationale described for the brain [37, 76]: a synthetic CT derived from MRI would simplify the clinical workflow and reduce radiation exposure. **CBCT-to-CT** synthesis is particularly relevant in the lung given the respiratory motion that characterizes thoracic treatments, which makes online adaptive planning especially challenging when relying on the inherently lower quality of CBCT images [76, 78]. Synthesizing CT-quality images from CBCT scans enables more accurate dose recalculation and adaptive replanning without requiring an additional CT acquisition.

B.3. Pelvis

Epidemiology: The pelvic region encompasses several of the most prevalent malignancies, including prostate, colon-rectal, cervical, and bladder cancer. In Europe, prostate cancer represented approximately 473,000 new cases in 2022, making it the most common cancer in men [39]. Colon-rectal cancer accounted for approximately 538,000 new cases in Europe in the same year, ranking second overall in incidence and representing the second most common cause of cancer death after lung cancer [39]. In the United States, prostate cancer is expected to account for 333,830 new cases and 36,320 deaths in 2026 [40]. Cervical cancer remains a significant global challenge, with approximately 662,000 new cases and 349,000 deaths recorded worldwide in 2022, with the vast majority occurring in low- and middle-income countries [39].

Imaging Modalities: MRI is the primary modality for diagnosis of prostate, rectal, and cervical cancer, due to its superior soft-tissue contrast in pelvis. Multi-parametric MRI (mpMRI), combining T2w, diffusion-weighted imaging (DWI), and dynamic contrast-enhanced (DCE) sequences, is the standard for prostate cancer localization and staging, and is increasingly used to guide targeted biopsy [84]. For rectal cancer, T2w MRI is essential for assessing mesorectal fascia involvement and guiding neoadjuvant treatment decisions. CT is routinely used for staging and radiotherapy planning across all pelvic malignancies, as it provides the information required for dose calculation. For bladder cancer, CT is the primary modality for loco-regional detection of lymph node involvement, while MRI is increasingly adopted for local T-staging, particularly in the assessment of muscle-invasive disease prior to surgical or radiotherapeutic planning. CBCT is acquired before each radiotherapeutic fraction for patient positioning verification and adaptive planning, though its soft-tissue contrast in the pelvis is substantially inferior to MRI or CT. Finally, FDG-PET/CT plays an important role in the diagnosis of colorectal and cervical cancers, particularly for detecting lymph node involvement and distant metastases [83].

Clinical Rationale for Image Synthesis Tasks: In pelvic oncology, MRI provides a superior soft-tissue contrast for local staging of prostate, rectal, and cervical cancers, while CT is required for radiotherapy dose planning and distant staging. **MRI-to-CT** synthesis is therefore clinically well-motivated, following the same rationale as in the brain and lung: a synthetic CT derived from MRI would enable an MRI-only radiotherapy planning workflow, eliminating the need for a separate CT acquisition [37, 76, 77]. **CBCT-to-CT** synthesis supports online adaptive radiotherapy planning, given that pelvic CBCT images suffer from inferior soft-tissue contrast compared to CT [76, 78]. Pelvis intra-MRI translation is clinically well-motivated, given the routine acquisition of multiple MRI sequences in prostate and rectal cancer protocols; however, no public paired intra-MRI dataset of the pelvic region was identified at the time of this study. Similarly, CT-to-PET synthesis for colorectal and cervical cancers shares the same clinical motivation as in the

lung [3], though no public paired dataset of pelvic FDG-PET/CT was identified. PET-to-CT and PET-to-MRI directions are not clinically meaningful in the pelvic context for the same reasons discussed above.

C. Datasets overview and dataset-specific pre-processing

This work evaluates medical image synthesis across multiple anatomical regions and imaging modalities, covering both inter-modality and intra-modality translations. The datasets have been identified through a systematic search of publicly available medical imaging repositories, and were selected based on their free accessibility and their exclusive coverage of oncological cases, in line with the clinical focus of this benchmark. Five datasets satisfied both criteria and have been retained for training and evaluation: SynthRAD2023 [42], SynthRAD2025 [38], BraTS2023 [43], autoPET [44], and ENHANCE.PET 1.6k [45]. Together, these collections cover a broad spectrum of cancer types, anatomical regions, and imaging modality pairs, making them a comprehensive and clinically grounded benchmarking of generative synthesis models in oncological imaging. Additional pre-processing steps beyond those provided with the original data releases are described in Section D.

C.1. SynthRAD23

The SynthRAD2023 dataset is provided by the SynthRAD Grand Challenge organizers [42] and covers two anatomical regions: brain and pelvis. Data were collected between 2018 and 2022 across three Dutch university medical centers. The patient cohort included 1,080 subjects in total, with ages ranging from 3 to 93 years (average 60), all treated with external beam radiotherapy (photon or proton) in the brain or pelvic region. For each anatomical region, two synthesis tasks are defined, yielding four experimental configurations in total: Task 1 consists of MR/CT image pairs and addresses T1-weighted MRI-to-CT synthesis; Task 2 consists of CBCT/CT image pairs and focuses on CBCT-to-CT translation. For brain images, gadolinium contrast agent was used in two of the three centers. For pelvis, two-thirds of MRI acquisitions were acquired with a T1-weighted spoiled gradient echo sequence and one-third with a T2-weighted fast spin echo sequence. Each of the four experimental configurations contains 180 subjects. It should be noted that these splits do not reflect the official challenge partitions. Since the ground-truth CTs for the official validation and test sets have not yet been released at the time of this study, only the 180 cases of the challenge training set have been used, for which ground truth scans were available. These were then redistributed into a custom 75%-25% train-test split applied independently for each task-anatomy combination, yielding 135 training and 45 test cases per subset, as reported in Table 1. Prior to release, the challenge organizers applied a standardized pre-processing pipeline comprising DICOM-to-NIfTI conversion; resampling to a grid of $1 \times 1 \times 1 \text{ mm}^3$ and $1 \times 1 \times 2.5 \text{ mm}^3$ for brain and pelvis regions, respectively; rigid inter-modality registration using Elastix [85]; and full anonymisation including defacing of brain images [42]. A binary body mask, generated via thresholding and hole-filling from the ITK toolkit and dilated to include a surrounding air margin, was provided alongside each image pair.

C.2. SynthRAD25

The SynthRAD2025 dataset [38] is the second edition of the SynthRAD Grand Challenge, extending its predecessor to three new anatomical regions: head/neck, thorax, and abdomen. As in SynthRAD2023, two synthesis tasks are defined for each region: T1-weighted MRI-to-CT (Task 1) and CBCT-to-CT (Task 2). The dataset comprises 2,362 cases in total, 890 MRI-CT pairs and 1,472 CBCT-CT pairs, collected at five European university medical centers across a wide range of scanners and acquisition protocols. For Task 1, 340, 280, and 270 cases were included for the head/neck, the thorax, and the abdomen cohort, respectively. For Task 2, all five centers contributed to each region, yielding 500, 495, and 477 CBCT-CT pairs for head/neck, thorax, and abdomen, respectively. For this benchmark, only two of the three anatomical regions were retained, to maintain consistency with the other datasets that cover brain, pelvis, and lung. Specifically, the thorax subset is referred to as 'Lung' throughout this work. The head/neck subset was additionally included because it encompasses the base of the skull and is therefore representative of brain anatomy. The abdomen region was excluded as it falls outside the anatomical scope of the benchmark. It should be noted that the official challenge validation and test sets were not employed, as the corresponding ground-truth CTs had not yet been released at the time of this study. Furthermore, data from one center was excluded, as it is distributed under a restricted license valid only during the official challenge. Available patients were redistributed into a custom 75%-25% train-test split applied independently for each task-anatomy combination, yielding the case counts reported in Table 1. As a result, the splits do not match with the official challenge partitions. All images were preprocessed by the challenge organisers through a pipeline comprising rigid registration of MRI/CBCT to the corresponding CT via

Elastix; automated defacing for patient anonymisation; resampling to a uniform voxel spacing of $1 \times 1 \times 3 \text{ mm}^3$; and body-mask generation.

C.3. BraTS23

The BraTS2023 dataset originates from the ASNR-MICCAI Brain Tumour Segmentation 2023 Glioma Challenge [43] and it is used in this benchmark for intra-MRI translation between T2-weighted and FLAIR contrasts. The task of BraTS2023 builds directly on the dataset introduced in the ASNR-MICCAI BraTS2021 benchmark [86]. Each patient scan includes four multi-parametric MRI sequences acquired across multiple institutions with heterogeneous equipment and imaging protocols: native T1-weighted, post-contrast T1-weighted, T2-weighted, and FLAIR. The full challenge dataset comprises 2,040 patients in total, distributed across a training set of 1,251 cases with publicly available ground-truth segmentation labels, a validation set of 219 cases with hidden labels and accessible only through the challenge evaluation platform, and a private test set of 570 cases used exclusively for the final challenge evaluation [86]. Since only the labeled subset can be used for supervised training, the 1,251 cases with ground-truth annotations were used in this work and divided into 938 training cases and 313 test cases (Table 1). All BraTS MRI scans were released with a standardized pre-processing pipeline by the challenge organizers [87, 88]. The pipeline comprises four sequential steps applied uniformly to all four MRI sequences: DICOM-to-NIfTI conversion; co-registration to the SRI24 multi-channel atlas of the normal adult human brain [89], ensuring a common spatial reference frame across modalities and subjects; resampling to a uniform isotropic resolution of $1 \times 1 \times 1 \text{ mm}^3$; and anonymisation.

C.4. AutoPET

The autoPET dataset is derived from a publicly available whole-body FDG-PET/CT collection [44] comprising 900 subjects and 1,014 studies acquired between 2014 and 2018 at the University Hospital of Tübingen. Preprocessing involved conversion to NIfTI format, followed by CT resampling to match the PET resolution. PET voxel values were converted into Standardized Uptake Values (SUV, [90]), defined as $SUV = \frac{r}{a' \cdot W}$ where r is the radioactivity concentration [kBq/ml] in a region of interest (ROI), a' is the decay-corrected injected activity [kBq], and W is the patient's weight [kg], representing the tracer's distribution volume. To account for modality-specific intensity variability, PET volumes were clipped and rescaled to the SUV range [0, 20], capturing the majority of tumor-related values. CT volumes were left unwindowed to preserve full dynamic range across anatomical structures. Most patients contribute a single acquisition, while a minority have more than one study. The original cohort includes patients diagnosed with malignant lymphoma, melanoma, and NSCLC lung cancer, as well as negative-control studies of subjects without confirmed malignancy. To maintain consistency with the oncological focus of this work, only NSCLC cases were retained. Subjects with lymphoma, melanoma, or no confirmed malignancy were excluded, yielding a final subset of 150 paired CT/FDG-PET cases. All volumes were cropped to the lung region, ensuring anatomical consistency with the body districts analysed across the benchmark.

C.5. EnhancePET

The ENHANCE.PET 1.6k dataset [45] is a large multi-centre whole-body FDG-PET/CT collection hosted on the AWS Open Data Registry, comprising approximately 1,600 subjects. Since 1,014 cases correspond to the autoPET cohort already described in Section C.4, those subjects were excluded to avoid any overlap between datasets. The remaining subjects, all diagnosed with lung cancer and acquired at two clinical sites, yielded a final cohort of 583 CT/FDG-PET pairs. Raw PET volumes were provided in Bq/mL units. Patient-specific acquisition parameters (body weighted injected dose) were retrieved from the metadata spreadsheet and used to convert each PET volume to SUV values according to:

$$SUV = PET_{\text{Bq/mL}} \times \frac{W \cdot 1000}{D \cdot 10^6} \quad (1)$$

where W is the patient body weight in kg and D is the injected dose in MBq, so that the resulting SUV images are expressed in g/mL. CT and PET volumes were acquired with different voxel spacings across scanners and centres. To harmonise the two modalities, the CT volumes were resampled to match the PET voxel grid using linear interpolation, filling regions outside the field of view with -1024 HU (the nominal HU for air). This degraded the CT to the coarser PET resolution, but ensured that both modalities share the same spatial grid before training. Lung masks provided with the dataset were resampled to the new spacing and used to crop both CT and PET volumes to the thoracic region.




Figure 10: Overview of the pre-processing pipeline. Each volume passes through eight sequential steps: body masking; voxel resampling; clipping; intensity normalization; spatial padding; foreground mask computation; mask intersection to obtain the common anatomical region of interest; and patch extraction.

Cropping was performed exclusively along the axial axis: the inferior and superior extents of the lung mask were identified as the first and last slices containing at least one lung voxel, and a margin of 20 mm was added in both directions to preserve the immediately surrounding anatomy. The resulting Z range was then applied identically to both modalities, so that the full field of view was retained in the transverse plane while the volume was restricted to the lung region.

D. Pre-processing Pipeline

All datasets share a common pre-processing pipeline implemented in PyTorch using MONAI transforms [91] (Figure 10). First, a body segmentation mask (if available) is applied independently to the source and target volumes, and background voxels are replaced with a modality-specific fill value. For CT, CBCT, and PET, the fill value is set to the minimum intensity observed within the foreground region, thereby avoiding artificial discontinuities at the mask boundary; for MRI it is set to zero. All volumes are then resampled to the target voxel spacing using trilinear interpolation, and scan intensities are linearly rescaled to zero mean and unit variance. For CT and CBCT, intensities are clipped to $[-1024, 3000]$ HU before rescaling; the lower bound represents air, while the upper bound covers the full range of clinically relevant tissue densities. The same approach is adopted for FDG-PET, with a clipping range $[0, 20]$ SUV encompassing normal tissues as well as highly metabolically active lesions. For MRI modalities, absolute intensity values are acquisition-dependent, making a fixed global range clipping unsuitable. Spatial padding is then applied to ensure volume dimensions are consistent with patch extraction. To identify voxels containing anatomically relevant tissue, a foreground mask is computed by thresholding each volume’s intensity values above a threshold of 0.1. For EnhancePET, a lower target threshold of 0.01 was used (Table 5). This reflects its more compressed intensity distribution of the PET signal after normalization, as selecting a threshold of 0.1 excludes a substantial fraction of metabolically active tissue. The two foreground masks are computed independently for the source and the target volume, and then intersected to get common region of interest.

All pre-processing operations are applied prior to patch extraction, where three patches of size $96 \times 96 \times 96$ voxels are randomly extracted from each training volume. To avoid patches focused on background regions, the extraction follows a foreground-guided sampling strategy: volume intensities are thresholded, and patch centers are placed only in correspondence of regions exhibiting above-threshold signals. Since the two modalities are co-registered, this strategy does not result in any loss of anatomically relevant information. Volumetric patches are then fed as input–target pairs to all AI models, as described in Section E. During inference, outputs are mapped back to the original image space through an invertible transformation pipeline that reverses the pre-processing steps applied, including intensity rescaling and spatial resampling.

E. Models

This work benchmarks seven generative models for paired medical image synthesis, spanning different architectural families: Generative Adversarial Networks (GANs) and latent generative models, which include diffusion-based and

Table 5

Voxel spacing (mm) after resampling and foreground intensity thresholds (applied after normalization) used to compute the foreground masks for source and target volumes.

Dataset	Voxel spacing (xxyyz)	Source threshold	Target threshold
Synthrad23 (Head/Neck)	$1.0 \times 1.0 \times 1.0$	0.1	0.1
Synthrad23 (Pelvis)	$1.0 \times 1.0 \times 2.5$	0.1	0.1
Synthrad25 (Lung)	$1.0 \times 1.0 \times 3.0$	0.1	0.1
Synthrad25 (Head/Neck)	$1.0 \times 1.0 \times 1.0$	0.1	0.1
BraTS23	$1.0 \times 1.0 \times 1.0$	0.1	0.1
autoPET	$2.0 \times 2.0 \times 3.0$	0.1	0.1
EnhancePET	$4.063 \times 4.063 \times 3.0$	0.1	0.01

flow matching approaches. All models are fully three-dimensional, operating directly on volumetric patches of size $96 \times 96 \times 96$ voxels, and are implemented in PyTorch (Figure 11).

E.1. Pix2Pix

Pix2Pix [11] is a conditional GAN composed of a U-Net generator and a PatchGAN discriminator that evaluates local patch synthesis conditioned on the source volume. Both the generator and the discriminator are optimized using the Adam optimizer with a learning rate of 10^{-4} , $\beta_1 = 0.5$, and $\beta_2 = 0.999$. The choice of $\beta_1 = 0.5$ follows the original Pix2Pix recommendation, which found lower momentum to stabilize GAN training. The learning rate schedule is decaying by a factor of 0.5 at epochs 200, 400, 600, and 800. Mixed-precision training with automatic gradient scaling is used throughout. To stabilize training, generated volumes are not used immediately to update the discriminator. Instead, they are stored in a pool of 50 previously generated volumes, from which one is randomly selected at each iteration. This exposes the discriminator to the history of generated images rather than only the most recent ones, reducing oscillations during training.

E.2. CycleGAN

CycleGAN [12] extends the Pix2Pix into a bidirectional image-translation framework, training two ResNet-based generators and two PatchGAN discriminators simultaneously. The implementation adopted in this work extends the original 2D framework to three-dimensional volumetric data, operating directly on 3D patches throughout all components of the architecture. All four networks are optimised independently using the Adam optimiser with a learning rate of 10^{-4} , $\beta_1 = 0.5$ and $\beta_2 = 0.999$. The learning rate schedule and mixed-precision training settings are identical to the one chosen for the Pix2Pix model.

E.3. SRGAN

SRGAN [21] is a multi-resolution 3D GAN specifically designed for volumetric medical image synthesis. The generator is a three-dimensional, multi-resolution Dense-Attention U-Net (3D-mDAUNet), organised as an encoder-decoder architecture with skip connections at each resolution level. The discriminator is a multi-resolution U-Net that produces voxel-wise real/fake predictions at four spatial scales, providing spatially dense adversarial feedback to the generator at each resolution level. Both the generator and the discriminator are optimised with the Adam optimiser with a learning rate of 10^{-4} and default momentum parameters ($\beta_1 = 0.9$, $\beta_2 = 0.999$). The learning rate is decayed by a factor of 0.5 at epochs 200, 400, 600, and 800 via a MultiStepLR scheduler, applied independently to each network. Automatic mixed-precision training with FP16 gradient scaling is used throughout, with separate gradient scalers for the generator and the discriminator to ensure numerical stability in the independent update steps.

E.4. Shared Variational Autoencoder

All diffusion and flow-based models in this work operate in a compressed latent space rather than directly in the voxel domain. This is made possible by a shared three-dimensional Variational Autoencoder (VAE) [93], which is pre-trained independently on each target modality and kept frozen during all subsequent model training. Here, the VAE learns a compact and structured latent representation of volumetric medical images, allowing the downstream models to operate on much smaller tensors without losing spatial detail. The VAE follows an encoder-decoder architecture based on the AutoencoderKL framework from the MONAI library [91]. Both the encoder and decoder are organised




Figure 11: Overview of the proposed benchmarking framework. The pipeline consists of four stages: (1) a configuration module, where the user specifies data, model, and training parameters and the dataset is split into training and test sets (75%–25%); (2) a data pipeline, which applies a sequence of preprocessing steps to produce paired source–target volumes; (3) a training pipeline, where GAN-based models operate in the image domain and latent generative models operate on VAE-encoded representations, following a unified training loop of forward pass, loss computation, backward pass, and checkpoint saving; and (4) an inference pipeline, where trained models generate full-resolution volumes via stitching, with model-specific inference strategies (single forward pass for GANs, DDIM sampling for diffusion models, Euler integration for Flow Matching), followed by quantitative evaluation using PSNR, SSIM, and NMSE.

into three hierarchical stages, with channel widths of 32, 64, and 128 respectively. For a patch of $96 \times 96 \times 96$ voxels, the encoder therefore produces a latent volume of $24 \times 24 \times 24$ voxels, corresponding to an $\times 4$ spatial compression factor. Both the autoencoder and the discriminator are optimized with the Adam optimizer with a learning rate of 10^{-4} . A warm-up schedule is applied for the first 10 epochs, during which the learning rate is reduced by a factor of 10, after which it remains constant. Training is performed for 300 epochs using mixed-precision.

E.5. LDM-Palette

LDM-Palette [46] performs the diffusion process entirely in the compressed latent space of the frozen VAE described in Section E.4. Both the source and target volumes are independently encoded into their respective latent representations using the frozen VAE encoder. The source latent is used as spatial conditioning by concatenating it with the noisy target latent channel-wise, before passing it to the denoising network. Since each latent has 3 channels, the concatenated input has 6 channels in total. This spatial concatenation conditioning avoids the need for cross-attention, making the conditioning computationally efficient. Noise is added to the target latent z_0 following the Denoising Diffusion Probabilistic Models (DDPM) formulation:

$$z_t = \sqrt{\bar{\alpha}_t} z_0 + \sqrt{1 - \bar{\alpha}_t} \epsilon, \quad \epsilon \sim \mathcal{N}(0, I) \quad (2)$$

where $\bar{\alpha}_t$ is the cumulative product of the noise schedule coefficients at timestep t . A scaled-linear beta schedule is used, with $\beta_{\text{start}} = 0.0015$, $\beta_{\text{end}} = 0.0205$, and $T = 1000$ total timesteps. The U-Net is optimised with AdamW with a learning rate of 1.5×10^{-5} , weight decay of 10^{-2} , and default momentum parameters. The learning rate is reduced by a factor of 0.5 when the training loss plateaus for 30 consecutive epochs via a ReduceLROnPlateau scheduler. Mixed-precision training with FP16 and gradient scaling is used throughout.

E.6. Brownian Bridge

The Brownian Bridge diffusion model [47] reformulates the diffusion process as a stochastic bridge that transitions directly between the source and target latent representations, without requiring a Gaussian noise prior at either endpoint. Unlike standard denoising diffusion models that start from pure noise, BBridge constructs a conditional process that interpolates between the two latent spaces, making the generative trajectory inherently conditioned on the source modality. Both the source and target volumes are first encoded into their respective latent representations using the frozen VAE encoder described in Section E.4. The Brownian Bridge forward process defines a noisy interpolated trajectory between the target latent z_0 and the source latent z_y at each discrete timestep t :

$$z_t = (1 - m_t) z_0 + m_t z_y + \sigma_t \epsilon, \quad \epsilon \sim \mathcal{N}(0, I) \quad (3)$$

where m_t is a mixing coefficient that increases linearly from 0.001 to 0.999 over $T = 1000$ timesteps, so that at $t = 0$ the process is close to the target latent and at $t = T$ it is close to the source latent. The noise variance at each step is defined as $\sigma_t^2 = 2(m_t - m_t^2)$, which peaks at the midpoint of the trajectory and vanishes at both endpoints, ensuring a smooth and bounded stochastic interpolation. The U-Net is optimised with AdamW with a learning rate of 1.5×10^{-5} , weight decay of 10^{-2} , and default momentum parameters. The learning rate is reduced by a factor of 0.5 when the training loss stops improving for 3000 consecutive iterations, via a ReduceLROnPlateau scheduler. An exponential moving average of the model weights is maintained with a decay of 0.995, starting from step 1000 and updated every 8 steps. Mixed-precision training with FP16 and gradient scaling is used throughout, with an initial scale of 65536.

E.7. LDM-Palette + ControlNet

ControlNet [34] augments a pre-trained diffusion model with a trainable auxiliary network that injects spatially structured conditioning information via residual connections, without modifying the backbone weights. In this work, the ControlNet framework is applied on top of the frozen LDM-Palette backbone described in Section E.5, extending the existing implementation proposed in [25], from 2D slice-wise processing to fully volumetric 3D inference in the latent space of the shared VAE, with only the ControlNet parameters updated during training. The ControlNet is optimised with AdamW with a learning rate of 1.5×10^{-5} , weight decay of 10^{-2} , and default momentum parameters. A cosine annealing schedule is applied over $T_{\text{max}} = 1000$ steps, decaying the learning rate to a minimum of 1.5×10^{-7} . Mixed-precision training with FP16 and gradient scaling is used throughout. As a training diagnostic, the magnitude of the ControlNet residuals is monitored every 50 epochs to verify that the module is contributing meaningfully to the denoising process without destabilising training. The diffusion schedule and inference procedure are identical to those of LDM-Palette, using DDIM sampling over 50 steps starting from Gaussian noise in the latent space.

E.8. Flow Matching

Flow Matching [48] offers a deterministic alternative to stochastic diffusion models by learning a time-varying velocity field that transports samples along straight-path trajectories directly from source to target. Unlike LDM-Palette and ControlNet, which learn to predict the noise added to the target latent at each timestep and require a stochastic

reverse process guided by a noise scheduler, Flow Matching learns to predict the velocity, that is the direction and rate of change needed to move from the source latent towards the target latent in a single continuous flow. The inference process is therefore fully deterministic and does not require any noise scheduler, as integration starts directly from the source latent rather than from Gaussian noise. The implementation adopted here extends the implementation proposed in [25] from 2D slice-wise processing to fully volumetric 3D inference in the latent space of the shared VAE. Unlike the other diffusion-based models, the Euler integration is implemented directly in the trainer rather than relying on an external inference scheduler, keeping the sampling procedure self-contained and independent of the diffusion library infrastructure. The velocity U-Net is optimised with AdamW with a learning rate of 1.5×10^{-5} , weight decay of 10^{-2} , and default momentum parameters. The learning rate is reduced by a factor of 0.5 when the training loss plateaus for 30 consecutive epochs via a ReduceLROnPlateau scheduler. Mixed-precision training with FP16 and gradient scaling is used throughout.

F. Visual Turing Test Platform

The Visual Turing Test was administered through a custom web platform specifically designed for the interactive review of 3D medical volumes (Figure 12).

The platform is implemented as a client-side JavaScript application served through JATOS (Just Another Tool for Online Studies), enabling deployment on any institutional server without external dependencies. Volumes are rendered using NiiVue v0.50.0 [92], which provides WebGL-accelerated multi-planar reformatting and 3D surface reconstruction directly in the browser. The platform architecture separates study configuration (JSON-based question definitions, volume paths, and randomization seeds) from the rendering and interaction logic, allowing researchers to create new reader studies by editing a single configuration file. Supported question types include: (i) single-volume binary classification (real vs. synthetic); (ii) side-by-side pairwise preference with optional synchronized slice navigation; and (iii) multi-volume ordinal ranking with drag-and-drop or button-based rank assignment. Sanity-check questions (identical volume pairs or triplets) can be inserted at arbitrary positions to monitor response consistency. All responses are stored server-side in CSV format, anonymized by participant ID, and include timestamps for response-time analysis. The platform source code, including deployment instructions, example study configurations, and sample volumes, is available alongside the benchmarking code and can be adapted for reader studies involving any NIfTI-compatible 3D medical imaging modality.

Upon accessing their personalised link, participants were presented with a landing page that collected the number of years of clinical experience (<5 years / 5–10 years / >10 years) and primary imaging modality (CT, MRI, PET, or all modalities). Each volume was displayed through a **multi-planar viewer** rendered by a grid layout providing three orthogonal anatomical planes (axial, sagittal, and coronal) alongside a 3D surface reconstruction. The three planar views were spatially linked, allowing simultaneous navigation across all anatomical planes.

Participants could adjust image appearance and navigate volumes via the following per-volume controls: a *brightness slider*, that shifts the centre of the displayed intensity window, effectively controlling overall image brightness; a *contrast slider*, that controls the width of the intensity window, affecting tonal contrast; and a *slice slider*, that positioned the axial crosshair along the superior–inferior axis, initialised at the mid-slice of each volume. The first two sliders were initialised using *auto-windowing*: the window was set to the full data range of the loaded volume (robust minimum and maximum), ensuring visually comparable display regardless of whether the image was a real acquisition (with native HU values) or a normalised model output. For CT brain tasks, a clinical preset of 50 HU for the brightness slider and 640 HU for the contrast slider was applied as the default.

Following the completion of the test, some participants provided feedback highlighting aspects of the study design that may have influenced their responses, and that should be considered when designing future reader studies for medical image synthesis. Specifically, participants could not adjust the window width and level of the displayed volumes; all images were presented in a grayscale colormap; no maximum intensity projection (MIP) views were provided; and all images were shown in isolation, whereas in clinical practice PET is interpreted as a hybrid examination in combination with CT or MRI. These factors were taken into account when interpreting the Turing test results, and future studies would benefit from incorporating adjustable windowing, clinically standard color maps, MIP reconstructions, and fused multi-modal displays to better align with the diagnostic reading environment.




Figure 12: The Visual Turing test platform. Each volume was displayed through a multi-planar viewer rendered by a grid layout providing three orthogonal anatomical planes (axial, sagittal, and coronal) alongside a 3D surface reconstruction. In Part 1, a single volume was displayed and participants were asked to classify the image as either *Real* or *AI-generated* using two mutually exclusive buttons positioned below the viewer. In Part 2, two volumes from the same patient were displayed side by side, each with independent brightness, contrast, and slice controls. Participants selected *Prefer Image A*, *Prefer Image B*, or *No Difference*. An optional toggle linked the slice positions of the two viewers, enabling direct anatomical comparison at the same relative depth across both images simultaneously. In Part 3, three volumes from the same patient were displayed simultaneously in a three-column grid, each with independent viewer controls. Participants assigned a rank from 1 (most realistic) to 3 (least realistic) to each image using circular rank buttons below each viewer.