

Table of contents

\USER

Auerbach

REMIND

LOC3

anat-loc_ses-lo1
anat-T1w_acq-MPR
fmap-acq-rest_dir-PA
fmap-acq-rest_dir-AP
func-bold_task-rest_run-01
func-bold_task-rest_run-02
fmap-acq-selfref_dir-PA
fmap-acq-selfref_dir-AP
func-bold_task-selfref_run-01
func-bold_task-selfref_run-02

\\USER\Auerbach\REMIND\LOC3\anat-loc_ses-lo1

TA: 0:12 PM: REF Voxel size: 0.5×0.5×7.0 mmPAT: Off Rel. SNR: 1.00 : fl

Properties

Prio recon	On
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	On
Load images to graphic segments	On
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	1
Dist. factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Slices	1
Dist. factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Transversal
Phase enc. dir.	A >> P
Slice group	3
Slices	1
Dist. factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Coronal
Phase enc. dir.	R >> L
AutoAlign	---
Phase oversampling	0 %
FoV read	250 mm
FoV phase	100.0 %
Slice thickness	7.0 mm
TR	7.5 ms
TE	3.69 ms
Averages	2
Concatenations	3
Filter	Prescan Normalize, Elliptical filter
Coil elements	HC1-7

Contrast - Common

TR	7.5 ms
TE	3.69 ms
TD	0 ms
MTC	Off
Magn. preparation	None
Flip angle	20 deg
Fat suppr.	None
Water suppr.	None
SWI	Off

Contrast - Dynamic

Averages	2
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1

Contrast - Dynamic

Multiple series	Each measurement
-----------------	------------------

Resolution - Common

FoV read	250 mm
FoV phase	100.0 %
Slice thickness	7.0 mm
Base resolution	256
Phase resolution	91 %
Phase partial Fourier	Off
Interpolation	On

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	On

Geometry - Common

Slice group	1
Slices	1
Dist. factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Slices	1
Dist. factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Transversal
Phase enc. dir.	A >> P
Slice group	3
Slices	1
Dist. factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Coronal
Phase enc. dir.	R >> L
FoV read	250 mm
FoV phase	100.0 %
Slice thickness	7.0 mm
TR	7.5 ms
Multi-slice mode	Sequential
Series	Interleaved
Concatenations	3

Geometry - AutoAlign

Slice group	1
Position	L0.0 A20.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Position	L0.0 A20.0 H0.0 mm

Geometry - AutoAlign

Orientation	Transversal
Phase enc. dir.	A >> P
Slice group	3
Position	L0.0 A20.0 H0.0 mm
Orientation	Coronal
Phase enc. dir.	R >> L
AutoAlign	---
Initial Position	L0.0 A20.0 H0.0
L	0.0 mm
A	20.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Saturation mode	Standard
Fat suppr.	None
Water suppr.	None
Special sat.	None

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Slice-sel.

System - Tx/Rx

Frequency 1H	123.259412 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	7.5 ms
Concatenations	3
Segments	1

Physio - Cardiac

Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	250 mm
FoV phase	100.0 %
Phase resolution	91 %

Physio - PACE

Resp. control	Off
Concatenations	3

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Liver registration	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Soft Tissue

Wash - In	Off
Wash - Out	Off
TTP	Off
PEI	Off
MIP - time	Off
Measurements	1

Inline - Composing

Distortion Corr.	Off
------------------	-----

Sequence - Part 1

Introduction	On
Dimension	2D
Phase stabilisation	Off
Asymmetric echo	Allowed
Contrasts	1
Flow comp.	No
Multi-slice mode	Sequential
Bandwidth	320 Hz/Px

Sequence - Part 2

Segments	1
Acoustic noise reduction	None
RF pulse type	Fast
Gradient mode	Fast
Excitation	Slice-sel.
RF spoiling	On

Sequence - Assistant

Mode	Off
Allowed delay	0 s

\\USER\Auerbach\REMIND\LOC3\anat-T1w_acq-MPR

TA: 4:30 PM: REF Voxel size: 1.0×1.0×1.0 mmPAT: 3 Rel. SNR: 1.00 : tfl

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	On
Load images to graphic segments	On
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Dist. factor	50 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	Head > Basis
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	176
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	1.00 mm
TR	2530.0 ms
TE	1.92 ms
Averages	1
Concatenations	1
Filter	Prescan Normalize
Coil elements	HC1-7

Contrast - Common

TR	2530.0 ms
TE	1.92 ms
Magn. preparation	Non-sel. IR
T1	1400 ms
Flip angle	7 deg
Fat suppr.	None
Water suppr.	None

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	256 mm
FoV phase	100.0 %
Slice thickness	1.00 mm
Base resolution	256
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Off
Slice partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	32
Accel. factor 3D	1
Reference scan mode	Integrated

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	On
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slab group	1
Slabs	1
Dist. factor	50 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice oversampling	0.0 %
Slices per slab	176
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	1.00 mm
TR	2530.0 ms
Multi-slice mode	Single shot
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	Head > Basis
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Navigator**System - Miscellaneous**

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off

System - Miscellaneous

Matrix Optimization	Off
AutoAlign	Head > Basis
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Sagittal
Rotation	0.00 deg
A >> P	256 mm
F >> H	256 mm
R >> L	176 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.259412 MHz
Correction factor	1
Gain	Low
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	2530.0 ms
Concatenations	1

Physio - Cardiac

Magn. preparation	Non-sel. IR
TI	1400 ms
Fat suppr.	None
Dark blood	Off
FoV read	256 mm
FoV phase	100.0 %
Phase resolution	100 %

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off

Inline - MIP

Save original images	On
----------------------	----

Inline - Composing

Distortion Corr.	Off
------------------	-----

Sequence - Part 1

Introduction	On
Dimension	3D
Elliptical scanning	Off
Reordering	Linear
Asymmetric echo	Allowed
Flow comp.	No
Multi-slice mode	Single shot
Echo spacing	5.9 ms
Bandwidth	260 Hz/Px

Sequence - Part 2

RF pulse type	Normal
Gradient mode	Normal
Excitation	Non-sel.
RF spoiling	On
Incr. Gradient spoiling	Off
Turbo factor	176

Sequence - Assistant

Mode	Off
------	-----

\\USER\Auerbach\REMIND\LOC3\ fmap-acq-rest_dir-PA

TA: 0:36 PM: REF Voxel size: 2.0×2.0×2.0 mmPAT: 2 Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Resolution - iPAT

Ref. lines PE	36
Reference scan mode	EPI/separate

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	P >> A
AutoAlign	---
Phase oversampling	0 %
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	6000 ms
TE	43.0 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	HC1-7

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	P >> A
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	6000 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	P >> A
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	-180.00 deg
Initial Orientation	Transversal

Contrast - Common

TR	6000 ms
TE	43.0 ms
MTC	Off
Magn. preparation	None
Fat suppr.	Fat sat.
Fat sat. mode	Weak

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	3
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	256 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
Base resolution	128
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

Accel. mode	GRAPPA
Accel. factor PE	2

Geometry - Saturation

Fat suppr.	Fat sat.
Fat sat. mode	Weak
Special sat.	None

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Performance
AutoAlign	---
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	180.00 deg
A >> P	256 mm
R >> L	256 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
--------------	----------

System - Tx/Rx

Frequency 1H	123.259412 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	6000 ms
Concatenations	1

Inline - Common

Subtract	Off
Measurements	3
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Distortion Corr.	Off
------------------	-----

Sequence - Part 1

Introduction	Off
Multi-slice mode	Interleaved
Free echo spacing	On
Echo spacing	0.57 ms
Bandwidth	2170 Hz/Px

Sequence - Part 2

EPI factor	128
RF pulse type	Normal
Gradient mode	Performance

\\USER\Auerbach\REMIND\LOC3\ fmap-acq-rest_dir-AP

TA: 0:36 PM: REF Voxel size: 2.0×2.0×2.0 mmPAT: 2 Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	R >> L
AutoAlign	---
Phase oversampling	0 %
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	6000 ms
TE	43.0 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	6000 ms
TE	43.0 ms
MTC	Off
Magn. preparation	None
Fat suppr.	Fat sat.
Fat sat. mode	Weak

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	3
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	256 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
Base resolution	128
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

Accel. mode	GRAPPA
Accel. factor PE	2

Resolution - iPAT

Ref. lines PE	36
Reference scan mode	EPI/separate

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	R >> L
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	6000 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	R >> L
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	-180.00 deg
Initial Orientation	Transversal

Geometry - Saturation

Fat suppr.	Fat sat.
Fat sat. mode	Weak
Special sat.	None

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Performance
AutoAlign	---
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	90.00 deg
R >> L	256 mm
A >> P	256 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
--------------	----------

System - Tx/Rx

Frequency 1H	123.259412 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	6000 ms
Concatenations	1

Inline - Common

Subtract	Off
Measurements	3
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Distortion Corr.	Off
------------------	-----

Sequence - Part 1

Introduction	Off
Multi-slice mode	Interleaved
Free echo spacing	On
Echo spacing	0.57 ms
Bandwidth	2170 Hz/Px

Sequence - Part 2

EPI factor	128
RF pulse type	Normal
Gradient mode	Performance

\\USER\Auerbach\REMIND\LOC3\func-bold_task-rest_run-01

TA: 5:13 PM: REF Voxel size: 2.0×2.0×2.0 mmPAT: 8 Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Resolution - iPAT

Accel. factor slice	4
Reference scan mode	EPI/separate

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	P >> A
AutoAlign	---
Phase oversampling	0 %
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	1200 ms
TE	30.0 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	HC1-7

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	P >> A
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	1200 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	P >> A
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	-180.00 deg
Initial Orientation	Transversal

Contrast - Common

TR	1200 ms
TE	30.0 ms
MTC	Off
Flip angle	61 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	250
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	256 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
Base resolution	128
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

Accel. mode	Slice accel.
Accel. factor PE	2
Ref. lines PE	36

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Performance
AutoAlign	---
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm

System - Adjustments

Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	180.00 deg
A >> P	256 mm
R >> L	256 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.259412 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	1200 ms
Concatenations	1

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	15
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Baseline
Meas[12]	Baseline
Meas[13]	Baseline
Meas[14]	Baseline
Meas[15]	Active
Motion correction	Off
Spatial filter	Off
Measurements	250
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.57 ms
Bandwidth	2170 Hz/Px

Sequence - Part 2

EPI factor	128
RF pulse type	Normal
Gradient mode	Performance
Excitation	Standard

Sequence - pTX Pulses**Sequence - Part 1**

Introduction	Off
--------------	-----

\\USER\Auerbach\REMIND\LOC3\func-bold_task-rest_run-02

TA: 5:13 PM: REF Voxel size: 2.0×2.0×2.0 mmPAT: 8 Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	P >> A
AutoAlign	---
Phase oversampling	0 %
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	1200 ms
TE	30.0 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	1200 ms
TE	30.0 ms
MTC	Off
Flip angle	61 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	250
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	256 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
Base resolution	128
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

Accel. mode	Slice accel.
Accel. factor PE	2
Ref. lines PE	36

Resolution - iPAT

Accel. factor slice	4
Reference scan mode	EPI/separate

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	P >> A
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	1200 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	P >> A
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	-180.00 deg
Initial Orientation	Transversal

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Performance
AutoAlign	---
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm

System - Adjustments

Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	180.00 deg
A >> P	256 mm
R >> L	256 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.259412 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	1200 ms
Concatenations	1

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	15
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Baseline
Meas[12]	Baseline
Meas[13]	Baseline
Meas[14]	Baseline
Meas[15]	Active
Motion correction	Off
Spatial filter	Off
Measurements	250
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.57 ms
Bandwidth	2170 Hz/Px

Sequence - Part 2

EPI factor	128
RF pulse type	Normal
Gradient mode	Performance
Excitation	Standard

Sequence - pTX Pulses**Sequence - Part 1**

Introduction	Off
--------------	-----

\\USER\Auerbach\REMIND\LOC3\fmmap-acq-selfref_dir-PA

TA: 0:36 PM: REF Voxel size: 2.0×2.0×2.0 mmPAT: 2 Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Resolution - iPAT

Ref. lines PE	36
Reference scan mode	EPI/separate

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	P >> A
AutoAlign	---
Phase oversampling	0 %
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	6000 ms
TE	43.0 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	HC1-7

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	P >> A
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	6000 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	P >> A
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	-180.00 deg
Initial Orientation	Transversal

Contrast - Common

TR	6000 ms
TE	43.0 ms
MTC	Off
Magn. preparation	None
Fat suppr.	Fat sat.
Fat sat. mode	Weak

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	3
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	256 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
Base resolution	128
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

Accel. mode	GRAPPA
Accel. factor PE	2

Geometry - Saturation

Fat suppr.	Fat sat.
Fat sat. mode	Weak
Special sat.	None

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Performance
AutoAlign	---
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	180.00 deg
A >> P	256 mm
R >> L	256 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
--------------	----------

System - Tx/Rx

Frequency 1H	123.259412 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	6000 ms
Concatenations	1

Inline - Common

Subtract	Off
Measurements	3
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Distortion Corr.	Off
------------------	-----

Sequence - Part 1

Introduction	Off
Multi-slice mode	Interleaved
Free echo spacing	On
Echo spacing	0.57 ms
Bandwidth	2170 Hz/Px

Sequence - Part 2

EPI factor	128
RF pulse type	Normal
Gradient mode	Performance

\\USER\Auerbach\REMIND\LOC3\fmmap-acq-selfref_dir-AP

TA: 0:36 PM: REF Voxel size: 2.0×2.0×2.0 mmPAT: 2 Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Resolution - iPAT

Ref. lines PE	36
Reference scan mode	EPI/separate

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	R >> L
AutoAlign	---
Phase oversampling	0 %
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	6000 ms
TE	43.0 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	HC1-7

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	R >> L
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	6000 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	R >> L
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	-180.00 deg
Initial Orientation	Transversal

Contrast - Common

TR	6000 ms
TE	43.0 ms
MTC	Off
Magn. preparation	None
Fat suppr.	Fat sat.
Fat sat. mode	Weak

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	3
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	256 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
Base resolution	128
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

Accel. mode	GRAPPA
Accel. factor PE	2

Geometry - Saturation

Fat suppr.	Fat sat.
Fat sat. mode	Weak
Special sat.	None

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Performance
AutoAlign	---
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	90.00 deg
R >> L	256 mm
A >> P	256 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
--------------	----------

System - Tx/Rx

Frequency 1H	123.259412 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	6000 ms
Concatenations	1

Inline - Common

Subtract	Off
Measurements	3
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Distortion Corr.	Off
------------------	-----

Sequence - Part 1

Introduction	Off
Multi-slice mode	Interleaved
Free echo spacing	On
Echo spacing	0.57 ms
Bandwidth	2170 Hz/Px

Sequence - Part 2

EPI factor	128
RF pulse type	Normal
Gradient mode	Performance

\\USER\Auerbach\REMIND\LOC3\func-bold_task-selfref_run-01

TA: 6:42 PM: REF Voxel size: 2.0×2.0×2.0 mmPAT: 8 Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Resolution - iPAT

Accel. factor slice	4
Reference scan mode	EPI/separate

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	P >> A
AutoAlign	---
Phase oversampling	0 %
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	1200 ms
TE	30.0 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	HC1-7

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	P >> A
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	1200 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	P >> A
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	-180.00 deg
Initial Orientation	Transversal

Contrast - Common

TR	1200 ms
TE	30.0 ms
MTC	Off
Flip angle	61 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	324
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	256 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
Base resolution	128
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

Accel. mode	Slice accel.
Accel. factor PE	2
Ref. lines PE	36

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Performance
AutoAlign	---
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm

System - Adjustments

Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	180.00 deg
A >> P	256 mm
R >> L	256 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.259412 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	1200 ms
Concatenations	1

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	15
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Baseline
Meas[12]	Baseline
Meas[13]	Baseline
Meas[14]	Baseline
Meas[15]	Active
Motion correction	Off
Spatial filter	Off
Measurements	324
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.57 ms
Bandwidth	2170 Hz/Px

Sequence - Part 2

EPI factor	128
RF pulse type	Normal
Gradient mode	Performance
Excitation	Standard

Sequence - pTX Pulses**Sequence - Part 1**

Introduction	Off
--------------	-----

\\USER\Auerbach\REMIND\LOC3\func-bold_task-selfref_run-02

TA: 6:42 PM: REF Voxel size: 2.0×2.0×2.0 mmPAT: 8 Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	P >> A
AutoAlign	---
Phase oversampling	0 %
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	1200 ms
TE	30.0 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	1200 ms
TE	30.0 ms
MTC	Off
Flip angle	61 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	324
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	256 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
Base resolution	128
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

Accel. mode	Slice accel.
Accel. factor PE	2
Ref. lines PE	36

Resolution - iPAT

Accel. factor slice	4
Reference scan mode	EPI/separate

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	P >> A
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	1200 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	P >> A
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	-180.00 deg
Initial Orientation	Transversal

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Performance
AutoAlign	---
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm

System - Adjustments

Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	180.00 deg
A >> P	256 mm
R >> L	256 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.259412 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	1200 ms
Concatenations	1

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	15
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Baseline
Meas[12]	Baseline
Meas[13]	Baseline
Meas[14]	Baseline
Meas[15]	Active
Motion correction	Off
Spatial filter	Off
Measurements	324
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.57 ms
Bandwidth	2170 Hz/Px

Sequence - Part 2

EPI factor	128
RF pulse type	Normal
Gradient mode	Performance
Excitation	Standard

Sequence - pTX Pulses**Sequence - Part 1**

Introduction	Off
--------------	-----