

Patient name : Fernandez^Claudia Ines  
 Patient ID : 5-21165-0  
 Plan file : Fernandez\_Claudia\_Ineset1  
 Clinic : Mevaterapia  
 Physician :  
 Physicist :  
 Plan Approved by : -----  
 Machine name:CETRO\_Fotones\_06MV (Linac Photons)  
 Calibration [cGy/MU]: 1.0000 (27/12/2017)  
 (calibration at reference SSD + dmax)

Target dose : 240.0 Fractions: 25 Total: 6000 cGy  
 Target percent selected : 100.0 (Normalization value: 107.0)

| FIELD #    | 1               | 2               | 3               | 4               | 5              | 6              |
|------------|-----------------|-----------------|-----------------|-----------------|----------------|----------------|
| ISOCENTER  | ISO1            | ISO1            | ISO1            | ISO1            | ISO1           | ISO1           |
| INPLANE    | 125             | 130             | 90              | 30              | 95             | 95             |
| CROSSPLANE | 155             | 150             | 155             | 105             | 50             | 55             |
| TABLE ANG  | 180             | 180             | 180             | 180             | 180            | 180            |
| GANTRY ANG | 135             | 320             | 135             | 350             | 170            | 350            |
| GANTRY ROT | 0               | 0               | 0               | 0               | 0              | 0              |
| COLLIM ANG | 270             | 90              | 270             | 180             | 270            | 90             |
| WEIGHT     | 0.550           | 0.650           | 0.100           | 0.130           | 0.630          | 0.565          |
| WEDGE FILT | EDW30           | NONE            | NONE            | NONE            | NONE           | NONE           |
| WEDGE POS  | IN              | ----            | ----            | ----            | ----           | ----           |
| ShieldFile | Fernandez_e4b61 | Fernandez_e4b72 | Fernandez_e4b63 | Fernandez_e4ba4 | Fernandez_e4b6 | Fernandez_e4b6 |
| TrayFactor | 1.000           | 1.000           | 1.000           | 1.000           | 1.000          | 1.000          |
| HEAD-JAW   | 35              | 40              | 35              | 15              | 80             | 80             |
| FEET-JAW   | 90              | 90              | 55              | 15              | 15             | 15             |
| RIGHT-JAW  | 5               | 145             | 5               | 80              | 45             | 5              |
| LEFT-JAW   | 150             | 5               | 150             | 25              | 5              | 50             |
| SSD        | 939.8           | 888.4           | 939.8           | 888.2           | 949.8          | 888.2          |
| Device     | 0               | 0               | 0               | 0               | 0              | 0              |
| MON. UNITS | 118.8           | 113.5           | 19.7            | 24.3            | 131.5          | 103.2          |

Dose Computation Mode: Convolution - Pencil Beam 8x8mm  
 Normalized at depth of maximum dose for each field  
 Density correction ON