Boston Scientific | IMAGEREADY™ MR-Conditional Systems

# ImageReady™ MR-Conditional* Model List*

| Therapy | Device Name | Device Model Numbers | MRI System Configuration* |
|---|---|---|---|
| **ICM[1]** | LUX-Dx™ | M301 | 1.5T or 3T |
| | LUX-Dx™ II+ | M312 | |
| **S-ICD[2]** | EMBLEM™ | A209 | 1.5T |
| | EMBLEM™MRI | A219 | |

The following S-ICD electrodes are labeled as MR-Conditional*

**S-ICD ELECTRODES:** 3010, 3400, 3401, 3501

## Brady Pulse Generators[4]

| Therapy | Device Name | Device Model Numbers | MRI System Configuration* |
|---|---|---|---|
| **Pacemakers** | ACCOLADE™ MRI | L310, L311, L331 | 1.5T or 3T |
| | FORMIO™ MRI | K289 | 1.5T |
| | VITALIO™ MRI | K285, K286, K287 | |
| | INGENIO™ MRI | K185, K186, K187 | |
| | ADVANTIO™ MRI | K085, K086, K087 | |
| **CRT-P** | VISIONIST™X4 | U228 | 1.5T or 3T |

The following Pacemaker and CRT-P leads and accessories are labeled as MR-Conditional*

**Right Atrium and Right Ventricle Leads and Accessories**

**INGEVITY™ MRI:** 7731, 7732, 7735, 7736, 7740, 7741, 7742
**INGEVITY+™:** 7840,7841, 7842
**Suture Sleeves for INGEVITY™ MRI /INGEVITY™+ leads:** 6402
**FINELINE™ II EZ:** 4469, 4470, 4471, 4472, 4473, 4474
**FINELINE™ II:** 4456, 4457, 4458, 4459, 4479, 4480
**Port Plugs:** 7145

**Left Ventricle Leads and Accessories**

**ACUITY™ X4:** 4671, 4672, 4674, 4675, 4677, 4678
**Suture Sleeves:** 4603
**Port Plugs:** 7148

Boston Scientific | IMAGEREADY™ MR-Conditional Systems

# ImageReady™ MR-Conditional* Model List*

The following ICD and CRT-D leads and accessories are labeled as MR-Conditional*

## Tachy Pulse Generators[3]

| Therapy | Device Name | Device Model Numbers | MRI System Configuration* |
|---|---|---|---|
| ICDs | RESONATE™ | D432, D433 | 1.5T or 3T |
| | MOMENTUM™ | D120, D121 | 1.5T |
| | AUTOGEN™ | D044, D045, D046, D047, D174, D175, D176, D177 | |
| | DYNAGEN™ | D020, D021, D022, D023, D150, D151, D152, D153 | |
| CRT-Ds | RESONATE™ | G447 | 1.5T or 3T |
| | MOMENTUM™ | G124, G125, G128, G138 | 1.5T |
| | AUTOGEN™ | G172, G173, G177, G179 | |
| | DYNAGEN™ | G150, G151, G156, G158 | |

### Right Atrium and Right Ventricle Leads and Accessories

**INGEVITY™ MRI:** 7735, 7736, 7740, 7741, 7742
**INGEVITY+™:** 7840,7841, 7842
**Suture Sleeves for INGEVITY™ MRI /INGEVITY+ leads:** 6402
**FINELINE™ II:** 4479, 4480
**FINELINE™ II EZ:** 4469, 4470, 4471, 4472, 4473, 4474
**Suture Sleeves for Fineline leads:** 6220, 6221
**Port Plugs:** 7145

### Left Ventricle Leads and Accessories

**ACUITY™ X4:** 4671, 4672, 4674, 4675, 4677, 4678
**Suture Sleeves:** 4603
**IS4/DF4 Port Plugs:** 7148
**ACUITY™ Spiral:** 4591, 4592, 4593 (1.5 T only)
**IS-1 Port Plugs:** 7145 (1.5 T only)

### Right Ventricle Leads and Accessories

**ENDOTAK RELIANCE 4-SITE™ (DF4):** 0262, 0263, 0265, 0266, 0272, 0273, 0275, 0276, 0282, 0283, 0285, 0286, 0292, 0293, 0295, 0296
**ENDOTAK RELIANCE™ (DF1):** 0170, 0171, 0172, 0173,0174, 0175, 0176, 0177, 0180, 0181, 0182, 0183, 0184, 0185, 0186, 0187 (1.5 T only)
**RELIANCE 4-FRONT™:** 0636, 0650, 0651, 0652, 0653, 0654, 0655, 0657, 0658, 0662, 0663, 0665, 0672, 0673, 0675, 0676, 0682, 0683, 0685, 0686, 0692, 0693, 0695, 0696
**Suture Sleeves for RELIANCE 4-FRONT™:** 6403
**DF-1 Lead Port Plug for ENDOTAK RELIANCE™ (IS-1) Leads- Single Coil:** 6996 (1.5 T only)

References:
1. MRI TECHNICAL GUIDE IMAGEREADY MR CONDITIONAL ICM SYSTEM 51583241-001 EN Europe 2023-05
2. MRI TECHNICAL GUIDE IMAGEREADY MR CONDITIONAL S-ICD SYSTEM 92346927-001 EN EU 2020-09
3. MRI TECHNICAL GUIDE IMAGEREADY MR CONDITIONAL DEFIBRILLATION SYSTEM 51164993-002 EN AUS 2021-10
4. MRI TECHNICAL GUIDE IMAGEREADY MR CONDITIONAL PACING SYSTEM 97036544-001 EN Australia 2023-04

*When conditions of use are met.

Visit http://www.bostonscientific.com/imageready for additional information cardiology/radiology checklists, patient resources, and the MRI technical manual.

Boston Scientific
Advancing science for life™

2025 Copyright © Boston Scientific Corporation or its affiliates. All rights reserved.

CRM-1889601-AB

CAUTION: The law restricts these devices to sale by or on the order of a physician. Indications, contraindications, warnings, and instructions for use can be found in the product labelling supplied with each device or at www.IFU-BSCI.com.

Boston Scientific Pty Ltd | PO Box 332 Botany NSW 1455 Australia | Tel +61 2 8063 8100
Boston Scientific NZ Ltd, 3 Diamond Street, Eden Terrace, Auckland 1021, New Zealand