



Shreeji Calibration Laboratory LLP

(FORMERLY KNOWN AS SHREEJI CALIBRATION LABORATORY)

" A Complete Calibration & Validation Center " " An NABL Accredited Laboratory "



CC-4263

CALIBRATION CERTIFICATE

1. NAME & ADDRESS OF CUSTOMER				UNIQUE LAB REPORT NO.		CC426326000001851F
M/S MAYA INTERNATIONAL EXPORT Gat No. 407/1/B Tanda, Gondia, Maharashtra-441614				CERTIFICATE NO.		260331001
				DATE OF CALIBRATION		31/03/2026
				DATE OF ISSUE		02/04/2026
				NEXT DUE DATE (as per agreed with customer)		30/03/2027
				PAGE		01 of 01
2. SRF NO. & DATE :- 260328001 & 28/03/2026						
3. CALIBRATION LOCATION :- ON SITE						
4. DESCRIPTION OF UUC						
Instrument Name	Make / Sr. No.	ID No.	Range	Least Count	Type	Location
Digital Temperature Controller	Shivam	MIE/BOD/01	10 to 40 °C	0.1°C	Pt-100	BOD Incubator (20°C)
5. CONDITION OF UUC & DATE OF RECEIPT :- GOOD & 31/03/2026						
6. CALIBRATION STANDARD USED:-						
Name	Make / ID No.	Certificate No:	Valid up to	Calibrated By		
Universal Calibrator	Masibus / SCL/UC/02	C&IJ/CAL/25-06/077	18/06/2026	CC-3959		
7. ENVIRONMENTAL CONDITION :-						
Ambient Temperature :- 25 ± 10 °C			Relative Humidity :- 50 ± 20 %RH			
8. METHOD USED FOR CALIBRATION :- SCL/SOP/ET/01 Based on ITS-90 & NABL-129						
9. DISCIPLINE :- Electro - Technical						
10. CALIBRATION RESULTS :-						
Sr. No.	Calibration Point	Standard Reading	UUC Reading	Error	Expanded Uncertainty (± °C)	
	°C	°C	°C	°C		
1	10.0	10.0	10.2	0.2	1.03	
2	15.0	15.0	15.3	0.3	1.03	
3	20.0	20.0	20.1	0.1	1.03	
4	30.0	30.0	30.2	0.2	1.03	
5	40.0	40.0	39.8	-0.2	1.03	

11. The reported uncertainty is the expended uncertainty in measurement obtained by multiplying the uncertainty by the coverage factor k=1.96, which corresponds to a confidence level 95% for normal distribution.

12. REMARK :-

a) Standard Reading are average of five readings.

***** End of Certificate *****

Calibrated By

Calibration Engineer
(Avinash Koladiya)

SCL/QF/7.8/02



Reviewed and Authorized By

Quality Manager
(Arpan Patel)

* NOTE : The Certificate is issued Subjected to Condition stated Overleaf.

Plot No.30, Kirti Nagar Society, Nr. Community Hall,

Utran, Surat, Gujarat - 394105

+91 98241 66343 / 92282 98749

shreeji.cal169@gmail.com



Shreeji Calibration Laboratory LLP

(FORMERLY KNOWN AS SHREEJI CALIBRATION LABORATORY)

" A Complete Calibration & Validation Center " " An NABL Accredited Laboratory "



CC-4263

CALIBRATION CERTIFICATE

1. NAME & ADDRESS OF CUSTOMER			UNIQUE LAB REPORT NO.		CC426326000001852F	
M/S MAYA INTERNATIONAL EXPORT Gat No. 407/1/B Tanda, Gondia, Maharashtra-441614			CERTIFICATE NO.		260331002	
			DATE OF CALIBRATION		31/03/2026	
			DATE OF ISSUE		02/04/2026	
			NEXT DUE DATE (as per agreed with customer)		30/03/2027	
			PAGE		01 of 01	
2. SRF NO. & DATE :- 260328001 & 28/03/2026						
3. CALIBRATION LOCATION :- ON SITE						
4. DESCRIPTION OF UUC						
Instrument Name	Make / Sr. No.	ID No.	Range	Least Count	Type	Location
Digital Temperature Controller	Shivam	MIE/BOD/02	10 to 40 °C	0.1°C	Pt-100	BOD Incubator (25°C)
5. CONDITION OF UUC & DATE OF RECEIPT :- GOOD & 31/03/2026						
6. CALIBRATION STANDARD USED:-						
Name	Make / ID No.	Certificate No:	Valid up to	Calibrated By		
Universal Calibrator	Masibus / SCL/UC/02	C&IJ/CAL/25-06/077	18/06/2026	CC-3959		
7. ENVIRONMENTAL CONDITION :-						
Ambient Temperature :- 25 ± 10 °C			Relative Humidity :- 50 ± 20 %RH			
8. METHOD USED FOR CALIBRATION :- SCL/SOP/ET/01 Based on ITS-90 & NABL-129						
9. DISCIPLINE :- Electro - Technical						
10. CALIBRATION RESULTS :-						
Sr. No.	Calibration Point	Standard Reading	UUC Reading	Error	Expanded Uncertainty (± °C)	
	°C	°C	°C	°C		
1	10.0	10.0	10.3	0.3	1.03	
2	15.0	15.0	15.2	0.2	1.03	
3	20.0	20.0	20.2	0.2	1.03	
4	30.0	30.0	30.1	0.1	1.03	
5	40.0	40.0	39.9	-0.1	1.03	

11. The reported uncertainty is the expended uncertainty in measurement obtained by multiplying the uncertainty by the coverage factor k=1.96, which corresponds to a confidence level 95% for normal distribution.

12. REMARK :-

a) Standard Reading are average of five readings.

***** End of Certificate *****

Calibrated By

Calibration Engineer
(Avinash Koladiya)

SCL/QF/7.8/02



Reviewed and Authorized By

Quality Manager
(Arpan Patel)

* NOTE : The Certificate is issued Subjected to Condition stated Overleaf.



Shreeji Calibration Laboratory LLP

(FORMERLY KNOWN AS SHREEJI CALIBRATION LABORATORY)

" A Complete Calibration & Validation Center " " An NABL Accredited Laboratory "



CC-4263

CALIBRATION CERTIFICATE

1. NAME & ADDRESS OF CUSTOMER				UNIQUE LAB REPORT NO.		CC426326000001853F
M/S MAYA INTERNATIONAL EXPORT Gat No. 407/1/B Tanda, Gondia, Maharashtra-441614				CERTIFICATE NO.		260331003
				DATE OF CALIBRATION		31/03/2026
				DATE OF ISSUE		02/04/2026
				NEXT DUE DATE (as per agreed with customer)		30/03/2027
				PAGE		01 of 01
2. SRF NO. & DATE :- 260328001 & 28/03/2026						
3. CALIBRATION LOCATION :- ON SITE						
4. DESCRIPTION OF UUC						
Instrument Name	Make / Sr. No.	ID No.	Range	Least Count	Type	Location
Digital Temperature Controller	Shivam	MIE/BI/01	Amb. to 110 °C	0.1 °C	Pt-100	Bacteriological Incubator
5. CONDITION OF UUC & DATE OF RECEIPT :- GOOD & 31/03/2026						
6. CALIBRATION STANDARD USED:-						
Name	Make / ID No.	Certificate No:	Valid up to	Calibrated By		
Universal Calibrator	Masibus / SCL/UC/02	C&IJ/CAL/25-06/077	18/06/2026	CC-3959		
7. ENVIRONMENTAL CONDITION :-						
Ambient Temperature :- 25 ± 10 °C			Relative Humidity :- 50 ± 20 %RH			
8. METHOD USED FOR CALIBRATION :-			SCL/SOP/ET/01 Based on ITS-90 & NABL-129			
9. DISCIPLINE :- Electro - Technical						
10. CALIBRATION RESULTS :-						
Sr. No.	Calibration Point	Standard Reading	UUC Reading	Error	Expanded Uncertainty (± °C)	
	°C	°C	°C	°C		
1	30.0	30.0	30.2	0.2	1.03	
2	50.0	50.0	50.3	0.3	1.03	
3	80.0	80.0	80.1	0.1	1.03	
4	100.0	100.0	100.2	0.2	1.03	
5	110.0	110.0	109.9	-0.1	1.03	

11. The reported uncertainty is the expended uncertainty in measurement obtained by multiplying the uncertainty by the coverage factor k=1.96, which corresponds to a confidence level 95% for normal distribution.

12. REMARK :-

a) Standard Reading are average of five readings.

***** End of Certificate *****

Calibrated By

Calibration Engineer
(Avinash Koladiya)
SCL/QF/7.8/02



Reviewed and Authorized By

Quality Manager
(Arpan Patel)

* NOTE : The Certificate is issued Subjected to Condition stated Overleaf.



Shreeji Calibration Laboratory LLP

(FORMERLY KNOWN AS SHREEJI CALIBRATION LABORATORY)

" A Complete Calibration & Validation Center " " An NABL Accredited Laboratory "



CC-4263

CALIBRATION CERTIFICATE

1. NAME & ADDRESS OF CUSTOMER			UNIQUE LAB REPORT NO.		CC426326000001854F	
M/S MAYA INTERNATIONAL EXPORT Gat No. 407/1/B Tanda, Gondia, Maharashtra-441614			CERTIFICATE NO.		260331004	
			DATE OF CALIBRATION		31/03/2026	
			DATE OF ISSUE		02/04/2026	
			NEXT DUE DATE (as per agreed with customer)		30/03/2027	
			PAGE		01 of 01	
2. SRF NO. & DATE :- 260328001 & 28/03/2026						
3. CALIBRATION LOCATION :- ON SITE						
4. DESCRIPTION OF UUC						
Instrument Name	Make / Sr. No.	ID No.	Range	Least Count	Type	Location
Digital Temperature Controller	Shivam	MIE/BI/02	Amb. to 110 °C	0.1 °C	Pt-100	Bacteriological Incubator
5. CONDITION OF UUC & DATE OF RECEIPT :- GOOD & 31/03/2026						
6. CALIBRATION STANDARD USED:-						
Name	Make / ID No.	Certificate No:	Valid up to	Calibrated By		
Universal Calibrator	Masibus / SCL/UC/02	C&IJ/CAL/25-06/077	18/06/2026	CC-3959		
7. ENVIRONMENTAL CONDITION :-						
Ambient Temperature :- 25 ± 10 °C		Relative Humidity :- 50 ± 20 %RH				
8. METHOD USED FOR CALIBRATION :- SCL/SOP/ET/01 Based on ITS-90 & NABL-129						
9. DISCIPLINE :- Electro - Technical						
10. CALIBRATION RESULTS :-						
Sr. No.	Calibration Point °C	Standard Reading °C	UUC Reading °C	Error °C	Expanded Uncertainty (± °C)	
1	30.0	30.0	30.1	0.1	1.03	
2	50.0	50.0	50.2	0.2	1.03	
3	80.0	80.0	80.3	0.3	1.03	
4	100.0	100.0	100.1	0.1	1.03	
5	110.0	110.0	109.8	-0.2	1.03	

11. The reported uncertainty is the expended uncertainty in measurement obtained by multiplying the uncertainty by the coverage factor k=1.96, which corresponds to a confidence level 95% for normal distribution.

12. REMARK :-

a) Standard Reading are average of five readings.

***** End of Certificate *****

Calibrated By

Calibration Engineer
(Avinash Koladiya)

SCL/QF/7.8/02



Reviewed and Authorized By

Quality Manager
(Arpan Patel)

* NOTE : The Certificate is issued Subjected to Condition stated Overleaf.



Shreeji Calibration Laboratory LLP

(FORMERLY KNOWN AS SHREEJI CALIBRATION LABORATORY)

" A Complete Calibration & Validation Center " " An NABL Accredited Laboratory "



CC-4263

CALIBRATION CERTIFICATE

1. NAME & ADDRESS OF CUSTOMER				UNIQUE LAB REPORT NO.		CC426326000001855F	
M/S MAYA INTERNATIONAL EXPORT Gat No. 407/1/B Tanda, Gondia, Maharashtra-441614				CERTIFICATE NO.		260331005	
				DATE OF CALIBRATION		31/03/2026	
				DATE OF ISSUE		02/04/2026	
				NEXT DUE DATE (as per agreed with customer)		30/03/2027	
				PAGE		01 of 01	
2. SRF NO. & DATE :- 260328001 & 28/03/2026							
3. CALIBRATION LOCATION :- ON SITE							
4. DESCRIPTION OF UUC							
Instrument Name	Make / Sr. No.	ID No.	Range	Least Count	Type	Location	
Digital Temperature Controller	Shivam	MIE/HAO/01	Amb. To 300°C	1°C	J-Type	Hot Air Oven	
5. CONDITION OF UUC & DATE OF RECEIPT :- GOOD & 31/03/2026							
6. CALIBRATION STANDARD USED:-							
Name	Make / ID No.	Certificate No:	Valid up to	Calibrated By			
Universal Calibrator	Masibus / SCL/UC/02	C&IJ/CAL/25-06/077	18/06/2026	CC-3959			
7. ENVIRONMENTAL CONDITION :-							
Ambient Temperature :- 25 ± 10 °C				Relative Humidity :- 50 ± 20 %RH			
8. METHOD USED FOR CALIBRATION :-				SCL/SOP/ET/01 Based on ITS-90 & NABL-129			
9. DISCIPLINE :- Electro - Technical							
10. CALIBRATION RESULTS :-							
Sr. No.	Calibration Point	Standard Reading	UUC Reading	Error	Expanded Uncertainty (± °C)		
	°C	°C	°C	°C			
1	50	50.0	51	1.0	1.03		
2	100	100.0	103	3.0	1.03		
3	200	200.0	202	2.0	1.03		
4	250	250.0	252	2.0	1.03		
5	300	300.0	298	-2.0	1.03		

11. The reported uncertainty is the expended uncertainty in measurement obtained by multiplying the uncertainty by the coverage factor k=1.96, which corresponds to a confidence level 95% for normal distribution.

12. REMARK :-

a) Standard Reading are average of five readings.

***** End of Certificate *****

Calibrated By

Calibration Engineer
(Avinash Koladiya)

SCL/QF/7.8/02



Reviewed and Authorized By

Quality Manager
(Arpan Patel)

* NOTE : The Certificate is issued Subjected to Condition stated Overleaf.



Shreeji Calibration Laboratory LLP

(FORMERLY KNOWN AS SHREEJI CALIBRATION LABORATORY)

" A Complete Calibration & Validation Center " " An NABL Accredited Laboratory "



CC-4263

CALIBRATION CERTIFICATE

1. NAME & ADDRESS OF CUSTOMER				UNIQUE LAB REPORT NO.		CC426326000001856F
M/S MAYA INTERNATIONAL EXPORT Gat No. 407/1/B Tanda, Gondia, Maharashtra-441614				CERTIFICATE NO.		260331006
				DATE OF CALIBRATION		31/03/2026
				DATE OF ISSUE		02/04/2026
				NEXT DUE DATE (as per agreed with customer)		30/03/2027
				PAGE		01 of 02
2. SRF NO. & DATE :- 260328001 & 28/03/2026						
3. CALIBRATION LOCATION :- ON SITE Stabilization Time:- 1 Hour						
4. DESCRIPTION OF UUC						
Instrument Name	Make / Sr. No.	ID No.	Range	Least Count	Type	Location
Single Pan Balance	U-Mex	MIE/SPB/01	0 to 200 g	0.1 mg	Analog	--
5. CONDITION OF UUC & DATE OF RECEIPT :- GOOD & 31/03/2026						
6. CALIBRATION STANDARD USED:-						
Name	Make / ID No.	Certificate No:	Valid up to	Calibrated By		
E2 Class Weight Box	Weightronics / SCL/WB/01	PL-611	15/11/2026	CC-3213		
7. ENVIRONMENTAL CONDITION :-						
Ambient Temperature :- 25 ± 10 °C			Relative Humidity :- 50 ± 20 %RH			
8. METHOD USED FOR CALIBRATION :- SCL/SOP/M/02 Based on OIML R76 and NABL-129						
9. DISCIPLINE :- Mechanical - (Weighing Scale & Balance)						
10. CALIBRATION RESULTS :-						
Sr. No.	Weight placed on PAN	Reference standards Value	Measured Mass Value	Error	Expanded Uncertainty (± mg)	
	g	g	g	g		
1	1	1.000015	1.0001	0.0001	0.85	
2	2	2.000022	2.0001	0.0001		
3	5	5.000032	5.0001	0.0001		
4	10	10.000043	10.0002	0.0002		
5	20	20.000046	20.0001	0.0001		
6	50	50.000054	50.0001	0.0000		
7	100	100.000942	100.0007	-0.0002		
8	120	120.000988	120.0008	-0.0002		
9	150	150.001077	150.0007	-0.0004		
10	200	200.000130	199.9999	-0.0002		

Calibrated By

Calibration Engineer
(Avinash Koladiya)

SCL/QF/7.8/02



Reviewed and Authorized By

Quality Manager
(Arpan Patel)

* NOTE : The Certificate is issued Subjected to Condition stated Overleaf.



Shreeji Calibration Laboratory LLP

(FORMERLY KNOWN AS SHREEJI CALIBRATION LABORATORY)

" A Complete Calibration & Validation Center " " An NABL Accredited Laboratory "



CC-4263

UNIQUE LAB REPORT NO.	CC426326000001856F
CERTIFICATE NO.	260331006
DATE OF CALIBRATION	31/03/2026
DATE OF ISSUE	02/04/2026
NEXT DUE DATE (as per agreed with customer)	30/03/2027
PAGE	02 of 02

Eccentricity Test

Sr. No.	Location on Pan	Observed Reading (g)	
1	1	50.0013	Minimum Value (g)
2	2	50.0019	50.0013
3	3	50.0018	
4	4	50.0014	Maximum Value (g)
5	5	50.0015	50.0019
6	5	50.0018	
7	4	50.0018	Difference (g)
8	3	50.0013	0.0006
9	2	50.0014	
10	1	50.0016	




Fig. Load Pattern For Eccentricity Test

Repeatability

Sr. No	At Full Load (200 g)		At Half Load (100 g)	
	Observed Reading (g)		Observed Reading (g)	
1	199.9998	Minimum Value (g)	100.0014	Minimum Value (g)
2	199.9992	199.9991	100.0015	100.0014
3	199.9991		100.0018	
4	199.9997	Maximum Value (g)	100.0021	Maximum Value (g)
5	199.9997	199.9998	100.0023	100.0023
6	199.9996		100.0016	
7	199.9994	Difference (g)	100.0017	Difference (g)
8	199.9993	0.0007	100.0019	0.0009
9	199.9991		100.0018	
10	199.9994		100.0021	

11. The reported uncertainty is the expended uncertainty in measurement obtained by multiplying the uncertainty by the coverage factor $k=1.96$, which corresponds to a confidence level 95% for normal distribution.

12. REMARK :-

a) Standard Reading are average of five readings.

***** End of Certificate *****

Calibrated By

Calibration Engineer
(Avinash Koladiya)

SCL/QF/7.8/02



Reviewed and Authorized By

Quality Manager
(Arpan Patel)

* NOTE : The Certificate is issued Subjected to Condition stated Overleaf.

Plot No.30, Kirti Nagar Society, Nr. Community Hall,
Utran, Surat, Gujarat - 394105
+91 98241 66343 / 92282 98749
shreeji.cal169@gmail.com



Shreeji Calibration Laboratory LLP

(FORMERLY KNOWN AS SHREEJI CALIBRATION LABORATORY)

" A Complete Calibration & Validation Center " " An NABL Accredited Laboratory "



CC-4263

CALIBRATION CERTIFICATE

1. NAME & ADDRESS OF CUSTOMER				UNIQUE LAB REPORT NO.		CC426326000001857F	
M/S MAYA INTERNATIONAL EXPORT Gat No. 407/1/B Tanda, Gondia, Maharashtra-441614				CERTIFICATE NO.		260331007	
				DATE OF CALIBRATION		31/03/2026	
				DATE OF ISSUE		02/04/2026	
				NEXT DUE DATE (as per agreed with customer)		30/03/2027	
				PAGE		01 of 01	
2. SRF NO. & DATE :- 260328001 & 28/03/2026							
3. CALIBRATION LOCATION :- ON SITE							
4. DESCRIPTION OF UUC							
Instrument Name	Make / Sr. No.	ID No.	Range	Least Count	Type	Location	
Pressure Gauge	-	MIE/PG/01	0 to 2.1 kg/cm ²	0.05 kg/cm ²	Bourdon Tube	Autoclave	
5. CONDITION OF UUC & DATE OF RECEIPT :- GOOD & 31/03/2026							
6. CALIBRATION STANDARD USED:-							
Name	Make / ID No.	Certificate No:	Valid up to		Calibrated By		
Digital Pressure Gauge	Vijay Enterprises / SCL/DPG/02	VILLP/25-26/M-0639	10/06/2026		CC-2695		
7. ENVIRONMENTAL CONDITION :-							
Ambient Temperature :- 25 ± 10 °C				Relative Humidity :- 50 ± 20 %RH			
8. METHOD USED FOR CALIBRATION :- SCL/SOP/M/01 Based on DKD-R61 and NABL-129							
9. DISCIPLINE :- Mechanical - Pressure							
10. CALIBRATION RESULTS :-							
Sr. No.	Set Pressure on UUC	Observed Average Reading on STD kg/cm ²		Average Error kg/cm ²		Hysteresis of Average	Expanded Uncertainty (± kg/cm ²)
	kg/cm ²	Up	Down	Up	Down	kg/cm ²	
1	0.00	0.00	0.00	0.00	0.00	0.00	0.08
2	0.50	0.54	0.53	-0.04	-0.03	0.01	0.08
3	1.00	1.03	1.02	-0.03	-0.02	0.01	0.08
4	1.50	1.51	1.50	-0.01	0.00	0.01	0.08
5	2.00	2.02	2.01	-0.02	-0.01	0.01	0.08
6	2.10	2.11	2.11	-0.01	-0.01	0.00	0.08

11. The reported uncertainty is the expended uncertainty in measurement obtained by multiplying the uncertainty by the coverage factor k=1.96, which corresponds to a confidence level 95% for normal distribution.

12. REMARK :-

- a) Standard Reading are average of five readings.
b) Pressure Conversion:- 1 kg/cm² = 0.9807 Bar

***** End of Certificate *****

Calibrated By

Calibration Engineer
(Avinash Koladiya)

SCL/QF/7.8/02



Reviewed and Authorized By

Quality Manager
(Arpan Patel)

* NOTE : The Certificate is issued Subjected to Condition stated Overleaf.



Shreeji Calibration Laboratory LLP

(FORMERLY KNOWN AS SHREEJI CALIBRATION LABORATORY)

" A Complete Calibration & Validation Center " " An NABL Accredited Laboratory "



CC-4263

CALIBRATION CERTIFICATE

1. NAME & ADDRESS OF CUSTOMER				UNIQUE LAB REPORT NO.		CC426326000001858F
M/S MAYA INTERNATIONAL EXPORT Gat No. 407/1/B Tanda, Gondia, Maharashtra-441614				CERTIFICATE NO.		260331008
				DATE OF CALIBRATION		31/03/2026
				DATE OF ISSUE		02/04/2026
				NEXT DUE DATE (as per agreed with customer)		30/03/2027
				PAGE		01 of 01
2. SRF NO. & DATE :- 260328001 & 28/03/2026						
3. CALIBRATION LOCATION :- IN LAB						
4. DESCRIPTION OF UUC						
Instrument Name	Make / Sr. No.	ID No.	Range	Least Count	Type	Location
Glass Thermometer	-	MIE/GT/01	-10 to 110°C	0.5°C	Liquid In Glass	-
5. CONDITION OF UUC & DATE OF RECEIPT :- GOOD & 31/03/2026						
6. CALIBRATION STANDARD USED:-						
Name	Make / ID No.	Certificate No:	Valid up to	Calibrated By		
Universal Calibrator	Masibus / SCL/UC/02	C&IJ/CAL/25-06/077	18/06/2026	CC-3959		
7. ENVIRONMENTAL CONDITION :-						
Ambient Temperature :- 25 ± 10 °C			Relative Humidity :- 50 ± 20 %RH			
8. METHOD USED FOR CALIBRATION :-			SCL/SOP/ET/01 Based on ITS-90 & NABL-129			
9. DISCIPLINE :- Electro - Technical						
10. CALIBRATION RESULTS :-						
Sr. No.	Calibration Point	UUC Reading	Standard Reading	Error	Expanded Uncertainty (± °C)	
	°C	°C	°C	°C		
1	40.0	41.0	41.74	0.74	1.30	
2	80.0	81.0	81.68	0.68	1.30	
3	100.0	101.0	101.75	0.75	1.30	

11. The reported uncertainty is the expended uncertainty in measurement obtained by multiplying the uncertainty by the coverage factor k=1.96, which corresponds to a confidence level 95% for normal distribution.

12. REMARK :-

- Standard Reading are average of five readings.
- Temperature Scale ITS-90

***** End of Certificate *****

Calibrated By

Calibration Engineer
(Avinash Koladiya)

SCL/QF/7.8/02



Reviewed and Authorized By

Quality Manager
(Arpan Patel)

* NOTE : The Certificate is issued Subjected to Condition stated Overleaf.



Shreeji Calibration Laboratory LLP

(FORMERLY KNOWN AS SHREEJI CALIBRATION LABORATORY)

" A Complete Calibration & Validation Center " " An NABL Accredited Laboratory "



CC-4263

CALIBRATION CERTIFICATE

1. NAME & ADDRESS OF CUSTOMER				UNIQUE LAB REPORT NO.		CC426326000001859F
M/S MAYA INTERNATIONAL EXPORT Gat No. 407/1/B Tanda, Gondia, Maharashtra-441614				CERTIFICATE NO.		260331009
				DATE OF CALIBRATION		31/03/2026
				DATE OF ISSUE		02/04/2026
				NEXT DUE DATE (as per agreed with customer)		30/03/2027
				PAGE		01 of 01
2. SRF NO. & DATE :- 260328001 & 28/03/2026						
3. CALIBRATION LOCATION :- IN LAB						
4. DESCRIPTION OF UUC						
Instrument Name	Make / Sr. No.	ID No.	Range	Least Count	Type	Location
Glass Thermometer	-	MIE/GT/02	0 to 250°C	1°C	Liquid In Glass	--
5. CONDITION OF UUC & DATE OF RECEIPT :- GOOD & 31/03/2026						
6. CALIBRATION STANDARD USED:-						
Name	Make / ID No.	Certificate No:	Valid up to	Calibrated By		
Universal Calibrator	Masibus / SCL/UC/02	C&IJ/CAL/25-06/077	18/06/2026	CC-3959		
7. ENVIRONMENTAL CONDITION :-						
Ambient Temperature :- 25 ± 10 °C			Relative Humidity :- 50 ± 20 %RH			
8. METHOD USED FOR CALIBRATION :- SCL/SOP/ET/01 Based on ITS-90 & NABL-129						
9. DISCIPLINE :- Electro - Technical						
10. CALIBRATION RESULTS :-						
Sr. No.	Calibration Point	UUC Reading	Standard Reading	Error	Expanded Uncertainty (± °C)	
	°C					
1	40	41	41.61	0.61	1.30	
2	100	101	101.46	0.46	1.30	
3	200	199	199.79	0.79	1.30	

11. The reported uncertainty is the expended uncertainty in measurement obtained by multiplying the uncertainty by the coverage factor k=1.96, which corresponds to a confidence level 95% for normal distribution.

12. REMARK :-

- Standard Reading are average of five readings.
- Temperature Scale ITS-90

***** End of Certificate *****


Calibrated By


Calibration Engineer
(Avinash Koladiya)

SCL/QF/7.8/02



Reviewed and Authorized By


Quality Manager
(Arpan Patel)

* NOTE : The Certificate is issued Subjected to Condition stated Overleaf.



Shreeji Calibration Laboratory LLP

(FORMERLY KNOWN AS SHREEJI CALIBRATION LABORATORY)

" A Complete Calibration & Validation Center " " An NABL Accredited Laboratory "



CC-4263

CALIBRATION CERTIFICATE

1. NAME & ADDRESS OF CUSTOMER		UNIQUE LAB REPORT NO.		CC426326000001860F		
M/S MAYA INTERNATIONAL EXPORT Gat No. 407/1/B Tanda, Gondia, Maharashtra-441614		CERTIFICATE NO.		260331010		
		DATE OF CALIBRATION		31/03/2026		
		DATE OF ISSUE		02/04/2026		
		NEXT DUE DATE (as per agreed with customer)		30/03/2027		
		PAGE		01 of 01		
2. SRF NO. & DATE :- 260328001 & 28/03/2026						
3. CALIBRATION LOCATION :- IN LAB						
4. DESCRIPTION OF UUC						
Instrument Name	Make / Sr. No.	ID No.	Range	Least Count	Type	Location
Glass Thermometer	-	MIE/GT/03	0 to 360°C	1°C	Liquid In Glass	--
5. CONDITION OF UUC & DATE OF RECEIPT :- GOOD & 31/03/2026						
6. CALIBRATION STANDARD USED:-						
Name	Make / ID No.	Certificate No:	Valid up to	Calibrated By		
Universal Calibrator	Masibus / SCL/UC/02	C&IJ/CAL/25-06/077	18/06/2026	CC-3959		
7. ENVIRONMENTAL CONDITION :-						
Ambient Temperature :- 25 ± 10 °C			Relative Humidity :- 50 ± 20 %RH			
8. METHOD USED FOR CALIBRATION :- SCL/SOP/ET/01 Based on ITS-90 & NABL-129						
9. DISCIPLINE :- Electro - Technical						
10. CALIBRATION RESULTS :-						
Sr. No.	Calibration Point	UUC Reading	Standard Reading	Error	Expanded Uncertainty (± °C)	
	°C	°C	°C	°C		
1	40	41	41.38	0.38	1.30	
2	100	101	101.62	0.62	1.30	
3	200	199	199.49	0.49	1.30	

11. The reported uncertainty is the expended uncertainty in measurement obtained by multiplying the uncertainty by the coverage factor k=1.96, which corresponds to a confidence level 95% for normal distribution.

12. REMARK :-

- Standard Reading are average of five readings.
- Temperature Scale ITS-90

***** End of Certificate *****

Calibrated By

Calibration Engineer
(Avinash Koladiya)

SCL/QF/7.8/02



Reviewed and Authorized By

Quality Manager
(Arpan Patel)

* NOTE : The Certificate is issued Subjected to Condition stated Overleaf.



Shreeji Calibration Laboratory LLP

(FORMERLY KNOWN AS SHREEJI CALIBRATION LABORATORY)

" A Complete Calibration & Validation Center " " An NABL Accredited Laboratory "

CALIBRATION CERTIFICATE

1. NAME & ADDRESS OF CUSTOMER				CERTIFICATE NO.		260331011
M/S MAYA INTERNATIONAL EXPORT Gat No. 407/1/B Tanda, Gondia, Maharashtra-441614				DATE OF CALIBRATION		31/03/2026
				DATE OF ISSUE		02/04/2026
				NEXT DUE DATE (as per agreed with customer)		30/03/2027
				PAGE		01 of 01
2. SRF NO. & DATE :- 260328001 & 28/03/2026						
3. CALIBRATION LOCATION :- ON SITE						
4. DESCRIPTION OF UUC						
Instrument Name	Make / Sr. No.	ID No.	Range	Least Count	Type	Location
Digital pH Meter	-	MIE/PH/01	0 to 14 pH	0.01 pH	Digital	--
5. CONDITION OF UUC & DATE OF RECEIPT :- GOOD & 31/03/2026						
6. CALIBRATION STANDARD USED:-						
Name	Make / ID No.	Certificate No:	Valid up to	Calibrated By		
Universal Calibrator	Masibus / SCL/UC/02	C&IJ/CAL/25-06/077	18/06/2026	CC-3959		
7. ENVIRONMENTAL CONDITION :-						
Ambient Temperature :- 25 ± 10 °C			Relative Humidity :- 50 ± 20 %RH			
8. METHOD USED FOR CALIBRATION :- SCL/SOP/11 Based on ITS-90 & NABL-129						
9. DISCIPLINE :- Electro - Technical						
10. CALIBRATION RESULTS :-						
Sr. No.	Calibration Point	Standard Reading		UUC Reading	Error	Expanded Uncertainty (± pH)
	pH	pH	mV	pH	pH	
1	0	0	400.00	0.01	0.0	0.18
2	4	4	171.43	4.03	0.03	0.18
3	7	7	0.00	7.04	0.04	0.18
4	10	10	-171.43	10.05	0.05	0.18
5	14	14	-400.00	14.06	0.06	0.18

11. The reported uncertainty is the expended uncertainty in measurement obtained by multiplying the uncertainty by the coverage factor k=1.96, which corresponds to a confidence level 95% for normal distribution.

12. REMARK :-

a) Standard Reading are average of five readings.

***** End of Certificate *****

Calibrated By

Calibration Engineer
(Avinash Koladiya)



Reviewed and Authorized By

Quality Manager
(Arpan Patel)

* NOTE : The Certificate is issued Subjected to Condition stated Overleaf.



Shreeji Calibration Laboratory LLP

(FORMERLY KNOWN AS SHREEJI CALIBRATION LABORATORY)

" A Complete Calibration & Validation Center " " An NABL Accredited Laboratory "

CALIBRATION CERTIFICATE

1. NAME & ADDRESS OF CUSTOMER				CERTIFICATE NO.		260331012
M/S MAYA INTERNATIONAL EXPORT Gat No. 407/1/B Tanda, Gondia, Maharashtra-441614				DATE OF CALIBRATION		31/03/2026
				DATE OF ISSUE		02/04/2026
				NEXT DUE DATE (as per agreed with customer)		30/03/2027
				PAGE		01 of 01
2. SRF NO. & DATE :- 260328001 & 28/03/2026						
3. CALIBRATION LOCATION :- ON SITE						
4. DESCRIPTION OF UUC						
Instrument Name	Make / Sr. No.	ID No.	Range	Least Count	Type	Location
Digital Turbidity Meter	-	MIE/DTM/01	0 to 1000 NTU	0.1 NTU	Digital	--
5. CONDITION OF UUC & DATE OF RECEIPT :- GOOD & 31/03/2026						
6. CALIBRATION STANDARD USED:-						
Name	Make / ID No.	Certificate No:	Valid up to	Calibrated By		
Universal Calibrator	Masibus / SCL/UC/02	C&IJ/CAL/25-06/077	18/06/2026	CC-3959		
7. ENVIRONMENTAL CONDITION :-						
Ambient Temperature :- 25 ± 10 °C		Relative Humidity :- 50 ± 20 %RH				
8. METHOD USED FOR CALIBRATION :- SCL/SOP/14 Based on ITS-90 & NABL-129						
9. DISCIPLINE :- Electro - Technical						
10. CALIBRATION RESULTS :-						
Sr. No.	Calibration Point	Standard Reading		UUC Reading	Error	Expanded Uncertainty (± %)
	NTU	NTU	mV	NTU	NTU	
1	10.0	10	10.00	10.1	0.1	1.12
2	100.0	100	100.00	100.1	0.1	1.12
3	200.0	200	200.00	200.2	0.2	1.12
4	500.0	500	500.00	500.3	0.3	1.12
5	1000.0	1000	1000.00	1000.2	0.2	1.12

11. The reported uncertainty is the expended uncertainty in measurement obtained by multiplying the uncertainty by the coverage factor k=1.96, which corresponds to a confidence level 95% for normal distribution.

12. REMARK :-

a) Standard Reading are average of five readings.

***** End of Certificate *****

Calibrated By

Calibration Engineer
(Avinash Koladiya)



Reviewed and Authorized By

Quality Manager
(Arpan Patel)

* NOTE : The Certificate is issued Subjected to Condition stated Overleaf.



Shreeji Calibration Laboratory LLP

(FORMERLY KNOWN AS SHREEJI CALIBRATION LABORATORY)

" A Complete Calibration & Validation Center " " An NABL Accredited Laboratory "

CALIBRATION CERTIFICATE

1. NAME & ADDRESS OF CUSTOMER				CERTIFICATE NO.		260331013
M/S MAYA INTERNATIONAL EXPORT Gat No. 407/1/B Tanda, Gondia, Maharashtra-441614				DATE OF CALIBRATION		31/03/2026
				DATE OF ISSUE		02/04/2026
				NEXT DUE DATE (as per agreed with customer)		30/03/2027
				PAGE		01 of 01
2. SRF NO. & DATE :- 260328001 & 28/03/2026						
3. CALIBRATION LOCATION :- ON SITE						
4. DESCRIPTION OF UUC						
Instrument Name	Make / Sr. No.	ID No.	Range	Least Count	Type	Location
Digital Spectrophotometer	-	MIE/SPM/01	330 to 970 nm	1 nm	Digital	--
5. CONDITION OF UUC & DATE OF RECEIPT :- GOOD & 31/03/2026						
6. CALIBRATION STANDARD USED:-						
Name	Make / ID No.	Certificate No:	Valid up to	Calibrated By		
Universal Calibrator	Masibus / SCL/UC/02	C&IJ/CAL/25-06/077	18/06/2026	CC-3959		
7. ENVIRONMENTAL CONDITION :-						
Ambient Temperature :- 25 ± 10 °C		Relative Humidity :- 50 ± 20 %RH				
8. METHOD USED FOR CALIBRATION :- SCL/SOP/12 Based on ITS-90 & NABL-129						
9. DISCIPLINE :- Electro - Technical						
10. CALIBRATION RESULTS :-						
Sr. No.	Calibration Point	Standard Reading		UUC Reading	Error	Expanded Uncertainty (± %)
	NM	NM	mV	NM	NM	
1	330	330	1.00	332	2	1.29
2	532	532	300.00	534	2	1.29
3	724	724	600.00	727	3	1.29
4	916	916	900.00	918	2	1.29
5	970	970	1000.00	968	-2	1.29

11. The reported uncertainty is the expended uncertainty in measurement obtained by multipling the uncertainty by the coverage factor k=1.96, which corresponds to a confidence level 95% for normal distribution.

12. REMARK :-

a) Standard Reading are average of five readings.


***** End of Certificate *****

Calibrated By


Calibration Engineer
(Avinash Koladiya)



Reviewed and Authorized By


Quality Manager
(Arpan Patel)

* NOTE : The Certificate is issued Subjected to Condition stated Overleaf.