

Experiment : 20240726 Inna - GFP		Sort Report		Report Date : 2024.07.26 at 12:11:05			
Specimen : 20240726 Sort				Device : 96 Well - Falcon			
Tube : 0 ug/ml Dox sample				User ID : Administrator			
Sort Layout : 96w - 0 Dox				Cytometer : FACS AriaIII (1)			
Application : FACSDiva Version 9.0.1							
Sort Settings							
Sort Setup	100 micron	Precision	Single Cell				
Frequency	30.0	Yield Mask	0				
Amplitude	5.9	Purity Mask	32				
Phase	0.00	Phase Mask	16				
Drop Delay	25.11	Single Cell	On				
Attenuation	Off	Plates Voltage	2,750				
Sweet Spot	On	Voltage Centering	42				
First Drop	205	Sheath Pressure	20.00				
Target Gap	11						
Side Stream Voltage (%)							
Far Left	Left	Right	Far Right				
18.00	0.00	0.00	0.00				
Neighboring Drop Charge (%)							
2nd	3rd	4th					
16.00	7.00	1.00					
Acquisition Counters							
Threshold Count	2735						
Processed Events Count(evt)	4731						
Electronic Aborts Count(evt)	9						
Sort Elapsed Time(hh:mm:ss)	00:00:53						
Sort Counters							
	1						
Sort Rate(evt/s)	0						
Conflicts Count(evt)	NA						
Conflicts Rate(evt/s)	0						
Efficiency(%)	NA						

Sort Layout						
	1	2	3	4	5	6
A			Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	
B			Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	
C			Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	
D			Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	
E			Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	
F			Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	
G			Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	
H			Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	
Sort Layout						
	7	8	9	10	11	12
A						
B						
C						
D						
E						
F						
G						
H						