

Experiment : 20240726 Inna - GFP		Sort Report		Report Date : 2024.07.26 at 10:57:16	
Specimen : 20240726 Sort				Device : 384 Well - Falcon	
Tube : 0 ug/ml Dox sample				User ID : Administrator	
Sort Layout : 384w - 0 Dox				Cytometer : FACS AriaIII (1)	
Application : FACSDiva Version 9.0.1					
Sort Settings					
Sort Setup	100 micron		Precision		Single Cell
Frequency	30.0		Yield Mask		0
Amplitude	5.9		Purity Mask		32
Phase	0.00		Phase Mask		16
Drop Delay	25.11		Single Cell		On
Attenuation	Off		Plates Voltage		2,750
Sweet Spot	On		Voltage Centering		42
First Drop	205		Sheath Pressure		20.00
Target Gap	11				
Side Stream Voltage (%)					
Far Left	Left		Right		Far Right
18.00	0.00		0.00		0.00
Neighboring Drop Charge (%)					
2nd			3rd		4th
16.00			7.00		1.00
Acquisition Counters					
Threshold Count					15394
Processed Events Count(evt)					15181
Electronic Aborts Count(evt)					22
Sort Elapsed Time(hh:mm:ss)					00:02:38
Sort Counters					
					1
Sort Rate(evt/s)					4
Conflicts Count(evt)					NA
Conflicts Rate(evt/s)					0
Efficiency(%)					NA

Sort Layout

	1	2	3	4	5	6	7	8	9	10	11	12
A	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1							
B	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1							
C	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1							
D	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1							
E	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1							
F	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1							
G	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1							
H	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1							
I	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1							
J	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1							
K	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1							
L	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1							
M	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1							
N	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1							
O	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1							
P	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1	Viable : 1 / 1							

Sort Layout

	13	14	15	16	17	18	19	20	21	22	23	24
A												
B												
C												
D												
E												
F												
G												
H												
I												
J												
K												
L												
M												
N												
O												
P												