Calibration Certificate

General Information

Certificate No.	CAL-2026-00418
Date of Issue	2026-10-05 (1405/07/13 SH)
Date of Calibration	2026-10-04 (1405/07/12 SH)
Recommended Next Calibration	2027-10-04 (1406/07/12 SH)
Customer	Ironwood Machinery Works
Address	Sample Industrial Park, 3rd Street, No. 12

Item Calibrated

Description	Industrial pressure transmitter
Manufacturer	Helix
Model	HP-200
Serial No.	HX-2203-131
Asset Tag	AT-0057
Range	0 - 200 bar
Resolution	0.1 bar

Environmental Conditions

Temperature	22.4 °C
Relative Humidity	46 %
Ambient Pressure	101.3 kPa

Reference Standards

Standard	ID	Certificate	Valid until
Digital pressure calibrator	REF-DPC-07	QC-88213	2027-02-01
Reference thermometer	REF-THM-02	QC-77410	2027-05-15

Measurement Results

All values in bar.

No.	Nominal	Measured	Error	Tolerance ±	Uncertainty ±U	Result
1	0.00	0.10	+0.10	0.50	0.15	**PASS**
2	50.00	49.90	-0.10	0.50	0.15	**PASS**
3	100.00	100.20	+0.20	0.50	0.15	**PASS**
4	150.00	150.80	+0.80	0.50	0.18	**FAIL**

No.	Nominal	Measured	Error	Tolerance ±	Uncertainty ±U	Result
5	200.00	200.40	+0.40	0.50	0.18	**PASS**

Decision rule: simple acceptance. A point passes when |error| ≤ tolerance; the uncertainty is reported but not added to the error. Expanded uncertainty U is stated for a coverage factor k = 2 (about 95 % confidence).

Overall result

FAIL

4 of 5 points within tolerance

Remarks

The 150 bar point was out of tolerance. The instrument must be adjusted and recalibrated.

Calibrated by

Sample Technician A

Signature

Approved by

Sample Technical Manager B

Signature

Fictional sample document generated for demonstration. All names and values are invented.