CONFIDENTIAL
UNCONTROLLED COPY

MASTER COPY
Sign. :
Date : 27/02/25

embio

STANDARD OPERATING PROCEDURE		Page 2 of 7
RAW MATERIAL SPECIFICATIONS AND TESTING METHODS		
TITLE: **SPECIFICATIONS AND TEST METHODS FOR N,N-DICYCLOHEXYL CARBODIIMIDE (DCC)**		Ref No.: **Q/SOP/TMD-387** Revision number: **2**
Prepared by: **Sr. Officer QC** Date: 21/02/25	Checked by: **Sr. Manager QC** Date: 21/02/25	Authorized by: **Head Quality** Date: 21/02/25
Date effective: 01/03/25	Review date: March. 2028	

4.0 A Specifications:

Sr. No	Tests	Method Reference	Limits
1	Description	In-house	Colourless / White to pale yellow solid or liquid
2	Identification by GC	In-house	The retention time of DCC peak in the chromatogram of sample preparation by GC should match with the retention time of working standard as obtained under purity by GC
3	Purity by GC (Area %)	In-house	Not less than 98.0%

4.0 B PRECAUTIONS

1. All glassware used should be cleaned and dried.
2. Pipettes, volumetric flasks should be of "Class A"
3. Analytical grade (AR) (Extra Pure) reagents to be used, unless otherwise specified.
4. Analytical balances and equipment's used should be calibrated.
5. Refer MSDS prior to handling API's, chemicals and reagents.