

Ruiz, Patrick

**From:** Mostaan, Niloofar  
**Sent:** Friday, May 17, 2024 8:58 AM  
**To:** DMP1 White Boxing; DMP1 Packaging; 5mm Lap Top Level; RSM-DMP1 Supervisors  
**Cc:** QC-Clean Rooms  
**Subject:** EB213 s/o1520624 Assembly Inspection completed

Hi Team,

Assembly Inspection on EB213 shop order 1520624 has been completed.

Failed devices been scrapped in RSL per Q-4018864 Rev. R

Completed Device Assembly Inspection Report	
Model:	EB213
Shop Order:	1520624

Devices Inspected by QC	569
Total Passed by QC	550
Total Failed by QC	19
Total Failed by PRD	0

Non-Conformances	Quantity	Scrap %
Visual Inspection	1	0.18%
Verify Insulation	4	0.70%
Fuse Switch	2	0.35%
Continuity Range	12	2.11%

"Visual Inspection" Details	Quantity
Scratched Logo	
Missing Silicone	
Trigger Insert Defect	
Incorect assembly	1

"Verify Insulation" Details	Quantity
Incorrect Insulation Orientation	
Insulation Damage	4

Statistics for Shop Order #1520624

Device SN = 569	Program Count = 571	Program Success = 569
CER Evaluation = 0	Lost Serial Number = 0	Verification Success = 550

Thank you,  
Niloo