Radiotherapy Dose Optimization with Clinical Knowledge Based Reinforcement Learning

 $\begin{array}{c} {\rm Paul~Dubois^{1,2,3[0009-0003-3856-8048]},} \\ {\rm Paul\text{-}Henry~Courn\`ede^{1[0000-0001-7679-6197]},~and} \\ {\rm Nikos~Paragios^{2[0000-0002-9668-4763]}} \end{array}$

¹ Biomathematics, MICS, CentraleSupélec, Université Paris-Saclay {p.dubois,paul-henry.cournede}@centralesupelec.fr https://biomathematics.mics.centralesupelec.fr/

² TheraPanacea, Paris, France {p.dubois,n.paragios}@therapanacea.eu https://www.therapanacea.eu/

³ Institut du Cancer de Montpellier (Val d'Aurelle), Montpellier, France paul.dubois@icm.unicancer.fr https://www.therapanacea.eu/

Abstract. Achieving optimal dose distribution in radiation therapy planning is a complex task, with contradicting goals. Yet, this step is crucial with profound implications for patient treatment.

The absence of universally agreed-upon constraints prioritization in radiation therapy planning complicates the definition of an optimal plan, requiring a delicate balance between multiple objectives. This balance usually ends up being done manually.

The optimization process is further hindered by complex mathematical aspects, involving non-convex multi-objective inverse problems with a vast solution space. Expert bias introduces variability in clinical practice, as treatment planning is shaped by the preferences and expertise of radiation oncologists and medical physicists.

To surmount these challenges, we propose a first step towards a fully automated approach, using an innovative deep-learning method that allows automatic navigation toward acceptable solution. We successfully trained an agent evaluating actions of a human dosimetrist, in order to reach a plan similar to past cases. As this is very new and ongoing research, we generated synthetic phantom patients and associated trust-able clinical doses. In future work, we hope to be able to apply this technique to real cases

Keywords: Radiotherapy \cdot Dose Optimization \cdot Reinforcement Learning \cdot Deep Learning.

1 Introduction

In contemporary radiation therapy, photon intensity modulated radiation therapy (IMRT) is a pivotal technique to attain precise and conformal dose distributions within target volumes [17]. This achievement owes its realization to the advent of the multileaf collimator (MLC) [5].

Radiation therapy is now a reliable treatment for oncology [13]. Despite this consensus, the way to deliver radiotherapy for its best result remains very dependent upon doctors. Moreover, there appears to be a large variability across physicians and centres, but in terms of 3D structures contouring and irradiation constraints priorities [3].

To achieve the best treatment, doctors must solve a complex inverse mathematical optimization problem with multiple trade-offs [10] [14]. A lack of standardized prioritization of constraints makes the optimization a real challenge. The standard procedure nowadays is to guide computer optimization manually: dosimetrists manually update the settings of an optimizing software (so-called Treatment Planning System) [1].

There have been many tries to create a metric that quantifies the quality of a treatment plan: Normal tissue complication probabilities (NTCP), Target coverage, Conformity index, Heterogeneity index (non-exhaustive list) [8] [7]. However, they have yet to satisfy all radio-oncologists, and the only reliable way to assess a doctor's plan is to check out the dose-volume histograms (DVHs) themselves.

As a result, Pareto surface exploration is doomed to failure due to the lack of impartial quantitative measurement for a particular plan [6]. Other meta-optimization techniques are similarly bounded for the same reason [15] [16]. An extra challenge to attend for those is the fact that not all cases have the same "difficulty." Hence, for an "easy" case, doctors will require an excellent dose (in terms of the metrics mentioned above), while they can be more permissive for "harder" cases. This makes the acceptability of a plan hard to define in general.

Reinforcement learning is a machine learning paradigm that trains agents to make sequential decisions in dynamic environments [2]. Through trial and error guided by rewards or penalties, agents learn to optimize their actions to achieve long-term objectives. The decisions taken by dosimetrists when optimizing treatment can be formalized as a reinforcement learning problem. Moreover, dosimetrists can guide the TPS towards an acceptable plan but usually struggle to explain their decision while interacting with the TPS. The difficulty in explaining why certain decisions are taken suggests using deep reinforcement learning over expert-based methods. This setup is similar to image recognition, where one can say a picture represents a car or a boat but struggles to explain why.

2 Materials and Methods

We introduce a new paradigm in reinforcement learning (RL) based on evaluating states rather than the reward.

2.1 Reinforcement Learning Reward




Fig. 1. Classical reinforcement learning reward for automatic dosimetry.

In classical RL, we want $V(S_t) = R_t + \gamma V(S_{t+1})$ (so the update is $V(S_t) \leftarrow (1 - \alpha)V(S_t) + \alpha \left[R_{t+1} + \gamma V(S_{t+1})\right]$). In the context of dose optimization, the reward R_t is defined as $R_t = \mathcal{E}(S_{t+1}) - \mathcal{E}(S_t)$. Where \mathcal{E} is a function that evaluates the quality of a state (such that higher is better; if lower is better, then swap s_t and S_{t+1}).

The evaluation \mathcal{E} can be one or a mixture of the metrics mentioned in the introduction (Section 1) [11] [12] [9]. This setup may leverage knowledge about which actions to perform instead of guessing randomly as a meta-optimizer would do. We can hope to gain some computation time.

However, this technique does not use past plans; it only needs the optimizer inputs (CT, structures contours). We propose using the availability of past treatment plans to better catch the complexity of decisions made by dosimetrists and better match their expectations of a fully automatic treatment planning system.

As developed in previous work, we can derive a distance between dose plans [add citation]. If we consider the clinical dose of past cases (used for training) as the best achievable one, we can evaluate a dose plan by computing its distance from the clinical dose plan.

Let D_t be the dose associated with S_t , and D_C the clinical dose. We then define $\mathcal{E}(S_t) = \mathcal{D}(D_t, D_C)$. Since in that case, lower is better, we will define the reward as

$$R_t = \mathcal{E}(S_t) - \mathcal{E}(S_{t+1}) = \mathcal{D}(D_t, D_C) - \mathcal{D}(D_{t+1}, D_C).$$

This reward can be interpreted as the "distance gained to the clinical dose".

2.2 Reward-Free Reinforcement Learning

Since the reward is based on an evaluation of the state, one may drop all the reward machinery and directly use the state evaluation. This allows us to capture

the signal better: it is not embedded in a reward system, and the network can learn faster.

The reason why the reward system is used in many games (such as go or chess) is that state evaluation is challenging, if not impossible; there is no function that, given the board, gives the quality of white/black's position. A reward is only given at the end of a game, usually +1 for winning and -1 for losing. The state's evaluation is deduced by the reinforcement learning agent while learning to maximize the reward.

However, in our dosimetrist case, we have precisely such an evaluation function. Embedding this evaluation in a reward by defining the reward as the evaluation difference turns out to dilute the signal, and the agent learns slower.

By predicting the next state evaluation instead of using the Bellman optimality equation, we lose the ability to make short-term concessions in order to obtain large rewards later on. In our case, making short-term concessions would mean that we trigger the weights in a way that the optimizer performs worse, but such that triggering a little more makes the optimizer perform better. Given the problem we are solving (optimization of a radiotherapy dose), this is highly unlikely.

Because optimizers have the terrible habit of getting stuck in a local minimum, we can not use the last optimization result as a "warm start" after modifying optimization weights. Hence, restarting optimization from zero is forced after every weight modification.

Given the statement of the optimization problem, and the fact that we can not use a warm start, making short-term compromises does not make much sense. It is, therefore, safe to predict the next state evaluation directly, without the use of the Bellman optimality equation.

2.3 Avoiding Over-fitting

We use a dense neural network, taking the DVHs and the current normalized weight values as inputs. Dense layers are very prone to overfitting. In order to force the network to actually predict the following evaluation for each possible action, without over-fitting, we incorporated a bottleneck in the network. Compressing the information stops the network from learning "by heart". Networks with such architecture show significantly better results on validation.

2.4 Avoiding Off-Distribution

We generated a training set of over 125k actions (this took five days on an NVIDIA GeForce GTX 1080). Despite this relatively large dataset, we have not explored exhaustively the state-actions space, and the network still gets off distribution. This can easily be spotted when the predicted distance is negative; we choose to ignore those predictions. In fact, we ignore all outlier predictions. The justification is that our set of actions is limited, no action will suddenly drastically improve the plan. It is the combination of several sequential actions that

allows good plan optimization. Therefore, while testing, we choose the action with the best prediction, while passing the outlier test just mentioned.

3 Results

We successfully managed to mimic the dose with RL.

4 Discussion

Appendix

As this is very new and ongoing research, we generated synthetic phantom patients and associated trustable clinical doses. In future work, we hope to be able to apply this technique to real cases.

Synthetic phantom patients

We generated 130 patients with oval axial section bodies. We set the body density to water density. We then added an ellipsoid PTV within the body, with a slightly different density (following $\mathcal{N}(1,0.05)$). Likewise, we generate five organs gravitating around the PTV, aligned on the axial section.




Fig. 2. first generated patient: Main axial slice (center of the PTV) **CT**. Structures are also contoured with legend.

Clinical dose

After generating the patient's CT and structures, we needed to generate a reference dose that our agent should mimic. We manually set weights and performed a standard optimization.




Fig. 3. First generated patient: main axial slice (center of the PTV) of the clinical dose; structures are contoured as before.

Optimization

We optimize the plan using the LBFGS optimizer (shown to be the most appropriate in [4]), with learning rate 0.03. Given the list of DVH constraints (e.g. for PTV, $D_{95} > 80~Gy$), we use a linear penalization of the overdose. Summing the multiplication of each cost term with corresponding weight gives our objective function.




Fig. 4. First generated patient: clinical dose DVH.

References

- 1. Treatment Planning System Basics | Oncology Medical Physics, https://oncologymedicalphysics.com/introduction-to-treatment-planning-systems/
- 2. Brooks, R.: What is reinforcement learning? (Dec 2021), https://online.york.ac.uk/what-is-reinforcement-learning/
- 3. Das, I.J., Compton, J.J., Bajaj, A., Johnstone, P.A.: Intra- and inter-physician variability in target volume delineation in radiation therapy. Journal of Radiation Research p. rrab080 (Sep 2021). https://doi.org/10.1093/jrr/rrab080, https://academic.oup.com/jrr/advance-article/doi/10.1093/jrr/rrab080/6367625
- 4. Dubois, P.: Radiotherapy Dosimetry: A Review on Open-Source Optimizer (May 2023), http://arxiv.org/abs/2305.18014, arXiv:2305.18014 [cs, eess]
- Galvin, J.M., Smith, A.R., Lally, B.: Characterization of a multileaf collimator system. International Journal of Radiation Oncology*Biology*Physics 25(2), 181–192 (Jan 1993). https://doi.org/10.1016/0360-3016(93)90339-W, https://linkinghub.elsevier.com/retrieve/pii/036030169390339W
- Huang, C., Yang, Y., Panjwani, N., Boyd, S., Xing, L.: Pareto Optimal Projection Search (POPS): Automated Radiation Therapy Treatment Planning by Direct Search of the Pareto Surface. IEEE Transactions on Biomedical Engineering 68(10), 2907–2917 (Oct 2021). https://doi.org/10.1109/TBME.2021.3055822, https://ieeexplore.ieee.org/document/9343695/
- 7. Li, X., Ge, Y., Wu, Q., Wang, C., Sheng, Y., Wang, W., Stephens, H., Yin, F.F., Wu, Q.J.: Input feature design and its impact on the performance of deep learning models for predicting fluence maps in intensity-modulated radiation therapy. Physics in Medicine & Biology 67(21), 215009 (Nov 2022). https://doi.org/10.1088/1361-6560/ac9882, https://iopscience.iop.org/article/10.1088/1361-6560/ac9882
- 8. Lyman, J.T.: Normal tissue complication probabilities: Variable dose per fraction. International Journal of Radiation Oncology*Biology*Physics 22(2), 247–250 (Jan 1992). https://doi.org/10.1016/0360-3016(92)90040-O, https://linkinghub.elsevier.com/retrieve/pii/036030169290040O
- 9. Moreau, G., François-Lavet, V., Desbordes, P., Macq, B.: Reinforcement Learning for Radiotherapy Dose Fractioning Automation. Biomedicines 9(2), 214 (Feb 2021). https://doi.org/10.3390/biomedicines9020214, https://www.mdpi.com/2227-9059/9/2/214
- T.: U., 10. Oelfke, Bortfeld. Inverse planning for photon and proton beams. Medical Dosimetry **26**(2). 113-124 (Jun 2001). https://doi.org/10.1016/S0958-3947(01)00057-7, https://linkinghub.elsevier.com/retrieve/pii/S0958394701000577
- 11. Shen, C., Chen, L., Jia, X.: A hierarchical deep reinforcement learning framework for intelligent automatic treatment planning of prostate cancer intensity modulated radiation therapy. Physics in Medicine & Biology 66(13), 134002 (Jul 2021). https://doi.org/10.1088/1361-6560/ac09a2, https://iopscience.iop.org/article/10.1088/1361-6560/ac09a2
- 12. Shen, C., Gonzalez, Y., Klages, P., Qin, N., Jung, H., Chen, L., Nguyen, D., Jiang, S.B., Jia, X.: Intelligent Inverse Treatment Planning via Deep Reinforcement Learning, a Proof-of-Principle Study in High Dose-rate Brachytherapy for Cervical Cancer. Physics in Medicine & Biology 64(11), 115013 (May 2019). https://doi.org/10.1088/1361-6560/ab18bf, http://arxiv.org/abs/1811.10102, arXiv:1811.10102 [physics]

- Valentini, V., Cellini, F., Minsky, B.D., Mattiucci, G.C., Balducci, M., D'Agostino, G., D'Angelo, E., Dinapoli, N., Nicolotti, N., Valentini, C., La Torre, G.: Survival after radiotherapy in gastric cancer: Systematic review and meta-analysis. Radiotherapy and Oncology 92(2), 176–183 (Aug 2009). https://doi.org/10.1016/j.radonc.2009.06.014, https://linkinghub.elsevier.com/retrieve/pii/S0167814009003247
- 14. Webb, S.: The physical basis of IMRT and inverse planning. The British Journal of Radiology **76**(910), 678–689 (Oct 2003). https://doi.org/10.1259/bjr/65676879, https://academic.oup.com/bjr/article/76/910/678-689/7470601
- Wu, X., Zhu, Y.: An optimization method for importance factors and beam weights based on genetic algorithms for radiotherapy treatment planning. Physics in Medicine and Biology 46(4), 1085–1099 (Apr 2001). https://doi.org/10.1088/0031-9155/46/4/313, https://iopscience.iop.org/article/10.1088/0031-9155/46/4/313
- Xing, L., Li, J.G., Donaldson, S., Le, Q.T., Boyer, A.L.: Optimization of importance factors in inverse planning. Physics in Medicine and Biology 44(10), 2525–2536 (Oct 1999). https://doi.org/10.1088/0031-9155/44/10/311, https://iopscience.iop.org/article/10.1088/0031-9155/44/10/311
- 17. Xu, D., Li, G., Li, H., Jia, F.: Comparison of IMRT versus 3D-CRT in the treatment of esophagus cancer: A systematic review and meta-analysis. Medicine 96(31), e7685 (Aug 2017). https://doi.org/10.1097/MD.00000000000007685, https://journals.lww.com/00005792-201708040-00033