Mid-PhD Defense

Paul Dubois

TheraPanacea
MICS, CentraleSupélec
Institut du Cancer de Montpellier

21st June 2023

Outline

- Introduction
  - Cancer treatments
  - Radiotherapy
  - Multi-Leaf Collimator
  - IMRT/VMAT scheme
    - Step-and-shot
    - Sliding-windows
  - VMAT scheme
  - Radiotherapy workflow
- Problem Statement
  - Optimization workflow
  - Fluence discretization
  - FMO problem
    - Formulation
    - Optimization
- Early results
- Future work
- Others
  - Courses
  - Doctoral training
- References

Automatic Dose Optimization for Radiotherapy



Problem Formulation

IMRT

Bixel values:

$$x_{i,j}^{\theta} \geq 0, \text{ for } \theta \in \Theta \text{ and } 1 \leq i, j \leq 20^{1}$$

usually concatenated to a single bixels-value vector x.

Dose calculation:

$\mathbf{y} = L\mathbf{x}$ with L (pre-calculated) dose-influence (DI) matrix

[1] 20x20 is a typical bixel discretization

Problem Formulation

IMRT (bis)

Objective for *maximum* constraint c on structure s, dose d:

$$f_c(\mathbf{y}) = \frac{1}{|\mathcal{V}|} \sum_{v \in \mathcal{V}} (\mathbf{y}_v - d)_+^2$$

(reverse sign for minimal constraint).

Final objective:

$$f(\mathbf{y}) = \sum_{c \in \mathcal{C}} w_c f_c(\mathbf{y})$$

with w_c the weight of constraint c.

Problem Optimization

Optimizer review



Figure: Typical prostate case.

`https://arxiv.org/abs/2305.18014`

Problem Optimization

Optimizer review (bis)



Figure: Typical prostate case.

`https://arxiv.org/abs/2305.18014`

Meta-Optimization

Usual optimization

$$\min_{\mathbf{x}} f(\mathbf{x}, w) \text{ s.t. } \mathbf{x} > 0$$

... and fine-tune w until the dose is clinically acceptable.

Meta optimization

$$\min_{w} \left\{ \min_{\mathbf{x}} f(\mathbf{x}, w) \text{ s.t. } \mathbf{x} > 0 \right\}$$

... still need to fine-tune the parameters (learning rate, momentum, etc...) of the meta-optimizer.

Dose Distances



Figure: Example of two doses that have the same clinical effect (measured from the DVHs), but very different voxel-wise dose values.