

A Preliminary Study on Automatic Motion Artifact Detection in Electrodermal Activity Data Using Machine Learning

Md-Billal Hossain, *Member, IEEE*, Hugo F. Posada-Quintero, *Member IEEE*, Youngsun Kong, *Member, IEEE*, Riley McNaboe and Ki H. Chon, *Senior Member, IEEE*

Abstract— The electrodermal activity (EDA) signal is a sensitive and non-invasive surrogate measure of sympathetic function. Use of EDA has increased in popularity in recent years for such applications as emotion and stress recognition; assessment of pain, fatigue, and sleepiness; diagnosis of depression and epilepsy; and other uses. Recently, there have been several studies using ambulatory EDA recordings, which are often quite useful for analysis of many physiological conditions. Because ambulatory monitoring uses wearable devices, EDA signals are often affected by noise and motion artifacts. An automated noise and motion artifact detection algorithm is therefore of utmost importance for accurate analysis and evaluation of EDA signals. In this paper, we present machine learning-based algorithms for motion artifact detection in EDA signals. With ten subjects, we collected two simultaneous EDA signals from the right and left hands, while instructing the subjects to move only the right hand. Using these data, we proposed a cross-correlation-based approach for non-biased labeling of EDA data segments. A set of statistical, spectral and model-based features were calculated which were then subjected to a feature selection algorithm. Finally, we trained and validated several machine learning methods using a leave-one-subject-out approach. The classification accuracy of the developed model was 83.85% with a standard deviation of 4.91%, which was better than a recent standard method that we considered for comparison to our algorithm.

I. INTRODUCTION

Electrodermal activity (EDA) represents changes in electrical conductance of the skin due to opening of sweat pores [1], [2]. It is believed that EDA represents sudomotor activities, as they are innervated by sympathetic C nerve fibers of the autonomic nervous system. Therefore, EDA has the potential to be used for the evaluation of sympathetic nervous function and cognitive arousal [3]–[7]. Due to ease of use and noninvasive data collection, EDA has been applied in many diverse areas including emotional arousal [8], decision-making [9], pain [10], [11], stress [12], autism [13], and panic disorder [14], as these are all related to elevation of sympathetic nervous activities.

While EDA has potential to be used as a peripheral sympathetic nervous activity marker, it has some limitations. For example, EDA can be affected by non-sympathetic factors such as atmospheric temperature and humidity [15]. Most of the traditional EDA studies were performed carefully in a

controlled laboratory setting with limited movement of the subjects in order to minimize the effect of motion artifacts and noise. However, recently there have been a large number of studies of ambulatory EDA data collection over long time periods using wearable devices [7]. Wearable devices are more prone to noise and artifacts [16] and EDA is not immune from this issue. Noise and motion artifacts can be generated from different sources such as poor contact between skin and the recording electrodes, movements that cause variations in the skin-electrode contact, intentional or unintentional touching of electrodes (e.g. autistic people might be prone to the latter), and contextual factors (e.g. temperature and humidity) that may cause excessive sweating. Therefore, for accurate evaluation of EDA signals, corrupted segments should be automatically identified and removed.

Despite an increasing volume of EDA research over the last decades, there has been only a handful of research papers on motion artifact detection in EDA. Many researchers used either exponential smoothing [17] or low pass filtering [18] to combat noise and motion artifacts. These techniques may smooth high variations in the signal; however, they cannot compensate for sudden and large-magnitude motion artifacts which are often present in EDA signals, especially in ambulatory recordings. There have been some heuristic methods [18] to remove motion artifacts in EDA. However, all the methods were developed on a specific dataset, hence, their performance becomes ineffective for untrained datasets. Taylor et al. [19] developed a machine learning algorithm using the support vector machine (SVM) classifier. They have used manually annotated EDA segments and extracted different statistical features to train an SVM classifier. Ian et al. [20] proposed a simple EDA quality assessment procedure based on some simple decision rules. While this method works for spiky and large amplitude motion artifacts, it fails in several other cases.

It should be noted that unlike other biosignals, such as electrocardiogram (ECG) and photoplethysmography (PPG), EDA does not exhibit periodicity. Hence, manual adjudication of clean versus noisy EDA can be rather tricky. In order to avoid this issue, we propose an automated correlation-based annotation criterion to determine if an EDA data segment is clean or noisy. Certainly, this is not a practical solution since this scheme will require a stationary reference signal. However, for our purpose in this work, this approach was used to determine clean versus noisy data. Finally, based on the

*Research supported by Office of Naval Research.

MB Hossain, H.F. Posada-Quintero, Y. Kong, R. McNaboe and Ki H. Chon are with the Department of Biomedical Engineering at University of

Connecticut, Storrs, CT 06269, USA. email: {md.b.hossain, hugo.posada-quintero, youngsun.kong, riley.mcnaboe, ki.chon} at uconn.edu

adjudicated data from the cross-correlation scheme, we developed a machine learning algorithm to identify the noisy EDA segments using several statistical and model-based features. This machine learning is the ultimate outcome of an automated way of adjudicating clean versus noisy data in practice.

II. METHODS AND MATERIALS

A. Data Collection

Ten healthy subjects, aged 20-35, participated in this study. Two channels of simultaneous EDA were collected from the right and left hands using ADInstrument's galvanic skin response modules. On each hand, a pair of stainless-steel electrodes were placed on index and middle fingers. The EDA signals were recorded at a sampling frequency of 1,000 Hz, then down-sampled to 8 Hz. The data collection protocols were designed so that the right hand made occasional movements to mimic regular motion artifacts people could create in their daily lives while the left hand was immobile in order to provide a reference EDA. All the experimental procedures were approved by the Institutional Review Board (IRB) for human subject research at The University of Connecticut and a written consent was collected from all the subjects before the experiment. A summary of the protocol is shown in Table I. The protocol consists of two parts, where the first part was done with no significant motion and the second

reference (from the left hand) EDA segments and the targeted EDA segments (from the right hand). If the correlation coefficient exceeded 0.85, we considered this targeted EDA segment as clean, otherwise it was considered noisy. Please note that the reference signal can be noisy occasionally because of intentional or unintentional movements. Therefore, we first manually checked the reference and marked the obviously noisy portions of EDA. For this purpose, we used three independent reviewers' annotations; any segment that was marked as noisy by any of the three reviewers was discarded from the analysis unless there was a very high correlation (> 0.95) between the reference and the target EDA segment. Fig. 1 shows an example of two simultaneous EDA channels (target EDA and reference EDA), their corresponding cross-correlation coefficient, and the observers' annotation for the reference EDA. It can be seen from Fig. 1 that the correlation coefficient follows the target EDA, meaning that when there is noise in the target EDA the correlation coefficient is low, and vice versa.

TABLE I. DATA COLLECTION PROTOCOL SUMMARY

Duration (second)	Activity	Remarks
Part I (Stress Test)		
120	Flat table, relaxing with eyes closed	Baseline
30	Start table tilt	Orthostatic Stress
120	Subject remains in tilted position	
150	Return table to flat position, subject relaxation	
120	Perform Stroop test	Cognitive Stress
Part II (Motion Artifact Test)		
60	Sitting up in a chair with no movement	Motion Artifact Induction
60	Sitting down and typing on the computer	
60	Sitting down and holding a mouse, clicking the mouse	
60	Standing up with the arm next to torso	
60	Standing up swinging arm by the side as if to simulate walking	
60	Standing up with arm straight out, moving arm up and down continuously	
60	Standing up with arm straight out moving the arm at the elbow allowing the wrist to come into the chest and straight out and do the same again.	
60	Standing up with arm straight, completing a circle with the fingertips by rolling the shoulder.	




Figure 1. Noisy EDA channel (right hand), clean or reference EDA (left hand), correlation coefficient and observers' annotation of reference EDA.

part was designed to mimic different motion artifacts.

B. Data Labeling

As mentioned earlier, EDA data quality labeling can be difficult given the non-conventional characteristics of EDA signals, hence, observers' adjudications can be incorrect. To avoid human adjudication bias, we propose a cross-correlation coefficient-based criterion to annotate EDA segments as either noisy or clean. EDA data were first segmented into 1692 non-overlapping 5-second windows. We computed the cross-correlation coefficient between each of the simultaneous

C. Feature Extraction

We computed a set of features from each of the 5 second EDA window segments. First, we computed different statistical features from the original EDA, and its first and second derivative. We modeled the EDA sequence using an autoregressive (AR) model and considered 2 AR parameters (a_1 and a_2) and the AR noise variance as features. The motivation behind including AR modelling is that when EDA data are corrupted by noise, there will be more residual noise in the AR model than in the clean data; similarly, both AR parameters will have greater values for the noisy than clean data.

We then used a high-resolution time frequency decomposition method called variable frequency complex demodulation (VFCDM) [21], which provides more dynamic features of the clean and corrupted EDA. VFCDM has been previously used for several biosignal applications and has been found to be effective in analyzing signal characteristics [22] and removing noise and artifacts [23], [24]. We decomposed the EDA data segments into 12 non-overlapping frequency bands using VFCDM. Finally, we reconstructed two signals using the first 3 modes of VFCDM for the first signal, and the rest of the modes for the second signal. We computed the mean, variance, ratio of the variances, and ranges (max-min) of these two signals using VFCDM. Note that modes (1-3) include most of the dynamic characteristics of the EDA. Therefore, these features may reflect the signal and noise strength in the EDA segment.

As in [19], we also used the discrete wavelet transform and computed several other features from the details and approximation coefficients. We used three-level wavelet decomposition using the Haar wavelet. By doing so, we may have included some redundant features. However, we were not concerned about including too many features because we used a feature selection algorithm to reduce the risk of overfitting, as described in the subsequent section. All features computed are shown in Table II.

TABLE II. SUMMARY OF THE FEATURES COMPUTED

Index	Category	Specific features
1-3	AR(2) Modelling	AR parameters, AR noise variance
4-9	Raw EDA	Mean, median, variance, Shannon entropy, range, and skewness
10-19	First and second derivative of EDA	Mean, variance, max, and min of the absolute value
20-43	Wavelet decomposition	Mean, median, variance, Shannon entropy, range, and number above zero of the wavelet coefficients
44-52	VFCDM decomposition	Mean, variance, range of the two intermediate reconstructed signals, and ratio of the variances.

D. Feature Selection and classification:

We used the random forests (RF) machine learning algorithm for feature selection [25]. RF is a popular feature selection algorithm because of its good predictive performance, low overfitting, and interpretability. Feature selection using RF is in the embedded methods category which is a fusion of filter and wrapper methods. The embedded methods are highly accurate, easily generalizable, and interpretable in terms of feature selection.

To perform feature and model selection we followed a subject-independent validation strategy. We performed a leave one subject out (LOSO) validation strategy, meaning in every fold we left one subject out for testing and used the rest for training. We tested several classifiers such as random

forests (RF), support vector machine (SVM) with linear and radial basis function (RBF) kernels, and K nearest neighbor (KNN). Among those tested, RF and SVM with RBF kernel showed the best performance. We obtained almost similar results using both RF and SVM with RBF kernel, hence, we present the results for only these two classifiers. For feature selection we used a group k-fold validation using the training data at each fold of LOSO validation. Again, we used a group k-fold to ensure the classifiers were subject-independent. To select the optimal parameter for each fold, we performed a grid-search cross-validation technique with the group k-fold. The parameters C and gamma for SVM were selected from parameter candidates of 1, 10, 100, and 1000, and 0.001, 0.01, 0.1, 1, respectively.

III. RESULTS

A. Classification results

Table III shows the classification results using both RF and SVM. We compared the classification performance with a previously published motion artifact detection algorithm, which showed promising results [20].

TABLE III. CLASSIFICATION RESULTS AND COMPARISON

Methods	Mean accuracy	Standard deviation
Ian et al. [20]	75.05%	9.73%
This work (RF)	83.40%	4.06%
This work (SVM)	83.85%	4.91%

As shown in Table III, SVM provided the highest detection accuracy. The performance of [20] is lower compared to SVM. The performance of RF and SVM was similar; SVM had slightly higher accuracy (83.85%) albeit at the expense of higher standard deviation. While RF and SVM have standard deviation less than 5% (4.06% and 4.91%, respectively), [20] showed a standard deviation of 9.73%. This suggests that the machine learning methods used in this work are more consistent for motion artifact detection.

B. Feature Analysis

We have tracked the features selected by RF at each fold and the most consistent features that were selected are presented in Table IV. As mentioned earlier, at every fold we performed a group five-fold validation on the training data to select the best feature combinations. For the RF classifier, the features selected did not change the classification accuracy significantly, however, for SVM the accuracy increased by around 3.2% (SVM accuracy was 80.65% before feature selection). Finally, we have statistically analyzed some of the features selected by the RF feature selection. We performed a pairwise t-test for the feature values of the two classes and found statistically significant difference in most of the selected features.

TABLE IV. FEATURES SELECTED CONSISTENTLY

Category	Features selected
AR modeling	Noise variance, AR parameters (a_1 and a_2)
Raw EDA	Mean, range 1 st derivative: max 2 nd derivative: range

VFCDM	Variance in first three modes and rest of the modes (4-12)
Wavelet	Mean of approximations Variances of details coefficients




Figure 2. Selected feature statistics for clean and noisy EDA.

Fig. 2 shows box plots of the distribution of four of the consistent features' distributions for the clean and noisy EDA. As shown in Fig. 2, there is a clear visual difference between the mean values of the features of two classes. As expected, the AR model had higher residual noise for the noisy EDA; the AR parameters also reflect this observation. The same is noted for the variance of $cd1$ (detail coefficients 1). The t-test also provided statistically significant differences ($p < 0.01$).

IV. CONCLUSIONS

We presented preliminary results of an automated motion artifact detection algorithm for EDA signals. We created an EDA database with simultaneous reference and target EDA channels and proposed a correlation-based criterion to define clean and noisy EDA segments without human annotations which could be biased and may differ from person to person. We have proposed several statistical, model-based and time-frequency features and applied a subject-independent machine learning algorithm for automated motion artifact detection in EDA signals. We also selected the consistent features using RF for subsequent classification analysis. The performance of the machine learning-based motion artifact detector is compared with a recently published promising method. It was observed that our approach performed better and with more consistency. While the proposed method showed promising performance, more data and further rigorous analysis is needed in the future to confirm the results.

REFERENCES

- [1] "Publication recommendations for electrodermal measurements," *Psychophysiology*, vol. 49, no. 8, pp. 1017–1034, 2012, doi: <https://doi.org/10.1111/j.1469-8986.2012.01384.x>.
- [2] J. T. Cacioppo, L. G. Tassinary, and G. G. Berntson, Eds., *Handbook of Psychophysiology*, 4th ed. Cambridge: Cambridge University Press, 2016.
- [3] M. Benedek and C. Kaernbach, "A continuous measure of phasic electrodermal activity," *J. Neurosci. Methods*, vol. 190, no. 1, pp. 80–91, Jun. 2010, doi: [10.1016/j.jneumeth.2010.04.028](https://doi.org/10.1016/j.jneumeth.2010.04.028).
- [4] P. H. Ellaway, A. Kuppaswamy, A. Nicotra, and C. J. Mathias, "Sweat production and the sympathetic skin response: Improving the clinical assessment of autonomic function," *Auton. Neurosci. Basic Clin.*, vol. 155, no. 1, pp. 109–114, Jun. 2010, doi: [10.1016/j.autneu.2010.01.008](https://doi.org/10.1016/j.autneu.2010.01.008).
- [5] C. Setz, B. Amrich, J. Schumm, R. L. Marca, G. Tröster, and U. Ehlert, "Discriminating Stress From Cognitive Load Using a Wearable EDA Device," *IEEE Trans. Inf. Technol. Biomed.*, vol. 14, no. 2, pp. 410–417, Mar. 2010, doi: [10.1109/TITB.2009.2036164](https://doi.org/10.1109/TITB.2009.2036164).
- [6] J. A. Healey and R. W. Picard, "Detecting stress during real-world driving tasks using physiological sensors," *IEEE Trans. Intell. Transp. Syst.*, vol. 6, no. 2, pp. 156–166, Jun. 2005.
- [7] H. F. Posada-Quintero and K. H. Chon, "Innovations in Electrodermal Activity Data Collection and Signal Processing: A Systematic Review," *Sensors*, vol. 20, no. 2, Art. no. 2, Jan. 2020.
- [8] M. M. Bradley and P. J. Lang, "Emotion and motivation," in *Handbook of psychophysiology*, 3rd ed. New York, NY, US: Cambridge University Press, 2007, pp. 581–607.
- [9] A. Bechara, H. Damasio, A. R. Damasio, and G. P. Lee, "Different Contributions of the Human Amygdala and Ventromedial Prefrontal Cortex to Decision-Making," *J. Neurosci.*, vol. 19, no. 13, pp. 5473–5481, Jul. 1999, doi: [10.1523/JNEUROSCI.19-13-05473.1999](https://doi.org/10.1523/JNEUROSCI.19-13-05473.1999).
- [10] H. F. Posada-Quintero et al., "Using electrodermal activity to validate multilevel pain stimulation in healthy volunteers evoked by thermal grills," *Am. J. Physiol.-Regul. Integr. Comp. Physiol.*, vol. 319, no. 3, pp. R366–R375, Jul. 2020, doi: [10.1152/ajpregu.00102.2020](https://doi.org/10.1152/ajpregu.00102.2020).
- [11] Y. Kong, H. F. Posada-Quintero, and K. H. Chon, "Pain Detection using a Smartphone in Real Time*," in *2020 42nd Annual International Conference of the IEEE Engineering in Medicine Biology Society (EMBC)*, Jul. 2020, pp. 4526–4529.
- [12] T. Reinhardt, C. Schmahl, S. Wüst, and M. Bohus, "Salivary cortisol, heart rate, electrodermal activity and subjective stress responses to the Mannheim Multicomponent Stress Test (MMST)," *Psychiatry Res.*, vol. 198, no. 1, pp. 106–111, Jun. 2012, doi: [10.1016/j.psychres.2011.12.009](https://doi.org/10.1016/j.psychres.2011.12.009).
- [13] E. B. Prince et al., "The relationship between autism symptoms and arousal level in toddlers with autism spectrum disorder, as measured by electrodermal activity," *Autism*, vol. 21, no. 4, pp. 504–508, May 2017.
- [14] A. E. Meuret et al., "Do Unexpected Panic Attacks Occur Spontaneously?," *Biol. Psychiatry*, vol. 70, no. 10, pp. 985–991, Nov. 2011, doi: [10.1016/j.biopsych.2011.05.027](https://doi.org/10.1016/j.biopsych.2011.05.027).
- [15] W. Boucsein, *Electrodermal Activity*, 2nd ed. Springer US, 2012.
- [16] J. Lázaro, N. Reljin, M. B. Hossain, Y. Noh, P. Laguna, and K. Chon, "Wearable Armband Device for Daily Life Electrocardiogram Monitoring," *IEEE Trans. Biomed. Eng.*, pp. 1–1, 2020, doi: [10.1109/TBME.2020.2987759](https://doi.org/10.1109/TBME.2020.2987759).
- [17] J. Hernandez, R. R. Morris, and R. W. Picard, "Call Center Stress Recognition with Person-Specific Models," in *Affective Computing and Intelligent Interaction*, Berlin, Heidelberg, 2011, pp. 125–134.
- [18] R. Kocielnik, N. Sidorova, F. M. Maggi, M. Ouwerkerk, and J. H. D. M. Westerink, "Smart technologies for long-term stress monitoring at work," in *Proceedings of the 26th IEEE International Symposium on Computer-Based Medical Systems*, Jun. 2013, pp. 53–58.
- [19] S. Taylor, N. Jaques, W. Chen, S. Fedor, A. Sano, and R. Picard, "Automatic identification of artifacts in electrodermal activity data," in *2015 37th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC)*, Aug. 2015, pp. 1934–1937.
- [20] I. R. Kleckner et al., "Simple, Transparent, and Flexible Automated Quality Assessment Procedures for Ambulatory Electrodermal Activity Data," *IEEE Trans. Biomed. Eng.*, vol. 65, no. 7, pp. 1460–1467, Jul. 2018.
- [21] H. Wang, K. Siu, K. Ju, and K. H. Chon, "A High Resolution Approach to Estimating Time-Frequency Spectra and Their Amplitudes," *Ann. Biomed. Eng.*, vol. 34, no. 2, pp. 326–338, Feb. 2006.
- [22] H. F. Posada-Quintero, J. P. Florian, A. D. Orjuela-Cañón, and K. H. Chon, "Highly sensitive index of sympathetic activity based on time-frequency spectral analysis of electrodermal activity," *Am. J. Physiol.-Regul. Integr. Comp. Physiol.*, vol. 311, no. 3, pp. R582–R591, Jul. 2016.
- [23] M.-B. Hossain, J. Lázaro, Y. Noh, and K. H. Chon, "Denosing Wearable Armband ECG Data Using the Variable Frequency Complex Demodulation Technique," in *2020 42nd Annual International Conference of the IEEE Engineering in Medicine Biology Society (EMBC)*, Jul. 2020, pp. 592–595.
- [24] M.-B. Hossain, S. K. Bashar, J. Lázaro, N. Reljin, Y. Noh, and K. H. Chon, "A robust ECG denosing technique using variable frequency complex demodulation," *Comput. Methods Programs Biomed.*, p. 105856, Nov. 2020, doi: [10.1016/j.cmpb.2020.105856](https://doi.org/10.1016/j.cmpb.2020.105856).
- [25] L. Breiman, "Random Forests," *Mach. Learn.*, vol. 45, no. 1, pp. 5–32, Oct. 2001, doi: [10.1023/A:1010933404324](https://doi.org/10.1023/A:1010933404324).