

Literature Review: Deep Learning for Seizure Detection and Prediction from Non-EEG Wearable Signals (2015–2025)(Exposé)

Paulin Saher
Universität zu Köln

December 16, 2025

Contents

1	Introduction	1
1.1	Problem Statement and Motivation	1
2	Theoretical Foundations for Ambulatory Seizure Monitoring	2
2.1	The Ambulatory Monitoring Paradigm: Requirements and Challenges . . .	2
2.2	Physiological Biosignals for Wearable Sensing	2
2.3	Seizure Manifestations and Corresponding Biosignals	2
2.4	Evaluation Frameworks: From Retrospective Analysis to Prospective Validity	3
2.5	Methodological Approach and Structure	3
2.6	Provisional Table of Contents	4
3	Modeling Landscape and Design Dimensions	5
3.1	Architectural Evolution: Feature Engineering vs. Deep Learning	6
3.2	Sensor Modalities and Fusion Strategies	6
3.3	Personalization and Domain Shift	6
3.4	Robustness and Generalization	6
3.5	Deployment Constraints: Latency, Energy, and Interpretability	6

1 Introduction

1.1 Problem Statement and Motivation

Epilepsy, a chronic neurological disorder affecting 60 million people worldwide, with approximately one-third of patients remaining drug-resistant (Hussein et al., 2018). Sveinson et al., 2020 found that 69% of Deaths due to a Generalized tonic-clonic seizure could have been prevented if not left unattended, highlighting the demand for reliable and timely prediction.

While reliable prediction is already possible through electroencephalography (EEG), its operational complexity render it unsuitable for ambulatory use. To address the need for ambulatory detection many are experimenting with utilizing autonomic and motion signals to detect and predict seizures.

However, translating and analysing these often noisy and complex time-series physiological signals has pushed traditional machine learning (ML) approaches to their limits. These approaches require manual engineering such as cleaning and extracting of the data, often failing to capture the intricate and non-linear complex temporal dynamic of these signals. It is this limitation that deep learning (DL) models are designed to overcome, with their ability of hierarchical and learned feature extraction uniquely addressing these limitations.

This review therefore synthesizes and critically evaluates recent advances in time-series deep learning—particularly convolutional neural networks (CNNs) and long short term memory (LSTMs)—applied to seizure detection and prediction using non-EEG wearable biosignals. Central to this analysis is the question of how architectural choices, training paradigms (e.g., transfer learning, personalization), and rigorous, prospective evaluation can converge to meet the stringent clinical requirements of high sensitivity and minimal false alarm rates (FAR) in real-world deployment.

2 Theoretical Foundations for Ambulatory Seizure Monitoring

2.1 The Ambulatory Monitoring Paradigm: Requirements and Challenges

Beniczky et al., 2021 The translation of seizure detection from the controlled clinical environment to an ambulatory setting redefines approaches and performance requirements. In clinical environments maximum sensitivity is prioritized over FAR, when engineering for ambulatory use a balance of both metrics is needed. Failing to detect a seizure (low sensitivity) poses a safety risk, While excessive false alarms (high FAR) leads to desensitizing and "alarm fatigue" and the device being ignored or not used at all. While more complex models can find more intricate patterns they also introduce the "black-box" problem wherein the model lacks Transparency. Therefore the design should incorporate explanations for decisions to ensure users trust and regulatory compliance.

2.2 Physiological Biosignals for Wearable Sensing

Wearable devices circumvent the impracticality of EEG by capturing peripheral biosignals that serve as proxies for central nervous system activity during seizures. These proxies will inherently relay incomplete and noisy signals so a combination of multiple sensors is required. The acceleration/attitude sensor excels at detecting muscle convulsions but is noisy in daily use, as it captures all movement. The electrodermal activity (EDA) sensor detects autonomic arousal by measuring the skin conductance associated with sweat gland activity.

2.3 Seizure Manifestations and Corresponding Biosignals

Wearable biosignals act as peripheral proxies for cerebral activity during seizures. Different seizure types produce distinct physiological signatures across these signals, informing detection approaches. This relationship centers on two primary manifestations: motor activity and autonomic activation.

Convulsive seizures, such as generalized tonic-clonic events, produce clear motor signatures captured by accelerometry (ACC), which measures movement and posture changes characteristic of ictal events (Wang et al., 2025; Wu et al., 2024).

Simultaneously, seizures drive autonomic responses recorded through complementary biosignals. Electrodermal activity (EDA) measures sympathetic nervous system arousal via skin conductance changes (Jain et al., 2017; Miron et al., 2025). Cardiac activity provides particularly strong autonomic proxies: electrocardiography (ECG) captures heart rate and rhythm alterations (Ghaderi, 2025; Leal et al., 2017), while heart rate variability (HRV) quantifies autonomic balance dynamics (Mason et al., 2024; Pavei et al., 2017).

In wearable applications, photoplethysmography (PPG) serves as a practical alternative for cardiac monitoring, deriving pulse rate and variability though with greater motion sensitivity than ECG (seizuresDetectionUsingMultimodal; sethFeasibilityCardiacBasedSeizur

Consequently, detection strategies combine these biosignals according to target seizure types, with multimodal fusion providing the most robust approach for ambulatory monitoring.

2.4 Evaluation Frameworks: From Retrospective Analysis to Prospective Validity

2.5 Methodological Approach and Structure

The review follows structured IS/medical literature guidance (concept matrix inspired by (Webster & Watson, 2002) and PRISMA principles). The methodological approach uses a multi-stage search and transparent extraction/synthesis procedures as outlined below.

1. Search strategy (Multi-stage, 2015–2025):

- *Scoping*: Google Scholar for broad coverage and identifying candidate papers.
- *Formal queries*: IEEE Xplore, PubMed, and Scopus using controlled keyword strings and deduplication.
- *Citation chasing*: Backward and forward citation tracking on key papers to ensure completeness.

2. Inclusion / Exclusion criteria:

- *Inclusion*: Studies using wearable/non-EEG autonomic signals for seizure detection or preictal prediction.
- *Exclusion*: EEG-only studies, invasive recordings, or papers limited to purely algorithmic simulations without human data.

Extracted dimensions populate a concept matrix with the following fields:

1. Setting: clinical vs ambulatory.
2. Modality: ECG/HRV, PPG, EDA, ACC.
3. Task: detection vs prediction.
4. Model class and fusion strategy.
5. Study phase: retrospective, pseudo-prospective, prospective.

6. Evaluation granularity: sample- vs alarm-based.
7. Windowing specifics (window length, stride, overlap).
8. Metrics: sensitivity, FAR/h, AUC, f1-score, latency.
9. Personalization.
10. Interpretability and safety considerations.
11. Data characteristics and dataset identity.
12. Missing data handling (imputation, exclusion) and synthetic data use.
13. Inference requirements (latency, computational load, hardware).

Prioritization Criteria: Studies including splits preserving patient-specific data, FAR/h reporting and external or cross validation will be prioritized in the synthesis. A PRISMA-style flow diagram will summarize and visualize screening and inclusion.

2.6 Provisional Table of Contents

Target Length: Approx. 15 Pages of Text

1. **Introduction** (approx. 1.5 pages)
 - 1.1 Problem Statement and Motivation
 - 1.2 Research Aim and Questions
 - 1.3 Structure of this Review
2. **Methodological Approach** (approx. 1.5 pages)
 - 2.1 Search Strategy (Multi-stage, 2015–2025)
 - 2.2 Inclusion and Exclusion Criteria
 - 2.3 Extraction Dimensions and Concept Matrix
 - 2.4 PRISMA Flow Diagram
3. **Theoretical Background and Clinical Concepts** (approx. 2 pages)
 - 3.1 Clinical Deployment Requirements (Sensitivity vs. FAR/h, Alarm Fatigue)
 - 3.2 Seizure Classification and Clinical Implications
 - 3.3 Autonomic Markers and Wearable Sensing (ECG/HRV, PPG, EDA, ACC)
 - 3.4 Study Phases: Retrospective vs. Pseudo-Prospective vs. Prospective
4. **Modeling Landscape and Design Dimensions** (approx. 5 pages)
 - 4.1 Architectural Evolution: Feature Engineering vs. Deep Learning
 - Traditional Pipelines (SVM, RF, kNN with HRV/EDA Features)
 - Deep Temporal Models (1D-CNN, LSTM, TCN)
 - Attention and Transformer Variants

- 4.2 Sensor Modalities and Fusion Strategies
 - Feature-Level, Decision-Level, and Representation-Level Fusion
 - Complementary Cues under Noise and Non-Stationarity
 - 4.3 Personalization and Domain Shift
 - Patient-Specific Fine-Tuning and Calibration
 - Cross-Dataset Evaluation and External Validation
 - 4.4 Robustness and Generalization
 - Artifact Handling and Augmentation Strategies
 - Out-of-Distribution Detection
 - 4.5 Deployment Constraints (Latency, Energy, Interpretability)
5. **Results: Synthesis and Comparative Analysis** (approx. 3 pages)
- 5.1 Performance Ranges by Study Phase and Setting
 - 5.2 Model Comparison: Feature-Based vs. Deep Learning
 - 5.3 Impact of Temporal Choices (Window Length, Stride, Context)
 - 5.4 Evidence on Robustness and Domain Shift
 - 5.5 Dataset Gaps, Class Imbalance, and Augmentation Effects
6. **Discussion** (approx. 1.5 pages)
- 6.1 Trade-offs: Complexity vs. Latency vs. Interpretability
 - 6.2 Clinical vs. Ambulatory Deployment Considerations
 - 6.3 Limitations of Current Evidence
7. **Conclusion and Future Work** (approx. 0.5 pages)
- 7.1 Summary of Key Findings
 - 7.2 Practical Recommendations
 - 7.3 Research Gaps and Future Directions

Note: The following sections provide preliminary background to contextualize the research questions. They will be expanded with systematic literature findings in the final paper.

3 Modeling Landscape and Design Dimensions

Detection and preictal prediction from ECG/HRV, PPG, EDA, and ACC are framed by windowing, labeling, and alarm logic, with alarm-based sensitivity, FAR/h, and latency as the outcomes used in the research questions.

3.1 Architectural Evolution: Feature Engineering vs. Deep Learning

Classical pipelines utilizing HRV, EDA, and ACC features combined with SVM, RF, or kNN remain competitive in retrospective settings (Mason et al., 2024; Seth et al., 2023). However, deep temporal models such as 1D-CNN, LSTM, and TCN often improve phase discrimination. Recently, attention variants and Transformers are being explored on these signals (Abtahi et al., 2025; Kalousios et al., 2024; Pordoy et al., 2025; Yang et al., 2022). A key inquiry of this review is determining how feature-based models compare to deep models and whether attention mechanisms add distinct value under ambulatory, alarm-based evaluation.

3.2 Sensor Modalities and Fusion Strategies

Fusion across ECG, EDA, ACC, and PPG at the feature, decision, or representation level seeks complementary cues and is relevant to maintaining performance under noise and non-stationarity (Pordoy et al., 2025; Yang et al., 2022).

3.3 Personalization and Domain Shift

Patient-specific fine-tuning and post-hoc calibration shift the operating point toward lower FAR/h and higher sensitivity in ambulatory use (Andrade et al., 2024). Thresholding and periodic recalibration are commonly used to handle drift.

3.4 Robustness and Generalization

Robustness requires explicit checks of domain shift across patient, device, site, and activity. Cross-dataset evaluation (e.g., training on TUH and testing on RPAH or EPILEPSIAE) exposes this shift and favors alarm-based over sample-based metrics (Andrade et al., 2024; Yang et al., 2022). Artifact handling, augmentation through time-warping, jitter, or scaling, and Out-of-Distribution (OOD) detection support generalization (Kalousios et al., 2024; Seth et al., 2023).

3.5 Deployment Constraints: Latency, Energy, and Interpretability

Wearable seizure detection imposes strict deployment constraints. Detection latency must remain below clinically relevant thresholds (typically seconds to allow caregiver response or patient warning), which limits model complexity and inference time. Energy consumption determines battery life; continuous inference on resource-constrained MCUs or mobile processors requires lightweight architectures such as quantized CNNs or efficient RNNs. Compact sequence models (e.g., depthwise-separable 1D-CNNs, temporal convolutional networks) and event-driven or on-demand processing reduce computational burden.

Interpretability remains critical for clinical adoption. Regulators and clinicians require transparency in alarm generation to trust automated decisions affecting patient safety. Post-hoc explanation methods (e.g., attention weights, Grad-CAM on temporal segments, SHAP values for HRV features) can highlight which signal segments or features triggered an alarm. Hybrid pipelines that combine interpretable HRV/EDA features with deep representations offer a middle ground between performance and explainability.

References

- Abtahi, A., Ryvlin, P., & Aminifar, A. (2025). Identification of Relevant ECG Features for Epileptic Seizure Prediction Using Interpretable Machine Learning. *IEEE Access*, 13, 111293–111303. DOI: [10.1109/ACCESS.2025.3583461](https://doi.org/10.1109/ACCESS.2025.3583461) (cit. on p. 6).
- Andrade, I., Teixeira, C., & Pinto, M. (2024). On the performance of seizure prediction machine learning methods across different databases: The sample and alarm-based perspectives. *Frontiers in Neuroscience*, 18. DOI: [10.3389/fnins.2024.1417748](https://doi.org/10.3389/fnins.2024.1417748) (cit. on p. 6).
- Beniczky, S., Wiebe, S., Jeppesen, J., Tatum, W. O., Brazdil, M., Wang, Y., Herman, S. T., & Ryvlin, P. (2021). Automated seizure detection using wearable devices: A clinical practice guideline of the International League Against Epilepsy and the International Federation of Clinical Neurophysiology. *Clinical Neurophysiology*, 132(5), 1173–1184. DOI: [10.1016/j.clinph.2020.12.009](https://doi.org/10.1016/j.clinph.2020.12.009) (cit. on p. 2).
- Ghaderi, F. (2025, April 11). *Advances in Machine Learning for Epileptic Seizure Prediction: A Review of ECG-Based Approaches*. 2025040942. DOI: [10.20944/preprints202504.0942.v1](https://doi.org/10.20944/preprints202504.0942.v1). (Cit. on p. 3).
- Hussein, A. F., Arunkumar, N., Gomes, C., Alzubaidi, A. K., Habash, Q. A., Santamaria-Granados, L., Mendoza-Moreno, J. F., & Ramirez-Gonzalez, G. (2018). Focal and Non-Focal Epilepsy Localization: A Review. *IEEE Access*, 6, 49306–49324. DOI: [10.1109/ACCESS.2018.2867078](https://doi.org/10.1109/ACCESS.2018.2867078) (cit. on p. 1).
- Jain, S., Oswal, U., Xu, K. S., Eriksson, B., & Haupt, J. (2017). A Compressed Sensing Based Decomposition of Electrodermal Activity Signals. *IEEE Transactions on Biomedical Engineering*, 64(9), 2142–2151. DOI: [10.1109/TBME.2016.2632523](https://doi.org/10.1109/TBME.2016.2632523) (cit. on p. 3).
- Kalousios, S., Müller, J., Yang, H., Eberlein, M., Uckermann, O., Schackert, G., Polanski, W. H., & Leonhardt, G. (2024). ECG-based epileptic seizure prediction: Challenges of current data-driven models. *Epilepsia Open*, 10(1), 143–154. DOI: [10.1002/epi4.13073](https://doi.org/10.1002/epi4.13073) (cit. on p. 6).
- Leal, A., da Graça Ruano, M., Henriques, J., de Carvalho, P., & Teixeira, C. (2017). On the viability of ECG features for seizure anticipation on long-term data. *2017 IEEE 3rd International Forum on Research and Technologies for Society and Industry (RTSI)*, 1–5. DOI: [10.1109/RTSI.2017.8065951](https://doi.org/10.1109/RTSI.2017.8065951) (cit. on p. 3).
- Mason, F., Scarabello, A., Taruffi, L., Pasini, E., Calandra-Buonaura, G., Vignatelli, L., & Bisulli, F. (2024). Heart Rate Variability as a Tool for Seizure Prediction: A Scoping Review. *Journal of Clinical Medicine*, 13(3), 747. DOI: [10.3390/jcm13030747](https://doi.org/10.3390/jcm13030747) (cit. on pp. 3, 6).
- Miron, G., Halimeh, M., Jeppesen, J., Loddenkemper, T., & Meisel, C. (2025). Autonomic biosignals, seizure detection, and forecasting. *Epilepsia*, 66 Suppl 3, 25–38. DOI: [10.1111/epi.18034](https://doi.org/10.1111/epi.18034) (cit. on p. 3).
- Pavei, J., Heinzen, R. G., Novakova, B., Walz, R., Serra, A. J., Reuber, M., Ponnusamy, A., & Marques, J. L. B. (2017). Early Seizure Detection Based on Cardiac Autonomic Regulation Dynamics. *Frontiers in Physiology*, 8. DOI: [10.3389/fphys.2017.00765](https://doi.org/10.3389/fphys.2017.00765) (cit. on p. 3).
- Pordoy, J., Jones, G., Matoorian, N., Evans, M., Dadashiserej, N., & Zolgharni, M. (2025). Enhanced Non-EEG Multimodal Seizure Detection: A Real-World Model for Identifying Generalised Seizures Across the Ictal State. *IEEE Journal of Biomedical*

- and *Health Informatics*, 29(5), 3329–3342. DOI: [10.1109/JBHI.2025.3532223](https://doi.org/10.1109/JBHI.2025.3532223) (cit. on p. 6).
- Seth, E. A., Watterson, J., Xie, J., Arulsamy, A., Md Yusof, H. H., Ngadimon, I. W., Khoo, C. S., Kadirvelu, A., & Shaikh, M. F. (2023). Feasibility of cardiac-based seizure detection and prediction: A systematic review of non-invasive wearable sensor-based studies. *Epilepsia Open*, 9(1), 41–59. DOI: [10.1002/epi4.12854](https://doi.org/10.1002/epi4.12854) (cit. on p. 6).
- Sveinsson, O., Andersson, T., Mattsson, P., Carlsson, S., & Tomson, T. (2020). Clinical risk factors in SUDEP. *Neurology*, 94(4), e419–e429. DOI: [10.1212/WNL.00000000000008741](https://doi.org/10.1212/WNL.00000000000008741) (cit. on p. 1).
- Wang, J., Hu, D., Lai, X., Jiang, T., Jiang, T., Gao, F., Vidal, P.-P., & Cao, J. (2025). Epileptic Seizure Detection Based on Attitude Angle Signal of Wearable Device. *IEEE Transactions on Instrumentation and Measurement*, 74, 1–10. DOI: [10.1109/TIM.2025.3529058](https://doi.org/10.1109/TIM.2025.3529058) (cit. on p. 3).
- Webster, J., & Watson, R. T. (2002). Analyzing the past to prepare for the future: Writing a literature review. *MIS quarterly*, xiii–xxiii. Retrieved December 3, 2025, from URL: <https://www.jstor.org/stable/4132319> (cit. on p. 3).
- Wu, D., Wei, J., Vidal, P.-P., Wang, D., Yuan, Y., Cao, J., & Jiang, T. (2024). A Novel Seizure Detection Method Based on the Feature Fusion of Multimodal Physiological Signals. *IEEE Internet of Things Journal*, 11(16), 27545–27556. DOI: [10.1109/JIOT.2024.3398418](https://doi.org/10.1109/JIOT.2024.3398418) (cit. on p. 3).
- Yang, Y., Truong, N. D., Eshraghian, J. K., Maher, C., Nikpour, A., & Kavehei, O. (2022). A Multimodal AI System for Out-of-Distribution Generalization of Seizure Identification. *IEEE Journal of Biomedical and Health Informatics*, 26(7), 3529–3538. DOI: [10.1109/JBHI.2022.3157877](https://doi.org/10.1109/JBHI.2022.3157877) (cit. on p. 6).