


Prepared By: Quality Manager		Document No:	TDPL/GE/FOM-009
Approved By: Lab Head		Version No:	2.0
		Version Date:	01/10/2024
		Revision No:	00
DOCUMENT TITLE:	EQAS Sample Processing Form	Revision Date:	00-00-0000

Name Of EQAS Provider: *Metropolis Healthcare Ltd, EQAS*

EQAS Lab ID: *837*

Department Name: *Biochemistry*

Temperature of EQAS sample at the time of Receiving: *4°C*

EQAS Sample No.: *240412 - Sample No. 1, 2, 3*

EQAS Accession/SIN No.:

Reconstitute By: *Raju*

Reconstitution Date: *15/02/2025*

Processed By: *Raju*

Reviewed By (Pathologist): *Dr. K. Anusha*

Result Share with QA department (Date & Signature): *15/02/2025*

Result Dispatched to EQAS provider (Date): *15/02/2025*

EQAS evaluation result received on: *20/02/2025*

Section Head Signature: *K. Anusha*

Department Head Signature: *K. Anusha*

Sample ID: MHLEQASSAMPLE1      Sample Type: Serum  
Patient ID:  
Name:  
Sex:  
Birth Date:

Instrument:	609005				
<u>Test Name</u>	<u>Result</u>	<u>Ref. Range</u>	<u>Flags</u>	<u>Dilution</u>	<u>Completed</u>
FRT4	No Value	1.19 ng/dL	NCR	1	15/02/25 06:20 PM
SNSE2	238.68 pg/mL		CEX LEX PEX CLX	1	15/02/25 06:46 PM

CEX = Curve expired ; LEX = Lot expired ; PEX = Pack expired ;  
CLX = Calibrator lot expired ; NCR = No curve

Technologist \_\_\_\_\_

Sample Report  
UniCel DxI 800, Access Immunoassay System  
S/N 609005, Version 7.1.0

Sample ID: EQAS240412NO-1  
Patient ID:  
Name:  
Sex:  
Birth Date:

Sample Type: Serum

Instrument: 609005

<u>Test Name</u>	<u>Result</u>	<u>Ref. Range</u>	<u>Flags</u>	<u>Dilution</u>	<u>Completed</u>
AMH	4.32 ng/mL			1	15/02/25 03:59 PM
Ferritin	77.30 ng/mL			1	15/02/25 03:50 PM
Insulin	12.65 uIU/mL		CEX PEX	1	15/02/25 03:58 PM
Testo	1.48 ng/mL		CEX PEX CLX	1	15/02/25 03:38 PM
TotT3	114.30 ng/dL			1	15/02/25 03:48 PM
TotT4	6.70 ug/dL			1	15/02/25 03:41 PM
VitD	18.95 ng/mL			1	15/02/25 04:00 PM
hFSH	27.74 mIU/mL			1	15/02/25 04:10 PM
hLH	15.58 mIU/mL		CEX CLX	1	15/02/25 04:11 PM

CEX = Curve expired ; PEX = Pack expired ; CLX = Calibrator lot expired

Technologist \_\_\_\_\_

Sample Report  
UniCel Dxl 800, Access Immunoassay System  
S/N 609005, Version 7.1.0

Sample ID: MHLEQASNO1      Sample Type: Serum  
Patient ID:  
Name:  
Sex:  
Birth Date:

Instrument: 609005

<u>Test Name</u>	<u>Result</u>	<u>Ref. Range</u>	<u>Flags</u>	<u>Dilution</u>	<u>Completed</u>
FT3	2.49 pg/mL		CEX LEX PEX CLX	1	15/02/25 03:37 PM
TSH3	3.121 uIU/mL		LEX PEX	1	15/02/25 03:36 PM

LEX = Lot expired ; PEX = Pack expired ; CEX = Curve expired ;  
CLX = Calibrator lot expired

Technologist \_\_\_\_\_

Sample Report  
UniCel Dxl 800, Access Immunoassay System  
S/N 609005, Version 7.1.0

Sample ID: EQASSAMPLE-2  
Patient ID:  
Name:  
Sex:  
Birth Date:

Sample Type: Serum

Instrument: 609005

<u>Test Name</u>	<u>Result</u>	<u>Ref. Range</u>	<u>Flags</u>	<u>Dilution</u>	<u>Completed</u>
AMH	3.91 ng/mL			1	15/02/25 04:04 PM
FT3	2.55 pg/mL		CEX LEX PEX CLX	1	15/02/25 03:49 PM
Ferritin	15.56 ng/mL			1	15/02/25 03:55 PM
Insulin	13.39 uIU/mL		CEX PEX	1	15/02/25 04:02 PM
PAPPA	358.86 ng/mL 358.16		CEX	1	15/02/25 04:25 PM
TSH3	2.881 uIU/mL		LEX PEX	1	15/02/25 03:50 PM
Testo	1.52 ng/mL		CEX PEX CLX	1	15/02/25 03:42 PM
TotT3	82.23 ng/dL			1	15/02/25 03:53 PM
TotT4	13.25 ug/dL			1	15/02/25 03:46 PM
VitdD	20.33 ng/mL			1	15/02/25 04:05 PM
hFSH	24.64 mIU/mL			1	15/02/25 04:14 PM
hLH	14.57 mIU/mL		CEX CLX	1	15/02/25 04:15 PM

CEX = Curve expired ; PEX = Pack expired ; CLX = Calibrator lot expired ;  
LEX = Lot expired

Technologist \_\_\_\_\_

Page 1 of 1  
Printed 28/05/25 10:58 AM

Sample Report  
UniCel Dxl 800, Access Immunoassay System  
S/N 609005, Version 7.1.0

Sample ID:	MHLEQASSAMPLE2	Sample Type:	Serum		
Patient ID:					
Name:					
Sex:					
Birth Date:					
Instrument:	609005				
<u>Test Name</u>	<u>Result</u>	<u>Ref. Range</u>	<u>Flags</u>	<u>Dilution</u>	<u>Completed</u>
<del>FRT4</del>	<del>No Value</del>		<del>NCR</del>	<del>1</del>	<del>15/02/25 06:21 PM</del>
PAPPA	392.97 ng/mL 358.16		CEX	1	15/02/25 06:56 PM
SNSE2	386.90 pg/mL		CEX/LEX PEX/CLX	1	15/02/25 06:46 PM

CEX = Curve expired ; LEX = Lot expired ; PEX = Pack expired ;  
CLX = Calibrator lot expired ; NCR = No curve

Technologist \_\_\_\_\_

Sample Report  
UniCel Dxl 800, Access Immunoassay System  
S/N 609005, Version 7.1.0

Sample ID: EQASSAMPLENO-3  
Patient ID:  
Name:  
Sex:  
Birth Date:

Sample Type: Serum

Instrument: 609005

<u>Test Name</u>	<u>Result</u>	<u>Ref. Range</u>	<u>Flags</u>	<u>Dilution</u>	<u>Completed</u>
PAPPA	36.74 ng/mL	136.74 mIU/mL	CEX	1	15/02/25 04:50 PM

f  
CEX = Curve expired

Technologist \_\_\_\_\_