



CHRISTIAN MEDICAL COLLEGE

VELLORE, TAMIL NADU - 632004. INDIA

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME



PC-1024

MONTHLY SUMMARY REPORT - APRIL 2025

LAB NO: 14855

LAB NAME: TRUSTLAB DIAGNOSTICS PVT LTD

CONSTITUENT GROUP

HbA1c

Date of Result Entered: 11/04/2025

Date of Result Published: 05/05/2025

PT Item: Lyophilized human whole blood based

S. No.	Analyte	Method / Principle	Analyser	No. of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	HbA1c	HPLC-Ion Exchange	Tosoh (G8, G11 - HPLC / HPLC-723G8)	363	5.5	5.39	0.3	5.6 %	0.33	0.03

u* - measurement of uncertainty

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (Action Signal)

Kindly note: for appeals regarding result analysis and evaluation process, please contact via phone/email as given below within one week of receiving the report.

Enquiries: for all communications, please use CMC-EQAS user ID.

Coordinator Contact Details:

E-Mail: clinqc@cmcvellore.ac.in

Contact No.: 0416-2283102

LAB ADDRESS

TRUSTLAB DIAGNOSTICS PVT LTD,
STREET NO.5, PRAKASH NAGAR, BEGUMPET,
TELANGANA

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS report is confidential.

CMC EQAS does not sub contract any components.

***** End of Report *****

Pamela Christudoss

Dr. Pamela Christudoss,
CMC EQAS Coordinator,
Christian Medical College, Vellore.