

Instrument : Quant Studio

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
X HBV viral load	Log IU/mL	11.34	6.26	4.31	90300344.60
@ HCV viral load	Log IU/mL	5.83		0.90	-1.62
@ HIV-1 viral load	Log IU/mL	2.64	3.04	-1.65	9996.92

Legend @ : Acceptable	Total Parameters	3
! : 2.0 < z score < 3.0 - Warning	Not Evaluated Parameters	0
X : z score ≥3 - Unacceptable	Evaluated Parameters	3
# : Not Evaluated	Outlier Parameters (X)	1
⊙ : Delayed Result Entry	EQAS Score Molecular Biology- Quant	66.67 %
* : Not considered for evaluation.		

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kiroli Road, Kurla (W),
Mumbai - 400070

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⊙ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification:

Corrective Action:

Reviewed by:

Dated:

Analyte	Instrument	Result Value	Standard Unit	Z-Score
X HBV viral load	Quant Studio	11.34	Log Copies/ mL	4.31
@ HCV viral load	Quant Studio	5.83	Log Copies/ mL	0.9
<p>Legend @ : z score \leq 2.0 - Acceptable</p> <p>! : 2.0 < z score < 3.0 - Warning</p> <p>X : z score \geq 3 - Unacceptable</p> <p># : Not Evaluated</p> <p>⊙ : Delayed Result Entry</p> <p>* : Not considered for evaluation.</p>				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Quant Studio

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ HBV viral load	Log IU/mL	3.43	1.72	1.32	90300344.60
* HCV viral load	Log IU/mL	0.00	0.00	*	-1.62
@ HIV-1 viral load	Log IU/mL	0.00	0.00	0.00	9996.92

Legend @ : Acceptable	Total Parameters	3
! : 2.0 < z score < 3.0 - Warning	Not Evaluated Parameters	0
X : z score ≥ 3 - Unacceptable	Evaluated Parameters	3
# : Not Evaluated	Outlier Parameters (X)	0
⌚ : Delayed Result Entry	EQAS Score Molecular Biology- Quant	100.00 %
* : Not considered for evaluation.		

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kiroi Road, Kurla (W),
Mumbai - 400070