

# Validation of the Polar H10 Accelerometer in a Sports-Based Environment<sup>†</sup>

Emily Lam<sup>1,2</sup>, Danielle DuPlessis<sup>1,3</sup> , Michael Hutchison<sup>4,5,6</sup>, Shannon Scratch<sup>1,3,6</sup> and Elaine Biddiss<sup>1,2,3,\*</sup>

<sup>1</sup> Bloorview Research Institute, Holland Bloorview Kids Rehabilitation Hospital, Toronto, ON M4G 1R8, Canada

<sup>2</sup> Institute of Biomedical Engineering, University of Toronto, Toronto, ON M5S 1A8, Canada

<sup>3</sup> Rehabilitation Sciences Institute, University of Toronto, Toronto, ON M5S 1A8, Canada

<sup>4</sup> Faculty of Kinesiology and Physical Education, University of Toronto, Toronto, ON M5S 1A8, Canada

<sup>5</sup> Department of Physical Therapy, University of Toronto, Toronto, ON M5S 1A8, Canada

<sup>6</sup> Department of Pediatrics, University of Toronto, Toronto, ON M5S 1A8, Canada

\* Correspondence: ebiddiss@hollandbloorview.ca

<sup>†</sup> Presented at the 9th International Electronic Conference on Sensors and Applications, 1–15 November 2022; Available online: <https://ecsa-9.sciforum.net/>.

**Abstract:** The Polar H10 is a low-cost wearable with a heart rate monitor and tri-axial accelerometer with potential for many applications. While the device's heart rate monitor has been widely studied, there is no research validating the accelerometer specifically. The purpose of this study was to conduct a validation of the Polar H10 accelerometer to establish static and dynamic validity during a sports-based task. Static validity was determined by computing the relative error when using a level guide to hold each axis of the Polar H10 against gravity. Fifteen healthy adults (8F/7M) participated in sports-based tasks while wearing the Polar H10 (Polar Electro, Kempele, Finland) and a comparison device, the MetaMotionR inertial measurement unit (MbleintLab Inc., San Francisco, CA, USA). Dynamic validity was characterized using Pearson's correlation coefficient and root mean square error (RMSE). Additionally, common features in human activity recognition (mean magnitude, root mean square, power, and signal magnitude area) were computed in 2 s windows and compared via RMSE and Wilcoxon rank sum tests. When held against gravity, the Polar H10 had relative errors ranging from 2.620% to 4.288%, suggesting high static validity. During sports-based tasks, the accelerometers had correlations between 0.888 and 0.954, indicating sufficient concurrent validity for all axes, as well as acceleration magnitude. The differences in acceleration features were minimal (RMSE for mean, root mean square, power, and signal magnitude were 0.003 G, 0.004 G, 0.112 G, and 0.017 G, respectively), but all reached significance ( $p < 0.001$ ). These results provide evidence for the use of the Polar H10 accelerometer to measure movement during sport-like activities.

**Keywords:** accelerometer; validation; heart rate monitor; activity recognition; physical activity



**Citation:** Lam, E.; DuPlessis, D.; Hutchison, M.; Scratch, S.; Biddiss, E. Validation of the Polar H10 Accelerometer in a Sports-Based Environment. *Eng. Proc.* **2022**, *27*, 71. <https://doi.org/10.3390/ecsa-9-13346>

Academic Editor: Francisco Falcone

Published: 1 November 2022

**Publisher's Note:** MDPI stays neutral with regard to jurisdictional claims in published maps and institutional affiliations.



**Copyright:** © 2022 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (<https://creativecommons.org/licenses/by/4.0/>).

## 1. Introduction

Usability [1–3] and cost [4] are barriers to the widespread uptake of wearable sensing systems. The growth of the commercial wearable market presents an opportunity to overcome both of these challenges. Consumer-grade devices can be more user-friendly than those employed in research, which often require longer set up times [5] and are more expensive. However, it is important to establish the accuracy and reliability of all wearable devices, as these factors directly impact data interpretation and the utility of technologies developed using these sensors (i.e., algorithms and machine learning classifiers) [1].

The Polar chest band is a consumer-grade low-cost wearable that has previously been validated for heart rate monitoring in adults [6–8] and children [9]. A recent version of the device, the Polar H10, also features a tri-axial accelerometer. While the use of the sensor (in conjunction with Polar Team Pro software) to measure speed and distance has been

investigated [10], the use of the accelerometer has yet to be validated. The purpose of this work was to conduct a validation of the Polar H10 tri-axial accelerometer under static and dynamic conditions that may support its use in research and technology development.

## 2. Methods

Fifteen healthy adults were recruited from Holland Bloorview Kids Rehabilitation Hospital via email lists and word of mouth. Individuals over 18 years of age with normal or corrected-to-normal vision and hearing were eligible for the study. Although no screened participants were excluded, individuals with pre-existing cardiovascular conditions, a recent musculoskeletal injury (up to 10 days before the study session), and/or a condition or injury that could be aggravated by exercise were not eligible for the study. This research was approved by the Bloorview Research Institute Ethics Board (eREB #2020-0305) and the University of Toronto. Written informed consent was obtained from each participant.

To collect raw accelerometer data, the Polar H10 (Polar Electro, Kempele, Finland) was connected to the Polar Sensor Logger app (Jukka Happonen, available on the Google Play Store) via Bluetooth to log data at 200 Hz with a range of 16 G.

Static validity of the Polar H10 accelerometer was assessed by holding the device against a level surface guide with each axis aligned to gravity for 30 s. Accelerometer data were averaged, then relative errors between the accelerometer values and expected value (1 G) were computed. Since Kolmogorov–Smirnov normality tests reached significance, Wilcoxon rank sum tests were used to evaluate whether the static validity of the accelerometer changed after the study session, during which the accelerometer experienced a period of high-intensity movement.

For testing under dynamic conditions, Polar H10 measurements were compared to the MetaMotionR (Mbleintlab Inc., San Francisco, CA, USA), an inertial measurement unit that had previously been validated against motion capture systems [11]. Data were logged directly on the MetaMotionR at 200 Hz with a range of 16 G. The MetaBase App (Mbleintlab Inc., USA) was used to configure the MetaMotionR and transfer data. The Polar H10 and MetaMotionR were synchronized by tapping them on a table three times. Participants wore the Polar H10 around the chest as described in the device user manual [12]. The MetaMotionR was placed just below the Polar H10 on the participants' torsos using a skin-safe adhesive (Cover-Roll Stretch, BSN Medical, Charlotte, NC, USA) directly on the skin, followed by the MetaMotionR with a stronger tape overtop (Leukotape P, BSN Medical, USA) to minimize loosening of the adhesives and sensors as the participants started to sweat.

After placing sensors and going through a brief warm-up, participants performed semi-structured tasks simulating anticipated movements in a sports-based assessment under development [13]. During the study, participants performed 34 shuttle-runs in an 8 m<sup>2</sup> to 9 m<sup>2</sup> space. The shuttle-run activity was layered with a cognitive task (e.g., touching buttons in a specific sequence) such that the following dynamic movements were elicited: walking, running, lunging, shuffling, stopping/pausing, changing direction, and accelerating/decelerating [14]. Each study session lasted approximately two hours.

Dynamic validity was assessed by calculating the Pearson's correlation coefficient between the Polar H10 and MetaMotionR for each acceleration vector and the overall magnitude under raw and filtered conditions [15–17]. Low-pass zero-lag 4th order Butterworth filters with differing cut-off frequencies (10 Hz, 20 Hz, 30 Hz, and 40 Hz) were examined. To explore the application of the sensor for activity recognition, commonly used features [1] were compared using Wilcoxon rank sum tests. Mean magnitude, root mean square (RMS) of magnitude, power of magnitude, and signal magnitude area (SMA) were calculated over 2 s intervals, and then compared using the root mean squared error (RMSE) between the features from each accelerometer and Wilcoxon rank sum tests.

### 3. Results and Discussion

#### 3.1. Participants

Fifteen individuals (8F/7M) participated in the study. Table 1 summarizes the age ranges of the participants. Using the Godin–Shephard Leisure-Time Physical Activity Questionnaire, two participants were identified as inactive (score of less than 14), four were moderately active (score between 15 and 23), and nine were active (a score over 24) [18].

**Table 1.** Participant age and biological sex.

Age (Years)	Sex	
	Female	Male
18–24	5	3
25–34	3	3
34–55	0	1

#### 3.2. Static Validity

The relative error between the Polar H10 reading and expected value (1 G) for each axis is outlined in Table 2. A Wilcoxon rank sum test comparing relative error before and after the high-intensity activity found a significant difference for the Z axis ( $p = 0.006$ ), but not the X ( $p = 0.36$ ) or Y ( $p = 0.29$ ) axes.

**Table 2.** Static relative error (%) for each accelerometer axis before and after high-intensity activity.

Axis	Before High-Intensity Activity	After High-Intensity Activity
X	3.053 (0.466)	2.929 (0.466)
Y	4.288 (0.513)	4.255 (0.505)
Z	2.862 (0.207)	2.620 (0.195)

Under static conditions, the Polar H10 accelerometer had mean relative errors of 2.620% to 4.288%. This indicates high static validity, as relative error was less than 5% for all axes [15,16]. While the difference between relative error before and after the sport-like activity was statistically significant for the Z-axis ( $p = 0.006$ ), the magnitude of the difference was minute (a change in RE of 0.252%) and is likely acceptable for most applications.

#### 3.3. Concurrent Validity during Motion

Table 3 highlights the concurrent validity and RMSE for raw accelerometer data during the sports-based task. The highest mean correlation was observed in the overall magnitude ( $r = 0.954$ ) while the lowest was in the y-axis ( $r = 0.888$ ). The lowest RMSE was observed in magnitude (0.141 G), while the highest was in the z-axis (0.273 G).

**Table 3.** Mean and standard deviations of correlation coefficient ( $r$ ) and root mean squared error (RMSE) during motion.

Axis	$r$	RMSE (G)
X	0.891 (0.113)	0.231 (0.119)
Y	0.888 (0.160)	0.182 (0.125)
Z	0.909 (0.085)	0.273 (0.139)
Magnitude	0.954 (0.046)	0.141 (0.069)

Low-pass filtering at different cut-off frequencies led to minor decreases in correlations (Table 4) and increases in RMSE (Table 5). While small, all these differences were significant (Wilcoxon signed-rank test,  $p < 0.001$ ). The largest change was seen at a cut-off frequency

of 10 Hz, where the correlation for acceleration magnitude decreased from 0.954 to 0.949 and RMSE increased from 0.141 G to 0.151 G.

**Table 4.** Mean and standard deviations of correlation coefficient (r) for accelerometer data low-pass filtered at different cut-off frequencies.

Axis	Cut-Off Frequency			
	40 Hz	30 Hz	20 Hz	10 Hz
X	0.891 (0.232)	0.890 (0.144)	0.889 (0.145)	0.887 (0.146)
Y	0.888 (0.160)	0.887 (0.160)	0.886 (0.160)	0.884 (0.162)
Z	0.908 (0.085)	0.908 (0.086)	0.906 (0.086)	0.904 (0.088)
Magnitude	0.954 (0.047)	0.953 (0.047)	0.952 (0.049)	0.959 (0.051)

**Table 5.** Comparison between Polar and MetaMotionR values for features commonly used in activity recognition.

Feature	Polar	MetaMotionR	RMSE
Mean Magnitude (G)	1.098 (0.104)	1.134 (0.113)	0.003
Root Mean Square of Magnitude (G)	1.163 (0.162)	1.215 (0.113)	0.004
Power of Magnitude (n.u.)	1.778 (0.698)	1.945 (0.805)	0.112
SMA (G)	1.524 (0.229)	1.618 (0.233)	0.017

Table 5 shows a comparison of commonly used activity recognition features calculated from each accelerometer. Distributions for all features were non-normal (Kolmogorov–Smirnov,  $p < 0.001$ ) and Wilcoxon rank sum tests reached significance ( $p < 0.001$  for all features). However, the RMSEs between the Polar and MetaMotionR features were less than 0.005 G for RMS and mean and 0.017 G for SMA. Power had the highest RMSE at 0.112 G.

During dynamic sport-like tasks, the Polar H10 and MetaMotionR had mean correlations between 0.888 and 0.954, reflecting a high to very high correlation between the accelerometer signals [19]. With correlations of over 0.7, concurrent validity was found to be sufficient for all axes and for the overall magnitude [15,16]. Filtering led to significant but small decreases in concurrent validity. Commonly used activity recognition features (mean magnitude, RMS of magnitude, power of magnitude, and SMA) calculated with a window size of 2 s were found to be significantly different between the two devices. However, the magnitudes of the differences were small, especially when considering that the sensors were placed in slightly different locations.

### 3.4. Limitations

In addition to the slight offset between sensors, correlations between the accelerometers may have been lowered if sensors (especially the tape-secured MetaMotionR) became loose during exercise. While participants were instructed to inform research staff when this happened, correlations decreased in trials leading up to sensor adjustments. In this study, a preliminary validation focused on the correlation between the Polar H10 and MetaMotionR accelerometers as the desired application of the device was to measure acceleration patterns during high-intensity activity. Further investigation is needed to validate the Polar H10 accelerometer for other purposes, such as measuring peak impacts during contact sports.

### 3.5. Comparisons with Previous Work

With mean correlations of over 0.887 relative to the MetaMotionR, the Polar H10 accelerometer performs similarly to other wearable-integrated accelerometers. When com-

paring the Gear S smartwatch (Samsung Electronics, Suwon City, Korea) accelerometer to the Actigraph GR3X+, Davoudi et al. [20] found that the devices had correlations over 0.89 during shaker table and treadmill tests and over 0.7 for other daily activities. Additionally, when comparing the MinimaxX S4 (Catapult Innovations, Melbourne, Australia) accelerometer to a motion analysis system, Wundersitz et al. [16] found that RMSE between the measures were 0.11 G and 0.23 G during jogging and running tasks. Similarly, the RMSEs reported in the current study ranged from 0.141 G to 0.273 G.

#### 4. Conclusions

The findings of this research suggest that the accelerometer contained within the Polar H10 can measure accelerations with adults during non-contact sports activities. While filtering the data may be beneficial in some applications, it had a minor impact on the concurrent validity and error. These findings open opportunities for future development and use of the Polar H10 as a reliable low-cost, consumer-grade heart rate monitor and accelerometer.

**Author Contributions:** Conceptualization, E.L., D.D., M.H., S.S. and E.B.; Methodology, E.L., D.D., S.S. and E.B.; Software, E.L.; Validation, E.L.; Formal Analysis, E.L., S.S. and E.B.; Investigation, E.L., D.D., S.S. and E.B.; Resources, E.L., D.D., M.H., S.S. and E.B.; Data Curation, E.L.; Writing—Original Draft Preparation, E.L., D.D., S.S. and E.B.; Writing—Review and Editing, E.L., D.D., M.H., S.S. and E.B.; Visualization, E.L.; S.S. and E.B.; Supervision, S.S. and E.B.; Project Administration, E.L., S.S. and E.B.; Funding Acquisition, E.L., S.S. and E.B. All authors have read and agreed to the published version of the manuscript.

**Funding:** This research was funded by the Ontario Graduate Scholarship (OGS), Canadian Institute of Health Research (CIHR) Canada Graduate Scholarship (CGS), the Holland Bloorview Center for Leadership, and the Natural Sciences and Engineering Research Council of Canada (NSERC) Discovery Grant (#491004).

**Institutional Review Board Statement:** The study was conducted in accordance with the Declaration of Helsinki, and approved by the Institutional Review Board of Holland Bloorview Kids Rehabilitation Hospital (2020-0305, approved 4 January 2021) and the University of Toronto (40679, approved 23 February 2021).

**Informed Consent Statement:** Informed consent was obtained from all subjects involved in the study.

**Data Availability Statement:** The data presented in this study are available on request from the corresponding author. The data are not publicly available due to ethical restrictions.

**Acknowledgments:** We thank Ajmal Khan for providing technical support and expertise and Azadeh Kushki for providing valuable input and guidance. This work was supported by the Holland Bloorview Kids Rehabilitation Hospital Foundation, the Canadian Institutes of Health Research, and the University of Toronto.

**Conflicts of Interest:** The authors declare no conflict of interest.

#### References

1. Avci, A.; Bosch, S.; Marin-Perianu, M.; Marin-Perianu, R.; Havinga, P. Activity Recognition Using Inertial Sensing for Healthcare, Wellbeing and Sports Applications: A Survey. In Proceedings of the 23th International Conference on Architecture of Computing Systems 2010, Hannover, Germany, 22–25 February 2010; pp. 1–10.
2. Motti, V.G.; Caine, K. Human Factors Considerations in the Design of Wearable Devices. *Proc. Hum. Factors Ergon. Soc. Annu. Meet.* **2014**, *58*, 1820–1824. [[CrossRef](#)]
3. Smuck, M.; Odonkor, C.A.; Wilt, J.K.; Schmidt, N.; Swiernik, M.A. The Emerging Clinical Role of Wearables: Factors for Successful Implementation in Healthcare. *npj Digit. Med.* **2021**, *4*, 45. [[CrossRef](#)] [[PubMed](#)]
4. Patel, S.; Park, H.; Bonato, P.; Chan, L.; Rodgers, M. A Review of Wearable Sensors and Systems with Application in Rehabilitation. *J. NeuroEng. Rehabil.* **2012**, *9*, 21. [[CrossRef](#)] [[PubMed](#)]
5. Peake, J.M.; Kerr, G.; Sullivan, J.P. A Critical Review of Consumer Wearables, Mobile Applications, and Equipment for Providing Biofeedback, Monitoring Stress, and Sleep in Physically Active Populations. *Front. Physiol.* **2018**, *9*, 743. [[CrossRef](#)] [[PubMed](#)]
6. Hinde, K.; White, G.; Armstrong, N. Wearable Devices Suitable for Monitoring Twenty Four Hour Heart Rate Variability in Military Populations. *Sensors* **2021**, *21*, 1061. [[CrossRef](#)] [[PubMed](#)]

7. Gilgen-Ammann, R.; Schweizer, T.; Wyss, T. RR Interval Signal Quality of a Heart Rate Monitor and an ECG Holter at Rest and during Exercise. *Eur. J. Appl. Physiol.* **2019**, *119*, 1525–1532. [[CrossRef](#)] [[PubMed](#)]
8. Müller, A.M.; Wang, N.X.; Yao, J.; Tan, C.S.; Low, I.C.C.; Lim, N.; Tan, J.; Tan, A.; Müller-Riemenschneider, F. Heart Rate Measures from Wrist-Worn Activity Trackers in a Laboratory and Free-Living Setting: Validation Study. *JMIR Mhealth Uhealth* **2019**, *7*, e14120. [[CrossRef](#)] [[PubMed](#)]
9. Speer, K.E.; Semple, S.; Naumovski, N.; McKune, A.J. Measuring Heart Rate Variability Using Commercially Available Devices in Healthy Children: A Validity and Reliability Study. *Eur. J. Investig. Health Psychol. Educ.* **2020**, *10*, 390–404. [[CrossRef](#)]
10. Fox, J.L.; O’Grady, C.J.; Scanlan, A.T.; Sargent, C.; Stanton, R. Validity of the Polar Team Pro Sensor for Measuring Speed and Distance Indoors. *J. Sci. Med. Sport* **2019**, *22*, 1260–1265. [[CrossRef](#)]
11. Beange, K.H.E.; Chan, A.D.C.; Graham, R.B. Evaluation of Wearable IMU Performance for Orientation Estimation and Motion Tracking. In Proceedings of the 2018 IEEE International Symposium on Medical Measurements and Applications (MeMeA), Rome, Italy, 11–13 June 2018; pp. 1–6.
12. Polar Electro. Polar H10 User Manual. Available online: [https://support.polar.com/e\\_manuals/h10-heart-rate-sensor/polar-h10-user-manual-english/manual.pdf](https://support.polar.com/e_manuals/h10-heart-rate-sensor/polar-h10-user-manual-english/manual.pdf) (accessed on 13 January 2021).
13. Duplessis, D. R2Play and Stakeholder Needs: Fostering User-Driven Technology to Support Return-to-Play Decision-Making. Master’s Thesis, ProQuest, University of Toronto, Toronto, ON, Canada, 2021.
14. Lam, E. Using Accelerometers to Score a Multi-Domain Return-to-Play Assessment for Youth Post-Concussion. Master’s Thesis, University of Toronto, Toronto, ON, Canada, 2021.
15. Wundersitz, D.W.T.; Gastin, P.B.; Robertson, S.; Davey, P.C.; Netto, K.J. Validation of a Trunk-Mounted Accelerometer to Measure Peak Impacts during Team Sport Movements. *Int. J. Sport. Med.* **2015**, *36*, 742–746. [[CrossRef](#)] [[PubMed](#)]
16. Wundersitz, D.W.T.; Gastin, P.B.; Richter, C.; Robertson, S.J.; Netto, K.J. Validity of a Trunk-Mounted Accelerometer to Assess Peak Accelerations during Walking, Jogging and Running. *Eur. J. Sport Sci.* **2015**, *15*, 382–390. [[CrossRef](#)] [[PubMed](#)]
17. Koldenhoven, R.M.; Hertel, J. Validation of a Wearable Sensor for Measuring Running Biomechanics. *DIB* **2018**, *2*, 74–78. [[CrossRef](#)] [[PubMed](#)]
18. Godin, G. The Godin-Shephard Leisure-Time Physical Activity Questionnaire. *Health Fit. J. Can.* **2011**, *4*, 18–22. [[CrossRef](#)]
19. Mukaka, M.M. A Guide to Appropriate Use of Correlation Coefficient in Medical Research. *Malawi Med. J.* **2012**, *24*, 69–71. [[CrossRef](#)] [[PubMed](#)]
20. Davoudi, A.; Wanigatunga, A.A.; Kheirhahan, M.; Corbett, D.B.; Mendoza, T.; Battula, M.; Ranka, S.; Fillingim, R.B.; Manini, T.M.; Rashidi, P. Accuracy of Samsung Gear S Smartwatch for Activity Recognition: Validation Study. *JMIR mHealth uHealth* **2019**, *7*, e11270. [[CrossRef](#)] [[PubMed](#)]