

# HS Circularity - Number Distribution



## Measurement Details

**Sample Name:** VINNAPAS 8031 H (WACKER)

**SOP Name:** MATERIE PRIME.vsop

**User Name:**

**Date:** 28 May 2014 10:33:41

**Edited:** True

**Edited On:** 28 May 2014 11:39:02

## Analysis

**SOP Analysis:** 3.0

**SOP Fill Holes:** True

**SOP Segmentation:** None

**SOP Trash Size:** 10

**SOP Optic(s) used:** 20x

## Result

**HS Circularity Minimum:** 0.017

**HS Circularity D[n, 0.1]:** 0.515

**HS Circularity Maximum:** 1.000

**HS Circularity D[n, 0.5]:** 0.846

**HS Circularity Mean:** 0.786

**HS Circularity D[n, 0.9]:** 0.955

**Particles Counted:** 8630

**HS Circularity STDV:** 0.181

**HS Circularity RSD (%):** 22.979

HS Circularity smoothed over 11 points



Record 162: VINNAPAS 8031 H (WACKER)

## Comments:

Morphologi

Malvern Instruments Worldwide

[www.malvern.com](http://www.malvern.com)