Software Requirements Specification (SRS): Smart Infusion Pump

Document Version: 1.2 | Author: Clinical Product Management

REQ-MED-01: The infusion pump shall deliver medication within +/- 5% accuracy under nominal conditions.

REQ-MED-02: The pump controller should quickly sound an acoustic alarm if occlusion pressure exceeds limit.

REQ-MED-03: The system shall maintain an audit trail of all bolus deliveries and operator overrides.

REQ-MED-04: The device shall automatically switch to battery backup power within [TBD: switchover_time_ms] on AC mains lo

REQ-MED-05: The operator interface shall provide user-friendly alerts where feasible during night mode.